

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062723\
 Data File : BP015757.D
 Acq On : 27 Jun 2023 20:52
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020769

Quant Time: Jun 27 22:12:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.058	152	165429	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	637330	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	336965	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.469	188	621969	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	587459	20.000	ng/u1	0.00
88) Perylene-d12	24.110	264	748609	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	38753	8.428	ng/uL	0.00
4) Pyridine-d5	3.828	84	244208	21.070	ng/u1	0.00
7) Phenol-d5	7.228	99	285533	20.173	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	184392	21.057	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	226995	21.245	ng/u1	0.00
15) 4-Methylphenol-d8	8.787	113	222216	19.873	ng/u1	0.00
21) Nitrobenzene-d5	9.240	128	109442	22.469	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	121011	21.287	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	212472	20.974	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	250491	19.249	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	539239	20.704	ng/u1	0.00
49) Acenaphthylene-d8	14.404	160	632871	21.616	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	56215	16.356	ng/u1	0.01
60) Fluorene-d10	15.704	176	435290	20.298	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	72529	18.962	ng/u1	0.00
73) Anthracene-d10	17.569	188	611487	21.523	ng/u1	0.00
81) Pyrene-d10	19.798	212	664667	17.327	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.939	264	780385	20.665	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.423	88	43264	8.743	ng/uL	99
5) Pyridine	3.846	79	262631	21.660	ng/u1	97
6) Benzaldehyde	7.205	77	117879	20.599	ng/u1	97
8) Phenol	7.258	94	301203	20.506	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.481	93	243096	20.801	ng/u1	98
12) 2-Chlorophenol	7.622	128	239178	21.070	ng/u1	99
13) 2-Methylphenol	8.516	108	227256	20.492	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.593	45	341879	21.418	ng/u1	99
16) Acetophenone	8.899	105	364180	19.813	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.875	70	199197	19.512	ng/u1	97
18) 4-Methylphenol	8.852	108	238224	19.993	ng/u1	99
19) Hexachloroethane	9.140	117	109396	20.866	ng/u1	97
22) Nitrobenzene	9.287	77	300245	21.951	ng/u1	99
23) Isophorone	9.810	82	530148	20.254	ng/u1	100
25) 2-Nitrophenol	10.005	139	129789	21.513	ng/u1	99
26) 2,4-Dimethylphenol	10.058	107	279394	21.066	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.293	93	312994	21.182	ng/u1	97
29) 2,4-Dichlorophenol	10.546	162	210542	20.905	ng/u1	100
30) Naphthalene	10.934	128	732407	21.139	ng/u1	99
32) 4-Chloroaniline	11.069	127	262099	19.386	ng/u1	100
33) Hexachlorobutadiene	11.204	225	161249	21.201	ng/u1	98
34) Caprolactam	11.857	113	52296	16.976	ng/u1	93
35) 4-Chloro-3-methylphenol	12.193	107	224152	19.014	ng/u1	99
36) 2-Methylnaphthalene	12.546	142	463370	20.316	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062723\
 Data File : BP015757.D
 Acq On : 27 Jun 2023 20:52
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020769

Quant Time: Jun 27 22:12:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	456790	19.633	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.910	216	256213	23.074	ng/ul	99
40) Hexachlorocyclopentadiene	12.875	237	68175	20.523	ng/ul	96
41) 2,4,6-Trichlorophenol	13.157	196	155314	22.132	ng/ul	97
42) 2,4,5-Trichlorophenol	13.246	196	162659	21.853	ng/ul	99
43) 1,1'-Biphenyl	13.546	154	602485	22.574	ng/ul	98
44) 2-Chloronaphthalene	13.593	162	474049	22.546	ng/ul	99
45) 2-Nitroaniline	13.816	65	141308	21.359	ng/ul	95
47) Dimethylphthalate	14.163	163	557998	20.702	ng/ul	99
48) 2,6-Dinitrotoluene	14.304	165	110871	20.278	ng/ul	98
50) Acenaphthylene	14.434	152	697059	21.608	ng/ul	100
51) 3-Nitroaniline	14.646	138	86328	20.135	ng/ul	90
52) Acenaphthene	14.775	153	482804	21.582	ng/ul	98
53) 2,4-Dinitrophenol	14.881	184	39241	13.528	ng/ul	87
55) 4-Nitrophenol	14.969	109	53741	15.089	ng/ul	97
56) Dibenzofuran	15.110	168	646001	20.802	ng/ul	99
57) 2,4-Dinitrotoluene	15.104	165	144658	19.415	ng/ul#	89
59) Diethylphthalate	15.522	149	540236	19.806	ng/ul	99
61) Fluorene	15.757	166	509524	20.228	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.745	204	258938	20.090	ng/ul	98
63) 4-Nitroaniline	15.816	138	50676	17.269	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.875	198	73395	18.965	ng/ul	93
67) N-Nitrosodiphenylamine	15.969	169	419201	22.514	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	154851	21.803	ng/ul	92
69) Hexachlorobenzene	16.769	284	174920	20.872	ng/ul	95
70) Atrazine	16.922	200	146751	20.590	ng/ul	99
71) Pentachlorophenol	17.134	266	80307	19.966	ng/ul	94
72) Phenanthrene	17.510	178	717376	21.231	ng/ul	99
74) Anthracene	17.610	178	723278	21.303	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.516	216	253044	25.003	ng/uL	98
76) Pentachlorobenzene	15.034	250	232017	22.715	ng/uL	98
77) Carbazole	17.887	167	617738	21.587	ng/ul	99
78) Di-n-butylphthalate	18.398	149	800290	21.095	ng/ul	100
80) Fluoranthene	19.475	202	806730	16.922	ng/ul	99
82) Pyrene	19.828	202	845789	17.392	ng/ul	97
83) Butylbenzylphthalate	20.675	149	361770	18.234	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.475	252	297361	22.392	ng/ul	100
85) Benzo(a)anthracene	21.545	228	883020	21.368	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.427	149	528532	19.384	ng/ul	99
87) Chrysene	21.598	228	839283	21.406	ng/ul	99
89) Di-n-octyl phthalate	22.392	149	945670	17.285	ng/ul	100
90) Benzo(b)fluoranthene	23.322	252	979499	20.363	ng/ul	99
91) Benzo(k)fluoranthene	23.369	252	952447	19.598	ng/ul	99
93) Benzo(a)pyrene	23.992	252	859727	20.455	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.804	276	1222994	21.435	ng/ul	98
95) Dibenzo(a,h)anthracene	26.810	278	1007326	21.494	ng/ul	97
96) Benzo(g,h,i)perylene	27.645	276	1014909	21.622	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062723\
 Data File : BP015757.D
 Acq On : 27 Jun 2023 20:52
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020769

Quant Time: Jun 27 22:12:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.743 ng/uL
RT: 3.423 min Scan# 40
Delta R.T. 0.001 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:	88	Resp:	43264
Ion	Ratio	Lower	Upper
88	100		
43	31.8	27.4	41.2
58	74.7	59.8	89.8



#5
Pyridine
Concen: 21.660 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:	79	Resp:	262631
Ion	Ratio	Lower	Upper
79	100		
52	78.8	65.5	98.3
51	38.3	31.8	47.8









#6
Benzaldehyde
Concen: 20.599 ng/ul
RT: 7.205 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion: 77 Resp: 117879
Ion Ratio Lower Upper
77 100
105 91.4 73.4 110.0
106 87.9 66.6 100.0



#8
Phenol
Concen: 20.506 ng/ul
RT: 7.258 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion: 94 Resp: 301203
Ion Ratio Lower Upper
94 100
65 31.6 26.4 39.6
66 45.8 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 20.801 ng/ul
RT: 7.481 min Scan# 730
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion: 93 Resp: 243096
Ion Ratio Lower Upper
93 100
63 80.4 62.8 94.2
95 32.2 25.6 38.4



#12
2-Chlorophenol
Concen: 21.070 ng/ul
RT: 7.622 min Scan# 754
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:128 Resp: 239178
Ion Ratio Lower Upper
128 100
64 52.7 41.2 61.8
130 32.5 26.0 39.0





#13

2-Methylphenol

Concen: 20.492 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :





SSTD020769

Tgt Ion:108 Resp: 227256

Ion Ratio Lower Upper

108 100

107 89.1 73.8 110.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.418 ng/ul

RT: 8.593 min Scan# 919

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion: 45 Resp: 341879


Ion Ratio Lower Upper

45 100

77 14.8 12.4 18.6

79 11.5 9.3 13.9









#16
Acetophenone
Concen: 19.813 ng/ul
RT: 8.899 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:105 Resp: 364180
Ion Ratio Lower Upper
105 100
77 83.8 65.0 97.4
51 32.4 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 19.512 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion: 70 Resp: 199197
Ion Ratio Lower Upper
70 100
42 62.3 48.2 72.2
101 9.8 7.6 11.4
130 19.1 13.9 20.9









#18
4-Methylphenol
Concen: 19.993 ng/ul
RT: 8.852 min Scan# 963
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:108 Resp: 238224
Ion Ratio Lower Upper
108 100
107 115.6 91.3 136.9



#19
Hexachloroethane
Concen: 20.866 ng/ul
RT: 9.140 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:117 Resp: 109396
Ion Ratio Lower Upper
117 100
201 99.9 84.2 126.2
199 64.3 51.8 77.8









#22
Nitrobenzene
Concen: 21.951 ng/ul
RT: 9.287 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion: 77 Resp: 300245
Ion Ratio Lower Upper
77 100
123 39.2 31.0 46.4
65 14.3 11.3 16.9



#23
Isophorone
Concen: 20.254 ng/ul
RT: 9.810 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion: 82 Resp: 530148
Ion Ratio Lower Upper
82 100
95 8.0 6.3 9.5
138 15.6 12.3 18.5









#25
2-Nitrophenol
Concen: 21.513 ng/ul
RT: 10.005 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:139 Resp: 129789
Ion Ratio Lower Upper
139 100
65 63.0 50.2 75.2
109 41.0 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 21.066 ng/ul
RT: 10.058 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:107 Resp: 279394
Ion Ratio Lower Upper
107 100
121 47.2 38.2 57.2
122 76.2 62.1 93.1





#27

Bis(2-Chloroethoxy)methane

Concen: 21.182 ng/ul

RT: 10.293 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion: 93 Resp: 312994

Ion Ratio Lower Upper

93 100

95 31.2 26.6 39.8

123 10.3 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 20.905 ng/ul

RT: 10.546 min Scan# 1251

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:162 Resp: 210542


Ion Ratio Lower Upper

162 100

164 65.9 53.0 79.6

98 39.2 31.3 46.9









#30
Naphthalene
Concen: 21.139 ng/ul
RT: 10.934 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:128 Resp: 732407
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 13.3 10.3 15.5



#32
4-Chloroaniline
Concen: 19.386 ng/ul
RT: 11.069 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:127 Resp: 262099
Ion Ratio Lower Upper
127 100
129 30.8 24.7 37.1





#33

Hexachlorobutadiene

Concen: 21.201 ng/ul

RT: 11.204 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:225 Resp: 161249

Ion Ratio Lower Upper

225 100

223 64.4 50.4 75.6

227 66.0 51.0 76.4



#34

Caprolactam

Concen: 16.976 ng/ul

RT: 11.857 min Scan# 1474

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:113 Resp: 52296


Ion Ratio Lower Upper

113 100

55 188.6 142.8 214.2

56 137.1 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 19.014 ng/ul
RT: 12.193 min Scan# 1531
Delta R.T. 0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:107 Resp: 224152
Ion Ratio Lower Upper
107 100
144 23.7 18.4 27.6
142 72.7 58.6 87.8



#36
2-Methylnaphthalene
Concen: 20.316 ng/ul
RT: 12.546 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:142 Resp: 463370
Ion Ratio Lower Upper
142 100
141 89.6 73.2 109.8





#37

1-Methylnaphthalene

Concen: 19.633 ng/ul

RT: 12.763 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:142 Resp: 456790

Ion Ratio Lower Upper

142 100

141 92.8 73.1 109.7

116 4.3 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 23.074 ng/ul

RT: 12.910 min Scan# 1653

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Tgt Ion:216 Resp: 256213




Ion Ratio Lower Upper


216 100

214 79.2 62.4 93.6

179 22.5 17.6 26.4

108 18.5 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 20.523 ng/ul
RT: 12.875 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:237 Resp: 68175
Ion Ratio Lower Upper
237 100
235 64.1 48.6 73.0
272 11.8 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 22.132 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:196 Resp: 155314
Ion Ratio Lower Upper
196 100
198 98.2 75.4 113.2
200 29.8 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 21.853 ng/ul
RT: 13.246 min Scan# 1710
Delta R.T. 0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:196 Resp: 162659
Ion Ratio Lower Upper
196 100
198 94.2 74.9 112.3
200 31.2 24.3 36.5



#43
1,1'-Biphenyl
Concen: 22.574 ng/ul
RT: 13.546 min Scan# 1761
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:154 Resp: 602485
Ion Ratio Lower Upper
154 100
153 39.7 32.8 49.2
76 13.1 10.7 16.1





#44

2-Chloronaphthalene

Concen: 22.546 ng/ul

RT: 13.593 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:162 Resp: 474049

Ion Ratio Lower Upper

162 100

164 32.4 26.1 39.1

127 41.1 33.4 50.0



#45

2-Nitroaniline

Concen: 21.359 ng/ul

RT: 13.816 min Scan# 1807

Delta R.T. 0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion: 65 Resp: 141308


Ion Ratio Lower Upper

65 100

92 67.4 49.5 74.3

138 91.9 71.4 107.2









#47
Dimethylphthalate
Concen: 20.702 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:163 Resp: 557998
Ion Ratio Lower Upper
163 100
194 3.7 3.1 4.7
164 10.1 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 20.278 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:165 Resp: 110871
Ion Ratio Lower Upper
165 100
89 65.8 54.1 81.1
121 23.6 18.2 27.2









#50
Acenaphthylene
Concen: 21.608 ng/ul
RT: 14.434 min Scan# 1912
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:152 Resp: 697059
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 20.135 ng/ul
RT: 14.646 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:138 Resp: 86328
Ion Ratio Lower Upper
138 100
108 13.1 9.1 13.7
92 132.4 96.8 145.2





#52

Acenaphthene

Concen: 21.582 ng/ul

RT: 14.775 min Scan# 1970

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:153 Resp: 482804

Ion Ratio Lower Upper

153 100

152 47.8 39.4 59.2

154 87.3 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 13.528 ng/ul

RT: 14.881 min Scan# 1988

Delta R.T. 0.012 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:184 Resp: 39241


Ion Ratio Lower Upper

184 100

63 74.4 71.4 107.0

154 67.6 61.1 91.7









#55
4-Nitrophenol
Concen: 15.089 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. 0.012 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:	109	Resp:	53741
Ion	Ratio	Lower	Upper
109	100		
139	105.8	81.3	121.9
65	126.8	99.9	149.9



#56
Dibenzofuran
Concen: 20.802 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:	168	Resp:	646001
Ion	Ratio	Lower	Upper
168	100		
139	40.8	32.2	48.4









#57
2,4-Dinitrotoluene
Concen: 19.415 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:165 Resp: 144658
Ion Ratio Lower Upper
165 100
63 80.0 56.8 85.2
182 3.0 2.0 3.0#



#59
Diethylphthalate
Concen: 19.806 ng/ul
RT: 15.522 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:149 Resp: 540236
Ion Ratio Lower Upper
149 100
177 21.7 16.8 25.2
150 12.4 10.3 15.5









#61
Fluorene
Concen: 20.228 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:166 Resp: 509524
Ion Ratio Lower Upper
166 100
165 96.3 78.7 118.1
167 14.0 10.5 15.7



#62
4-Chlorophenyl-phenylether
Concen: 20.090 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:204 Resp: 258938
Ion Ratio Lower Upper
204 100
206 31.4 26.0 39.0
141 66.8 51.8 77.8









#63
4-Nitroaniline
Concen: 17.269 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:138 Resp: 50676
Ion Ratio Lower Upper
138 100
92 56.8 50.1 75.1
108 103.8 85.5 128.3



#66
4,6-Dinitro-2-methylphenol
Concen: 18.965 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:198 Resp: 73395
Ion Ratio Lower Upper
198 100
182 4.1 3.4 5.2
77 35.1 24.4 36.6





#67

N-Nitrosodiphenylamine

Concen: 22.514 ng/ul

RT: 15.969 min Scan# 2173

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:169 Resp: 419201

Ion Ratio Lower Upper

169 100

168 66.7 54.2 81.4

167 36.2 29.7 44.5



#68

4-Bromophenyl-phenylether

Concen: 21.803 ng/ul

RT: 16.645 min Scan# 2288

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:248 Resp: 154851


Ion Ratio Lower Upper

248 100

250 95.0 84.6 126.8

141 81.1 67.8 101.6





#69

Hexachlorobenzene

Concen: 20.872 ng/ul

RT: 16.769 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:284 Resp: 174920

Ion Ratio Lower Upper

284 100

142 39.0 27.8 41.8

249 31.2 26.1 39.1



#70

Atrazine

Concen: 20.590 ng/ul

RT: 16.922 min Scan# 2335

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:200 Resp: 146751


Ion Ratio Lower Upper

200 100

173 25.0 19.8 29.8

215 46.5 36.8 55.2





#71

Pentachlorophenol

Concen: 19.966 ng/ul

RT: 17.134 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:266 Resp: 80307

Ion Ratio Lower Upper

266 100

264 62.3 53.4 80.2

268 59.1 51.4 77.2



#72

Phenanthrene

Concen: 21.231 ng/ul

RT: 17.510 min Scan# 2435

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:178 Resp: 717376


Ion Ratio Lower Upper

178 100

179 15.9 12.4 18.6

176 19.2 15.6 23.4





#74

Anthracene

Concen: 21.303 ng/uI

RT: 17.610 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:178 Resp: 723278

Ion Ratio Lower Upper

178 100

179 15.0 12.1 18.1

176 18.6 14.6 22.0



#75

1,2,3,4-Tetrachlorobenzene

Concen: 25.003 ng/uI

RT: 13.516 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:216 Resp: 253044


Ion Ratio Lower Upper

216 100

214 75.7 62.0 93.0

218 48.7 38.4 57.6





#76

Pentachlorobenzene

Concen: 22.715 ng/uL

RT: 15.034 min Scan# 2014

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:250 Resp: 232017

Ion Ratio Lower Upper

250 100

248 60.9 50.7 76.1

252 62.9 51.2 76.8



#77

Carbazole

Concen: 21.587 ng/uL

RT: 17.887 min Scan# 2499

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:167 Resp: 617738


Ion Ratio Lower Upper

167 100

166 21.9 17.4 26.0

139 13.4 11.0 16.4





#78

Di-n-butylphthalate

Concen: 21.095 ng/ul

RT: 18.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:149 Resp: 800290

Ion Ratio Lower Upper

149 100

150 9.3 7.5 11.3

104 6.1 4.6 7.0



#80

Fluoranthene

Concen: 16.922 ng/ul

RT: 19.475 min Scan# 2769

Delta R.T. 0.000 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:202 Resp: 806730


Ion Ratio Lower Upper

202 100

101 10.7 8.3 12.5

100 8.1 6.6 9.8





#82

Pyrene

Concen: 17.392 ng/ul

RT: 19.828 min Scan# 2829

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:202 Resp: 845789

Ion Ratio Lower Upper

202 100

101 12.8 9.3 13.9

100 10.4 7.6 11.4



#83

Butylbenzylphthalate

Concen: 18.234 ng/ul

RT: 20.675 min Scan# 2973

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:149 Resp: 361770


Ion Ratio Lower Upper

149 100

91 78.9 59.8 89.6

206 20.3 16.4 24.6









#84
3,3'-Dichlorobenzidine
Concen: 22.392 ng/ul
RT: 21.475 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:252 Resp: 297361
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.6
126 10.7 8.6 13.0



#85
Benzo(a)anthracene
Concen: 21.368 ng/ul
RT: 21.545 min Scan# 3121
Delta R.T. 0.000 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:228 Resp: 883020
Ion Ratio Lower Upper
228 100
229 19.2 15.3 22.9
226 26.3 21.4 32.0





#86

Bis(2-ethylhexyl)phthalate

Concen: 19.384 ng/ul

RT: 21.427 min Scan# 3102

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:149 Resp: 528532

Ion Ratio Lower Upper

149 100

167 26.5 21.0 31.6

279 4.2 3.8 5.6



#87

Chrysene

Concen: 21.406 ng/ul

RT: 21.598 min Scan# 3130

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:228 Resp: 839283


Ion Ratio Lower Upper

228 100

226 30.0 24.3 36.5

229 19.9 15.5 23.3





#89

Di-n-octyl phthalate

Concen: 17.285 ng/ul

RT: 22.392 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52





Instrument :

BNA_P

ClientSampleId :

SSTD020769

Tgt Ion:149 Resp: 945670



#90

Benzo(b)fluoranthene

Concen: 20.363 ng/ul

RT: 23.322 min Scan# 3423

Delta R.T. -0.006 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:252 Resp: 979499


Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 9.0 7.5 11.3









#91
Benzo(k)fluoranthene
Concen: 19.598 ng/ul
RT: 23.369 min Scan# 34
Delta R.T. -0.012 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52




Instrument :
BNA_P
ClientSampleId :
SSTD020769


Tgt Ion:252 Resp: 952447
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 9.0 7.3 10.9



#93
Benzo(a)pyrene
Concen: 20.455 ng/ul
RT: 23.992 min Scan# 3537
Delta R.T. -0.012 min
Lab File: BP015757.D
Acq: 27 Jun 2023 20:52

Tgt Ion:252 Resp: 859727
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 10.2 8.2 12.2





#94

Indeno(1,2,3-cd)pyrene

Concen: 21.435 ng/ul

RT: 26.804 min Scan# 4015

Delta R.T. -0.012 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52

Instrument :

BNA_P

ClientSampleId :

SSTD020769





Tgt Ion:276 Resp: 1222994

Ion Ratio Lower Upper

276 100

138 21.2 15.7 23.5

277 25.2 19.6 29.4



#95

Dibenzo(a,h)anthracene

Concen: 21.494 ng/ul

RT: 26.810 min Scan# 4016

Delta R.T. -0.017 min

Lab File: BP015757.D

Acq: 27 Jun 2023 20:52




Tgt Ion:278 Resp: 1007326


Ion Ratio Lower Upper

278 100

139 15.5 10.5 15.7


279 23.7 18.4 27.6





#96
 Benzo(g,h,i)perylene
 Concen: 21.622 ng/ul
 RT: 27.645 min Scan# 41
 Delta R.T. -0.012 min
 Lab File: BP015757.D
 Acq: 27 Jun 2023 20:52

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020769



Tgt Ion:276 Resp: 1014909
 Ion Ratio Lower Upper
 276 100
 138 19.0 15.1 22.7
 277 23.8 18.8 28.2

