

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062723\
 Data File : BPCCC757.D
 Acq On : 27 Jun 2023 21:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jun 27 22:12:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.057	152	180799	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	727478	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	416178	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.469	188	770033	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	571425	20.000	ng/u1	0.00
88) Perylene-d12	24.110	264	717102	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	40642	8.088	ng/uL	0.00
4) Pyridine-d5	3.822	84	261902	20.675	ng/u1	0.00
7) Phenol-d5	7.228	99	317952	20.554	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	204511	21.369	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	246745	21.130	ng/u1	0.00
15) 4-Methylphenol-d8	8.787	113	253386	20.734	ng/u1	0.00
21) Nitrobenzene-d5	9.240	128	125633	22.597	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	139372	21.479	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.516	165	248710	21.509	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	301245	20.281	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	684604	21.283	ng/u1	0.00
49) Acenaphthylene-d8	14.404	160	780168	21.575	ng/u1	0.00
54) 4-Nitrophenol-d4	14.951	143	71857	16.928	ng/u1	0.00
60) Fluorene-d10	15.704	176	542779	20.493	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	90629	19.138	ng/u1	0.00
73) Anthracene-d10	17.569	188	742525	21.110	ng/u1	0.00
81) Pyrene-d10	19.798	212	717151	19.220	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.939	264	744516	20.582	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	42736	7.903	ng/uL	100
5) Pyridine	3.846	79	274618	20.723	ng/u1	95
6) Benzaldehyde	7.205	77	129828	20.758	ng/u1	97
8) Phenol	7.257	94	332401	20.706	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.481	93	272727	21.353	ng/u1	100
12) 2-Chlorophenol	7.622	128	257581	20.762	ng/u1	99
13) 2-Methylphenol	8.516	108	250814	20.694	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.593	45	387506	22.213	ng/u1	99
16) Acetophenone	8.899	105	417151	20.765	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.875	70	228513	20.481	ng/u1	98
18) 4-Methylphenol	8.852	108	272526	20.927	ng/u1	99
19) Hexachloroethane	9.140	117	118887	20.749	ng/u1	92
22) Nitrobenzene	9.287	77	339865	21.769	ng/u1	98
23) Isophorone	9.810	82	632328	21.164	ng/u1	100
25) 2-Nitrophenol	10.004	139	150221	21.814	ng/u1	99
26) 2,4-Dimethylphenol	10.057	107	321406	21.231	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.287	93	366098	21.706	ng/u1	99
29) 2,4-Dichlorophenol	10.546	162	246299	21.425	ng/u1	99
30) Naphthalene	10.934	128	837248	21.171	ng/u1	100
32) 4-Chloroaniline	11.063	127	307797	19.945	ng/u1	100
33) Hexachlorobutadiene	11.204	225	183381	21.123	ng/u1	98
34) Caprolactam	11.857	113	68203	19.396	ng/u1	93
35) 4-Chloro-3-methylphenol	12.187	107	276800	20.570	ng/u1	98
36) 2-Methylnaphthalene	12.540	142	542605	20.842	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062723\
 Data File : BPCCC757.D
 Acq On : 27 Jun 2023 21:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 22 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Jun 27 22:12:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	548717	20.661	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.910	216	308036	22.461	ng/ul	99
40) Hexachlorocyclopentadiene	12.869	237	82979	20.225	ng/ul	96
41) 2,4,6-Trichlorophenol	13.157	196	191195	22.060	ng/ul	99
42) 2,4,5-Trichlorophenol	13.240	196	200405	21.799	ng/ul	99
43) 1,1'-Biphenyl	13.540	154	732368	22.217	ng/ul	98
44) 2-Chloronaphthalene	13.592	162	570849	21.983	ng/ul	98
45) 2-Nitroaniline	13.810	65	177840	21.765	ng/ul	97
47) Dimethylphthalate	14.163	163	706626	21.226	ng/ul	99
48) 2,6-Dinitrotoluene	14.304	165	142181	21.055	ng/ul	95
50) Acenaphthylene	14.434	152	854718	21.452	ng/ul	99
51) 3-Nitroaniline	14.645	138	109765	20.728	ng/ul	96
52) Acenaphthene	14.775	153	585310	21.184	ng/ul	99
53) 2,4-Dinitrophenol	14.875	184	51514	14.379	ng/ul	95
55) 4-Nitrophenol	14.963	109	67138	15.262	ng/ul	99
56) Dibenzofuran	15.110	168	798459	20.818	ng/ul	100
57) 2,4-Dinitrotoluene	15.104	165	188893	20.527	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.345	232	158502	20.073	ng/ul	99
59) Diethylphthalate	15.522	149	702101	20.841	ng/ul	99
61) Fluorene	15.757	166	636430	20.457	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.745	204	327267	20.559	ng/ul	98
63) 4-Nitroaniline	15.816	138	60343	16.650	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.875	198	89980	18.779	ng/ul	96
67) N-Nitrosodiphenylamine	15.969	169	530607	23.018	ng/ul	98
68) 4-Bromophenyl-phenylether	16.645	248	196635	22.362	ng/ul	96
69) Hexachlorobenzene	16.769	284	220766	21.278	ng/ul	97
70) Atrazine	16.922	200	189378	21.461	ng/ul	98
71) Pentachlorophenol	17.133	266	99524	19.986	ng/ul	95
72) Phenanthrene	17.510	178	881696	21.077	ng/ul	99
74) Anthracene	17.610	178	888370	21.134	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.510	216	306179	24.436	ng/uL	98
76) Pentachlorobenzene	15.028	250	290417	22.965	ng/uL	99
77) Carbazole	17.886	167	724465	20.448	ng/ul	99
78) Di-n-butylphthalate	18.398	149	1030206	21.934	ng/ul	100
80) Fluoranthene	19.474	202	895160	19.304	ng/ul	99
82) Pyrene	19.827	202	912802	19.297	ng/ul	97
83) Butylbenzylphthalate	20.674	149	389184	20.167	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.474	252	280849	21.742	ng/ul	98
85) Benzo(a)anthracene	21.539	228	852189	21.200	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.427	149	563695	21.254	ng/ul	100
87) Chrysene	21.598	228	813238	21.324	ng/ul	99
89) Di-n-octyl phthalate	22.392	149	945170	18.035	ng/ul	100
90) Benzo(b)fluoranthene	23.321	252	936456	20.324	ng/ul	99
91) Benzo(k)fluoranthene	23.368	252	916391	19.684	ng/ul	98
93) Benzo(a)pyrene	23.992	252	841925	20.911	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.804	276	1177741	21.548	ng/ul	98
95) Dibenzo(a,h)anthracene	26.815	278	957261	21.323	ng/ul	97
96) Benzo(g,h,i)perylene	27.645	276	972359	21.625	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 27 22:12:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 27 00:56:58 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.903 ng/uL
RT: 3.417 min Scan# 39
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 42736
Ion Ratio Lower Upper
88 100
43 34.4 27.4 41.2
58 75.2 59.8 89.8



#5
Pyridine
Concen: 20.723 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. 0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion: 79 Resp: 274618
Ion Ratio Lower Upper
79 100
52 75.2 65.5 98.3
51 39.2 31.8 47.8









#6
Benzaldehyde
Concen: 20.758 ng/ul
RT: 7.205 min Scan# 683
Delta R.T. 0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 129828
Ion Ratio Lower Upper
77 100
105 90.0 73.4 110.0
106 87.0 66.6 100.0



#8
Phenol
Concen: 20.706 ng/ul
RT: 7.257 min Scan# 692
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion: 94 Resp: 332401
Ion Ratio Lower Upper
94 100
65 32.0 26.4 39.6
66 45.3 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 21.353 ng/ul
RT: 7.481 min Scan# 730
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 272727
Ion Ratio Lower Upper
93 100
63 78.2 62.8 94.2
95 31.9 25.6 38.4



#12
2-Chlorophenol
Concen: 20.762 ng/ul
RT: 7.622 min Scan# 754
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:128 Resp: 257581
Ion Ratio Lower Upper
128 100
64 52.8 41.2 61.8
130 32.1 26.0 39.0





#13

2-Methylphenol

Concen: 20.694 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 250814

Ion Ratio Lower Upper

108 100

107 88.9 73.8 110.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.213 ng/ul

RT: 8.593 min Scan# 919

Delta R.T. -0.000 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion: 45 Resp: 387506


Ion Ratio Lower Upper

45 100

77 14.7 12.4 18.6

79 11.5 9.3 13.9









#16
Acetophenone
Concen: 20.765 ng/ul
RT: 8.899 min Scan# 91
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 417151
Ion Ratio Lower Upper
105 100
77 84.8 65.0 97.4
51 31.0 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.481 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion: 70 Resp: 228513
Ion Ratio Lower Upper
70 100
42 61.2 48.2 72.2
101 9.1 7.6 11.4
130 18.9 13.9 20.9









#18
4-Methylphenol
Concen: 20.927 ng/ul
RT: 8.852 min Scan# 963
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 272526
Ion Ratio Lower Upper
108 100
107 115.6 91.3 136.9



#19
Hexachloroethane
Concen: 20.749 ng/ul
RT: 9.140 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:117 Resp: 118887
Ion Ratio Lower Upper
117 100
201 95.7 84.2 126.2
199 59.3 51.8 77.8









#22
Nitrobenzene
Concen: 21.769 ng/ul
RT: 9.287 min Scan# 1037
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 339865
Ion Ratio Lower Upper
77 100
123 39.7 31.0 46.4
65 14.7 11.3 16.9



#23
Isophorone
Concen: 21.164 ng/ul
RT: 9.810 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion: 82 Resp: 632328
Ion Ratio Lower Upper
82 100
95 7.9 6.3 9.5
138 15.6 12.3 18.5









#25
2-Nitrophenol
Concen: 21.814 ng/ul
RT: 10.004 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 150221
Ion Ratio Lower Upper
139 100
65 63.6 50.2 75.2
109 38.3 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 21.231 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:107 Resp: 321406
Ion Ratio Lower Upper
107 100
121 47.7 38.2 57.2
122 78.6 62.1 93.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.706 ng/ul
 RT: 10.287 min Scan# 1207
 Delta R.T. -0.006 min
 Lab File: BPCCC757.D
 Acq: 27 Jun 2023 21:26




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 366098
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.6 39.8
 123 10.7 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.425 ng/ul
 RT: 10.546 min Scan# 1251
 Delta R.T. -0.006 min
 Lab File: BPCCC757.D
 Acq: 27 Jun 2023 21:26

Tgt Ion:162 Resp: 246299
 Ion Ratio Lower Upper
 162 100
 164 65.6 53.0 79.6
 98 39.4 31.3 46.9









#30
Naphthalene
Concen: 21.171 ng/ul
RT: 10.934 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Instrument :
BNA_P
ClientSampleId :

Tgt Ion:128 Resp: 837248




Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	12.8
127	12.9	10.3	15.5




#32
4-Chloroaniline
Concen: 19.945 ng/ul
RT: 11.063 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:127 Resp: 307797

Ion	Ratio	Lower	Upper
127	100		
129	30.7	24.7	37.1









#33
Hexachlorobutadiene
Concen: 21.123 ng/ul
RT: 11.204 min Scan# 1363
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:225 Resp: 183381
Ion Ratio Lower Upper
225 100
223 62.6 50.4 75.6
227 61.7 51.0 76.4



#34
Caprolactam
Concen: 19.396 ng/ul
RT: 11.857 min Scan# 1474
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:113 Resp: 68203
Ion Ratio Lower Upper
113 100
55 192.0 142.8 214.2
56 151.3 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 20.570 ng/ul
RT: 12.187 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 276800
Ion Ratio Lower Upper
107 100
144 23.3 18.4 27.6
142 70.9 58.6 87.8



#36
2-Methylnaphthalene
Concen: 20.842 ng/ul
RT: 12.540 min Scan# 1590
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:142 Resp: 542605
Ion Ratio Lower Upper
142 100
141 89.1 73.2 109.8









#37
1-Methylnaphthalene
Concen: 20.661 ng/ul
RT: 12.757 min Scan# 1627
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 548717
Ion Ratio Lower Upper
142 100
141 93.8 73.1 109.7
116 4.5 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.461 ng/ul
RT: 12.910 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:216 Resp: 308036
Ion Ratio Lower Upper
216 100
214 78.3 62.4 93.6
179 21.7 17.6 26.4
108 18.1 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 20.225 ng/ul
RT: 12.869 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 82979
Ion Ratio Lower Upper
237 100
235 64.4 48.6 73.0
272 11.9 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 22.060 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:196 Resp: 191195
Ion Ratio Lower Upper
196 100
198 93.9 75.4 113.2
200 30.0 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 21.799 ng/ul

RT: 13.240 min Scan# 1710

Delta R.T. -0.000 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:196 Resp: 200405

Ion Ratio Lower Upper

196 100

198 94.9 74.9 112.3

200 30.2 24.3 36.5



#43

1,1'-Biphenyl

Concen: 22.217 ng/ul

RT: 13.540 min Scan# 1760

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:154 Resp: 732368


Ion Ratio Lower Upper

154 100

153 39.8 32.8 49.2


76 13.6 10.7 16.1








#44
2-Chloronaphthalene
Concen: 21.983 ng/ul
RT: 13.592 min Scan# 1769
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:162 Resp: 570849
Ion Ratio Lower Upper
162 100
164 34.0 26.1 39.1
127 40.9 33.4 50.0



#45
2-Nitroaniline
Concen: 21.765 ng/ul
RT: 13.810 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion: 65 Resp: 177840
Ion Ratio Lower Upper
65 100
92 67.2 49.5 74.3
138 90.3 71.4 107.2









#47
Dimethylphthalate
Concen: 21.226 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 706626
Ion Ratio Lower Upper
163 100
194 3.7 3.1 4.7
164 9.9 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 21.055 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:165 Resp: 142181
Ion Ratio Lower Upper
165 100
89 63.5 54.1 81.1
121 20.9 18.2 27.2









#50
Acenaphthylene
Concen: 21.452 ng/ul
RT: 14.434 min Scan# 1912
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 854718
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 12.8 10.5 15.7



#51
3-Nitroaniline
Concen: 20.728 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:138 Resp: 109765
Ion Ratio Lower Upper
138 100
108 10.8 9.1 13.7
92 126.2 96.8 145.2





#52

Acenaphthene

Concen: 21.184 ng/ul

RT: 14.775 min Scan# 1970

Delta R.T. -0.000 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:153 Resp: 585310

Ion Ratio Lower Upper

153 100

152 48.2 39.4 59.2

154 90.2 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 14.379 ng/ul

RT: 14.875 min Scan# 1987

Delta R.T. 0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:184 Resp: 51514


Ion Ratio Lower Upper

184 100

63 81.2 71.4 107.0

154 75.0 61.1 91.7









#55
4-Nitrophenol
Concen: 15.262 ng/ul
RT: 14.963 min Scan# 2002
Delta R.T. 0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 67138
Ion Ratio Lower Upper
109 100
139 102.5 81.3 121.9
65 127.2 99.9 149.9



#56
Dibenzofuran
Concen: 20.818 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:168 Resp: 798459
Ion Ratio Lower Upper
168 100
139 40.6 32.2 48.4









#57
2,4-Dinitrotoluene
Concen: 20.527 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 188893
Ion Ratio Lower Upper
165 100
63 77.4 56.8 85.2
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.073 ng/ul
RT: 15.345 min Scan# 2067
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26


Tgt Ion:232 Resp: 158502
Ion Ratio Lower Upper
232 100
131 48.0 38.2 57.2
130 2.7 2.2 3.2
166 33.0 25.6 38.4








#59
Diethylphthalate
Concen: 20.841 ng/ul
RT: 15.522 min Scan# 2097
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26


Instrument :
BNA_P
ClientSampleId :





Tgt Ion:149 Resp: 702101
Ion Ratio Lower Upper
149 100
177 21.6 16.8 25.2
150 12.4 10.3 15.5




#61
Fluorene
Concen: 20.457 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26



Tgt Ion:166 Resp: 636430
Ion Ratio Lower Upper
166 100
165 96.4 78.7 118.1
167 13.9 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.559 ng/ul
RT: 15.745 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 327267
Ion Ratio Lower Upper
204 100
206 33.4 26.0 39.0
141 66.3 51.8 77.8



#63
4-Nitroaniline
Concen: 16.650 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:138 Resp: 60343
Ion Ratio Lower Upper
138 100
92 58.8 50.1 75.1
108 99.5 85.5 128.3









#66
4,6-Dinitro-2-methylphenol
Concen: 18.779 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 89980
Ion Ratio Lower Upper
198 100
182 5.0 3.4 5.2
77 27.9 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 23.018 ng/ul
RT: 15.969 min Scan# 2173
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:169 Resp: 530607
Ion Ratio Lower Upper
169 100
168 66.2 54.2 81.4
167 36.6 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 22.362 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 196635
Ion Ratio Lower Upper
248 100
250 99.6 84.6 126.8
141 83.3 67.8 101.6



#69
Hexachlorobenzene
Concen: 21.278 ng/ul
RT: 16.769 min Scan# 2309
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:284 Resp: 220766
Ion Ratio Lower Upper
284 100
142 37.4 27.8 41.8
249 31.9 26.1 39.1





#70

Atrazine

Concen: 21.461 ng/ul

RT: 16.922 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 189378

Ion Ratio Lower Upper

200 100

173 26.4 19.8 29.8

215 47.1 36.8 55.2



#71

Pentachlorophenol

Concen: 19.986 ng/ul

RT: 17.133 min Scan# 2371

Delta R.T. 0.000 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:266 Resp: 99524


Ion Ratio Lower Upper

266 100

264 60.2 53.4 80.2

268 62.4 51.4 77.2





#72

Phenanthrene

Concen: 21.077 ng/ul

RT: 17.510 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 881696

Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 18.9 15.6 23.4



#74

Anthracene

Concen: 21.134 ng/ul

RT: 17.610 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:178 Resp: 888370


Ion Ratio Lower Upper

178 100

179 14.9 12.1 18.1

176 18.5 14.6 22.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 24.436 ng/uL

RT: 13.510 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:216 Resp: 306179

Ion Ratio Lower Upper

216 100

214 79.4 62.0 93.0

218 49.5 38.4 57.6



#76

Pentachlorobenzene

Concen: 22.965 ng/uL

RT: 15.028 min Scan# 2013

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:250 Resp: 290417


Ion Ratio Lower Upper

250 100

248 63.9 50.7 76.1


252 64.9 51.2 76.8








#77
 Carbazole
 Concen: 20.448 ng/ul
 RT: 17.886 min Scan# 2499
 Delta R.T. -0.000 min
 Lab File: BPCCC757.D
 Acq: 27 Jun 2023 21:26

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion:167 Resp: 724465
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.4 26.0
 139 13.3 11.0 16.4



#78
 Di-n-butylphthalate
 Concen: 21.934 ng/ul
 RT: 18.398 min Scan# 2586
 Delta R.T. -0.006 min
 Lab File: BPCCC757.D
 Acq: 27 Jun 2023 21:26

Tgt Ion:149 Resp: 1030206
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 6.0 4.6 7.0









#80
Fluoranthene
Concen: 19.304 ng/ul
RT: 19.474 min Scan# 21
Delta R.T. -0.000 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 895160
Ion Ratio Lower Upper
202 100
101 10.3 8.3 12.5
100 7.9 6.6 9.8



#82
Pyrene
Concen: 19.297 ng/ul
RT: 19.827 min Scan# 2829
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:202 Resp: 912802
Ion Ratio Lower Upper
202 100
101 12.7 9.3 13.9
100 10.5 7.6 11.4





#83

Butylbenzylphthalate

Concen: 20.167 ng/ul

RT: 20.674 min Scan# 2973

Delta R.T. -0.006 min

Lab File: BPCCC757.D





Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:	149	Resp:	389184
Ion Ratio	Lower	Upper	
149	100		
91	78.3	59.8	89.6
206	20.8	16.4	24.6



#84

3,3'-Dichlorobenzidine

Concen: 21.742 ng/ul




RT: 21.474 min Scan# 3109


Delta R.T. 0.000 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Tgt Ion:	252	Resp:	280849
Ion Ratio	Lower	Upper	
252	100		
254	65.7	51.0	76.6
126	10.7	8.6	13.0





#85

Benzo(a)anthracene

Concen: 21.200 ng/ul

RT: 21.539 min Scan# 3121

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 852189

Ion Ratio Lower Upper

228 100

229 19.7 15.3 22.9

226 26.5 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.254 ng/ul

RT: 21.427 min Scan# 3101

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:149 Resp: 563695


Ion Ratio Lower Upper

149 100

167 26.6 21.0 31.6

279 4.7 3.8 5.6





#87

Chrysene

Concen: 21.324 ng/ul

RT: 21.598 min Scan# 3131

Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 813238

Ion Ratio Lower Upper

228 100

226 29.8 24.3 36.5

229 19.2 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 18.035 ng/ul




RT: 22.392 min Scan# 3265


Delta R.T. -0.006 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Tgt Ion:149 Resp: 945170









#90
Benzo(b)fluoranthene
Concen: 20.324 ng/ul
RT: 23.321 min Scan# 3423
Delta R.T. -0.006 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 936456
Ion Ratio Lower Upper
252 100
253 21.4 17.3 25.9
125 8.9 7.5 11.3



#91
Benzo(k)fluoranthene
Concen: 19.684 ng/ul
RT: 23.368 min Scan# 3431
Delta R.T. -0.012 min
Lab File: BPCCC757.D
Acq: 27 Jun 2023 21:26

Tgt Ion:252 Resp: 916391
Ion Ratio Lower Upper
252 100
253 22.2 17.0 25.4
125 9.6 7.3 10.9





#93

Benzo(a)pyrene

Concen: 20.911 ng/ul

RT: 23.992 min Scan# 3538

Delta R.T. -0.012 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 841925

Ion Ratio Lower Upper

252 100

253 21.1 17.7 26.5

125 10.1 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.548 ng/ul

RT: 26.804 min Scan# 4015

Delta R.T. -0.012 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26




Tgt Ion:276 Resp: 1177741


Ion Ratio Lower Upper

276 100

138 21.6 15.7 23.5

277 24.7 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 21.323 ng/ul

RT: 26.815 min Scan# 4018

Delta R.T. -0.012 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 957261

Ion Ratio Lower Upper

278 100

139 14.0 10.5 15.7

279 24.5 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 21.625 ng/ul

RT: 27.645 min Scan# 4158

Delta R.T. -0.012 min

Lab File: BPCCC757.D

Acq: 27 Jun 2023 21:26

Tgt Ion:276 Resp: 972359

Ion Ratio Lower Upper

276 100

138 18.3 15.1 22.7

277 23.4 18.8 28.2

