

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062823\
 Data File : BP015768.D
 Acq On : 28 Jun 2023 08:43
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771

Quant Time: Jun 28 22:02:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.057	152	202847	20.000	ng/u1	0.00
20) Naphthalene-d8	10.881	136	874159	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.704	164	553475	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.469	188	1204054	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	922962	20.000	ng/u1	0.00
88) Perylene-d12	24.104	264	903966	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	44142	7.829	ng/uL	0.00
4) Pyridine-d5	3.822	84	284707	20.033	ng/u1	0.00
7) Phenol-d5	7.222	99	367672	21.184	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	234512	21.840	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	279728	21.351	ng/u1	0.00
15) 4-Methylphenol-d8	8.787	113	294500	21.479	ng/u1	0.00
21) Nitrobenzene-d5	9.240	128	142107	21.272	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	161475	20.710	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.516	165	293598	21.131	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	397547	22.273	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	899056	21.016	ng/u1	0.00
49) Acenaphthylene-d8	14.404	160	1020236	21.215	ng/u1	0.00
54) 4-Nitrophenol-d4	14.945	143	114713	20.320	ng/u1	0.00
60) Fluorene-d10	15.698	176	752989	21.377	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.857	200	141016	19.044	ng/u1	0.00
73) Anthracene-d10	17.569	188	1164021	21.164	ng/u1	0.00
81) Pyrene-d10	19.792	212	1256594	20.850	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.933	264	939332	20.599	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	47415	7.815	ng/uL	97
5) Pyridine	3.846	79	298209	20.057	ng/u1	94
6) Benzaldehyde	7.199	77	157905	22.503	ng/u1	96
8) Phenol	7.252	94	382776	21.253	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.475	93	313335	21.866	ng/u1	98
12) 2-Chlorophenol	7.616	128	293368	21.077	ng/u1	98
13) 2-Methylphenol	8.516	108	292779	21.530	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.593	45	457761	23.388	ng/u1	99
16) Acetophenone	8.899	105	479576	21.278	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.875	70	270215	21.586	ng/u1	98
18) 4-Methylphenol	8.852	108	321224	21.986	ng/u1	97
19) Hexachloroethane	9.140	117	134218	20.878	ng/u1	98
22) Nitrobenzene	9.287	77	389296	20.751	ng/u1	98
23) Isophorone	9.810	82	755734	21.050	ng/u1	99
25) 2-Nitrophenol	9.999	139	177019	21.392	ng/u1	95
26) 2,4-Dimethylphenol	10.057	107	373115	20.511	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.287	93	448064	22.108	ng/u1	97
29) 2,4-Dichlorophenol	10.546	162	297012	21.501	ng/u1	99
30) Naphthalene	10.934	128	993632	20.909	ng/u1	100
32) 4-Chloroaniline	11.063	127	415436	22.403	ng/u1	97
33) Hexachlorobutadiene	11.204	225	200243	19.195	ng/u1	99
34) Caprolactam	11.857	113	99432	23.532	ng/u1	92
35) 4-Chloro-3-methylphenol	12.187	107	350405	21.671	ng/u1	98
36) 2-Methylnaphthalene	12.540	142	668114	21.357	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062823\
 Data File : BP015768.D
 Acq On : 28 Jun 2023 08:43
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771

Quant Time: Jun 28 22:02:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	672122	21.061	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.910	216	374157	20.514	ng/ul	98
40) Hexachlorocyclopentadiene	12.869	237	89737	16.446	ng/ul	96
41) 2,4,6-Trichlorophenol	13.157	196	240166	20.836	ng/ul	97
42) 2,4,5-Trichlorophenol	13.240	196	260083	21.273	ng/ul	97
43) 1,1'-Biphenyl	13.540	154	912136	20.807	ng/ul	98
44) 2-Chloronaphthalene	13.587	162	715242	20.711	ng/ul	99
45) 2-Nitroaniline	13.810	65	243838	22.439	ng/ul	93
47) Dimethylphthalate	14.163	163	931636	21.043	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	193740	21.573	ng/ul	98
50) Acenaphthylene	14.434	152	1120130	21.139	ng/ul	99
51) 3-Nitroaniline	14.639	138	155199	22.038	ng/ul	90
52) Acenaphthene	14.769	153	772533	21.025	ng/ul	98
53) 2,4-Dinitrophenol	14.869	184	83802	17.589	ng/ul	92
55) 4-Nitrophenol	14.963	109	106911	18.275	ng/ul	89
56) Dibenzofuran	15.110	168	1080058	21.174	ng/ul	98
57) 2,4-Dinitrotoluene	15.098	165	274230	22.408	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.345	232	221398	21.083	ng/ul	99
59) Diethylphthalate	15.516	149	963578	21.507	ng/ul	99
61) Fluorene	15.757	166	881265	21.300	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.745	204	436241	20.606	ng/ul	98
63) 4-Nitroaniline	15.816	138	105344	21.856	ng/ul	86
66) 4,6-Dinitro-2-methylph...	15.869	198	145107	19.368	ng/ul	99
67) N-Nitrosodiphenylamine	15.963	169	750608	20.824	ng/ul	98
68) 4-Bromophenyl-phenylether	16.645	248	278273	20.239	ng/ul	92
69) Hexachlorobenzene	16.769	284	326117	20.101	ng/ul	97
70) Atrazine	16.922	200	290740	21.072	ng/ul	98
71) Pentachlorophenol	17.128	266	160038	20.553	ng/ul	98
72) Phenanthrene	17.510	178	1377104	21.053	ng/ul	99
74) Anthracene	17.604	178	1416625	21.553	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.510	216	382452	19.520	ng/ul	100
76) Pentachlorobenzene	15.028	250	384803	19.460	ng/ul	99
77) Carbazole	17.880	167	1234960	22.293	ng/ul	99
78) Di-n-butylphthalate	18.398	149	1665108	22.672	ng/ul	99
80) Fluoranthene	19.469	202	1557650	20.796	ng/ul	97
82) Pyrene	19.822	202	1568725	20.532	ng/ul#	93
83) Butylbenzylphthalate	20.669	149	733919	23.545	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.469	252	461438	22.116	ng/ul	99
85) Benzo(a)anthracene	21.539	228	1356100	20.887	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.421	149	1058352	24.706	ng/ul	100
87) Chrysene	21.598	228	1288168	20.912	ng/ul	99
89) Di-n-octyl phthalate	22.386	149	1721124	26.053	ng/ul	100
90) Benzo(b)fluoranthene	23.315	252	1216653	20.947	ng/ul	99
91) Benzo(k)fluoranthene	23.368	252	1238202	21.099	ng/ul	99
93) Benzo(a)pyrene	23.986	252	1055845	20.804	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.798	276	1338401	19.426	ng/ul	95
95) Dibenzo(a,h)anthracene	26.809	278	1104791	19.522	ng/ul	98
96) Benzo(g,h,i)perylene	27.639	276	1103083	19.461	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062823\
 Data File : BP015768.D
 Acq On : 28 Jun 2023 08:43
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771

Quant Time: Jun 28 22:02:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.815 ng/uL
RT: 3.417 min Scan# 39
Delta R.T. -0.005 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 88 Resp: 47415
Ion Ratio Lower Upper
88 100
43 29.9 27.4 41.2
58 76.0 59.8 89.8



#5
Pyridine
Concen: 20.057 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion: 79 Resp: 298209
Ion Ratio Lower Upper
79 100
52 76.0 65.5 98.3
51 37.7 31.8 47.8









#6
Benzaldehyde
Concen: 22.503 ng/ul
RT: 7.199 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 77 Resp: 157905
Ion Ratio Lower Upper
77 100
105 96.7 73.4 110.0
106 86.6 66.6 100.0



#8
Phenol
Concen: 21.253 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion: 94 Resp: 382776
Ion Ratio Lower Upper
94 100
65 30.5 26.4 39.6
66 43.2 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 21.866 ng/ul
RT: 7.475 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 93 Resp: 313335
Ion Ratio Lower Upper
93 100
63 80.6 62.8 94.2
95 31.9 25.6 38.4



#12
2-Chlorophenol
Concen: 21.077 ng/ul
RT: 7.616 min Scan# 753
Delta R.T. -0.012 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:128 Resp: 293368
Ion Ratio Lower Upper
128 100
64 53.0 41.2 61.8
130 32.1 26.0 39.0









#13
2-Methylphenol
Concen: 21.530 ng/ul
RT: 8.516 min Scan# 906
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:108 Resp: 292779
Ion Ratio Lower Upper
108 100
107 90.1 73.8 110.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.388 ng/ul
RT: 8.593 min Scan# 919
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion: 45 Resp: 457761
Ion Ratio Lower Upper
45 100
77 15.3 12.4 18.6
79 11.1 9.3 13.9









#16
Acetophenone
Concen: 21.278 ng/ul
RT: 8.899 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:105 Resp: 479576
Ion Ratio Lower Upper
105 100
77 83.2 65.0 97.4
51 31.0 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 21.586 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion: 70 Resp: 270215
Ion Ratio Lower Upper
70 100
42 59.6 48.2 72.2
101 9.4 7.6 11.4
130 19.6 13.9 20.9









#18
4-Methylphenol
Concen: 21.986 ng/ul
RT: 8.852 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:108 Resp: 321224
Ion Ratio Lower Upper
108 100
107 111.3 91.3 136.9



#19
Hexachloroethane
Concen: 20.878 ng/ul
RT: 9.140 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:117 Resp: 134218
Ion Ratio Lower Upper
117 100
201 102.6 84.2 126.2
199 63.5 51.8 77.8









#22
Nitrobenzene
Concen: 20.751 ng/ul
RT: 9.287 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 77 Resp: 389296
Ion Ratio Lower Upper
77 100
123 40.4 31.0 46.4
65 13.7 11.3 16.9



#23
Isophorone
Concen: 21.050 ng/ul
RT: 9.810 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion: 82 Resp: 755734
Ion Ratio Lower Upper
82 100
95 7.2 6.3 9.5
138 15.6 12.3 18.5









#25
2-Nitrophenol
Concen: 21.392 ng/ul
RT: 9.999 min Scan# 1159
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:139 Resp: 177019
Ion Ratio Lower Upper
139 100
65 58.5 50.2 75.2
109 35.8 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 20.511 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:107 Resp: 373115
Ion Ratio Lower Upper
107 100
121 47.9 38.2 57.2
122 81.3 62.1 93.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.108 ng/ul
 RT: 10.287 min Scan# 1207
 Delta R.T. -0.006 min
 Lab File: BP015768.D
 Acq: 28 Jun 2023 08:43




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771


Tgt Ion: 93 Resp: 448064
 Ion Ratio Lower Upper
 93 100
 95 31.5 26.6 39.8
 123 10.0 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.501 ng/ul
 RT: 10.546 min Scan# 1251
 Delta R.T. -0.006 min
 Lab File: BP015768.D
 Acq: 28 Jun 2023 08:43

Tgt Ion: 162 Resp: 297012
 Ion Ratio Lower Upper
 162 100
 164 65.5 53.0 79.6
 98 39.6 31.3 46.9









#30
Naphthalene
Concen: 20.909 ng/ul
RT: 10.934 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:128 Resp: 993632
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 22.403 ng/ul
RT: 11.063 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:127 Resp: 415436
Ion Ratio Lower Upper
127 100
129 32.7 24.7 37.1





#33

Hexachlorobutadiene

Concen: 19.195 ng/ul

RT: 11.204 min Scan# 1363

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:225 Resp: 200243

Ion Ratio Lower Upper

225 100

223 62.3 50.4 75.6

227 63.1 51.0 76.4



#34

Caprolactam

Concen: 23.532 ng/ul

RT: 11.857 min Scan# 1474

Delta R.T. 0.000 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:113 Resp: 99432


Ion Ratio Lower Upper

113 100

55 172.5 142.8 214.2

56 129.7 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 21.671 ng/ul
RT: 12.187 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:107 Resp: 350405
Ion Ratio Lower Upper
107 100
144 25.1 18.4 27.6
142 73.7 58.6 87.8



#36
2-Methylnaphthalene
Concen: 21.357 ng/ul
RT: 12.540 min Scan# 1590
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:142 Resp: 668114
Ion Ratio Lower Upper
142 100
141 89.1 73.2 109.8









#37
1-Methylnaphthalene
Concen: 21.061 ng/ul
RT: 12.757 min Scan# 1627
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:142 Resp: 672122
Ion Ratio Lower Upper
142 100
141 92.4 73.1 109.7
116 4.5 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.514 ng/ul
RT: 12.910 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:216 Resp: 374157
Ion Ratio Lower Upper
216 100
214 76.8 62.4 93.6
179 22.0 17.6 26.4
108 17.8 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 16.446 ng/ul
RT: 12.869 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:237 Resp: 89737
Ion Ratio Lower Upper
237 100
235 63.9 48.6 73.0
272 13.9 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 20.836 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:196 Resp: 240166
Ion Ratio Lower Upper
196 100
198 97.0 75.4 113.2
200 32.3 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 21.273 ng/ul

RT: 13.240 min Scan# 1710

Delta R.T. 0.000 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:196 Resp: 260083

Ion Ratio Lower Upper

196 100

198 96.6 74.9 112.3

200 31.1 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.807 ng/ul

RT: 13.540 min Scan# 1760

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:154 Resp: 912136


Ion Ratio Lower Upper

154 100

153 39.6 32.8 49.2

76 14.7 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.711 ng/ul

RT: 13.587 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:162 Resp: 715242

Ion Ratio Lower Upper

162 100

164 32.9 26.1 39.1

127 41.1 33.4 50.0



#45

2-Nitroaniline

Concen: 22.439 ng/ul

RT: 13.810 min Scan# 1806

Delta R.T. 0.000 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion: 65 Resp: 243838


Ion Ratio Lower Upper

65 100

92 67.8 49.5 74.3

138 95.3 71.4 107.2









#47
Dimethylphthalate
Concen: 21.043 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:163 Resp: 931636
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.0 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 21.573 ng/ul
RT: 14.298 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:165 Resp: 193740
Ion Ratio Lower Upper
165 100
89 65.5 54.1 81.1
121 22.1 18.2 27.2





#50

Acenaphthylene

Concen: 21.139 ng/ul

RT: 14.434 min Scan# 1912

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:152 Resp: 1120130

Ion Ratio Lower Upper

152 100

151 19.8 16.2 24.2

153 13.0 10.5 15.7



#51

3-Nitroaniline

Concen: 22.038 ng/ul

RT: 14.639 min Scan# 1947

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:138 Resp: 155199


Ion Ratio Lower Upper

138 100

108 12.5 9.1 13.7

92 133.1 96.8 145.2





#52

Acenaphthene

Concen: 21.025 ng/ul

RT: 14.769 min Scan# 1970

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:153 Resp: 772533

Ion Ratio Lower Upper

153 100

152 46.8 39.4 59.2

154 87.7 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 17.589 ng/ul

RT: 14.869 min Scan# 1986

Delta R.T. 0.000 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:184 Resp: 83802


Ion Ratio Lower Upper

184 100

63 81.2 71.4 107.0

154 70.5 61.1 91.7









#55
4-Nitrophenol
Concen: 18.275 ng/ul
RT: 14.963 min Scan# 2003
Delta R.T. 0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:109 Resp: 106911
Ion Ratio Lower Upper
109 100
139 116.0 81.3 121.9
65 134.5 99.9 149.9



#56
Dibenzofuran
Concen: 21.174 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:168 Resp: 1080058
Ion Ratio Lower Upper
168 100
139 38.9 32.2 48.4









#57
2,4-Dinitrotoluene
Concen: 22.408 ng/ul
RT: 15.098 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:165 Resp: 274230
Ion Ratio Lower Upper
165 100
63 69.4 56.8 85.2
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.083 ng/ul
RT: 15.345 min Scan# 2067
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:232 Resp: 221398
Ion Ratio Lower Upper
232 100
131 46.2 38.2 57.2
130 2.8 2.2 3.2
166 31.9 25.6 38.4









#59
Diethylphthalate
Concen: 21.507 ng/ul
RT: 15.516 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:149 Resp: 963578
Ion Ratio Lower Upper
149 100
177 21.6 16.8 25.2
150 12.5 10.3 15.5



#61
Fluorene
Concen: 21.300 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:166 Resp: 881265
Ion Ratio Lower Upper
166 100
165 97.4 78.7 118.1
167 13.0 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.606 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:204 Resp: 436241
Ion Ratio Lower Upper
204 100
206 33.5 26.0 39.0
141 63.5 51.8 77.8



#63
4-Nitroaniline
Concen: 21.856 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:138 Resp: 105344
Ion Ratio Lower Upper
138 100
92 55.6 50.1 75.1
108 89.2 85.5 128.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.368 ng/ul
RT: 15.869 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:198 Resp: 145107
Ion Ratio Lower Upper
198 100
182 4.7 3.4 5.2
77 30.0 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 20.824 ng/ul
RT: 15.963 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:169 Resp: 750608
Ion Ratio Lower Upper
169 100
168 66.2 54.2 81.4
167 36.4 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 20.239 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:248 Resp: 278273
Ion Ratio Lower Upper
248 100
250 98.2 84.6 126.8
141 77.4 67.8 101.6



#69
Hexachlorobenzene
Concen: 20.101 ng/ul
RT: 16.769 min Scan# 2309
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:284 Resp: 326117
Ion Ratio Lower Upper
284 100
142 36.9 27.8 41.8
249 31.2 26.1 39.1









#70
Atrazine
Concen: 21.072 ng/ul
RT: 16.922 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:200 Resp: 290740
Ion Ratio Lower Upper
200 100
173 26.0 19.8 29.8
215 47.0 36.8 55.2



#71
Pentachlorophenol
Concen: 20.553 ng/ul
RT: 17.128 min Scan# 2370
Delta R.T. -0.006 min
Lab File: BP015768.D
Acq: 28 Jun 2023 08:43

Tgt Ion:266 Resp: 160038
Ion Ratio Lower Upper
266 100
264 64.2 53.4 80.2
268 65.3 51.4 77.2





#72

Phenanthrene

Concen: 21.053 ng/ul

RT: 17.510 min Scan# 2435

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:178 Resp: 1377104

Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6

176 19.1 15.6 23.4



#74

Anthracene

Concen: 21.553 ng/ul

RT: 17.604 min Scan# 2451

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:178 Resp: 1416625


Ion Ratio Lower Upper

178 100

179 15.3 12.1 18.1

176 18.5 14.6 22.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.520 ng/uL

RT: 13.510 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:216 Resp: 382452

Ion Ratio Lower Upper

216 100

214 77.4 62.0 93.0

218 48.0 38.4 57.6



#76

Pentachlorobenzene

Concen: 19.460 ng/uL

RT: 15.028 min Scan# 2013

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:250 Resp: 384803


Ion Ratio Lower Upper

250 100

248 63.1 50.7 76.1

252 64.5 51.2 76.8





#77

Carbazole

Concen: 22.293 ng/ul

RT: 17.880 min Scan# 2499

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:167 Resp: 1234960

Ion Ratio Lower Upper

167 100

166 21.2 17.4 26.0

139 13.1 11.0 16.4



#78

Di-n-butylphthalate

Concen: 22.672 ng/ul

RT: 18.398 min Scan# 2586

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:149 Resp: 1665108


Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3

104 5.6 4.6 7.0





#80

Fluoranthene

Concen: 20.796 ng/ul

RT: 19.469 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:202 Resp: 1557650

Ion Ratio Lower Upper

202 100

101 11.8 8.3 12.5

100 9.3 6.6 9.8



#82

Pyrene

Concen: 20.532 ng/ul

RT: 19.822 min Scan# 2828

Delta R.T. -0.012 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:202 Resp: 1568725


Ion Ratio Lower Upper

202 100

101 14.8 9.3 13.9#

100 11.9 7.6 11.4#





#83

Butylbenzylphthalate

Concen: 23.545 ng/ul

RT: 20.669 min Scan# 2973

Delta R.T. -0.012 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:149 Resp: 733919

Ion Ratio Lower Upper

149 100

91 79.3 59.8 89.6

206 20.0 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 22.116 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:252 Resp: 461438


Ion Ratio Lower Upper

252 100

254 64.8 51.0 76.6

126 11.5 8.6 13.0





#85

Benzo(a)anthracene

Concen: 20.887 ng/ul

RT: 21.539 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:228 Resp: 1356100

Ion Ratio Lower Upper

228 100

229 19.4 15.3 22.9

226 26.8 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 24.706 ng/ul

RT: 21.421 min Scan# 3100

Delta R.T. -0.012 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:149 Resp: 1058352


Ion Ratio Lower Upper

149 100

167 26.4 21.0 31.6

279 3.9 3.8 5.6





#87

Chrysene

Concen: 20.912 ng/ul

RT: 21.598 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:228 Resp: 1288168

Ion Ratio Lower Upper

228 100

226 29.7 24.3 36.5

229 19.3 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 26.053 ng/ul




RT: 22.386 min Scan# 3264


Delta R.T. -0.012 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Tgt Ion:149 Resp: 1721124





#90

Benzo(b)fluoranthene

Concen: 20.947 ng/ul

RT: 23.315 min Scan# 3422

Delta R.T. -0.012 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:252 Resp: 1216653

Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 9.8 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 21.099 ng/ul

RT: 23.368 min Scan# 3431

Delta R.T. -0.012 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:252 Resp: 1238202


Ion Ratio Lower Upper

252 100

253 21.4 17.0 25.4

125 9.8 7.3 10.9





#93

Benzo(a)pyrene

Concen: 20.804 ng/ul

RT: 23.986 min Scan# 3536

Delta R.T. -0.018 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:252 Resp: 1055845

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 11.0 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.426 ng/ul

RT: 26.798 min Scan# 4014

Delta R.T. -0.018 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43




Tgt Ion:276 Resp: 1338401


Ion Ratio Lower Upper

276 100

138 23.1 15.7 23.5

277 25.6 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 19.522 ng/ul

RT: 26.809 min Scan# 4016

Delta R.T. -0.018 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:278 Resp: 1104791

Ion Ratio Lower Upper

278 100

139 15.7 10.5 15.7

279 22.9 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 19.461 ng/ul

RT: 27.639 min Scan# 4157

Delta R.T. -0.018 min

Lab File: BP015768.D

Acq: 28 Jun 2023 08:43

Tgt Ion:276 Resp: 1103083

Ion Ratio Lower Upper

276 100

138 20.2 15.1 22.7

277 23.1 18.8 28.2

