

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062823\
 Data File : BP015778.D
 Acq On : 28 Jun 2023 14:24
 Operator : MA/JU
 Sample : 03142-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCFS0

Quant Time: Jun 28 23:57:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 27 00:56:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	158069	20.000	ng/ul	0.00
20) Naphthalene-d8	10.892	136	593173	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.710	164	323020	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.474	188	646196	20.000	ng/ul	0.00
79) Chrysene-d12	21.562	240	619758	20.000	ng/ul	0.00
88) Perylene-d12	24.109	264	759786	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	9417	2.143	ng/uL	0.00
4) Pyridine-d5	3.810	84	364	0.033	ng/ul	-0.02
7) Phenol-d5	7.251	99	119001	8.799	ng/ul	0.02
9) Bis-(2-Chloroethyl)eth...	7.387	67	98838	11.812	ng/ul	0.00
11) 2-Chlorophenol-d4	7.598	132	116453	11.407	ng/ul	0.00
15) 4-Methylphenol-d8	8.822	113	37605	3.520	ng/ul	0.04
21) Nitrobenzene-d5	9.263	128	51134	11.280	ng/ul	0.02
24) 2-Nitrophenol-d4	9.992	143	51131	9.664	ng/ul	0.02
28) 2,4-Dichlorophenol-d3	10.586	165	68173	7.231	ng/ul	0.06
31) 4-Chloroaniline-d4	11.116	131	13199	1.090	ng/ul	0.08
46) Dimethylphthalate-d6	14.127	166	301568	12.079	ng/ul	0.00
49) Acenaphthylene-d8	14.557	160	731	0.026	ng/ul	0.15
54) 4-Nitrophenol-d4	14.945	143	140	0.042	ng/ul	0.00
60) Fluorene-d10	15.716	176	253645	12.338	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.845	200	41	0.010	ng/ul	-0.01
73) Anthracene-d10	17.580	188	374699	12.694	ng/ul	0.00
81) Pyrene-d10	19.798	212	475027	11.738	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.939	264	486071	12.682	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062823\
Data File : BP015778.D
Acq On : 28 Jun 2023 14:24
Operator : MA/JU
Sample : 03142-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCFS0

Quant Time: Jun 28 23:57:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 27 00:56:58 2023
Response via : Initial Calibration

