

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063023\
 Data File : BP015886.D
 Acq On : 01 Jul 2023 15:58
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020780

Quant Time: Jul 02 02:42:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 02:29:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	286473	20.000	ng/u1	0.00
20) Naphthalene-d8	10.875	136	1180917	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.704	164	708307	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.469	188	1557167	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.557	240	1395279	20.000	ng/u1	# 0.00
88) Perylene-d12	24.110	264	1649035	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	64396	8.088	ng/uL	0.00
4) Pyridine-d5	3.817	84	426678	21.258	ng/u1	0.00
7) Phenol-d5	7.228	99	534371	21.801	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.375	67	345152	22.761	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	401640	21.707	ng/u1	0.00
15) 4-Methylphenol-d8	8.787	113	401763	20.749	ng/u1	0.00
21) Nitrobenzene-d5	9.234	128	194262	21.525	ng/u1	0.00
24) 2-Nitrophenol-d4	9.963	143	218171	20.713	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	390829	20.822	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	549626	22.795	ng/u1	0.00
46) Dimethylphthalate-d6	14.110	166	1084002	19.800	ng/u1	0.00
49) Acenaphthylene-d8	14.398	160	1296043	21.059	ng/u1	0.00
54) 4-Nitrophenol-d4	14.969	143	142256	19.690	ng/u1	0.00
60) Fluorene-d10	15.704	176	949715	21.068	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	150648	15.731	ng/u1	0.00
73) Anthracene-d10	17.569	188	1525797	21.451	ng/u1	0.00
81) Pyrene-d10	19.798	212	1724330	18.926	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.939	264	1688440	20.297	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.411	88	71551	8.350	ng/uL	98
5) Pyridine	3.840	79	445473	21.216	ng/u1	95
6) Benzaldehyde	7.193	77	225489	22.754	ng/u1	95
8) Phenol	7.258	94	558019	21.938	ng/u1#	88
10) Bis(2-Chloroethyl)ether	7.469	93	450522	22.262	ng/u1	100
12) 2-Chlorophenol	7.611	128	426466	21.695	ng/u1	98
13) 2-Methylphenol	8.516	108	409774	21.337	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.581	45	656501	23.751	ng/u1	98
16) Acetophenone	8.893	105	652879	20.511	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.869	70	350628	19.833	ng/u1	97
18) 4-Methylphenol	8.852	108	442702	21.455	ng/u1	98
19) Hexachloroethane	9.128	117	185887	20.475	ng/u1	90
22) Nitrobenzene	9.281	77	527599	20.817	ng/u1	96
23) Isophorone	9.805	82	996009	20.536	ng/u1	98
25) 2-Nitrophenol	9.999	139	236395	21.147	ng/u1	93
26) 2,4-Dimethylphenol	10.057	107	483117	19.659	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.281	93	593204	21.666	ng/u1	97
29) 2,4-Dichlorophenol	10.546	162	392181	21.016	ng/u1	99
30) Naphthalene	10.928	128	1353404	21.082	ng/u1	100
32) 4-Chloroaniline	11.063	127	563730	22.503	ng/u1	100
33) Hexachlorobutadiene	11.193	225	252554	17.921	ng/u1	99
34) Caprolactam	11.863	113	115396	20.216	ng/u1	96
35) 4-Chloro-3-methylphenol	12.199	107	443789	20.317	ng/u1	96
36) 2-Methylnaphthalene	12.534	142	875334	20.713	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063023\
 Data File : BP015886.D
 Acq On : 01 Jul 2023 15:58
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020780

Quant Time: Jul 02 02:42:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 02:29:46 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	877986	20.365	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.910	216	465408	19.939	ng/ul	99
40) Hexachlorocyclopentadiene	12.863	237	51819	7.421	ng/ul	97
41) 2,4,6-Trichlorophenol	13.157	196	305451	20.707	ng/ul	95
42) 2,4,5-Trichlorophenol	13.246	196	334976	21.409	ng/ul	97
43) 1,1'-Biphenyl	13.540	154	1167096	20.803	ng/ul	98
44) 2-Chloronaphthalene	13.587	162	919693	20.809	ng/ul	99
45) 2-Nitroaniline	13.816	65	307634	22.122	ng/ul	90
47) Dimethylphthalate	14.157	163	1124523	19.848	ng/ul	98
48) 2,6-Dinitrotoluene	14.304	165	239611	20.849	ng/ul	96
50) Acenaphthylene	14.428	152	1451577	21.406	ng/ul	99
51) 3-Nitroaniline	14.651	138	228632	25.368	ng/ul	95
52) Acenaphthene	14.769	153	996376	21.189	ng/ul	97
53) 2,4-Dinitrophenol	14.887	184	85868	14.083	ng/ul	88
55) 4-Nitrophenol	14.987	109	111367	14.875	ng/ul#	84
56) Dibenzofuran	15.104	168	1372764	21.030	ng/ul	96
57) 2,4-Dinitrotoluene	15.110	165	336085	21.459	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.351	232	275941	20.533	ng/ul	98
59) Diethylphthalate	15.516	149	1173403	20.466	ng/ul	99
61) Fluorene	15.757	166	1122057	21.192	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.745	204	536039	19.786	ng/ul	96
63) 4-Nitroaniline	15.822	138	185329	30.045	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.881	198	154869	15.984	ng/ul#	100
67) N-Nitrosodiphenylamine	15.963	169	943917	20.249	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	336232	18.909	ng/ul	90
69) Hexachlorobenzene	16.769	284	393564	18.758	ng/ul	95
70) Atrazine	16.922	200	354519	19.868	ng/ul	98
71) Pentachlorophenol	17.134	266	206685	20.525	ng/ul	99
72) Phenanthrene	17.510	178	1786424	21.118	ng/ul	99
74) Anthracene	17.604	178	1831534	21.547	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.510	216	473408	18.684	ng/ul	99
76) Pentachlorobenzene	15.028	250	464226	18.153	ng/ul	99
77) Carbazole	17.886	167	1692850	23.628	ng/ul	99
78) Di-n-butylphthalate	18.398	149	2181736	22.970	ng/ul	99
80) Fluoranthene	19.475	202	2147263	18.964	ng/ul#	94
82) Pyrene	19.828	202	2218419	19.207	ng/ul#	92
83) Butylbenzylphthalate	20.669	149	1044980	22.176	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.474	252	591098	18.740	ng/ul	99
85) Benzo(a)anthracene	21.539	228	2100837	21.404	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.422	149	1525704	23.559	ng/ul	99
87) Chrysene	21.598	228	1968113	21.135	ng/ul	99
89) Di-n-octyl phthalate	22.386	149	2658255	22.058	ng/ul	100
90) Benzo(b)fluoranthene	23.374	252	2159302	20.379	ng/ul	98
91) Benzo(k)fluoranthene	23.374	252	2159264	20.169	ng/ul#	97
93) Benzo(a)pyrene	23.998	252	1909001	20.619	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.815	276	2550644	20.294	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.821	278	2105858	20.399	ng/ul#	95
96) Benzo(g,h,i)perylene	27.657	276	2084855	20.163	ng/ul#	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063023\
 Data File : BP015886.D
 Acq On : 01 Jul 2023 15:58
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020780

Quant Time: Jul 02 02:42:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Jul 02 02:29:46 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.350 ng/uL
 RT: 3.411 min Scan# 38
 Delta R.T. 0.000 min
 Lab File: BP015886.D
 Acq: 01 Jul 2023 15:58




Instrument :
 BNA_P
 ClientSampleId :
 SST020780


Tgt Ion:	88	Resp:	71551
Ion	Ratio	Lower	Upper
88	100		
43	34.0	27.4	41.2
58	77.4	59.8	89.8



#5
 Pyridine
 Concen: 21.216 ng/uL
 RT: 3.840 min Scan# 111
 Delta R.T. 0.000 min
 Lab File: BP015886.D
 Acq: 01 Jul 2023 15:58

Tgt Ion:	79	Resp:	445473
Ion	Ratio	Lower	Upper
79	100		
52	77.0	65.5	98.3
51	37.9	31.8	47.8









#6
Benzaldehyde
Concen: 22.754 ng/ul
RT: 7.193 min Scan# 681
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion: 77 Resp: 225489
Ion Ratio Lower Upper
77 100
105 93.3 73.4 110.0
106 90.3 66.6 100.0



#8
Phenol
Concen: 21.938 ng/ul
RT: 7.258 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion: 94 Resp: 558019
Ion Ratio Lower Upper
94 100
65 28.0 26.4 39.6
66 37.1 37.3 55.9#









#10
Bis(2-Chloroethyl)ether
Concen: 22.262 ng/ul
RT: 7.469 min Scan# 728
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion: 93 Resp: 450522
Ion Ratio Lower Upper
93 100
63 78.5 62.8 94.2
95 31.3 25.6 38.4



#12
2-Chlorophenol
Concen: 21.695 ng/ul
RT: 7.611 min Scan# 752
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:128 Resp: 426466
Ion Ratio Lower Upper
128 100
64 52.5 41.2 61.8
130 31.0 26.0 39.0





#13

2-Methylphenol

Concen: 21.337 ng/ul

RT: 8.516 min Scan# 905

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :





SSTD020780

Tgt Ion:108 Resp: 409774

Ion Ratio Lower Upper

108 100

107 89.9 73.8 110.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.751 ng/ul

RT: 8.581 min Scan# 917

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion: 45 Resp: 656501


Ion Ratio Lower Upper

45 100

77 14.8 12.4 18.6

79 11.0 9.3 13.9









#16
Acetophenone
Concen: 20.511 ng/ul
RT: 8.893 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:105 Resp: 652879
Ion Ratio Lower Upper
105 100
77 83.4 65.0 97.4
51 30.2 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 19.833 ng/ul
RT: 8.869 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion: 70 Resp: 350628
Ion Ratio Lower Upper
70 100
42 57.2 48.2 72.2
101 9.3 7.6 11.4
130 18.1 13.9 20.9









#18
4-Methylphenol
Concen: 21.455 ng/ul
RT: 8.852 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:108 Resp: 442702
Ion Ratio Lower Upper
108 100
107 112.5 91.3 136.9



#19
Hexachloroethane
Concen: 20.475 ng/ul
RT: 9.128 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:117 Resp: 185887
Ion Ratio Lower Upper
117 100
201 93.5 84.2 126.2
199 58.3 51.8 77.8





#22

Nitrobenzene

Concen: 20.817 ng/ul

RT: 9.281 min Scan# 1036

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion: 77 Resp: 527599

Ion Ratio Lower Upper

77 100

123 41.5 31.0 46.4

65 13.1 11.3 16.9



#23

Isophorone

Concen: 20.536 ng/ul

RT: 9.805 min Scan# 1125

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion: 82 Resp: 996009


Ion Ratio Lower Upper

82 100

95 6.5 6.3 9.5

138 15.7 12.3 18.5









#25
2-Nitrophenol
Concen: 21.147 ng/ul
RT: 9.999 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:139 Resp: 236395
Ion Ratio Lower Upper
139 100
65 59.1 50.2 75.2
109 32.6 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 19.659 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:107 Resp: 483117
Ion Ratio Lower Upper
107 100
121 49.8 38.2 57.2
122 82.2 62.1 93.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.666 ng/ul
 RT: 10.281 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BP015886.D
 Acq: 01 Jul 2023 15:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020780


Tgt Ion: 93 Resp: 593204
 Ion Ratio Lower Upper
 93 100
 95 31.6 26.6 39.8
 123 10.2 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.016 ng/ul
 RT: 10.546 min Scan# 1251
 Delta R.T. 0.000 min
 Lab File: BP015886.D
 Acq: 01 Jul 2023 15:58

Tgt Ion:162 Resp: 392181
 Ion Ratio Lower Upper
 162 100
 164 66.1 53.0 79.6
 98 40.2 31.3 46.9









#30
Naphthalene
Concen: 21.082 ng/ul
RT: 10.928 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:128 Resp: 1353404
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.8 10.3 15.5



#32
4-Chloroaniline
Concen: 22.503 ng/ul
RT: 11.063 min Scan# 1339
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:127 Resp: 563730
Ion Ratio Lower Upper
127 100
129 30.8 24.7 37.1





#33

Hexachlorobutadiene

Concen: 17.921 ng/ul

RT: 11.193 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:225 Resp: 252554

Ion Ratio Lower Upper

225 100

223 64.8 50.4 75.6

227 63.4 51.0 76.4



#34

Caprolactam

Concen: 20.216 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:113 Resp: 115396


Ion Ratio Lower Upper

113 100

55 189.2 142.8 214.2

56 145.5 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 20.317 ng/ul
RT: 12.199 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:107 Resp: 443789
Ion Ratio Lower Upper
107 100
144 24.5 18.4 27.6
142 76.3 58.6 87.8



#36
2-Methylnaphthalene
Concen: 20.713 ng/ul
RT: 12.534 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:142 Resp: 875334
Ion Ratio Lower Upper
142 100
141 88.6 73.2 109.8





#37

1-Methylnaphthalene

Concen: 20.365 ng/ul

RT: 12.757 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:142 Resp: 877986

Ion Ratio Lower Upper

142 100

141 92.3 73.1 109.7

116 4.1 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.939 ng/ul

RT: 12.910 min Scan# 1653

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Tgt Ion:216 Resp: 465408




Ion Ratio Lower Upper


216 100

214 78.0 62.4 93.6

179 20.3 17.6 26.4

108 18.6 15.5 23.3





#40

Hexachlorocyclopentadiene

Concen: 7.421 ng/ul

RT: 12.863 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:237 Resp: 51819

Ion Ratio Lower Upper

237 100

235 63.3 48.6 73.0

272 13.3 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 20.707 ng/ul

RT: 13.157 min Scan# 1695

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:196 Resp: 305451


Ion Ratio Lower Upper

196 100

198 98.9 75.4 113.2

200 31.7 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 21.409 ng/ul

RT: 13.246 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:196 Resp: 334976

Ion Ratio Lower Upper

196 100

198 97.7 74.9 112.3

200 31.0 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.803 ng/ul

RT: 13.540 min Scan# 1760

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:154 Resp: 1167096


Ion Ratio Lower Upper

154 100

153 40.4 32.8 49.2

76 15.4 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.809 ng/ul

RT: 13.587 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:162 Resp: 919693

Ion Ratio Lower Upper

162 100

164 32.7 26.1 39.1

127 40.5 33.4 50.0



#45

2-Nitroaniline

Concen: 22.122 ng/ul

RT: 13.816 min Scan# 1807

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion: 65 Resp: 307634


Ion Ratio Lower Upper

65 100

92 69.6 49.5 74.3

138 98.0 71.4 107.2









#47
Dimethylphthalate
Concen: 19.848 ng/ul
RT: 14.157 min Scan# 1865
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:163 Resp: 1124523
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.5 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 20.849 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:165 Resp: 239611
Ion Ratio Lower Upper
165 100
89 64.3 54.1 81.1
121 21.4 18.2 27.2





#50

Acenaphthylene

Concen: 21.406 ng/ul

RT: 14.428 min Scan# 1910

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:152 Resp: 1451577

Ion Ratio Lower Upper

152 100

151 20.0 16.2 24.2

153 13.4 10.5 15.7



#51

3-Nitroaniline

Concen: 25.368 ng/ul

RT: 14.651 min Scan# 1949

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:138 Resp: 228632


Ion Ratio Lower Upper

138 100

108 11.4 9.1 13.7

92 127.1 96.8 145.2





#52

Acenaphthene

Concen: 21.189 ng/ul

RT: 14.769 min Scan# 1968

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:153 Resp: 996376

Ion Ratio Lower Upper

153 100

152 47.8 39.4 59.2

154 86.4 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 14.083 ng/ul

RT: 14.887 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:184 Resp: 85868


Ion Ratio Lower Upper

184 100

63 77.9 71.4 107.0

154 65.7 61.1 91.7









#55
4-Nitrophenol
Concen: 14.875 ng/ul
RT: 14.987 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:109 Resp: 111367
Ion Ratio Lower Upper
109 100
139 124.3 81.3 121.9#
65 136.9 99.9 149.9



#56
Dibenzofuran
Concen: 21.030 ng/ul
RT: 15.104 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:168 Resp: 1372764
Ion Ratio Lower Upper
168 100
139 37.7 32.2 48.4









#57
2,4-Dinitrotoluene
Concen: 21.459 ng/ul
RT: 15.110 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:165 Resp: 336085
Ion Ratio Lower Upper
165 100
63 74.2 56.8 85.2
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.533 ng/ul
RT: 15.351 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:232 Resp: 275941
Ion Ratio Lower Upper
232 100
131 46.1 38.2 57.2
130 2.6 2.2 3.2
166 30.9 25.6 38.4









#59
Diethylphthalate
Concen: 20.466 ng/ul
RT: 15.516 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:149 Resp: 1173403
Ion Ratio Lower Upper
149 100
177 21.8 16.8 25.2
150 12.7 10.3 15.5



#61
Fluorene
Concen: 21.192 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:166 Resp: 1122057
Ion Ratio Lower Upper
166 100
165 97.8 78.7 118.1
167 13.3 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 19.786 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:204 Resp: 536039
Ion Ratio Lower Upper
204 100
206 33.2 26.0 39.0
141 60.1 51.8 77.8



#63
4-Nitroaniline
Concen: 30.045 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:138 Resp: 185329
Ion Ratio Lower Upper
138 100
92 54.5 50.1 75.1
108 79.8 85.5 128.3#









#66
4,6-Dinitro-2-methylphenol
Concen: 15.984 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:198 Resp: 154869
Ion Ratio Lower Upper
198 100
182 5.3 3.4 5.2#
77 30.5 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 20.249 ng/ul
RT: 15.963 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:169 Resp: 943917
Ion Ratio Lower Upper
169 100
168 67.0 54.2 81.4
167 36.9 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 18.909 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:248 Resp: 336232
Ion Ratio Lower Upper
248 100
250 97.8 84.6 126.8
141 73.2 67.8 101.6



#69
Hexachlorobenzene
Concen: 18.758 ng/ul
RT: 16.769 min Scan# 2309
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:284 Resp: 393564
Ion Ratio Lower Upper
284 100
142 39.7 27.8 41.8
249 31.4 26.1 39.1









#70
Atrazine
Concen: 19.868 ng/ul
RT: 16.922 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:200 Resp: 354519
Ion Ratio Lower Upper
200 100
173 26.5 19.8 29.8
215 46.3 36.8 55.2



#71
Pentachlorophenol
Concen: 20.525 ng/ul
RT: 17.134 min Scan# 2371
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:266 Resp: 206685
Ion Ratio Lower Upper
266 100
264 64.8 53.4 80.2
268 64.2 51.4 77.2





#72

Phenanthrene

Concen: 21.118 ng/ul

RT: 17.510 min Scan# 2434

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:178 Resp: 1786424

Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 19.3 15.6 23.4



#74

Anthracene

Concen: 21.547 ng/ul

RT: 17.604 min Scan# 2451

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:178 Resp: 1831534


Ion Ratio Lower Upper

178 100

179 15.5 12.1 18.1

176 18.7 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.684 ng/uL
RT: 13.510 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:216 Resp: 473408
Ion Ratio Lower Upper
216 100
214 76.5 62.0 93.0
218 47.5 38.4 57.6



#76
Pentachlorobenzene
Concen: 18.153 ng/uL
RT: 15.028 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:250 Resp: 464226
Ion Ratio Lower Upper
250 100
248 63.2 50.7 76.1
252 63.3 51.2 76.8









#77
Carbazole
Concen: 23.628 ng/ul
RT: 17.886 min Scan# 2498
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58




Instrument :
BNA_P
ClientSampleId :
SSTD020780


Tgt Ion:167 Resp: 1692850
Ion Ratio Lower Upper
167 100
166 21.8 17.4 26.0
139 13.3 11.0 16.4



#78
Di-n-butylphthalate
Concen: 22.970 ng/ul
RT: 18.398 min Scan# 2586
Delta R.T. 0.000 min
Lab File: BP015886.D
Acq: 01 Jul 2023 15:58

Tgt Ion:149 Resp: 2181736
Ion Ratio Lower Upper
149 100
150 9.0 7.5 11.3
104 5.3 4.6 7.0





#80

Fluoranthene

Concen: 18.964 ng/ul

RT: 19.475 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:202 Resp: 2147263

Ion Ratio Lower Upper

202 100

101 13.1 8.3 12.5#

100 10.0 6.6 9.8#



#82

Pyrene

Concen: 19.207 ng/ul

RT: 19.828 min Scan# 2829

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:202 Resp: 2218419


Ion Ratio Lower Upper

202 100

101 15.0 9.3 13.9#

100 12.3 7.6 11.4#





#83

Butylbenzylphthalate

Concen: 22.176 ng/ul

RT: 20.669 min Scan# 2971

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:149 Resp: 1044980

Ion Ratio Lower Upper

149 100

91 78.7 59.8 89.6

206 19.0 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 18.740 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:252 Resp: 591098


Ion Ratio Lower Upper

252 100

254 64.4 51.0 76.6

126 12.1 8.6 13.0





#85

Benzo(a)anthracene

Concen: 21.404 ng/ul

RT: 21.539 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:228 Resp: 2100837

Ion Ratio Lower Upper

228 100

229 20.1 15.3 22.9

226 27.4 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.559 ng/ul

RT: 21.422 min Scan# 3100

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:149 Resp: 1525704


Ion Ratio Lower Upper

149 100

167 26.9 21.0 31.6

279 4.0 3.8 5.6





#87

Chrysene

Concen: 21.135 ng/ul

RT: 21.598 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:228 Resp: 1968113

Ion Ratio Lower Upper

228 100

226 30.8 24.3 36.5

229 19.9 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 22.058 ng/ul




RT: 22.386 min Scan# 3264


Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Tgt Ion:149 Resp: 2658255





#90

Benzo(b)fluoranthene

Concen: 20.379 ng/ul

RT: 23.374 min Scan# 3421

Delta R.T. 0.053 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:252 Resp: 2159302

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 11.3 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 20.169 ng/ul

RT: 23.374 min Scan# 3432

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:252 Resp: 2159264


Ion Ratio Lower Upper

252 100

253 21.8 17.0 25.4

125 11.3 7.3 10.9#





#93

Benzo(a)pyrene

Concen: 20.619 ng/ul

RT: 23.998 min Scan# 3538

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:252 Resp: 1909001

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 12.8 8.2 12.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.294 ng/ul

RT: 26.815 min Scan# 4017

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58




Tgt Ion:276 Resp: 2550644


Ion Ratio Lower Upper

276 100

138 27.0 15.7 23.5#

277 25.2 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 20.399 ng/ul

RT: 26.821 min Scan# 410

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Instrument :

BNA_P

ClientSampleId :

SSTD020780





Tgt Ion:278 Resp: 2105858

Ion Ratio Lower Upper

278 100

139 17.5 10.5 15.7#

279 23.9 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 20.163 ng/ul

RT: 27.657 min Scan# 4160

Delta R.T. 0.000 min

Lab File: BP015886.D

Acq: 01 Jul 2023 15:58

Tgt Ion:276 Resp: 2084855

Ion Ratio Lower Upper

276 100

138 23.7 15.1 22.7#

277 23.7 18.8 28.2

