

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070623\
 Data File : BP016025.D
 Acq On : 06 Jul 2023 09:50
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Jul 06 23:23:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 06 23:18:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	203727	20.000	ng/u1	0.00
20) Naphthalene-d8	10.852	136	925368	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.675	164	585247	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	1299801	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.539	240	1194495	20.000	ng/u1	0.00
88) Perylene-d12	24.080	264	1208405	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	46415	8.197	ng/uL	0.00
4) Pyridine-d5	3.805	84	297900	20.870	ng/u1	0.00
7) Phenol-d5	7.216	99	386693	22.184	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.358	67	255896	23.729	ng/u1	0.00
11) 2-Chlorophenol-d4	7.563	132	288423	21.920	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	308040	22.370	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	149986	21.208	ng/u1	0.00
24) 2-Nitrophenol-d4	9.946	143	175622	21.277	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.504	165	301203	20.478	ng/u1	0.00
31) 4-Chloroaniline-d4	11.022	131	448022	23.712	ng/u1	0.00
46) Dimethylphthalate-d6	14.087	166	975704	21.570	ng/u1	0.00
49) Acenaphthylene-d8	14.375	160	1096474	21.563	ng/u1	0.00
54) 4-Nitrophenol-d4	14.963	143	82145	13.761	ng/u1	0.00
60) Fluorene-d10	15.675	176	800892	21.502	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	162523	20.331	ng/u1	0.00
73) Anthracene-d10	17.545	188	1289347	21.716	ng/u1	0.00
81) Pyrene-d10	19.775	212	1474481	18.904	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.910	264	1271783	20.863	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	50494	8.286	ng/uL	98
5) Pyridine	3.822	79	313326	20.983	ng/u1	98
6) Benzaldehyde	7.187	77	102497	14.544	ng/u1	96
8) Phenol	7.246	94	413059	22.835	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.452	93	341677	23.741	ng/u1	98
12) 2-Chlorophenol	7.593	128	307970	22.030	ng/u1	97
13) 2-Methylphenol	8.499	108	317770	23.267	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.563	45	523782	26.646	ng/u1	98
16) Acetophenone	8.875	105	532337	23.517	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.846	70	300582	23.908	ng/u1	98
18) 4-Methylphenol	8.840	108	350940	23.916	ng/u1	97
19) Hexachloroethane	9.099	117	137922	21.362	ng/u1	93
22) Nitrobenzene	9.263	77	425486	21.425	ng/u1	95
23) Isophorone	9.787	82	853338	22.454	ng/u1	98
25) 2-Nitrophenol	9.975	139	188581	21.528	ng/u1#	90
26) 2,4-Dimethylphenol	10.040	107	390824	20.296	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.263	93	499316	23.274	ng/u1	99
29) 2,4-Dichlorophenol	10.534	162	309341	21.155	ng/u1	99
30) Naphthalene	10.899	128	1074394	21.358	ng/u1	100
32) 4-Chloroaniline	11.046	127	471677	24.028	ng/u1	97
33) Hexachlorobutadiene	11.163	225	193079	17.484	ng/u1	99
34) Caprolactam	11.875	113	115976	25.928	ng/u1	93
35) 4-Chloro-3-methylphenol	12.181	107	367218	21.454	ng/u1	97
36) 2-Methylnaphthalene	12.516	142	707904	21.377	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070623\
 Data File : BP016025.D
 Acq On : 06 Jul 2023 09:50
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Jul 06 23:23:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 06 23:18:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.728	142	722144	21.376	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.887	216	374606	19.424	ng/ul	99
40) Hexachlorocyclopentadiene	12.834	237	76122	13.194	ng/ul	97
41) 2,4,6-Trichlorophenol	13.140	196	248819	20.415	ng/ul	99
42) 2,4,5-Trichlorophenol	13.234	196	274382	21.224	ng/ul	99
43) 1,1'-Biphenyl	13.516	154	975260	21.039	ng/ul	97
44) 2-Chloronaphthalene	13.563	162	760745	20.832	ng/ul	99
45) 2-Nitroaniline	13.798	65	266082	23.157	ng/ul	99
47) Dimethylphthalate	14.140	163	1007682	21.525	ng/ul	99
48) 2,6-Dinitrotoluene	14.287	165	207517	21.853	ng/ul	98
50) Acenaphthylene	14.404	152	1217049	21.721	ng/ul	99
51) 3-Nitroaniline	14.640	138	149745	20.109	ng/ul	95
52) Acenaphthene	14.740	153	840980	21.645	ng/ul	99
53) 2,4-Dinitrophenol	14.881	184	98289	19.509	ng/ul	85
55) 4-Nitrophenol	14.981	109	128670	20.800	ng/ul	85
56) Dibenzofuran	15.081	168	1164427	21.589	ng/ul	97
57) 2,4-Dinitrotoluene	15.092	165	296529	22.915	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.328	232	232362	20.925	ng/ul	98
59) Diethylphthalate	15.492	149	1048643	22.135	ng/ul	99
61) Fluorene	15.734	166	956310	21.859	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.716	204	464595	20.754	ng/ul	95
63) 4-Nitroaniline	15.816	138	99477	19.518	ng/ul#	83
66) 4,6-Dinitro-2-methylph...	15.869	198	165270	20.434	ng/ul#	98
67) N-Nitrosodiphenylamine	15.939	169	816857	20.993	ng/ul	99
68) 4-Bromophenyl-phenylether	16.616	248	287946	19.400	ng/ul	93
69) Hexachlorobenzene	16.745	284	336827	19.232	ng/ul	98
70) Atrazine	16.904	200	301856	20.266	ng/ul	96
71) Pentachlorophenol	17.116	266	149009	17.727	ng/ul	97
72) Phenanthrene	17.486	178	1535983	21.752	ng/ul	100
74) Anthracene	17.581	178	1567898	22.098	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.487	216	387367	18.315	ng/uL	98
76) Pentachlorobenzene	15.004	250	396772	18.587	ng/uL	98
77) Carbazole	17.869	167	1445853	24.177	ng/ul	99
78) Di-n-butylphthalate	18.369	149	1894708	23.898	ng/ul	99
80) Fluoranthene	19.451	202	1820217	18.777	ng/ul#	95
82) Pyrene	19.810	202	1897594	19.190	ng/ul#	92
83) Butylbenzylphthalate	20.645	149	890344	22.070	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.451	252	580083	21.483	ng/ul	99
85) Benzo(a)anthracene	21.522	228	1740694	20.716	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.392	149	1305381	23.546	ng/ul#	99
87) Chrysene	21.580	228	1729359	21.693	ng/ul	99
89) Di-n-octyl phthalate	22.351	149	2163678	24.500	ng/ul	100
90) Benzo(b)fluoranthene	23.292	252	1618326	20.843	ng/ul#	97
91) Benzo(k)fluoranthene	23.339	252	1700248	21.673	ng/ul#	97
93) Benzo(a)pyrene	23.963	252	1419912	20.929	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.774	276	1564359	16.985	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.780	278	1308627	17.298	ng/ul#	95
96) Benzo(g,h,i)perylene	27.621	276	1229932	16.232	ng/ul#	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.286 ng/uL
RT: 3.393 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:	88	Resp:	50494
Ion	Ratio	Lower	Upper
88	100		
43	33.6	27.4	41.2
58	72.9	59.8	89.8



#5
Pyridine
Concen: 20.983 ng/uL
RT: 3.822 min Scan# 108
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:	79	Resp:	313326
Ion	Ratio	Lower	Upper
79	100		
52	83.2	65.5	98.3
51	37.1	31.8	47.8









#6
Benzaldehyde
Concen: 14.544 ng/ul
RT: 7.187 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 77 Resp: 102497
Ion Ratio Lower Upper
77 100
105 86.5 73.4 110.0
106 84.9 66.6 100.0



#8
Phenol
Concen: 22.835 ng/ul
RT: 7.246 min Scan# 690
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion: 94 Resp: 413059
Ion Ratio Lower Upper
94 100
65 31.4 26.4 39.6
66 39.2 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 23.741 ng/ul
RT: 7.452 min Scan# 725
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 93 Resp: 341677
Ion Ratio Lower Upper
93 100
63 80.1 62.8 94.2
95 32.6 25.6 38.4



#12
2-Chlorophenol
Concen: 22.030 ng/ul
RT: 7.593 min Scan# 749
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:128 Resp: 307970
Ion Ratio Lower Upper
128 100
64 54.6 41.2 61.8
130 32.6 26.0 39.0









#13
2-Methylphenol
Concen: 23.267 ng/ul
RT: 8.499 min Scan# 903
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:108 Resp: 317770
Ion Ratio Lower Upper
108 100
107 90.0 73.8 110.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 26.646 ng/ul
RT: 8.563 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion: 45 Resp: 523782
Ion Ratio Lower Upper
45 100
77 14.8 12.4 18.6
79 11.1 9.3 13.9









#16
Acetophenone
Concen: 23.517 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:105 Resp: 532337
Ion Ratio Lower Upper
105 100
77 84.1 65.0 97.4
51 30.8 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 23.908 ng/ul
RT: 8.846 min Scan# 962
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion: 70 Resp: 300582
Ion Ratio Lower Upper
70 100
42 61.6 48.2 72.2
101 9.3 7.6 11.4
130 18.1 13.9 20.9









#18
4-Methylphenol
Concen: 23.916 ng/ul
RT: 8.840 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:108 Resp: 350940
Ion Ratio Lower Upper
108 100
107 110.4 91.3 136.9



#19
Hexachloroethane
Concen: 21.362 ng/ul
RT: 9.099 min Scan# 1005
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:117 Resp: 137922
Ion Ratio Lower Upper
117 100
201 97.6 84.2 126.2
199 60.8 51.8 77.8









#22
Nitrobenzene
Concen: 21.425 ng/ul
RT: 9.263 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 77 Resp: 425486
Ion Ratio Lower Upper
77 100
123 42.3 31.0 46.4
65 13.6 11.3 16.9



#23
Isophorone
Concen: 22.454 ng/ul
RT: 9.787 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion: 82 Resp: 853338
Ion Ratio Lower Upper
82 100
95 6.7 6.3 9.5
138 15.0 12.3 18.5









#25
2-Nitrophenol
Concen: 21.528 ng/ul
RT: 9.975 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:139 Resp: 188581
Ion Ratio Lower Upper
139 100
65 57.1 50.2 75.2
109 29.9 31.3 46.9#



#26
2,4-Dimethylphenol
Concen: 20.296 ng/ul
RT: 10.040 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:107 Resp: 390824
Ion Ratio Lower Upper
107 100
121 49.5 38.2 57.2
122 82.3 62.1 93.1









#27
Bis(2-Chloroethoxy)methane
Concen: 23.274 ng/ul
RT: 10.263 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 93 Resp: 499316
Ion Ratio Lower Upper
93 100
95 32.5 26.6 39.8
123 10.9 8.6 13.0



#29
2,4-Dichlorophenol
Concen: 21.155 ng/ul
RT: 10.534 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion: 162 Resp: 309341
Ion Ratio Lower Upper
162 100
164 66.6 53.0 79.6
98 38.4 31.3 46.9









#30
Naphthalene
Concen: 21.358 ng/ul
RT: 10.899 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:128 Resp: 1074394
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 24.028 ng/ul
RT: 11.046 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:127 Resp: 471677
Ion Ratio Lower Upper
127 100
129 32.8 24.7 37.1









#33
Hexachlorobutadiene
Concen: 17.484 ng/ul
RT: 11.163 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:225 Resp: 193079
Ion Ratio Lower Upper
225 100
223 63.0 50.4 75.6
227 62.8 51.0 76.4



#34
Caprolactam
Concen: 25.928 ng/ul
RT: 11.875 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:113 Resp: 115976
Ion Ratio Lower Upper
113 100
55 176.2 142.8 214.2
56 130.3 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 21.454 ng/ul
RT: 12.181 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:107 Resp: 367218
Ion Ratio Lower Upper
107 100
144 25.2 18.4 27.6
142 75.6 58.6 87.8



#36
2-Methylnaphthalene
Concen: 21.377 ng/ul
RT: 12.516 min Scan# 1586
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:142 Resp: 707904
Ion Ratio Lower Upper
142 100
141 89.7 73.2 109.8









#37
1-Methylnaphthalene
Concen: 21.376 ng/ul
RT: 12.728 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:142 Resp: 722144
Ion Ratio Lower Upper
142 100
141 92.3 73.1 109.7
116 3.9 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.424 ng/ul
RT: 12.887 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:216 Resp: 374606
Ion Ratio Lower Upper
216 100
214 79.0 62.4 93.6
179 20.7 17.6 26.4
108 19.0 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 13.194 ng/ul
RT: 12.834 min Scan# 1640
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:237 Resp: 76122
Ion Ratio Lower Upper
237 100
235 63.1 48.6 73.0
272 12.8 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 20.415 ng/ul
RT: 13.140 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:196 Resp: 248819
Ion Ratio Lower Upper
196 100
198 94.5 75.4 113.2
200 30.6 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 21.224 ng/ul
RT: 13.234 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:196 Resp: 274382
Ion Ratio Lower Upper
196 100
198 94.6 74.9 112.3
200 30.8 24.3 36.5



#43
1,1'-Biphenyl
Concen: 21.039 ng/ul
RT: 13.516 min Scan# 1756
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:154 Resp: 975260
Ion Ratio Lower Upper
154 100
153 39.6 32.8 49.2
76 15.2 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.832 ng/ul

RT: 13.563 min Scan# 1764

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:162 Resp: 760745

Ion Ratio Lower Upper

162 100

164 32.4 26.1 39.1

127 40.8 33.4 50.0



#45

2-Nitroaniline

Concen: 23.157 ng/ul

RT: 13.798 min Scan# 1804

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion: 65 Resp: 266082


Ion Ratio Lower Upper

65 100

92 62.4 49.5 74.3

138 88.4 71.4 107.2





#47

Dimethylphthalate

Concen: 21.525 ng/ul

RT: 14.140 min Scan# 1862

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:163 Resp: 1007682

Ion Ratio Lower Upper

163 100

194 3.9 3.1 4.7

164 10.0 7.7 11.5



#48

2,6-Dinitrotoluene

Concen: 21.853 ng/ul

RT: 14.287 min Scan# 1887

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion:165 Resp: 207517


Ion Ratio Lower Upper

165 100

89 66.0 54.1 81.1

121 22.2 18.2 27.2









#50
Acenaphthylene
Concen: 21.721 ng/ul
RT: 14.404 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:152 Resp: 1217049
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 12.9 10.5 15.7



#51
3-Nitroaniline
Concen: 20.109 ng/ul
RT: 14.640 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:138 Resp: 149745
Ion Ratio Lower Upper
138 100
108 9.9 9.1 13.7
92 126.8 96.8 145.2





#52

Acenaphthene

Concen: 21.645 ng/ul

RT: 14.740 min Scan# 1964

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:153 Resp: 840980

Ion Ratio Lower Upper

153 100

152 48.4 39.4 59.2

154 88.6 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 19.509 ng/ul

RT: 14.881 min Scan# 1988

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion:184 Resp: 98289


Ion Ratio Lower Upper

184 100

63 75.8 71.4 107.0

154 63.1 61.1 91.7









#55
4-Nitrophenol
Concen: 20.800 ng/ul
RT: 14.981 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:109 Resp: 128670
Ion Ratio Lower Upper
109 100
139 116.8 81.3 121.9
65 142.8 99.9 149.9



#56
Dibenzofuran
Concen: 21.589 ng/ul
RT: 15.081 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:168 Resp: 1164427
Ion Ratio Lower Upper
168 100
139 38.3 32.2 48.4









#57
2,4-Dinitrotoluene
Concen: 22.915 ng/ul
RT: 15.092 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:165 Resp: 296529
Ion Ratio Lower Upper
165 100
63 69.8 56.8 85.2
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.925 ng/ul
RT: 15.328 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:232 Resp: 232362
Ion Ratio Lower Upper
232 100
131 48.8 38.2 57.2
130 2.5 2.2 3.2
166 30.3 25.6 38.4









#59
Diethylphthalate
Concen: 22.135 ng/ul
RT: 15.492 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:149 Resp: 1048643
Ion Ratio Lower Upper
149 100
177 21.6 16.8 25.2
150 12.4 10.3 15.5



#61
Fluorene
Concen: 21.859 ng/ul
RT: 15.734 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:166 Resp: 956310
Ion Ratio Lower Upper
166 100
165 98.4 78.7 118.1
167 13.0 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.754 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:204 Resp: 464595
Ion Ratio Lower Upper
204 100
206 32.5 26.0 39.0
141 59.3 51.8 77.8



#63
4-Nitroaniline
Concen: 19.518 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:138 Resp: 99477
Ion Ratio Lower Upper
138 100
92 56.9 50.1 75.1
108 84.0 85.5 128.3#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.434 ng/ul
 RT: 15.869 min Scan# 2156
 Delta R.T. 0.000 min
 Lab File: BP016025.D
 Acq: 06 Jul 2023 09:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792


Tgt Ion:198 Resp: 165270
 Ion Ratio Lower Upper
 198 100
 182 5.2 3.4 5.2#
 77 31.4 24.4 36.6



#67
 N-Nitrosodiphenylamine
 Concen: 20.993 ng/ul
 RT: 15.939 min Scan# 2168
 Delta R.T. 0.000 min
 Lab File: BP016025.D
 Acq: 06 Jul 2023 09:50

Tgt Ion:169 Resp: 816857
 Ion Ratio Lower Upper
 169 100
 168 66.7 54.2 81.4
 167 37.0 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 19.400 ng/ul
RT: 16.616 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:248 Resp: 287946
Ion Ratio Lower Upper
248 100
250 98.3 84.6 126.8
141 77.6 67.8 101.6



#69
Hexachlorobenzene
Concen: 19.232 ng/ul
RT: 16.745 min Scan# 2305
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:284 Resp: 336827
Ion Ratio Lower Upper
284 100
142 36.2 27.8 41.8
249 31.3 26.1 39.1





#70

Atrazine

Concen: 20.266 ng/ul

RT: 16.904 min Scan# 2332

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:200 Resp: 301856

Ion Ratio Lower Upper

200 100

173 26.1 19.8 29.8

215 49.0 36.8 55.2



#71

Pentachlorophenol

Concen: 17.727 ng/ul

RT: 17.116 min Scan# 2368

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion:266 Resp: 149009


Ion Ratio Lower Upper

266 100

264 61.9 53.4 80.2

268 64.8 51.4 77.2





#72

Phenanthrene

Concen: 21.752 ng/ul

RT: 17.486 min Scan# 2431

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:178 Resp: 1535983

Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 19.4 15.6 23.4



#74

Anthracene

Concen: 22.098 ng/ul

RT: 17.581 min Scan# 2447

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion:178 Resp: 1567898


Ion Ratio Lower Upper

178 100

179 15.5 12.1 18.1

176 18.8 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.315 ng/uL
RT: 13.487 min Scan# 1751
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:216 Resp: 387367
Ion Ratio Lower Upper
216 100
214 79.9 62.0 93.0
218 48.9 38.4 57.6



#76
Pentachlorobenzene
Concen: 18.587 ng/uL
RT: 15.004 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:250 Resp: 396772
Ion Ratio Lower Upper
250 100
248 60.9 50.7 76.1
252 64.3 51.2 76.8









#77
 Carbazole
 Concen: 24.177 ng/ul
 RT: 17.869 min Scan# 2496
 Delta R.T. 0.000 min
 Lab File: BP016025.D
 Acq: 06 Jul 2023 09:50

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Tgt Ion:167 Resp: 1445853




Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.4	26.0
139	13.4	11.0	16.4




#78
 Di-n-butylphthalate
 Concen: 23.898 ng/ul
 RT: 18.369 min Scan# 2581
 Delta R.T. 0.000 min
 Lab File: BP016025.D
 Acq: 06 Jul 2023 09:50

Tgt Ion:149 Resp: 1894708

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.5	11.3
104	5.6	4.6	7.0









#80
Fluoranthene
Concen: 18.777 ng/ul
RT: 19.451 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:202 Resp: 1820217
Ion Ratio Lower Upper
202 100
101 12.6 8.3 12.5#
100 9.9 6.6 9.8#



#82
Pyrene
Concen: 19.190 ng/ul
RT: 19.810 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:202 Resp: 1897594
Ion Ratio Lower Upper
202 100
101 14.8 9.3 13.9#
100 11.9 7.6 11.4#









#83
Butylbenzylphthalate
Concen: 22.070 ng/ul
RT: 20.645 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:149 Resp: 890344
Ion Ratio Lower Upper
149 100
91 80.1 59.8 89.6
206 19.1 16.4 24.6



#84
3,3'-Dichlorobenzidine
Concen: 21.483 ng/ul
RT: 21.451 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:252 Resp: 580083
Ion Ratio Lower Upper
252 100
254 63.5 51.0 76.6
126 12.3 8.6 13.0





#85

Benzo(a)anthracene

Concen: 20.716 ng/ul

RT: 21.522 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:228 Resp: 1740694

Ion Ratio Lower Upper

228 100

229 20.1 15.3 22.9

226 27.6 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.546 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion:149 Resp: 1305381


Ion Ratio Lower Upper

149 100

167 26.3 21.0 31.6

279 3.7 3.8 5.6#





#87

Chrysene

Concen: 21.693 ng/ul

RT: 21.580 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:228 Resp: 1729359

Ion Ratio Lower Upper

228 100

226 29.3 24.3 36.5

229 19.1 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 24.500 ng/ul




RT: 22.351 min Scan# 3258


Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Tgt Ion:149 Resp: 2163678









#90
Benzo(b)fluoranthene
Concen: 20.843 ng/ul
RT: 23.292 min Scan# 3418
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:252 Resp: 1618326
Ion Ratio Lower Upper
252 100
253 22.3 17.3 25.9
125 11.4 7.5 11.3#



#91
Benzo(k)fluoranthene
Concen: 21.673 ng/ul
RT: 23.339 min Scan# 3426
Delta R.T. 0.000 min
Lab File: BP016025.D
Acq: 06 Jul 2023 09:50

Tgt Ion:252 Resp: 1700248
Ion Ratio Lower Upper
252 100
253 21.7 17.0 25.4
125 11.4 7.3 10.9#





#93

Benzo(a)pyrene

Concen: 20.929 ng/ul

RT: 23.963 min Scan# 3532

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:252 Resp: 1419912

Ion Ratio Lower Upper

252 100

253 22.8 17.7 26.5

125 12.8 8.2 12.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.985 ng/ul

RT: 26.774 min Scan# 4010

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50




Tgt Ion:276 Resp: 1564359


Ion Ratio Lower Upper

276 100

138 26.5 15.7 23.5#

277 25.7 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 17.298 ng/ul

RT: 26.780 min Scan# 4011

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:278 Resp: 1308627

Ion Ratio Lower Upper

278 100

139 17.1 10.5 15.7#

279 23.7 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 16.232 ng/ul

RT: 27.621 min Scan# 4154

Delta R.T. 0.000 min

Lab File: BP016025.D

Acq: 06 Jul 2023 09:50

Tgt Ion:276 Resp: 1229932

Ion Ratio Lower Upper

276 100

138 23.4 15.1 22.7#

277 23.6 18.8 28.2

