

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070623\
 Data File : BP016039.D
 Acq On : 06 Jul 2023 18:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793

Quant Time: Jul 06 23:29:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 06 23:18:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	183479	20.000	ng/u1	0.00
20) Naphthalene-d8	10.851	136	830761	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.675	164	534633	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	1214753	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.539	240	1183294	20.000	ng/u1	# 0.00
88) Perylene-d12	24.086	264	1302989	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	40925	8.025	ng/uL	0.00
4) Pyridine-d5	3.805	84	260283	20.247	ng/u1	0.00
7) Phenol-d5	7.216	99	346583	22.077	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.357	67	227312	23.404	ng/u1	0.00
11) 2-Chlorophenol-d4	7.563	132	259075	21.862	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	271946	21.928	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	137234	21.615	ng/u1	0.00
24) 2-Nitrophenol-d4	9.946	143	151061	20.386	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	263559	19.960	ng/u1	0.00
31) 4-Chloroaniline-d4	11.022	131	403175	23.768	ng/u1	0.00
46) Dimethylphthalate-d6	14.086	166	887610	21.480	ng/u1	0.00
49) Acenaphthylene-d8	14.375	160	988103	21.271	ng/u1	0.00
54) 4-Nitrophenol-d4	14.969	143	74150	13.598	ng/u1	0.00
60) Fluorene-d10	15.675	176	738375	21.701	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	135992	18.204	ng/u1	0.00
73) Anthracene-d10	17.545	188	1197625	21.584	ng/u1	0.00
81) Pyrene-d10	19.774	212	1393406	18.034	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.910	264	1353033	20.585	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	45434	8.279	ng/uL	99
5) Pyridine	3.822	79	281552	20.936	ng/u1	97
6) Benzaldehyde	7.193	77	87805	13.834	ng/u1	98
8) Phenol	7.246	94	365734	22.450	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.452	93	305219	23.548	ng/u1	99
12) 2-Chlorophenol	7.593	128	280010	22.241	ng/u1	97
13) 2-Methylphenol	8.499	108	282369	22.957	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.563	45	468521	26.465	ng/u1	99
16) Acetophenone	8.875	105	477414	23.418	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.846	70	267483	23.623	ng/u1	99
18) 4-Methylphenol	8.840	108	308843	23.370	ng/u1	97
19) Hexachloroethane	9.099	117	122545	21.075	ng/u1	92
22) Nitrobenzene	9.263	77	373069	20.925	ng/u1	98
23) Isophorone	9.781	82	760320	22.284	ng/u1	99
25) 2-Nitrophenol	9.981	139	162009	20.601	ng/u1	95
26) 2,4-Dimethylphenol	10.046	107	353341	20.439	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.263	93	440288	22.859	ng/u1	100
29) 2,4-Dichlorophenol	10.540	162	271994	20.719	ng/u1	99
30) Naphthalene	10.904	128	963969	21.345	ng/u1	100
32) 4-Chloroaniline	11.051	127	421544	23.919	ng/u1	98
33) Hexachlorobutadiene	11.157	225	169689	17.116	ng/u1	99
34) Caprolactam	11.869	113	100866	25.118	ng/u1	96
35) 4-Chloro-3-methylphenol	12.187	107	334710	21.782	ng/u1	97
36) 2-Methylnaphthalene	12.516	142	641165	21.566	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070623\
 Data File : BP016039.D
 Acq On : 06 Jul 2023 18:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793

Quant Time: Jul 06 23:29:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 06 23:18:26 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.734	142	654729	21.588	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.887	216	331303	18.805	ng/ul	99
40) Hexachlorocyclopentadiene	12.834	237	96324	18.276	ng/ul	95
41) 2,4,6-Trichlorophenol	13.145	196	219433	19.708	ng/ul	99
42) 2,4,5-Trichlorophenol	13.234	196	245089	20.753	ng/ul	99
43) 1,1'-Biphenyl	13.510	154	880087	20.783	ng/ul	97
44) 2-Chloronaphthalene	13.563	162	689129	20.658	ng/ul	99
45) 2-Nitroaniline	13.804	65	239295	22.797	ng/ul	92
47) Dimethylphthalate	14.139	163	923118	21.586	ng/ul	99
48) 2,6-Dinitrotoluene	14.286	165	186702	21.522	ng/ul	97
50) Acenaphthylene	14.404	152	1109390	21.674	ng/ul	100
51) 3-Nitroaniline	14.645	138	128336	18.865	ng/ul	91
52) Acenaphthene	14.739	153	770629	21.712	ng/ul	98
53) 2,4-Dinitrophenol	14.886	184	72642	15.784	ng/ul	91
55) 4-Nitrophenol	14.992	109	112842	19.968	ng/ul	94
56) Dibenzofuran	15.086	168	1061766	21.549	ng/ul	97
57) 2,4-Dinitrotoluene	15.098	165	271303	22.950	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.333	232	213330	21.030	ng/ul	99
59) Diethylphthalate	15.492	149	954132	22.047	ng/ul	99
61) Fluorene	15.733	166	876895	21.942	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.716	204	419004	20.490	ng/ul	97
63) 4-Nitroaniline	15.822	138	90566	19.452	ng/ul	88
66) 4,6-Dinitro-2-methylph...	15.875	198	134490	17.793	ng/ul#	97
67) N-Nitrosodiphenylamine	15.939	169	744972	20.486	ng/ul	99
68) 4-Bromophenyl-phenylether	16.616	248	264566	19.072	ng/ul	91
69) Hexachlorobenzene	16.745	284	310359	18.962	ng/ul	97
70) Atrazine	16.904	200	275541	19.794	ng/ul	96
71) Pentachlorophenol	17.122	266	141089	17.960	ng/ul	98
72) Phenanthrene	17.486	178	1434550	21.738	ng/ul	99
74) Anthracene	17.580	178	1456370	21.963	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.487	216	352734	17.845	ng/uL	98
76) Pentachlorobenzene	15.004	250	357302	17.910	ng/uL	98
77) Carbazole	17.875	167	1335292	23.891	ng/ul	99
78) Di-n-butylphthalate	18.369	149	1696631	22.898	ng/ul	100
80) Fluoranthene	19.451	202	1733692	18.054	ng/ul#	93
82) Pyrene	19.810	202	1800875	18.385	ng/ul#	92
83) Butylbenzylphthalate	20.645	149	825005	20.644	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.457	252	538304	20.124	ng/ul	98
85) Benzo(a)anthracene	21.521	228	1708793	20.529	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.392	149	1199891	21.848	ng/ul	99
87) Chrysene	21.580	228	1726217	21.858	ng/ul	99
89) Di-n-octyl phthalate	22.351	149	2088435	21.932	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	1786836	21.342	ng/ul	99
91) Benzo(k)fluoranthene	23.345	252	1798341	21.259	ng/ul#	98
93) Benzo(a)pyrene	23.968	252	1535610	20.991	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.774	276	1830505	18.432	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.780	278	1523505	18.677	ng/ul#	95
96) Benzo(g,h,i)perylene	27.615	276	1433368	17.544	ng/ul#	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070623\
Data File : BP016039.D
Acq On : 06 Jul 2023 18:26
Operator : MA/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020793

Quant Time: Jul 06 23:29:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 06 23:18:26 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.279 ng/uL
RT: 3.387 min Scan# 34
Delta R.T. -0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 88 Resp: 45434
Ion Ratio Lower Upper
88 100
43 35.3 27.4 41.2
58 74.8 59.8 89.8



#5
Pyridine
Concen: 20.936 ng/uL
RT: 3.822 min Scan# 108
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion: 79 Resp: 281552
Ion Ratio Lower Upper
79 100
52 79.4 65.5 98.3
51 37.8 31.8 47.8









#6
Benzaldehyde
Concen: 13.834 ng/ul
RT: 7.193 min Scan# 681
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 77 Resp: 87805
Ion Ratio Lower Upper
77 100
105 88.3 73.4 110.0
106 83.3 66.6 100.0



#8
Phenol
Concen: 22.450 ng/ul
RT: 7.246 min Scan# 690
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion: 94 Resp: 365734
Ion Ratio Lower Upper
94 100
65 30.7 26.4 39.6
66 39.7 37.3 55.9









#10
Bis(2-Chloroethyl)ether
Concen: 23.548 ng/ul
RT: 7.452 min Scan# 725
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 93 Resp: 305219
Ion Ratio Lower Upper
93 100
63 78.1 62.8 94.2
95 33.0 25.6 38.4



#12
2-Chlorophenol
Concen: 22.241 ng/ul
RT: 7.593 min Scan# 749
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:128 Resp: 280010
Ion Ratio Lower Upper
128 100
64 54.0 41.2 61.8
130 30.7 26.0 39.0









#13
2-Methylphenol
Concen: 22.957 ng/ul
RT: 8.499 min Scan# 903
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:108 Resp: 282369
Ion Ratio Lower Upper
108 100
107 90.7 73.8 110.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 26.465 ng/ul
RT: 8.563 min Scan# 914
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion: 45 Resp: 468521
Ion Ratio Lower Upper
45 100
77 14.6 12.4 18.6
79 11.4 9.3 13.9









#16
Acetophenone
Concen: 23.418 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:105 Resp: 477414
Ion Ratio Lower Upper
105 100
77 86.2 65.0 97.4
51 31.4 25.1 37.7



#17
N-Nitroso-di-n-propylamine
Concen: 23.623 ng/ul
RT: 8.846 min Scan# 962
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion: 70 Resp: 267483
Ion Ratio Lower Upper
70 100
42 60.3 48.2 72.2
101 9.0 7.6 11.4
130 18.9 13.9 20.9





#18

4-Methylphenol

Concen: 23.370 ng/ul

RT: 8.840 min Scan# 9

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :





SSTD020793

Tgt Ion:108 Resp: 308843

Ion Ratio Lower Upper

108 100

107 111.1 91.3 136.9



#19

Hexachloroethane

Concen: 21.075 ng/ul

RT: 9.099 min Scan# 1005

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:117 Resp: 122545


Ion Ratio Lower Upper

117 100

201 95.5 84.2 126.2

199 60.8 51.8 77.8









#22
Nitrobenzene
Concen: 20.925 ng/ul
RT: 9.263 min Scan# 1033
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 77 Resp: 373069
Ion Ratio Lower Upper
77 100
123 39.5 31.0 46.4
65 13.2 11.3 16.9



#23
Isophorone
Concen: 22.284 ng/ul
RT: 9.781 min Scan# 1121
Delta R.T. -0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion: 82 Resp: 760320
Ion Ratio Lower Upper
82 100
95 6.8 6.3 9.5
138 15.2 12.3 18.5









#25
2-Nitrophenol
Concen: 20.601 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:139 Resp: 162009
Ion Ratio Lower Upper
139 100
65 61.2 50.2 75.2
109 33.9 31.3 46.9



#26
2,4-Dimethylphenol
Concen: 20.439 ng/ul
RT: 10.046 min Scan# 1166
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:107 Resp: 353341
Ion Ratio Lower Upper
107 100
121 49.4 38.2 57.2
122 83.5 62.1 93.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.859 ng/ul
 RT: 10.263 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BP016039.D
 Acq: 06 Jul 2023 18:26




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793


Tgt Ion: 93 Resp: 440288
 Ion Ratio Lower Upper
 93 100
 95 33.3 26.6 39.8
 123 10.2 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 20.719 ng/ul
 RT: 10.540 min Scan# 1250
 Delta R.T. 0.006 min
 Lab File: BP016039.D
 Acq: 06 Jul 2023 18:26

Tgt Ion: 162 Resp: 271994
 Ion Ratio Lower Upper
 162 100
 164 65.5 53.0 79.6
 98 39.0 31.3 46.9





#30

Naphthalene

Concen: 21.345 ng/ul

RT: 10.904 min Scan# 1312

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:128 Resp: 963969

Ion Ratio Lower Upper

128 100

129 10.6 8.6 12.8

127 12.9 10.3 15.5



#32

4-Chloroaniline

Concen: 23.919 ng/ul

RT: 11.051 min Scan# 1337

Delta R.T. 0.006 min

Lab File: BP016039.D




Acq: 06 Jul 2023 18:26


Tgt Ion:127 Resp: 421544

Ion Ratio Lower Upper

127 100

129 32.0 24.7 37.1









#33
Hexachlorobutadiene
Concen: 17.116 ng/ul
RT: 11.157 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:225 Resp: 169689
Ion Ratio Lower Upper
225 100
223 63.6 50.4 75.6
227 64.2 51.0 76.4



#34
Caprolactam
Concen: 25.118 ng/ul
RT: 11.869 min Scan# 1476
Delta R.T. -0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:113 Resp: 100866
Ion Ratio Lower Upper
113 100
55 178.3 142.8 214.2
56 136.2 116.9 175.3









#35
4-Chloro-3-methylphenol
Concen: 21.782 ng/ul
RT: 12.187 min Scan# 1529
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:107 Resp: 334710
Ion Ratio Lower Upper
107 100
144 24.2 18.4 27.6
142 76.0 58.6 87.8



#36
2-Methylnaphthalene
Concen: 21.566 ng/ul
RT: 12.516 min Scan# 1586
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:142 Resp: 641165
Ion Ratio Lower Upper
142 100
141 88.9 73.2 109.8









#37
1-Methylnaphthalene
Concen: 21.588 ng/ul
RT: 12.734 min Scan# 1623
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:142 Resp: 654729
Ion Ratio Lower Upper
142 100
141 91.8 73.1 109.7
116 4.1 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.805 ng/ul
RT: 12.887 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:216 Resp: 331303
Ion Ratio Lower Upper
216 100
214 77.7 62.4 93.6
179 21.0 17.6 26.4
108 19.3 15.5 23.3









#40
Hexachlorocyclopentadiene
Concen: 18.276 ng/ul
RT: 12.834 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:237 Resp: 96324
Ion Ratio Lower Upper
237 100
235 65.0 48.6 73.0
272 13.3 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 19.708 ng/ul
RT: 13.145 min Scan# 1693
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:196 Resp: 219433
Ion Ratio Lower Upper
196 100
198 94.8 75.4 113.2
200 31.3 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 20.753 ng/ul

RT: 13.234 min Scan# 1708

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:196 Resp: 245089

Ion Ratio Lower Upper

196 100

198 92.6 74.9 112.3

200 29.5 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.783 ng/ul

RT: 13.510 min Scan# 1755

Delta R.T. -0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:154 Resp: 880087


Ion Ratio Lower Upper

154 100

153 39.8 32.8 49.2

76 15.8 10.7 16.1









#44
2-Chloronaphthalene
Concen: 20.658 ng/ul
RT: 13.563 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:162 Resp: 689129
Ion Ratio Lower Upper
162 100
164 32.3 26.1 39.1
127 41.1 33.4 50.0



#45
2-Nitroaniline
Concen: 22.797 ng/ul
RT: 13.804 min Scan# 1805
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion: 65 Resp: 239295
Ion Ratio Lower Upper
65 100
92 69.0 49.5 74.3
138 95.9 71.4 107.2









#47
Dimethylphthalate
Concen: 21.586 ng/ul
RT: 14.139 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:163 Resp: 923118
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.1 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 21.522 ng/ul
RT: 14.286 min Scan# 1887
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:165 Resp: 186702
Ion Ratio Lower Upper
165 100
89 70.0 54.1 81.1
121 23.4 18.2 27.2









#50
Acenaphthylene
Concen: 21.674 ng/ul
RT: 14.404 min Scan# 1907
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:152 Resp: 1109390
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 12.9 10.5 15.7



#51
3-Nitroaniline
Concen: 18.865 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:138 Resp: 128336
Ion Ratio Lower Upper
138 100
108 10.6 9.1 13.7
92 132.1 96.8 145.2





#52

Acenaphthene

Concen: 21.712 ng/ul

RT: 14.739 min Scan# 1964

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:153 Resp: 770629

Ion Ratio Lower Upper

153 100

152 47.7 39.4 59.2

154 87.9 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 15.784 ng/ul

RT: 14.886 min Scan# 1989

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:184 Resp: 72642


Ion Ratio Lower Upper

184 100

63 83.5 71.4 107.0

154 66.5 61.1 91.7









#55
4-Nitrophenol
Concen: 19.968 ng/ul
RT: 14.992 min Scan# 20
Delta R.T. 0.012 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:109 Resp: 112842
Ion Ratio Lower Upper
109 100
139 106.2 81.3 121.9
65 116.9 99.9 149.9



#56
Dibenzofuran
Concen: 21.549 ng/ul
RT: 15.086 min Scan# 2023
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:168 Resp: 1061766
Ion Ratio Lower Upper
168 100
139 38.7 32.2 48.4





#57

2,4-Dinitrotoluene

Concen: 22.950 ng/ul

RT: 15.098 min Scan# 2024

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:165 Resp: 271303

Ion Ratio Lower Upper

165 100

63 64.7 56.8 85.2

182 2.9 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.030 ng/ul

RT: 15.333 min Scan# 2065

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Tgt Ion:232 Resp: 213330




Ion Ratio Lower Upper


232 100

131 46.4 38.2 57.2

130 2.7 2.2 3.2

166 31.6 25.6 38.4









#59
Diethylphthalate
Concen: 22.047 ng/ul
RT: 15.492 min Scan# 2092
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:149 Resp: 954132
Ion Ratio Lower Upper
149 100
177 21.7 16.8 25.2
150 12.5 10.3 15.5



#61
Fluorene
Concen: 21.942 ng/ul
RT: 15.733 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:166 Resp: 876895
Ion Ratio Lower Upper
166 100
165 99.5 78.7 118.1
167 13.7 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.490 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:204 Resp: 419004
Ion Ratio Lower Upper
204 100
206 32.6 26.0 39.0
141 61.7 51.8 77.8



#63
4-Nitroaniline
Concen: 19.452 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:138 Resp: 90566
Ion Ratio Lower Upper
138 100
92 61.7 50.1 75.1
108 88.5 85.5 128.3









#66
4,6-Dinitro-2-methylphenol
Concen: 17.793 ng/ul
RT: 15.875 min Scan# 2156
Delta R.T. 0.006 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:198 Resp: 134490
Ion Ratio Lower Upper
198 100
182 5.7 3.4 5.2#
77 32.0 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 20.486 ng/ul
RT: 15.939 min Scan# 2168
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:169 Resp: 744972
Ion Ratio Lower Upper
169 100
168 67.1 54.2 81.4
167 36.6 29.7 44.5









#68
4-Bromophenyl-phenylether
Concen: 19.072 ng/ul
RT: 16.616 min Scan# 2283
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:248 Resp: 264566
Ion Ratio Lower Upper
248 100
250 96.7 84.6 126.8
141 76.7 67.8 101.6



#69
Hexachlorobenzene
Concen: 18.962 ng/ul
RT: 16.745 min Scan# 2305
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:284 Resp: 310359
Ion Ratio Lower Upper
284 100
142 36.2 27.8 41.8
249 30.7 26.1 39.1





#70

Atrazine

Concen: 19.794 ng/ul

RT: 16.904 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:200 Resp: 275541

Ion Ratio Lower Upper

200 100

173 27.0 19.8 29.8

215 48.2 36.8 55.2



#71

Pentachlorophenol

Concen: 17.960 ng/ul

RT: 17.122 min Scan# 2369

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:266 Resp: 141089


Ion Ratio Lower Upper

266 100

264 63.4 53.4 80.2

268 63.8 51.4 77.2





#72

Phenanthrene

Concen: 21.738 ng/ul

RT: 17.486 min Scan# 2431

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:178 Resp: 1434550

Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 19.2 15.6 23.4



#74

Anthracene

Concen: 21.963 ng/ul

RT: 17.580 min Scan# 2447

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:178 Resp: 1456370


Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 18.5 14.6 22.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 17.845 ng/uL

RT: 13.487 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:216 Resp: 352734

Ion Ratio Lower Upper

216 100

214 80.0 62.0 93.0

218 47.9 38.4 57.6



#76

Pentachlorobenzene

Concen: 17.910 ng/uL

RT: 15.004 min Scan# 2009

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:250 Resp: 357302


Ion Ratio Lower Upper

250 100

248 63.7 50.7 76.1

252 61.5 51.2 76.8





#77

Carbazole

Concen: 23.891 ng/ul

RT: 17.875 min Scan# 2496

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:167 Resp: 1335292

Ion Ratio Lower Upper

167 100

166 21.4 17.4 26.0

139 13.2 11.0 16.4



#78

Di-n-butylphthalate

Concen: 22.898 ng/ul

RT: 18.369 min Scan# 2581

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:149 Resp: 1696631


Ion Ratio Lower Upper

149 100

150 9.3 7.5 11.3

104 5.8 4.6 7.0









#80
Fluoranthene
Concen: 18.054 ng/ul
RT: 19.451 min Scan# 2765
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:202 Resp: 1733692
Ion Ratio Lower Upper
202 100
101 13.2 8.3 12.5#
100 10.3 6.6 9.8#



#82
Pyrene
Concen: 18.385 ng/ul
RT: 19.810 min Scan# 2826
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:202 Resp: 1800875
Ion Ratio Lower Upper
202 100
101 14.5 9.3 13.9#
100 12.3 7.6 11.4#





#83

Butylbenzylphthalate

Concen: 20.644 ng/ul

RT: 20.645 min Scan# 2968

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:149 Resp: 825005

Ion Ratio Lower Upper

149 100

91 80.8 59.8 89.6

206 19.3 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 20.124 ng/ul

RT: 21.457 min Scan# 3106

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:252 Resp: 538304


Ion Ratio Lower Upper

252 100

254 63.0 51.0 76.6

126 12.6 8.6 13.0





#85

Benzo(a)anthracene

Concen: 20.529 ng/ul

RT: 21.521 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:228 Resp: 1708793

Ion Ratio Lower Upper

228 100

229 19.3 15.3 22.9

226 27.8 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.848 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:149 Resp: 1199891


Ion Ratio Lower Upper

149 100

167 26.4 21.0 31.6

279 3.9 3.8 5.6









#87
Chrysene
Concen: 21.858 ng/ul
RT: 21.580 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:228 Resp: 1726217
Ion Ratio Lower Upper
228 100
226 29.6 24.3 36.5
229 19.8 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 21.932 ng/ul
RT: 22.351 min Scan# 3258
Delta R.T. -0.000 min
Lab File: BP016039.D
Acq: 06 Jul 2023 18:26

Tgt Ion:149 Resp: 2088435





#90

Benzo(b)fluoranthene

Concen: 21.342 ng/ul

RT: 23.345 min Scan# 3418

Delta R.T. 0.053 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:252 Resp: 1786836

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 11.1 7.5 11.3



#91

Benzo(k)fluoranthene

Concen: 21.259 ng/ul

RT: 23.345 min Scan# 3427

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:252 Resp: 1798341


Ion Ratio Lower Upper

252 100

253 21.6 17.0 25.4

125 11.1 7.3 10.9#





#93

Benzo(a)pyrene

Concen: 20.991 ng/ul

RT: 23.968 min Scan# 3532

Delta R.T. 0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:252 Resp: 1535610

Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 12.5 8.2 12.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.432 ng/ul

RT: 26.774 min Scan# 4010

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26




Tgt Ion:276 Resp: 1830505


Ion Ratio Lower Upper

276 100

138 26.7 15.7 23.5#

277 24.6 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 18.677 ng/ul

RT: 26.780 min Scan# 4011

Delta R.T. -0.000 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:278 Resp: 1523505

Ion Ratio Lower Upper

278 100

139 17.7 10.5 15.7#

279 22.7 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 17.544 ng/ul

RT: 27.615 min Scan# 4153

Delta R.T. -0.006 min

Lab File: BP016039.D

Acq: 06 Jul 2023 18:26

Tgt Ion:276 Resp: 1433368

Ion Ratio Lower Upper

276 100

138 23.2 15.1 22.7#

277 23.4 18.8 28.2

