

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP070820\
 Data File : BP002675.D
 Acq On : 08 Jul 2020 15:10
 Operator : CG/JU
 Sample : L3194-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TUNNEL-WALL-WASHWATER(EXH)

Quant Time: Jul 08 15:46:21 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP062920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 29 16:51:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	139293	20.00	ng	0.00
21) Naphthalene-d8	10.34	136	605223	20.00	ng	0.00
39) Acenaphthene-d10	14.22	164	429836	20.00	ng	0.00
64) Phenanthrene-d10	16.97	188	1043913	20.00	ng	0.00
76) Chrysene-d12	21.09	240	1222264	20.00	ng	0.00
86) Perylene-d12	23.29	264	1400210	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	361636	43.96	ng	0.01
7) Phenol-d6	6.79	99	321908	28.03	ng	0.02
23) Nitrobenzene-d5	8.71	82	892485	75.98	ng	-0.01
42) 2,4,6-Tribromophenol	15.72	330	578440	110.64	ng	0.00
45) 2-Fluorobiphenyl	12.83	172	2235860	76.56	ng	0.00
79) Terphenyl-d14	19.57	244	3810727	67.30	ng	0.00
Target Compounds						
32) Benzoic acid	9.70	122	208868	34.041	ng	98
50) Dimethylphthalate	13.68	163	469046	13.893	ng	99
70) Pentachlorophenol	16.65	266	2014	2.933	ng	97
84) Bis(2-ethylhexyl)phthalate	21.03	149	1137932	21.818	ng	94
85) Di-n-octyl phthalate	21.89	149	1775323	19.272	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP070820\
Data File : BP002675.D
Acq On : 08 Jul 2020 15:10
Operator : CG/JU
Sample : L3194-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TUNNEL-WALL-WASHWATER(EXH)

Quant Time: Jul 08 15:46:21 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP062920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 29 16:51:57 2020
Response via : Initial Calibration

