

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070823\
 Data File : BP016089.D
 Acq On : 08 Jul 2023 13:00
 Operator : MA/JU
 Sample : 03332-14
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YBTY7

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/10/2023
 Supervised By :mohammad ahmed 07/10/2023

Quant Time: Jul 08 23:20:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 08 10:11:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	111188	20.000	ng/u1	0.02
20) Naphthalene-d8	10.881	136	590601	20.000	ng/u1	0.03
38) Acenaphthene-d10	14.687	164	414951	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.451	188	939129	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.545	240	1068771	20.000	ng/u1	# 0.00
88) Perylene-d12	24.074	264	1201522	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	9177m	2.970	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.287	99	106945m	11.242	ng/u1	0.07
9) Bis-(2-Chloroethyl)eth...	7.399	67	89932m	15.280	ng/u1	0.04
11) 2-Chlorophenol-d4	7.617	132	99238	13.819	ng/u1	0.05
15) 4-Methylphenol-d8	8.869	113	98585m	13.118	ng/u1	0.10
21) Nitrobenzene-d5	9.299	128	47344	10.489	ng/u1	0.08
24) 2-Nitrophenol-d4	10.046	143	56043m	10.639	ng/u1	0.10
28) 2,4-Dichlorophenol-d3	10.663	165	95050m	10.125	ng/u1	0.16
31) 4-Chloroaniline-d4	11.152	131	109254m	9.060	ng/u1	0.13
46) Dimethylphthalate-d6	14.122	166	464196	14.473	ng/u1	0.04
49) Acenaphthylene-d8	14.387	160	481102	13.344	ng/u1	0.01
54) 4-Nitrophenol-d4	15.157	143	33755m	7.975	ng/u1	0.19
60) Fluorene-d10	15.704	176	402630	15.246	ng/u1	0.03
65) 4,6-Dinitro-2-methylph...	15.945	200	32636m	5.651	ng/u1	0.09
73) Anthracene-d10	17.569	188	646291	15.066	ng/u1	0.02
81) Pyrene-d10	19.786	212	925605	13.263	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.904	264	946130	15.610	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070823\
Data File : BP016089.D
Acq On : 08 Jul 2023 13:00
Operator : MA/JU
Sample : 03332-14
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YBTY7

Quant Time: Jul 08 23:20:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 08 10:11:43 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/10/2023
Supervised By :mohammad ahmed 07/10/2023

