

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070823\
 Data File : BP016083.D
 Acq On : 08 Jul 2023 09:35
 Operator : MA/JU
 Sample : 03332-13
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YBTY6

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/10/2023
 Supervised By :mohammad ahmed 07/10/2023

Quant Time: Jul 08 22:47:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 08 10:11:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	152036m	20.000	ng/u1	0.02
20) Naphthalene-d8	10.875	136	688390	20.000	ng/u1	# 0.02
38) Acenaphthene-d10	14.687	164	448277	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.457	188	972642	20.000	ng/u1	# 0.01
79) Chrysene-d12	21.551	240	973726	20.000	ng/u1	# 0.01
88) Perylene-d12	24.080	264	986427	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	10533	2.493	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.281	99	136373	10.483	ng/u1	0.06
9) Bis-(2-Chloroethyl)eth...	7.387	67	117365m	14.583	ng/u1	0.03
11) 2-Chlorophenol-d4	7.605	132	140966	14.355	ng/u1	0.04
15) 4-Methylphenol-d8	8.852	113	129155m	12.568	ng/u1	0.08
21) Nitrobenzene-d5	9.281	128	60401	11.481	ng/u1	0.06
24) 2-Nitrophenol-d4	10.058	143	68358m	11.133	ng/u1	0.11
28) 2,4-Dichlorophenol-d3	10.663	165	117807m	10.767	ng/u1	0.16
31) 4-Chloroaniline-d4	11.140	131	128554m	9.146	ng/u1	0.12
46) Dimethylphthalate-d6	14.122	166	500176	14.436	ng/u1	0.04
49) Acenaphthylene-d8	14.387	160	554683	14.241	ng/u1	0.01
54) 4-Nitrophenol-d4	15.175	143	24757m	5.415	ng/u1	0.21
60) Fluorene-d10	15.704	176	457855	16.049	ng/u1	0.03
65) 4,6-Dinitro-2-methylph...	15.975	200	27799m	4.647	ng/u1	0.12
73) Anthracene-d10	17.569	188	705141	15.871	ng/u1	0.02
81) Pyrene-d10	19.786	212	951831	14.970	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.910	264	807636	16.231	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070823\
Data File : BP016083.D
Acq On : 08 Jul 2023 09:35
Operator : MA/JU
Sample : 03332-13
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YBTY6

Quant Time: Jul 08 22:47:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 08 10:11:43 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/10/2023
Supervised By :mohammad ahmed 07/10/2023

