

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070823\
 Data File : BP016101.D
 Acq On : 08 Jul 2023 20:24
 Operator : MA/JU
 Sample : 03332-15
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YBTY8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/10/2023
 Supervised By :mohammad ahmed 07/10/2023

Quant Time: Jul 09 00:03:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 08 10:11:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.022	152	216276	20.000	ng/u1	0.00
20) Naphthalene-d8	10.851	136	986426	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.675	164	650436	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1482404	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.545	240	1387990	20.000	ng/u1	# 0.00
88) Perylene-d12	24.074	264	1416095	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	30312	5.043	ng/uL	0.00
4) Pyridine-d5	3.799	84	389369	25.696	ng/u1	0.00
7) Phenol-d5	7.222	99	551302	29.792	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.352	67	405712m	35.438	ng/u1	0.00
11) 2-Chlorophenol-d4	7.558	132	458461	32.820	ng/u1	0.00
15) 4-Methylphenol-d8	8.775	113	502792	34.394	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	240070	31.845	ng/u1	0.01
24) 2-Nitrophenol-d4	9.963	143	260985	29.662	ng/u1	0.02
28) 2,4-Dichlorophenol-d3	10.546	165	361311	23.044	ng/u1	0.04
31) 4-Chloroaniline-d4	11.051	131	615715	30.570	ng/u1	0.03
46) Dimethylphthalate-d6	14.092	166	1776120	35.329	ng/u1	0.00
49) Acenaphthylene-d8	14.369	160	1917958	33.937	ng/u1	0.00
54) 4-Nitrophenol-d4	15.016	143	166021m	25.025	ng/u1	0.05
60) Fluorene-d10	15.681	176	1484964	35.873	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.869	200	195733m	21.470	ng/u1	0.02
73) Anthracene-d10	17.545	188	2333721	34.465	ng/u1	0.00
81) Pyrene-d10	19.775	212	2874947	31.721	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	2472136	34.607	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070823\
Data File : BP016101.D
Acq On : 08 Jul 2023 20:24
Operator : MA/JU
Sample : 03332-15
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YBTY8

Quant Time: Jul 09 00:03:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 08 10:11:43 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/10/2023
Supervised By :mohammad ahmed 07/10/2023

