

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071023\
 Data File : BP016176.D
 Acq On : 11 Jul 2023 02:35
 Operator : MA/JU
 Sample : 03404-17
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YBW53

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/11/2023
 Supervised By :mohammad ahmed 07/11/2023

Quant Time: Jul 11 03:29:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 10 23:19:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.022	152	178363	20.000	ng/u1	0.00
20) Naphthalene-d8	10.852	136	832937	20.000	ng/u1	0.02
38) Acenaphthene-d10	14.669	164	538420	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.434	188	1167115	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.551	240	1094036	20.000	ng/u1	# 0.00
88) Perylene-d12	24.080	264	1144331	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	21667	4.371	ng/uL	0.00
4) Pyridine-d5	3.805	84	118558	9.487	ng/u1	0.01
7) Phenol-d5	7.234	99	336135	22.026	ng/u1	0.03
9) Bis-(2-Chloroethyl)eth...	7.358	67	251836m	26.673	ng/u1	0.01
11) 2-Chlorophenol-d4	7.569	132	287100	24.922	ng/u1	0.02
15) 4-Methylphenol-d8	8.799	113	287509	23.848	ng/u1	0.04
21) Nitrobenzene-d5	9.240	128	137762	21.642	ng/u1	0.03
24) 2-Nitrophenol-d4	9.987	143	156034m	21.002	ng/u1	0.05
28) 2,4-Dichlorophenol-d3	10.593	165	253119m	19.119	ng/u1	0.08
31) 4-Chloroaniline-d4	11.105	131	224936m	13.226	ng/u1	0.08
46) Dimethylphthalate-d6	14.099	166	1110876	26.694	ng/u1	0.02
49) Acenaphthylene-d8	14.369	160	1227084	26.230	ng/u1	0.00
54) 4-Nitrophenol-d4	15.128	143	89603m	16.316	ng/u1	0.14
60) Fluorene-d10	15.681	176	980135	28.603	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.910	200	87847m	12.239	ng/u1	0.05
73) Anthracene-d10	17.545	188	1523404	28.575	ng/u1	0.00
81) Pyrene-d10	19.775	212	1890416	26.462	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	1584845	27.455	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071023\
Data File : BP016176.D
Acq On : 11 Jul 2023 02:35
Operator : MA/JU
Sample : 03404-17
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YBW53

Quant Time: Jul 11 03:29:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 10 23:19:51 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/11/2023
Supervised By :mohammad ahmed 07/11/2023

