

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071123\
 Data File : BP016188.D
 Acq On : 11 Jul 2023 18:51
 Operator : MA/JU
 Sample : SSTD08099
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080643

Quant Time: Jul 12 00:36:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 12 00:24:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	164297	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	772382	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	491191	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.381	188	1113781	20.000	ng/u1	0.00
79) Chrysene-d12	21.474	240	1108458	20.000	ng/u1	0.00
88) Perylene-d12	23.974	264	1397584	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	147139	32.221	ng/uL	0.00
4) Pyridine-d5	3.728	84	1030264	89.501	ng/u1	-0.01
7) Phenol-d5	7.140	99	1300085	92.483	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.287	67	830087	95.445	ng/u1	0.00
11) 2-Chlorophenol-d4	7.487	132	967623	91.186	ng/u1	-0.01
15) 4-Methylphenol-d8	8.699	113	1059174	95.376	ng/u1	-0.01
21) Nitrobenzene-d5	9.146	128	499260	84.580	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.875	143	571392	82.939	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.422	165	986402	80.347	ng/u1	-0.03
31) 4-Chloroaniline-d4	10.940	131	1425125	90.366	ng/u1	-0.02
46) Dimethylphthalate-d6	14.034	166	3069490	80.850	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	3435576	80.500	ng/u1	0.00
54) 4-Nitrophenol-d4	14.881	143	529902	105.768	ng/u1	-0.02
60) Fluorene-d10	15.616	176	2486648	79.546	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.781	200	593091	86.587	ng/u1	-0.02
73) Anthracene-d10	17.486	188	3973774	78.108	ng/u1	0.00
81) Pyrene-d10	19.716	212	4785360	66.114	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.810	264	5594548	79.355	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.323	88	159508	32.458	ng/uL	97
5) Pyridine	3.746	79	1059868	88.012	ng/u1	98
6) Benzaldehyde	7.111	77	835788	147.057	ng/u1	97
8) Phenol	7.163	94	1377729	94.444	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.387	93	1077406	92.827	ng/u1	99
12) 2-Chlorophenol	7.522	128	983625	87.249	ng/u1	98
13) 2-Methylphenol	8.428	108	1028315	93.364	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.481	45	1707707	107.724	ng/u1	96
16) Acetophenone	8.810	105	1677956	91.916	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.787	70	938923	92.605	ng/u1	98
18) 4-Methylphenol	8.763	108	1118455	94.512	ng/u1	98
19) Hexachloroethane	9.028	117	412711	79.262	ng/u1	99
22) Nitrobenzene	9.193	77	1373793	82.877	ng/u1	97
23) Isophorone	9.722	82	2674003	84.297	ng/u1	99
25) 2-Nitrophenol	9.910	139	601306	82.240	ng/u1	98
26) 2,4-Dimethylphenol	9.963	107	1242708	77.316	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.193	93	1555506	86.865	ng/u1	100
29) 2,4-Dichlorophenol	10.446	162	980067	80.298	ng/u1	98
30) Naphthalene	10.828	128	3232747	76.991	ng/u1	100
32) 4-Chloroaniline	10.969	127	1407284	85.888	ng/u1	98
33) Hexachlorobutadiene	11.093	225	578621	62.776	ng/u1	99
34) Caprolactam	11.799	113	179748	48.145	ng/u1	98
35) 4-Chloro-3-methylphenol	12.104	107	1189041	83.226	ng/u1	98
36) 2-Methylnaphthalene	12.446	142	2223929	80.458	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071123\
 Data File : BP016188.D
 Acq On : 11 Jul 2023 18:51
 Operator : MA/JU
 Sample : SST08099
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST080643


Quant Time: Jul 12 00:36:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 12 00:24:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	2257039	80.044	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.816	216	1167660	72.138	ng/ul	97
40) Hexachlorocyclopentadiene	12.769	237	497681	102.776	ng/ul	99
41) 2,4,6-Trichlorophenol	13.069	196	801306	78.333	ng/ul	98
42) 2,4,5-Trichlorophenol	13.157	196	853326	78.646	ng/ul	96
43) 1,1'-Biphenyl	13.451	154	2930327	75.319	ng/ul	100
44) 2-Chloronaphthalene	13.498	162	2316545	75.584	ng/ul	99
45) 2-Nitroaniline	13.734	65	900560	93.384	ng/ul	99
47) Dimethylphthalate	14.081	163	3057903	77.828	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	669148	83.959	ng/ul	97
50) Acenaphthylene	14.345	152	3890764	82.738	ng/ul	99
51) 3-Nitroaniline	14.563	138	676660	108.267	ng/ul	99
52) Acenaphthene	14.681	153	2511942	77.032	ng/ul	100
53) 2,4-Dinitrophenol	14.798	184	416473	98.494	ng/ul	99
55) 4-Nitrophenol	14.892	109	443415	85.406	ng/ul	98
56) Dibenzofuran	15.022	168	3375232	74.561	ng/ul	99
57) 2,4-Dinitrotoluene	15.034	165	944556	86.969	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.257	232	731426	78.481	ng/ul	97
59) Diethylphthalate	15.440	149	3147102	79.151	ng/ul	100
61) Fluorene	15.669	166	2774541	75.565	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.657	204	1342456	71.453	ng/ul	99
63) 4-Nitroaniline	15.745	138	646112	151.047	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.798	198	618679	89.271	ng/ul#	95
67) N-Nitrosodiphenylamine	15.881	169	2413350	72.380	ng/ul	100
68) 4-Bromophenyl-phenylether	16.557	248	871900	68.553	ng/ul	98
69) Hexachlorobenzene	16.681	284	1018438	67.863	ng/ul	97
70) Atrazine	16.845	200	988635	77.460	ng/ul	98
71) Pentachlorophenol	17.039	266	587866	81.618	ng/ul	97
72) Phenanthrene	17.428	178	4590531	75.868	ng/ul	100
74) Anthracene	17.522	178	4728905	77.779	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.416	216	1229454	67.838	ng/uL	99
76) Pentachlorobenzene	14.940	250	1106770	60.508	ng/uL	100
77) Carbazole	17.804	167	4425856	86.367	ng/ul	98
78) Di-n-butylphthalate	18.316	149	5638296	82.994	ng/ul	99
80) Fluoranthene	19.392	202	5570178	61.923	ng/ul	100
82) Pyrene	19.745	202	5838777	63.631	ng/ul	99
83) Butylbenzylphthalate	20.598	149	2790247	74.535	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.386	252	2286072	91.233	ng/ul	97
85) Benzo(a)anthracene	21.457	228	5878605	75.391	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.339	149	4122558	80.132	ng/ul	100
87) Chrysene	21.516	228	5704855	77.114	ng/ul	98
89) Di-n-octyl phthalate	22.286	149	7290025	71.375	ng/ul	100
90) Benzo(b)fluoranthene	23.204	252	6497746	72.357	ng/ul	99
91) Benzo(k)fluoranthene	23.257	252	6458122	71.178	ng/ul	98
93) Benzo(a)pyrene	23.863	252	6403555	81.608	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.633	276	8240920	77.365	ng/ul	98
95) Dibenzo(a,h)anthracene	26.633	278	6761759	77.283	ng/ul	98
96) Benzo(g,h,i)perylene	27.433	276	6899865	78.737	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 12 00:36:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 12 00:24:54 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 32.458 ng/uL
RT: 3.323 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion: 88 Resp: 159508
Ion Ratio Lower Upper
88 100
43 31.1 27.5 41.3
58 75.7 62.2 93.4



#5
Pyridine
Concen: 88.012 ng/uL
RT: 3.746 min Scan# 95
Delta R.T. -0.012 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion: 79 Resp: 1059868
Ion Ratio Lower Upper
79 100
52 81.9 67.4 101.0
51 37.9 31.3 46.9









#6
Benzaldehyde
Concen: 147.057 ng/ul
RT: 7.111 min Scan# 60
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion: 77 Resp: 835788
Ion Ratio Lower Upper
77 100
105 90.4 74.6 111.8
106 87.9 68.3 102.5



#8
Phenol
Concen: 94.444 ng/ul
RT: 7.163 min Scan# 676
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion: 94 Resp: 1377729
Ion Ratio Lower Upper
94 100
65 30.1 24.2 36.2
66 39.4 31.4 47.2









#10
Bis(2-Chloroethyl)ether
Concen: 92.827 ng/ul
RT: 7.387 min Scan# 714
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion: 93 Resp: 1077406
Ion Ratio Lower Upper
93 100
63 81.0 65.7 98.5
95 32.3 26.4 39.6



#12
2-Chlorophenol
Concen: 87.249 ng/ul
RT: 7.522 min Scan# 737
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:128 Resp: 983625
Ion Ratio Lower Upper
128 100
64 54.0 44.6 67.0
130 32.1 26.5 39.7





#13

2-Methylphenol

Concen: 93.364 ng/ul

RT: 8.428 min Scan# 891

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :





SSTD080643

Tgt Ion:108 Resp: 1028315

Ion Ratio Lower Upper

108 100

107 88.7 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 107.724 ng/ul

RT: 8.481 min Scan# 900

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion: 45 Resp: 1707707


Ion Ratio Lower Upper

45 100

77 14.2 12.7 19.1

79 10.3 9.3 13.9









#16
Acetophenone
Concen: 91.916 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:105 Resp: 1677956
Ion Ratio Lower Upper
105 100
77 83.9 68.6 102.8
51 30.6 25.0 37.4



#17
N-Nitroso-di-n-propylamine
Concen: 92.605 ng/ul
RT: 8.787 min Scan# 952
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion: 70 Resp: 938923
Ion Ratio Lower Upper
70 100
42 61.2 50.5 75.7
101 8.9 7.8 11.6
130 18.3 14.2 21.4









#18
4-Methylphenol
Concen: 94.512 ng/ul
RT: 8.763 min Scan# 94
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:108 Resp: 1118455
Ion Ratio Lower Upper
108 100
107 113.5 92.5 138.7



#19
Hexachloroethane
Concen: 79.262 ng/ul
RT: 9.028 min Scan# 993
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:117 Resp: 412711
Ion Ratio Lower Upper
117 100
201 93.1 73.5 110.3
199 57.3 46.6 69.8









#22
Nitrobenzene
Concen: 82.877 ng/ul
RT: 9.193 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion: 77 Resp: 1373793
Ion Ratio Lower Upper
77 100
123 40.8 30.5 45.7
65 13.4 10.7 16.1



#23
Isophorone
Concen: 84.297 ng/ul
RT: 9.722 min Scan# 1111
Delta R.T. 0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion: 82 Resp: 2674003
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.0
138 15.0 12.4 18.6









#25
2-Nitrophenol
Concen: 82.240 ng/ul
RT: 9.910 min Scan# 1143
Delta R.T. -0.012 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:139 Resp: 601306
Ion Ratio Lower Upper
139 100
65 58.4 48.6 72.8
109 33.0 26.7 40.1



#26
2,4-Dimethylphenol
Concen: 77.316 ng/ul
RT: 9.963 min Scan# 1152
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:107 Resp: 1242708
Ion Ratio Lower Upper
107 100
121 48.7 37.4 56.2
122 80.7 66.0 99.0





#27

Bis(2-Chloroethoxy)methane

Concen: 86.865 ng/ul

RT: 10.193 min Scan# 1193

Delta R.T. -0.012 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion: 93 Resp: 1555506

Ion Ratio Lower Upper

93 100

95 32.4 26.2 39.2

123 9.9 8.0 12.0



#29

2,4-Dichlorophenol

Concen: 80.298 ng/ul

RT: 10.446 min Scan# 1234

Delta R.T. -0.035 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion: 162 Resp: 980067


Ion Ratio Lower Upper

162 100

164 65.3 54.4 81.6

98 40.4 32.6 49.0









#30
Naphthalene
Concen: 76.991 ng/ul
RT: 10.828 min Scan# 1299
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:128 Resp: 3232747
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 85.888 ng/ul
RT: 10.969 min Scan# 1323
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:127 Resp: 1407284
Ion Ratio Lower Upper
127 100
129 31.9 24.6 37.0





#33

Hexachlorobutadiene

Concen: 62.776 ng/ul

RT: 11.093 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:225 Resp: 578621

Ion Ratio Lower Upper

225 100

223 62.4 50.6 75.8

227 63.2 50.6 75.8



#34

Caprolactam

Concen: 48.145 ng/ul

RT: 11.799 min Scan# 1464

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:113 Resp: 179748


Ion Ratio Lower Upper

113 100

55 196.6 154.7 232.1

56 142.4 115.0 172.4









#35
4-Chloro-3-methylphenol
Concen: 83.226 ng/ul
RT: 12.104 min Scan# 1516
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:107 Resp: 1189041
Ion Ratio Lower Upper
107 100
144 23.8 20.0 30.0
142 74.3 60.6 91.0



#36
2-Methylnaphthalene
Concen: 80.458 ng/ul
RT: 12.446 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:142 Resp: 2223929
Ion Ratio Lower Upper
142 100
141 88.6 72.5 108.7









#37
1-Methylnaphthalene
Concen: 80.044 ng/ul
RT: 12.663 min Scan# 1611
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:142 Resp: 2257039
Ion Ratio Lower Upper
142 100
141 92.6 73.4 110.2
116 4.0 3.5 5.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 72.138 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:216 Resp: 1167660
Ion Ratio Lower Upper
216 100
214 78.9 61.3 91.9
179 20.8 16.9 25.3
108 18.7 16.6 24.8









#40
Hexachlorocyclopentadiene
Concen: 102.776 ng/ul
RT: 12.769 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:237 Resp: 497681
Ion Ratio Lower Upper
237 100
235 63.0 49.4 74.0
272 12.6 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 78.333 ng/ul
RT: 13.069 min Scan# 1680
Delta R.T. -0.012 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:196 Resp: 801306
Ion Ratio Lower Upper
196 100
198 91.7 75.1 112.7
200 30.8 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 78.646 ng/ul

RT: 13.157 min Scan# 1

Delta R.T. -0.024 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:196 Resp: 853326

Ion Ratio Lower Upper

196 100

198 97.8 75.0 112.6

200 31.9 24.7 37.1



#43

1,1'-Biphenyl

Concen: 75.319 ng/ul

RT: 13.451 min Scan# 1745

Delta R.T. 0.000 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:154 Resp: 2930327


Ion Ratio Lower Upper

154 100

153 39.6 31.8 47.6

76 15.4 12.7 19.1





#44

2-Chloronaphthalene

Concen: 75.584 ng/ul

RT: 13.498 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:162 Resp: 2316545

Ion Ratio Lower Upper

162 100

164 32.7 26.0 39.0

127 40.6 32.1 48.1



#45

2-Nitroaniline

Concen: 93.384 ng/ul

RT: 13.734 min Scan# 1793

Delta R.T. -0.018 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion: 65 Resp: 900560


Ion Ratio Lower Upper

65 100

92 65.2 51.6 77.4

138 90.7 71.2 106.8





#47

Dimethylphthalate

Concen: 77.828 ng/ul

RT: 14.081 min Scan# 1852

Delta R.T. 0.000 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:163 Resp: 3057903

Ion Ratio Lower Upper

163 100

194 3.8 2.9 4.3

164 10.0 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 83.959 ng/ul

RT: 14.222 min Scan# 1876

Delta R.T. -0.012 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:165 Resp: 669148


Ion Ratio Lower Upper

165 100

89 67.2 56.0 84.0

121 22.5 18.4 27.6









#50
Acenaphthylene
Concen: 82.738 ng/ul
RT: 14.345 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:152 Resp: 3890764
Ion Ratio Lower Upper
152 100
151 19.8 15.4 23.2
153 13.1 10.2 15.4



#51
3-Nitroaniline
Concen: 108.267 ng/ul
RT: 14.563 min Scan# 1934
Delta R.T. -0.024 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:138 Resp: 676660
Ion Ratio Lower Upper
138 100
108 10.7 8.8 13.2
92 129.6 102.6 153.8









#52
Acenaphthene
Concen: 77.032 ng/ul
RT: 14.681 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:153 Resp: 2511942
Ion Ratio Lower Upper
153 100
152 47.9 38.4 57.6
154 88.2 70.3 105.5



#53
2,4-Dinitrophenol
Concen: 98.494 ng/ul
RT: 14.798 min Scan# 1974
Delta R.T. -0.035 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:184 Resp: 416473
Ion Ratio Lower Upper
184 100
63 81.0 65.4 98.2
154 65.6 51.8 77.8









#55
4-Nitrophenol
Concen: 85.406 ng/ul
RT: 14.892 min Scan# 1995
Delta R.T. -0.029 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:109 Resp: 443415
Ion Ratio Lower Upper
109 100
139 120.0 92.5 138.7
65 142.3 112.9 169.3



#56
Dibenzofuran
Concen: 74.561 ng/ul
RT: 15.022 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:168 Resp: 3375232
Ion Ratio Lower Upper
168 100
139 39.0 30.8 46.2





#57

2,4-Dinitrotoluene

Concen: 86.969 ng/ul

RT: 15.034 min Scan# 2014

Delta R.T. -0.012 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:165 Resp: 944556

Ion Ratio Lower Upper

165 100

63 60.5 47.8 71.6

182 2.6 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 78.481 ng/ul

RT: 15.257 min Scan# 2052

Delta R.T. -0.012 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Tgt Ion:232 Resp: 731426




Ion Ratio Lower Upper


232 100

131 51.0 39.0 58.4

130 2.8 2.0 3.0

166 34.1 25.6 38.4









#59
Diethylphthalate
Concen: 79.151 ng/ul
RT: 15.440 min Scan# 2082
Delta R.T. 0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:149 Resp: 3147102
Ion Ratio Lower Upper
149 100
177 21.8 17.4 26.0
150 12.3 9.9 14.9



#61
Fluorene
Concen: 75.565 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:166 Resp: 2774541
Ion Ratio Lower Upper
166 100
165 96.7 78.6 117.8
167 13.7 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 71.453 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:204 Resp: 1342456
Ion Ratio Lower Upper
204 100
206 33.2 26.3 39.5
141 61.4 48.6 73.0



#63
4-Nitroaniline
Concen: 151.047 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:138 Resp: 646112
Ion Ratio Lower Upper
138 100
92 57.3 47.9 71.9
108 90.3 68.6 103.0









#66
4,6-Dinitro-2-methylphenol
Concen: 89.271 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. -0.018 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:198 Resp: 618679
Ion Ratio Lower Upper
198 100
182 3.8 4.9 7.3#
77 31.2 27.1 40.7



#67
N-Nitrosodiphenylamine
Concen: 72.380 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:169 Resp: 2413350
Ion Ratio Lower Upper
169 100
168 66.4 53.0 79.4
167 36.4 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 68.553 ng/ul
RT: 16.557 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:248 Resp: 871900
Ion Ratio Lower Upper
248 100
250 97.8 80.6 120.8
141 78.6 62.6 94.0



#69
Hexachlorobenzene
Concen: 67.863 ng/ul
RT: 16.681 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:284 Resp: 1018438
Ion Ratio Lower Upper
284 100
142 38.0 32.8 49.2
249 31.3 25.6 38.4





#70

Atrazine

Concen: 77.460 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:200 Resp: 988635

Ion Ratio Lower Upper

200 100

173 26.4 20.7 31.1

215 47.4 36.9 55.3



#71

Pentachlorophenol

Concen: 81.618 ng/ul

RT: 17.039 min Scan# 2355

Delta R.T. -0.018 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:266 Resp: 587866


Ion Ratio Lower Upper

266 100

264 63.5 48.9 73.3

268 63.5 49.7 74.5





#72

Phenanthrene

Concen: 75.868 ng/ul

RT: 17.428 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:178 Resp: 4590531

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 19.6 15.5 23.3



#74

Anthracene

Concen: 77.779 ng/ul

RT: 17.522 min Scan# 2437

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:178 Resp: 4728905


Ion Ratio Lower Upper

178 100

179 15.8 12.2 18.4

176 19.2 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 67.838 ng/uL

RT: 13.416 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:216 Resp: 1229454

Ion Ratio Lower Upper

216 100

214 79.3 63.6 95.4

218 48.4 39.4 59.0



#76

Pentachlorobenzene

Concen: 60.508 ng/uL

RT: 14.940 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:250 Resp: 1106770


Ion Ratio Lower Upper

250 100

248 63.6 50.9 76.3

252 63.4 50.3 75.5









#77
 Carbazole
 Concen: 86.367 ng/ul
 RT: 17.804 min Scan# 2485
 Delta R.T. -0.018 min
 Lab File: BP016188.D
 Acq: 11 Jul 2023 18:51




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080643


Tgt Ion:167 Resp: 4425856
 Ion Ratio Lower Upper
 167 100
 166 21.9 16.5 24.7
 139 13.2 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 82.994 ng/ul
 RT: 18.316 min Scan# 2572
 Delta R.T. -0.006 min
 Lab File: BP016188.D
 Acq: 11 Jul 2023 18:51

Tgt Ion:149 Resp: 5638296
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.2
 104 6.0 4.4 6.6





#80

Fluoranthene

Concen: 61.923 ng/ul

RT: 19.392 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:202 Resp: 5570178

Ion Ratio Lower Upper

202 100

101 12.9 10.4 15.6

100 10.0 8.1 12.1



#82

Pyrene

Concen: 63.631 ng/ul

RT: 19.745 min Scan# 2815

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:202 Resp: 5838777


Ion Ratio Lower Upper

202 100

101 15.4 12.6 19.0

100 12.8 10.9 16.3





#83

Butylbenzylphthalate

Concen: 74.535 ng/ul

RT: 20.598 min Scan# 2960

Delta R.T. 0.000 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:149 Resp: 2790247

Ion Ratio Lower Upper

149 100

91 79.8 64.6 96.8

206 20.7 15.1 22.7



#84

3,3'-Dichlorobenzidine

Concen: 91.233 ng/ul

RT: 21.386 min Scan# 3094

Delta R.T. -0.012 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:252 Resp: 2286072


Ion Ratio Lower Upper

252 100

254 65.2 50.2 75.2

126 12.0 9.5 14.3





#85

Benzo(a)anthracene

Concen: 75.391 ng/ul

RT: 21.457 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:228 Resp: 5878605

Ion Ratio Lower Upper

228 100

229 20.1 15.5 23.3

226 27.9 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 80.132 ng/ul

RT: 21.339 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:149 Resp: 4122558


Ion Ratio Lower Upper

149 100

167 26.3 21.0 31.4

279 3.9 3.4 5.0









#87
Chrysene
Concen: 77.114 ng/ul
RT: 21.516 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51




Instrument :
BNA_P
ClientSampleId :
SSTD080643


Tgt Ion:228 Resp: 5704855
Ion Ratio Lower Upper
228 100
226 30.8 23.5 35.3
229 20.3 16.1 24.1



#89
Di-n-octyl phthalate
Concen: 71.375 ng/ul
RT: 22.286 min Scan# 3247
Delta R.T. 0.000 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:149 Resp: 7290025





#90

Benzo(b)fluoranthene

Concen: 72.357 ng/ul

RT: 23.204 min Scan# 3404

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:252 Resp: 6497746

Ion Ratio Lower Upper

252 100

253 22.7 18.0 27.0

125 10.8 9.1 13.7



#91

Benzo(k)fluoranthene

Concen: 71.178 ng/ul

RT: 23.257 min Scan# 3412

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:252 Resp: 6458122


Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3

125 10.5 9.3 13.9





#93

Benzo(a)pyrene

Concen: 81.608 ng/ul

RT: 23.863 min Scan# 3516

Delta R.T. -0.006 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51

Instrument :

BNA_P

ClientSampleId :

SSTD080643





Tgt Ion:252 Resp: 6403555

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 12.0 10.4 15.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 77.365 ng/ul

RT: 26.633 min Scan# 3986

Delta R.T. 0.012 min

Lab File: BP016188.D

Acq: 11 Jul 2023 18:51




Tgt Ion:276 Resp: 8240920


Ion Ratio Lower Upper

276 100

138 24.8 20.4 30.6

277 25.9 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 77.283 ng/ul
RT: 26.633 min Scan# 3984
Delta R.T. 0.012 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Instrument :
BNA_P
ClientSampleId :
SSTD080643

Tgt Ion:278 Resp: 6761759
Ion Ratio Lower Upper
278 100
139 15.3 14.2 21.2
279 23.6 18.8 28.2



#96
Benzo(g,h,i)perylene
Concen: 78.737 ng/ul
RT: 27.433 min Scan# 4122
Delta R.T. -0.012 min
Lab File: BP016188.D
Acq: 11 Jul 2023 18:51

Tgt Ion:276 Resp: 6899865
Ion Ratio Lower Upper
276 100
138 22.3 18.9 28.3
277 24.0 19.0 28.4

