

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071123\
 Data File : BP016190.D
 Acq On : 11 Jul 2023 20:00
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV645

Quant Time: Jul 12 01:32:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 12 01:28:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	143993	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	692150	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	467258	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.381	188	1083501	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1130198	20.000	ng/u1	0.00
88) Perylene-d12	23.986	264	1355673	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	33687	8.541	ng/uL	0.00
4) Pyridine-d5	3.740	84	199221	20.309	ng/u1	0.00
7) Phenol-d5	7.152	99	284576	21.966	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	207785	23.970	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	212932	22.643	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	230697	22.330	ng/u1	0.00
21) Nitrobenzene-d5	9.158	128	109946	23.950	ng/u1	0.00
24) 2-Nitrophenol-d4	9.887	143	125809	23.096	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.452	165	225083	24.051	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	321050	21.507	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	800404	22.560	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	869392	22.346	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	71538	14.274	ng/u1	0.00
60) Fluorene-d10	15.616	176	658853	22.369	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	136353	22.427	ng/u1	0.00
73) Anthracene-d10	17.486	188	1061331	22.352	ng/u1	0.00
81) Pyrene-d10	19.722	212	1311141	22.149	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.810	264	1454228	22.520	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.323	88	36436	8.222	ng/uL	95
5) Pyridine	3.758	79	205471	19.738	ng/u1	97
6) Benzaldehyde	7.122	77	190736	22.421	ng/u1	97
8) Phenol	7.181	94	309017	22.003	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.387	93	249889	22.197	ng/u1	98
12) 2-Chlorophenol	7.528	128	225854	23.117	ng/u1	98
13) 2-Methylphenol	8.434	108	239159	22.732	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.499	45	391473	21.224	ng/u1	97
16) Acetophenone	8.816	105	407276	23.354	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.787	70	233742	24.528	ng/u1	98
18) 4-Methylphenol	8.781	108	253554	22.023	ng/u1	97
19) Hexachloroethane	9.028	117	95457	21.871	ng/u1	99
22) Nitrobenzene	9.205	77	323325	23.592	ng/u1	97
23) Isophorone	9.716	82	637941	23.331	ng/u1	99
25) 2-Nitrophenol	9.922	139	137856	23.456	ng/u1	99
26) 2,4-Dimethylphenol	9.981	107	303843	23.693	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.205	93	371137	23.575	ng/u1	99
29) 2,4-Dichlorophenol	10.481	162	231840	24.197	ng/u1	99
30) Naphthalene	10.834	128	798350	22.012	ng/u1	99
32) 4-Chloroaniline	10.987	127	345688	23.338	ng/u1	98
33) Hexachlorobutadiene	11.093	225	141481	22.128	ng/u1	99
34) Caprolactam	11.804	113	93984	27.034	ng/u1	93
35) 4-Chloro-3-methylphenol	12.122	107	290519	24.300	ng/u1	99
36) 2-Methylnaphthalene	12.451	142	564275	23.125	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071123\
 Data File : BP016190.D
 Acq On : 11 Jul 2023 20:00
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV645

Quant Time: Jul 12 01:32:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 12 01:28:58 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	542998	21.766	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.822	216	297549	22.464	ng/ul	99
40) Hexachlorocyclopentadiene	12.769	237	74048	16.491	ng/ul	97
41) 2,4,6-Trichlorophenol	13.081	196	196490	23.400	ng/ul	98
42) 2,4,5-Trichlorophenol	13.181	196	204154	23.297	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	773408	22.709	ng/ul	99
44) 2-Chloronaphthalene	13.504	162	595202	22.251	ng/ul	98
45) 2-Nitroaniline	13.746	65	213519	24.672	ng/ul	98
47) Dimethylphthalate	14.081	163	820680	22.813	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	174976	26.373	ng/ul	98
50) Acenaphthylene	14.345	152	976764	21.616	ng/ul	99
51) 3-Nitroaniline	14.581	138	179327	25.636	ng/ul	98
52) Acenaphthene	14.681	153	661152	21.827	ng/ul	100
53) 2,4-Dinitrophenol	14.834	184	78831	20.492	ng/ul	97
55) 4-Nitrophenol	14.928	109	108746	20.023	ng/ul	83
56) Dibenzofuran	15.022	168	944333	22.975	ng/ul	98
57) 2,4-Dinitrotoluene	15.045	165	243068	24.735	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.269	232	203617	25.577	ng/ul	96
59) Diethylphthalate	15.434	149	846048	22.945	ng/ul	100
61) Fluorene	15.669	166	770502	22.700	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.657	204	366936	22.422	ng/ul	98
63) 4-Nitroaniline	15.763	138	158630	24.947	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.816	198	138871	21.884	ng/ul#	98
67) N-Nitrosodiphenylamine	15.881	169	668670	23.000	ng/ul	98
68) 4-Bromophenyl-phenylether	16.557	248	230454	22.474	ng/ul	96
69) Hexachlorobenzene	16.681	284	275047	22.482	ng/ul	96
70) Atrazine	16.845	200	257574	22.938	ng/ul	99
71) Pentachlorophenol	17.057	266	121214	20.079	ng/ul	99
72) Phenanthrene	17.428	178	1266677	22.443	ng/ul	100
74) Anthracene	17.522	178	1282856	22.568	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.422	216	302265	20.782	ng/uL	97
76) Pentachlorobenzene	14.940	250	303827	22.557	ng/uL	100
77) Carbazole	17.816	167	1182024	23.391	ng/ul	98
78) Di-n-butylphthalate	18.316	149	1510921	23.168	ng/ul	100
80) Fluoranthene	19.398	202	1534464	22.179	ng/ul	100
82) Pyrene	19.751	202	1634010	22.012	ng/ul	99
83) Butylbenzylphthalate	20.598	149	741587	23.133	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.398	252	656429	25.333	ng/ul	100
85) Benzo(a)anthracene	21.457	228	1661022	22.736	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.339	149	1131147	23.734	ng/ul	99
87) Chrysene	21.522	228	1629136	21.946	ng/ul	98
89) Di-n-octyl phthalate	22.286	149	1988440	23.805	ng/ul	100
90) Benzo(b)fluoranthene	23.204	252	1800245	23.310	ng/ul	99
91) Benzo(k)fluoranthene	23.257	252	1766603	22.462	ng/ul	99
93) Benzo(a)pyrene	23.869	252	1592064	20.977	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.615	276	2152715	23.216	ng/ul	99
95) Dibenzo(a,h)anthracene	26.621	278	1754696	23.214	ng/ul	100
96) Benzo(g,h,i)perylene	27.445	276	1718010	22.886	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071123\
 Data File : BP016190.D
 Acq On : 11 Jul 2023 20:00
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV645

Quant Time: Jul 12 01:32:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 12 01:28:58 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.222 ng/uL
RT: 3.323 min Scan# 21
Delta R.T. 0.001 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion: 88 Resp: 36436
Ion Ratio Lower Upper
88 100
43 32.3 27.5 41.3
58 73.1 62.2 93.4



#5
Pyridine
Concen: 19.738 ng/uL
RT: 3.758 min Scan# 97
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion: 79 Resp: 205471
Ion Ratio Lower Upper
79 100
52 87.5 67.4 101.0
51 38.2 31.3 46.9









#6
Benzaldehyde
Concen: 22.421 ng/ul
RT: 7.122 min Scan# 609
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion: 77 Resp: 190736
Ion Ratio Lower Upper
77 100
105 89.5 74.6 111.8
106 87.8 68.3 102.5



#8
Phenol
Concen: 22.003 ng/ul
RT: 7.181 min Scan# 679
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion: 94 Resp: 309017
Ion Ratio Lower Upper
94 100
65 30.6 24.2 36.2
66 40.7 31.4 47.2









#10
Bis(2-Chloroethyl)ether
Concen: 22.197 ng/ul
RT: 7.387 min Scan# 715
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion: 93 Resp: 249889
Ion Ratio Lower Upper
93 100
63 80.5 65.7 98.5
95 32.5 26.4 39.6



#12
2-Chlorophenol
Concen: 23.117 ng/ul
RT: 7.528 min Scan# 738
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:128 Resp: 225854
Ion Ratio Lower Upper
128 100
64 57.1 44.6 67.0
130 33.9 26.5 39.7





#13

2-Methylphenol

Concen: 22.732 ng/ul

RT: 8.434 min Scan# 892

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :





SICV645

Tgt Ion:108 Resp: 239159

Ion Ratio Lower Upper

108 100

107 92.7 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.224 ng/ul

RT: 8.499 min Scan# 903

Delta R.T. 0.012 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion: 45 Resp: 391473


Ion Ratio Lower Upper

45 100

77 14.5 12.7 19.1

79 12.5 9.3 13.9









#16
Acetophenone
Concen: 23.354 ng/ul
RT: 8.816 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:105 Resp: 407276
Ion Ratio Lower Upper
105 100
77 90.6 68.6 102.8
51 33.8 25.0 37.4



#17
N-Nitroso-di-n-propylamine
Concen: 24.528 ng/ul
RT: 8.787 min Scan# 952
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion: 70 Resp: 233742
Ion Ratio Lower Upper
70 100
42 61.6 50.5 75.7
101 9.0 7.8 11.6
130 18.4 14.2 21.4





#18

4-Methylphenol

Concen: 22.023 ng/ul

RT: 8.781 min Scan# 91

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :





SICV645

Tgt Ion:108 Resp: 253554

Ion Ratio Lower Upper

108 100

107 112.3 92.5 138.7



#19

Hexachloroethane

Concen: 21.871 ng/ul

RT: 9.028 min Scan# 993

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:117 Resp: 95457


Ion Ratio Lower Upper

117 100

201 93.5 73.5 110.3

199 57.9 46.6 69.8









#22
Nitrobenzene
Concen: 23.592 ng/ul
RT: 9.205 min Scan# 1023
Delta R.T. 0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion: 77 Resp: 323325
Ion Ratio Lower Upper
77 100
123 40.2 30.5 45.7
65 12.8 10.7 16.1



#23
Isophorone
Concen: 23.331 ng/ul
RT: 9.716 min Scan# 1110
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion: 82 Resp: 637941
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.0
138 15.0 12.4 18.6









#25
2-Nitrophenol
Concen: 23.456 ng/ul
RT: 9.922 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:139 Resp: 137856
Ion Ratio Lower Upper
139 100
65 61.7 48.6 72.8
109 33.5 26.7 40.1



#26
2,4-Dimethylphenol
Concen: 23.693 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:107 Resp: 303843
Ion Ratio Lower Upper
107 100
121 49.0 37.4 56.2
122 82.6 66.0 99.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 23.575 ng/ul
 RT: 10.205 min Scan# 1193
 Delta R.T. 0.000 min
 Lab File: BP016190.D
 Acq: 11 Jul 2023 20:00




Instrument :
 BNA_P
 ClientSampleId :
 SICV645


Tgt Ion:	93	Resp:	371137
Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.2	39.2
123	11.0	8.0	12.0



#29
 2,4-Dichlorophenol
 Concen: 24.197 ng/ul
 RT: 10.481 min Scan# 1240
 Delta R.T. 0.000 min
 Lab File: BP016190.D
 Acq: 11 Jul 2023 20:00

Tgt Ion:	162	Resp:	231840
Ion	Ratio	Lower	Upper
162	100		
164	66.7	54.4	81.6
98	40.7	32.6	49.0









#30
Naphthalene
Concen: 22.012 ng/ul
RT: 10.834 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:128 Resp: 798350
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 13.6 10.6 16.0



#32
4-Chloroaniline
Concen: 23.338 ng/ul
RT: 10.987 min Scan# 1326
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:127 Resp: 345688
Ion Ratio Lower Upper
127 100
129 31.7 24.6 37.0









#33
Hexachlorobutadiene
Concen: 22.128 ng/ul
RT: 11.093 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:225 Resp: 141481
Ion Ratio Lower Upper
225 100
223 62.3 50.6 75.8
227 63.6 50.6 75.8



#34
Caprolactam
Concen: 27.034 ng/ul
RT: 11.804 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:113 Resp: 93984
Ion Ratio Lower Upper
113 100
55 187.2 154.7 232.1
56 131.4 115.0 172.4









#35
4-Chloro-3-methylphenol
Concen: 24.300 ng/ul
RT: 12.122 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:107 Resp: 290519
Ion Ratio Lower Upper
107 100
144 24.7 20.0 30.0
142 76.8 60.6 91.0



#36
2-Methylnaphthalene
Concen: 23.125 ng/ul
RT: 12.451 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:142 Resp: 564275
Ion Ratio Lower Upper
142 100
141 88.9 72.5 108.7





#37

1-Methylnaphthalene

Concen: 21.766 ng/ul

RT: 12.663 min Scan# 1611

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:142 Resp: 542998

Ion Ratio Lower Upper

142 100

141 90.7 73.4 110.2

116 4.3 3.5 5.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.464 ng/ul

RT: 12.822 min Scan# 1638

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Tgt Ion:216 Resp: 297549




Ion Ratio Lower Upper


216 100

214 77.6 61.3 91.9

179 20.9 16.9 25.3

108 19.9 16.6 24.8





#40

Hexachlorocyclopentadiene

Concen: 16.491 ng/ul

RT: 12.769 min Scan# 1629

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:237 Resp: 74048

Ion Ratio Lower Upper

237 100

235 59.1 49.4 74.0

272 12.0 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 23.400 ng/ul

RT: 13.081 min Scan# 1682

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:196 Resp: 196490


Ion Ratio Lower Upper

196 100

198 92.6 75.1 112.7

200 29.0 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 23.297 ng/ul
RT: 13.181 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:196 Resp: 204154
Ion Ratio Lower Upper
196 100
198 94.3 75.0 112.6
200 30.1 24.7 37.1



#43
1,1'-Biphenyl
Concen: 22.709 ng/ul
RT: 13.451 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:154 Resp: 773408
Ion Ratio Lower Upper
154 100
153 39.4 31.8 47.6
76 16.2 12.7 19.1





#44

2-Chloronaphthalene

Concen: 22.251 ng/ul

RT: 13.504 min Scan# 1754

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:162 Resp: 595202

Ion Ratio Lower Upper

162 100

164 32.3 26.0 39.0

127 38.6 32.1 48.1



#45

2-Nitroaniline

Concen: 24.672 ng/ul

RT: 13.746 min Scan# 1795

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion: 65 Resp: 213519


Ion Ratio Lower Upper

65 100

92 66.6 51.6 77.4

138 87.7 71.2 106.8









#47
Dimethylphthalate
Concen: 22.813 ng/ul
RT: 14.081 min Scan# 1852
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:163 Resp: 820680
Ion Ratio Lower Upper
163 100
194 3.8 2.9 4.3
164 9.7 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 26.373 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. -0.012 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:165 Resp: 174976
Ion Ratio Lower Upper
165 100
89 68.1 56.0 84.0
121 22.7 18.4 27.6









#50
Acenaphthylene
Concen: 21.616 ng/ul
RT: 14.345 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:152 Resp: 976764
Ion Ratio Lower Upper
152 100
151 19.7 15.4 23.2
153 13.0 10.2 15.4



#51
3-Nitroaniline
Concen: 25.636 ng/ul
RT: 14.581 min Scan# 1937
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:138 Resp: 179327
Ion Ratio Lower Upper
138 100
108 9.9 8.8 13.2
92 126.2 102.6 153.8









#52
Acenaphthene
Concen: 21.827 ng/ul
RT: 14.681 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:153 Resp: 661152
Ion Ratio Lower Upper
153 100
152 48.6 38.4 57.6
154 88.2 70.3 105.5



#53
2,4-Dinitrophenol
Concen: 20.492 ng/ul
RT: 14.834 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:184 Resp: 78831
Ion Ratio Lower Upper
184 100
63 79.1 65.4 98.2
154 67.1 51.8 77.8









#55
4-Nitrophenol
Concen: 20.023 ng/ul
RT: 14.928 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:109 Resp: 108746
Ion Ratio Lower Upper
109 100
139 99.8 92.5 138.7
65 118.1 112.9 169.3



#56
Dibenzofuran
Concen: 22.975 ng/ul
RT: 15.022 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:168 Resp: 944333
Ion Ratio Lower Upper
168 100
139 39.5 30.8 46.2









#57
2,4-Dinitrotoluene
Concen: 24.735 ng/ul
RT: 15.045 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:165 Resp: 243068
Ion Ratio Lower Upper
165 100
63 54.7 47.8 71.6
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 25.577 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:232 Resp: 203617
Ion Ratio Lower Upper
232 100
131 44.9 39.0 58.4
130 2.3 2.0 3.0
166 30.4 25.6 38.4









#59
Diethylphthalate
Concen: 22.945 ng/ul
RT: 15.434 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:149 Resp: 846048
Ion Ratio Lower Upper
149 100
177 21.8 17.4 26.0
150 12.5 9.9 14.9



#61
Fluorene
Concen: 22.700 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:166 Resp: 770502
Ion Ratio Lower Upper
166 100
165 97.5 78.6 117.8
167 13.1 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 22.422 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:204 Resp: 366936
Ion Ratio Lower Upper
204 100
206 33.9 26.3 39.5
141 62.5 48.6 73.0



#63
4-Nitroaniline
Concen: 24.947 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:138 Resp: 158630
Ion Ratio Lower Upper
138 100
92 56.4 47.9 71.9
108 87.4 68.6 103.0









#66
4,6-Dinitro-2-methylphenol
Concen: 21.884 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:198 Resp: 138871
Ion Ratio Lower Upper
198 100
182 4.6 4.9 7.3#
77 33.0 27.1 40.7



#67
N-Nitrosodiphenylamine
Concen: 23.000 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:169 Resp: 668670
Ion Ratio Lower Upper
169 100
168 67.4 53.0 79.4
167 37.6 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 22.474 ng/ul
RT: 16.557 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:248 Resp: 230454
Ion Ratio Lower Upper
248 100
250 96.0 80.6 120.8
141 75.1 62.6 94.0



#69
Hexachlorobenzene
Concen: 22.482 ng/ul
RT: 16.681 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:284 Resp: 275047
Ion Ratio Lower Upper
284 100
142 38.0 32.8 49.2
249 30.3 25.6 38.4





#70

Atrazine

Concen: 22.938 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:200 Resp: 257574

Ion Ratio Lower Upper

200 100

173 26.4 20.7 31.1

215 46.5 36.9 55.3



#71

Pentachlorophenol

Concen: 20.079 ng/ul

RT: 17.057 min Scan# 2358

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:266 Resp: 121214


Ion Ratio Lower Upper

266 100

264 61.4 48.9 73.3

268 61.3 49.7 74.5





#72

Phenanthrene

Concen: 22.443 ng/ul

RT: 17.428 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:178 Resp: 1266677

Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 19.7 15.5 23.3



#74

Anthracene

Concen: 22.568 ng/ul

RT: 17.522 min Scan# 2437

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:178 Resp: 1282856


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.4 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.782 ng/uL

RT: 13.422 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:216 Resp: 302265

Ion Ratio Lower Upper

216 100

214 76.4 63.6 95.4

218 47.0 39.4 59.0



#76

Pentachlorobenzene

Concen: 22.557 ng/uL

RT: 14.940 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:250 Resp: 303827


Ion Ratio Lower Upper

250 100

248 63.0 50.9 76.3

252 63.1 50.3 75.5





#77

Carbazole

Concen: 23.391 ng/ul

RT: 17.816 min Scan# 2488

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:167 Resp: 1182024

Ion Ratio Lower Upper

167 100

166 22.0 16.5 24.7

139 13.1 10.1 15.1



#78

Di-n-butylphthalate

Concen: 23.168 ng/ul

RT: 18.316 min Scan# 2572

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:149 Resp: 1510921


Ion Ratio Lower Upper

149 100

150 9.4 7.4 11.2

104 5.7 4.4 6.6





#80

Fluoranthene

Concen: 22.179 ng/ul

RT: 19.398 min Scan# 2756

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:202 Resp: 1534464

Ion Ratio Lower Upper

202 100

101 12.9 10.4 15.6

100 10.0 8.1 12.1



#82

Pyrene

Concen: 22.012 ng/ul

RT: 19.751 min Scan# 2816

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:202 Resp: 1634010


Ion Ratio Lower Upper

202 100

101 15.3 12.6 19.0

100 13.1 10.9 16.3





#83

Butylbenzylphthalate

Concen: 23.133 ng/ul

RT: 20.598 min Scan# 2960

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:149 Resp: 741587

Ion Ratio Lower Upper

149 100

91 80.1 64.6 96.8

206 20.4 15.1 22.7



#84

3,3'-Dichlorobenzidine

Concen: 25.333 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:252 Resp: 656429


Ion Ratio Lower Upper

252 100

254 62.9 50.2 75.2

126 12.0 9.5 14.3









#85
Benzo(a)anthracene
Concen: 22.736 ng/ul
RT: 21.457 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:228 Resp: 1661022
Ion Ratio Lower Upper
228 100
229 19.3 15.5 23.3
226 27.5 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 23.734 ng/ul
RT: 21.339 min Scan# 3086
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:149 Resp: 1131147
Ion Ratio Lower Upper
149 100
167 25.6 21.0 31.4
279 3.8 3.4 5.0





#87

Chrysene

Concen: 21.946 ng/ul

RT: 21.522 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:228 Resp: 1629136

Ion Ratio Lower Upper

228 100

226 30.2 23.5 35.3

229 19.2 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 23.805 ng/ul




RT: 22.286 min Scan# 3247


Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Tgt Ion:149 Resp: 1988440









#90
Benzo(b)fluoranthene
Concen: 23.310 ng/ul
RT: 23.204 min Scan# 3404
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00




Instrument :
BNA_P
ClientSampleId :
SICV645


Tgt Ion:252 Resp: 1800245
Ion Ratio Lower Upper
252 100
253 22.2 18.0 27.0
125 11.5 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 22.462 ng/ul
RT: 23.257 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BP016190.D
Acq: 11 Jul 2023 20:00

Tgt Ion:252 Resp: 1766603
Ion Ratio Lower Upper
252 100
253 22.3 17.5 26.3
125 11.4 9.3 13.9





#93

Benzo(a)pyrene

Concen: 20.977 ng/ul

RT: 23.869 min Scan# 3516

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:252 Resp: 1592064

Ion Ratio Lower Upper

252 100

253 22.2 17.7 26.5

125 13.3 10.4 15.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.216 ng/ul

RT: 26.615 min Scan# 3983

Delta R.T. -0.006 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00




Tgt Ion:276 Resp: 2152715


Ion Ratio Lower Upper

276 100

138 26.2 20.4 30.6

277 25.3 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 23.214 ng/ul

RT: 26.621 min Scan# 3984

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Instrument :

BNA_P

ClientSampleId :

SICV645





Tgt Ion:278 Resp: 1754696

Ion Ratio Lower Upper

278 100

139 17.7 14.2 21.2

279 23.7 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 22.886 ng/ul

RT: 27.445 min Scan# 4124

Delta R.T. 0.000 min

Lab File: BP016190.D

Acq: 11 Jul 2023 20:00

Tgt Ion:276 Resp: 1718010

Ion Ratio Lower Upper

276 100

138 21.9 18.9 28.3

277 23.9 19.0 28.4

