

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071223\
 Data File : BP016193.D
 Acq On : 12 Jul 2023 09:37
 Operator : MA/JU
 Sample : SST01096
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010647

Quant Time: Jul 13 01:51:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 13 01:50:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	156147	20.000	ng/ul	0.00
20) Naphthalene-d8	10.787	136	733793	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.616	164	490448	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.387	188	1147045	20.000	ng/ul	0.00
79) Chrysene-d12	21.486	240	1214731	20.000	ng/ul	0.01
88) Perylene-d12	23.986	264	1487055	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	16058	3.639	ng/uL	0.00
4) Pyridine-d5	3.752	84	83129	7.554	ng/ul	0.01
7) Phenol-d5	7.175	99	111658	7.917	ng/ul	0.02
9) Bis-(2-Chloroethyl)eth...	7.305	67	83809	8.788	ng/ul	0.01
11) 2-Chlorophenol-d4	7.511	132	92804	9.115	ng/ul	0.02
15) 4-Methylphenol-d8	8.728	113	83404	7.341	ng/ul	0.02
21) Nitrobenzene-d5	9.175	128	43971	8.759	ng/ul	0.02
24) 2-Nitrophenol-d4	9.905	143	48293	8.467	ng/ul	0.02
28) 2,4-Dichlorophenol-d3	10.487	165	66215	6.665	ng/ul	0.04
31) 4-Chloroaniline-d4	10.993	131	124907	7.926	ng/ul	0.03
46) Dimethylphthalate-d6	14.040	166	357483	9.555	ng/ul	0.00
49) Acenaphthylene-d8	14.316	160	389226	9.493	ng/ul	0.00
54) 4-Nitrophenol-d4	14.951	143	48483	9.018	ng/ul	0.05
60) Fluorene-d10	15.622	176	296205	9.460	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.828	200	43537	6.609	ng/ul	0.03
73) Anthracene-d10	17.498	188	476018	9.487	ng/ul	0.01
81) Pyrene-d10	19.733	212	589224	9.348	ng/ul	0.01
92) Benzo(a)pyrene-d12	23.822	264	671522	9.509	ng/ul	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.323	88	17665	3.793	ng/uL	89
5) Pyridine	3.776	79	84365	7.484	ng/ul	97
6) Benzaldehyde	7.152	77	82492	9.482	ng/ul	94
8) Phenol	7.205	94	112154	7.528	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.399	93	109324	8.891	ng/ul	99
12) 2-Chlorophenol	7.546	128	93883	8.887	ng/ul	99
13) 2-Methylphenol	8.458	108	95302	8.375	ng/ul	94
14) 2,2'-oxybis(1-Chloropr...	8.511	45	188870	9.469	ng/ul#	96
16) Acetophenone	8.840	105	163371	8.720	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.793	70	91039	9.079	ng/ul	96
18) 4-Methylphenol	8.811	108	106522	8.560	ng/ul	99
19) Hexachloroethane	9.028	117	45450	9.438	ng/ul	97
22) Nitrobenzene	9.222	77	135143	9.410	ng/ul	98
23) Isophorone	9.728	82	263440	9.165	ng/ul	99
25) 2-Nitrophenol	9.946	139	51224	8.333	ng/ul	98
26) 2,4-Dimethylphenol	10.005	107	125520	9.308	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.234	93	145890	8.787	ng/ul	99
29) 2,4-Dichlorophenol	10.522	162	92024	8.999	ng/ul	98
30) Naphthalene	10.840	128	371962	9.659	ng/ul	99
32) 4-Chloroaniline	11.016	127	135718	8.521	ng/ul	100
33) Hexachlorobutadiene	11.093	225	66868	9.628	ng/ul	99
34) Caprolactam	11.810	113	31852	8.719	ng/ul	85
35) 4-Chloro-3-methylphenol	12.146	107	111322	8.903	ng/ul	99
36) 2-Methylnaphthalene	12.469	142	242299	9.384	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071223\
 Data File : BP016193.D
 Acq On : 12 Jul 2023 09:37
 Operator : MA/JU
 Sample : SSTD01096
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD010647


Quant Time: Jul 13 01:51:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 13 01:50:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	255370	9.617	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.840	216	131027	9.380	ng/ul	97
40) Hexachlorocyclopentadiene	12.769	237	44471	8.110	ng/ul	99
41) 2,4,6-Trichlorophenol	13.104	196	73133	8.308	ng/ul	98
42) 2,4,5-Trichlorophenol	13.210	196	73285	8.141	ng/ul	97
43) 1,1'-Biphenyl	13.457	154	343991	9.615	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	268684	9.559	ng/ul	99
45) 2-Nitroaniline	13.781	65	77027	8.450	ng/ul	97
47) Dimethylphthalate	14.087	163	365400	9.670	ng/ul	99
48) 2,6-Dinitrotoluene	14.251	165	62484	8.818	ng/ul	97
50) Acenaphthylene	14.346	152	451352	9.585	ng/ul	99
51) 3-Nitroaniline	14.616	138	55809	7.715	ng/ul#	82
52) Acenaphthene	14.687	153	307382	9.651	ng/ul	99
53) 2,4-Dinitrophenol	14.857	184	58558	11.745	ng/ul	95
55) 4-Nitrophenol	14.963	109	23891	4.873	ng/ul	98
56) Dibenzofuran	15.040	168	421305	9.712	ng/ul	98
57) 2,4-Dinitrotoluene	15.063	165	97480	9.358	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.281	232	78557	9.226	ng/ul	98
59) Diethylphthalate	15.440	149	378776	9.639	ng/ul	98
61) Fluorene	15.681	166	351317	9.694	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.663	204	173086	9.875	ng/ul	97
63) 4-Nitroaniline	15.798	138	17878	2.762	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.845	198	44305	6.467	ng/ul#	94
67) N-Nitrosodiphenylamine	15.893	169	292984	9.636	ng/ul	98
68) 4-Bromophenyl-phenylether	16.563	248	102906	9.440	ng/ul	98
69) Hexachlorobenzene	16.687	284	124689	9.498	ng/ul	99
70) Atrazine	16.851	200	107281	8.898	ng/ul	98
71) Pentachlorophenol	17.075	266	45513	6.914	ng/ul	96
72) Phenanthrene	17.434	178	570872	9.552	ng/ul	99
74) Anthracene	17.534	178	568644	9.469	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.434	216	142017	9.464	ng/ul	98
76) Pentachlorobenzene	14.951	250	137648	9.684	ng/ul	97
77) Carbazole	17.839	167	498175	9.098	ng/ul	98
78) Di-n-butylphthalate	18.328	149	633235	9.232	ng/ul	99
80) Fluoranthene	19.410	202	685110	9.295	ng/ul	99
82) Pyrene	19.763	202	733422	9.320	ng/ul	98
83) Butylbenzylphthalate	20.598	149	304949	8.985	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.416	252	240232	8.464	ng/ul	96
85) Benzo(a)anthracene	21.469	228	743069	9.394	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.339	149	461247	9.098	ng/ul	99
87) Chrysene	21.528	228	776191	9.686	ng/ul	100
89) Di-n-octyl phthalate	22.286	149	824410	8.851	ng/ul	100
90) Benzo(b)fluoranthene	23.216	252	777217	9.150	ng/ul	99
91) Benzo(k)fluoranthene	23.269	252	862408	9.977	ng/ul	99
93) Benzo(a)pyrene	23.874	252	792095	9.669	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.639	276	973219	9.425	ng/ul	98
95) Dibenzo(a,h)anthracene	26.645	278	768565	9.262	ng/ul	97
96) Benzo(g,h,i)perylene	27.474	276	744599	9.140	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 13 01:51:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jul 13 01:50:08 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.793 ng/uL
RT: 3.323 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion: 88 Resp: 17665
Ion Ratio Lower Upper
88 100
43 32.1 22.3 33.5
58 68.2 63.1 94.7



#5
Pyridine
Concen: 7.484 ng/uL
RT: 3.776 min Scan# 100
Delta R.T. 0.018 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion: 79 Resp: 84365
Ion Ratio Lower Upper
79 100
52 82.3 67.7 101.5
51 42.5 31.7 47.5









#6
Benzaldehyde
Concen: 9.482 ng/ul
RT: 7.152 min Scan# 61
Delta R.T. 0.030 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion: 77 Resp: 82492
Ion Ratio Lower Upper
77 100
105 91.7 69.4 104.2
106 89.7 66.2 99.4



#8
Phenol
Concen: 7.528 ng/ul
RT: 7.205 min Scan# 683
Delta R.T. 0.024 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion: 94 Resp: 112154
Ion Ratio Lower Upper
94 100
65 30.6 25.5 38.3
66 41.0 32.7 49.1









#10
Bis(2-Chloroethyl)ether
Concen: 8.891 ng/ul
RT: 7.399 min Scan# 714
Delta R.T. 0.012 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion: 93 Resp: 109324
Ion Ratio Lower Upper
93 100
63 81.0 63.5 95.3
95 31.6 25.1 37.7



#12
2-Chlorophenol
Concen: 8.887 ng/ul
RT: 7.546 min Scan# 741
Delta R.T. 0.018 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion: 128 Resp: 93883
Ion Ratio Lower Upper
128 100
64 55.6 44.1 66.1
130 30.9 24.3 36.5





#13

2-Methylphenol

Concen: 8.375 ng/ul

RT: 8.458 min Scan# 89

Delta R.T. 0.024 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :





SSTD010647

Tgt Ion:108 Resp: 95302

Ion Ratio Lower Upper

108 100

107 93.6 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.469 ng/ul

RT: 8.511 min Scan# 905

Delta R.T. 0.024 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion: 45 Resp: 188870


Ion Ratio Lower Upper

45 100

77 16.2 12.5 18.7

79 13.6 8.7 13.1#









#16
Acetophenone
Concen: 8.720 ng/ul
RT: 8.840 min Scan# 90
Delta R.T. 0.024 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:105 Resp: 163371
Ion Ratio Lower Upper
105 100
77 89.5 70.2 105.2
51 35.5 26.3 39.5



#17
N-Nitroso-di-n-propylamine
Concen: 9.079 ng/ul
RT: 8.793 min Scan# 953
Delta R.T. 0.012 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion: 70 Resp: 91039
Ion Ratio Lower Upper
70 100
42 60.0 50.8 76.2
101 9.1 7.0 10.6
130 17.9 13.3 19.9









#18
4-Methylphenol
Concen: 8.560 ng/ul
RT: 8.811 min Scan# 91
Delta R.T. 0.030 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:108 Resp: 106522
Ion Ratio Lower Upper
108 100
107 112.2 89.3 133.9



#19
Hexachloroethane
Concen: 9.438 ng/ul
RT: 9.028 min Scan# 993
Delta R.T. 0.000 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:117 Resp: 45450
Ion Ratio Lower Upper
117 100
201 98.3 76.5 114.7
199 56.8 47.9 71.9









#22
Nitrobenzene
Concen: 9.410 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. 0.024 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion: 77 Resp: 135143
Ion Ratio Lower Upper
77 100
123 41.1 31.8 47.8
65 12.6 10.4 15.6



#23
Isophorone
Concen: 9.165 ng/ul
RT: 9.728 min Scan# 1112
Delta R.T. 0.012 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion: 82 Resp: 263440
Ion Ratio Lower Upper
82 100
95 7.3 5.4 8.2
138 14.9 12.3 18.5









#25
2-Nitrophenol
Concen: 8.333 ng/ul
RT: 9.946 min Scan# 1145
Delta R.T. 0.024 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:	139	Resp:	51224
Ion Ratio	Lower	Upper	
139	100		
65	61.6	50.0	75.0
109	32.8	28.2	42.2



#26
2,4-Dimethylphenol
Concen: 9.308 ng/ul
RT: 10.005 min Scan# 1159
Delta R.T. 0.024 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:	107	Resp:	125520
Ion Ratio	Lower	Upper	
107	100		
121	50.9	37.2	55.8
122	82.3	64.7	97.1









#27
Bis(2-Chloroethoxy)methane
Concen: 8.787 ng/ul
RT: 10.234 min Scan# 1193
Delta R.T. 0.030 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion: 93 Resp: 145890
Ion Ratio Lower Upper
93 100
95 31.8 25.9 38.9
123 10.5 8.6 13.0



#29
2,4-Dichlorophenol
Concen: 8.999 ng/ul
RT: 10.522 min Scan# 1247
Delta R.T. 0.042 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion: 162 Resp: 92024
Ion Ratio Lower Upper
162 100
164 66.7 54.3 81.5
98 38.9 31.8 47.8





#30

Naphthalene

Concen: 9.659 ng/ul

RT: 10.840 min Scan# 1301

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:128 Resp: 371962

Ion Ratio Lower Upper

128 100

129 11.2 8.9 13.3

127 13.2 10.9 16.3



#32

4-Chloroaniline

Concen: 8.521 ng/ul

RT: 11.016 min Scan# 1331

Delta R.T. 0.030 min

Lab File: BP016193.D




Acq: 12 Jul 2023 09:37


Tgt Ion:127 Resp: 135718

Ion Ratio Lower Upper

127 100

129 32.4 26.1 39.1





#33

Hexachlorobutadiene

Concen: 9.628 ng/ul

RT: 11.093 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:225 Resp: 66868

Ion Ratio Lower Upper

225 100

223 63.7 49.8 74.6

227 63.8 51.7 77.5



#34

Caprolactam

Concen: 8.719 ng/ul

RT: 11.810 min Scan# 1466

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:113 Resp: 31852


Ion Ratio Lower Upper

113 100

55 202.9 147.8 221.8

56 151.3 103.8 155.6









#35
4-Chloro-3-methylphenol
Concen: 8.903 ng/ul
RT: 12.146 min Scan# 11
Delta R.T. 0.024 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:107 Resp: 111322
Ion Ratio Lower Upper
107 100
144 25.0 20.3 30.5
142 75.6 60.8 91.2



#36
2-Methylnaphthalene
Concen: 9.384 ng/ul
RT: 12.469 min Scan# 1578
Delta R.T. 0.018 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:142 Resp: 242299
Ion Ratio Lower Upper
142 100
141 89.4 71.9 107.9





#37

1-Methylnaphthalene

Concen: 9.617 ng/ul

RT: 12.675 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:142 Resp: 255370

Ion Ratio Lower Upper

142 100

141 92.5 73.9 110.9

116 4.2 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.380 ng/ul

RT: 12.840 min Scan# 1641

Delta R.T. 0.018 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Tgt Ion:216 Resp: 131027




Ion Ratio Lower Upper


216 100

214 76.4 62.9 94.3

179 20.4 17.8 26.6

108 18.7 16.0 24.0





#40

Hexachlorocyclopentadiene

Concen: 8.110 ng/ul

RT: 12.769 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:237 Resp: 44471

Ion Ratio Lower Upper

237 100

235 60.8 48.2 72.2

272 13.3 9.5 14.3



#41

2,4,6-Trichlorophenol

Concen: 8.308 ng/ul

RT: 13.104 min Scan# 1686

Delta R.T. 0.024 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:196 Resp: 73133


Ion Ratio Lower Upper

196 100

198 102.0 79.2 118.8

200 30.0 24.2 36.4





#42

2,4,5-Trichlorophenol

Concen: 8.141 ng/ul

RT: 13.210 min Scan# 1704

Delta R.T. 0.030 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:196 Resp: 73285

Ion Ratio Lower Upper

196 100

198 99.2 77.4 116.2

200 29.1 25.3 37.9



#43

1,1'-Biphenyl

Concen: 9.615 ng/ul

RT: 13.457 min Scan# 1746

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:154 Resp: 343991


Ion Ratio Lower Upper

154 100

153 38.5 31.1 46.7

76 15.7 12.9 19.3





#44

2-Chloronaphthalene

Concen: 9.559 ng/ul

RT: 13.516 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:162 Resp: 268684

Ion Ratio Lower Upper

162 100

164 31.7 26.0 39.0

127 38.9 31.2 46.8



#45

2-Nitroaniline

Concen: 8.450 ng/ul

RT: 13.781 min Scan# 1801

Delta R.T. 0.030 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion: 65 Resp: 77027


Ion Ratio Lower Upper

65 100

92 69.1 54.1 81.1

138 87.8 73.4 110.0









#47
Dimethylphthalate
Concen: 9.670 ng/ul
RT: 14.087 min Scan# 1852
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:163 Resp: 365400
Ion Ratio Lower Upper
163 100
194 3.6 3.0 4.6
164 9.7 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 8.818 ng/ul
RT: 14.251 min Scan# 1881
Delta R.T. 0.018 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:165 Resp: 62484
Ion Ratio Lower Upper
165 100
89 65.8 54.8 82.2
121 23.0 19.4 29.0









#50
Acenaphthylene
Concen: 9.585 ng/ul
RT: 14.346 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:152 Resp: 451352
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.8
153 13.3 10.2 15.4



#51
3-Nitroaniline
Concen: 7.715 ng/ul
RT: 14.616 min Scan# 1943
Delta R.T. 0.030 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:138 Resp: 55809
Ion Ratio Lower Upper
138 100
108 58.3 9.6 14.4#
92 147.6 109.0 163.4





#52

Acenaphthene

Concen: 9.651 ng/ul

RT: 14.687 min Scan# 1954

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:153 Resp: 307382

Ion Ratio Lower Upper

153 100

152 46.8 37.9 56.9

154 87.7 69.3 103.9



#53

2,4-Dinitrophenol

Concen: 11.745 ng/ul

RT: 14.857 min Scan# 1984

Delta R.T. 0.024 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:184 Resp: 58558


Ion Ratio Lower Upper

184 100

63 82.6 67.0 100.6

154 57.6 51.5 77.3





#55

4-Nitrophenol

Concen: 4.873 ng/ul

RT: 14.963 min Scan# 2010

Delta R.T. 0.030 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:109 Resp: 23891

Ion Ratio Lower Upper

109 100

139 91.4 76.8 115.2

65 115.3 92.2 138.4



#56

Dibenzofuran

Concen: 9.712 ng/ul

RT: 15.040 min Scan# 2015

Delta R.T. 0.012 min

Lab File: BP016193.D




Acq: 12 Jul 2023 09:37


Tgt Ion:168 Resp: 421305

Ion Ratio Lower Upper

168 100

139 39.2 30.6 45.8





#57

2,4-Dinitrotoluene

Concen: 9.358 ng/ul

RT: 15.063 min Scan# 2015

Delta R.T. 0.024 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:165 Resp: 97480

Ion Ratio Lower Upper

165 100

63 67.9 56.9 85.3

182 2.6 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 9.226 ng/ul

RT: 15.281 min Scan# 2056

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Tgt Ion:232 Resp: 78557




Ion Ratio Lower Upper


232 100

131 49.7 38.6 57.8

130 2.9 2.2 3.4

166 32.2 25.2 37.8









#59
Diethylphthalate
Concen: 9.639 ng/ul
RT: 15.440 min Scan# 2082
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:149 Resp: 378776
Ion Ratio Lower Upper
149 100
177 21.9 17.4 26.0
150 11.4 10.2 15.2



#61
Fluorene
Concen: 9.694 ng/ul
RT: 15.681 min Scan# 2124
Delta R.T. 0.012 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:166 Resp: 351317
Ion Ratio Lower Upper
166 100
165 99.1 78.6 117.8
167 13.4 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 9.875 ng/ul
RT: 15.663 min Scan# 2119
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:204 Resp: 173086
Ion Ratio Lower Upper
204 100
206 32.5 25.4 38.2
141 57.8 48.2 72.2



#63
4-Nitroaniline
Concen: 2.762 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. 0.030 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:138 Resp: 17878
Ion Ratio Lower Upper
138 100
92 63.4 46.6 70.0
108 88.6 68.3 102.5









#66
4,6-Dinitro-2-methylphenol
Concen: 6.467 ng/ul
RT: 15.845 min Scan# 2147
Delta R.T. 0.030 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:198 Resp: 44305
Ion Ratio Lower Upper
198 100
182 9.0 4.3 6.5#
77 36.0 26.3 39.5



#67
N-Nitrosodiphenylamine
Concen: 9.636 ng/ul
RT: 15.893 min Scan# 2160
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:169 Resp: 292984
Ion Ratio Lower Upper
169 100
168 65.9 54.2 81.4
167 37.9 30.2 45.4









#68
4-Bromophenyl-phenylether
Concen: 9.440 ng/ul
RT: 16.563 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:248 Resp: 102906
Ion Ratio Lower Upper
248 100
250 97.2 78.4 117.6
141 80.3 61.4 92.2



#69
Hexachlorobenzene
Concen: 9.498 ng/ul
RT: 16.687 min Scan# 2295
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:284 Resp: 124689
Ion Ratio Lower Upper
284 100
142 38.4 30.2 45.4
249 31.6 25.0 37.4





#70

Atrazine

Concen: 8.898 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:200 Resp: 107281

Ion Ratio Lower Upper

200 100

173 26.9 21.8 32.8

215 46.6 38.7 58.1



#71

Pentachlorophenol

Concen: 6.914 ng/ul

RT: 17.075 min Scan# 2361

Delta R.T. 0.018 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:266 Resp: 45513


Ion Ratio Lower Upper

266 100

264 57.6 49.8 74.8

268 63.3 49.5 74.3





#72

Phenanthrene

Concen: 9.552 ng/ul

RT: 17.434 min Scan# 2421

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:178 Resp: 570872

Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 19.0 15.6 23.4



#74

Anthracene

Concen: 9.469 ng/ul

RT: 17.534 min Scan# 2439

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:178 Resp: 568644


Ion Ratio Lower Upper

178 100

179 14.4 11.9 17.9

176 18.7 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.464 ng/uL

RT: 13.434 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:216 Resp: 142017

Ion Ratio Lower Upper

216 100

214 76.5 62.6 94.0

218 48.5 38.5 57.7



#76

Pentachlorobenzene

Concen: 9.684 ng/uL

RT: 14.951 min Scan# 2000

Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:250 Resp: 137648


Ion Ratio Lower Upper

250 100

248 60.5 50.5 75.7

252 62.5 51.7 77.5





#77

Carbazole

Concen: 9.098 ng/ul

RT: 17.839 min Scan# 2488

Delta R.T. 0.018 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:167 Resp: 498175

Ion Ratio Lower Upper

167 100

166 21.7 16.8 25.2

139 13.6 10.2 15.4



#78

Di-n-butylphthalate

Concen: 9.232 ng/ul

RT: 18.328 min Scan# 2574

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:149 Resp: 633235


Ion Ratio Lower Upper

149 100

150 8.9 7.4 11.0

104 5.4 4.6 7.0









#80
Fluoranthene
Concen: 9.295 ng/ul
RT: 19.410 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:202 Resp: 685110
Ion Ratio Lower Upper
202 100
101 12.8 10.4 15.6
100 10.6 8.2 12.4



#82
Pyrene
Concen: 9.320 ng/ul
RT: 19.763 min Scan# 2818
Delta R.T. 0.012 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:202 Resp: 733422
Ion Ratio Lower Upper
202 100
101 16.4 12.4 18.6
100 13.5 10.5 15.7









#83
Butylbenzylphthalate
Concen: 8.985 ng/ul
RT: 20.598 min Scan# 2959
Delta R.T. 0.006 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37




Instrument :
BNA_P
ClientSampleId :
SSTD010647


Tgt Ion:149 Resp: 304949
Ion Ratio Lower Upper
149 100
91 84.6 67.1 100.7
206 18.3 15.0 22.4



#84
3,3'-Dichlorobenzidine
Concen: 8.464 ng/ul
RT: 21.416 min Scan# 3099
Delta R.T. 0.018 min
Lab File: BP016193.D
Acq: 12 Jul 2023 09:37

Tgt Ion:252 Resp: 240232
Ion Ratio Lower Upper
252 100
254 66.3 50.6 76.0
126 12.4 9.4 14.0





#85

Benzo(a)anthracene

Concen: 9.394 ng/ul

RT: 21.469 min Scan# 3106

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:228 Resp: 743069

Ion Ratio Lower Upper

228 100

229 19.7 15.5 23.3

226 26.4 21.5 32.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.098 ng/ul

RT: 21.339 min Scan# 3086

Delta R.T. 0.000 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:149 Resp: 461247


Ion Ratio Lower Upper

149 100

167 25.4 21.0 31.4

279 3.6 3.0 4.6





#87

Chrysene

Concen: 9.686 ng/ul

RT: 21.528 min Scan# 3116

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:228 Resp: 776191

Ion Ratio Lower Upper

228 100

226 30.2 24.0 36.0

229 19.6 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 8.851 ng/ul




RT: 22.286 min Scan# 3247


Delta R.T. 0.006 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Tgt Ion:149 Resp: 824410





#90

Benzo(b)fluoranthene

Concen: 9.150 ng/ul

RT: 23.216 min Scan# 3402

Delta R.T. 0.018 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:252 Resp: 777217

Ion Ratio Lower Upper

252 100

253 22.2 17.3 25.9

125 11.6 9.4 14.0



#91

Benzo(k)fluoranthene

Concen: 9.977 ng/ul

RT: 23.269 min Scan# 3414

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:252 Resp: 862408


Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.2

125 11.1 9.3 13.9





#93

Benzo(a)pyrene

Concen: 9.669 ng/ul

RT: 23.874 min Scan# 3515

Delta R.T. 0.012 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:252 Resp: 792095

Ion Ratio Lower Upper

252 100

253 22.6 17.2 25.8

125 13.1 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.425 ng/ul

RT: 26.639 min Scan# 3982

Delta R.T. 0.030 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37




Tgt Ion:276 Resp: 973219


Ion Ratio Lower Upper

276 100

138 25.1 20.8 31.2

277 24.2 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 9.262 ng/ul

RT: 26.645 min Scan# 3988

Delta R.T. 0.024 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Instrument :

BNA_P

ClientSampleId :

SSTD010647





Tgt Ion:278 Resp: 768565

Ion Ratio Lower Upper

278 100

139 18.3 13.9 20.9

279 24.4 18.2 27.4



#96

Benzo(g,h,i)perylene

Concen: 9.140 ng/ul

RT: 27.474 min Scan# 4129

Delta R.T. 0.030 min

Lab File: BP016193.D

Acq: 12 Jul 2023 09:37

Tgt Ion:276 Resp: 744599

Ion Ratio Lower Upper

276 100

138 24.0 19.0 28.4

277 23.3 19.0 28.4

