

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071425\
 Data File : BP025126.D
 Acq On : 14 Jul 2025 19:09
 Operator : CG/JU
 Sample : Q2553-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 AOC-203

Quant Time: Jul 14 19:38:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 03 16:05:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.437	152	356736	20.000	ng	0.00
21) Naphthalene-d8	10.178	136	1359325	20.000	ng	0.00
39) Acenaphthene-d10	14.084	164	848645	20.000	ng	0.00
64) Phenanthrene-d10	16.895	188	1826157	20.000	ng	-0.01
76) Chrysene-d12	21.324	240	2432841	20.000	ng	-0.02
86) Perylene-d12	24.448	264	2988381	20.000	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.096	112	1448557	71.948	ng	0.00
7) Phenol-d6	6.625	99	1121205	42.674	ng	0.00
23) Nitrobenzene-d5	8.566	82	2104039	102.940	ng	0.00
42) 2,4,6-Tribromophenol	15.607	330	1717749	171.341	ng	-0.01
45) 2-Fluorobiphenyl	12.690	172	5059470	92.359	ng	0.00
79) Terphenyl-d14	19.672	244	9975517	90.117	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	10.231	128	775167	11.075	ng	99
37) 2-Methylnaphthalene	11.848	142	210250	4.257	ng	99
38) 1-Methylnaphthalene	12.078	142	459616	9.542	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071425\
Data File : BP025126.D
Acq On : 14 Jul 2025 19:09
Operator : CG/JU
Sample : Q2553-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
AOC-203

Quant Time: Jul 14 19:38:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

