

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071425\
 Data File : BP025121.D
 Acq On : 14 Jul 2025 15:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 16:32:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 03 16:05:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	37.085	7.3	120	0.00
3	Pyridine	40.000	38.382	4.0	121	0.00
4	n-Nitrosodimethylamine	40.000	38.408	4.0	124	0.00
5 S	2-Fluorophenol	80.000	79.708	0.4	128	0.00
6	Aniline	40.000	39.442	1.4	126	0.00
7 S	Phenol-d6	80.000	80.291	-0.4	130	0.00
8	2-Chlorophenol	40.000	40.508	-1.3	131	0.00
9	Benzaldehyde	40.000	39.275	1.8	134	0.00
10 C	Phenol	40.000	39.511	1.2	129	0.00
11	bis(2-Chloroethyl)ether	40.000	38.961	2.6	129	0.00
12	1,3-Dichlorobenzene	40.000	39.341	1.6	130	0.00
13 C	1,4-Dichlorobenzene	40.000	39.496	1.3	130	0.00
14	1,2-Dichlorobenzene	40.000	39.436	1.4	129	0.00
15	Benzyl Alcohol	40.000	41.768	-4.4	138	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.210	7.0	125	0.00
17	2-Methylphenol	40.000	40.796	-2.0	131	0.00
18	Hexachloroethane	40.000	41.537	-3.8	135	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.621	0.9	129	0.00
20	3+4-Methylphenols	40.000	40.815	-2.0	133	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	132	0.00
22	Acetophenone	40.000	38.627	3.4	129	0.00
23 S	Nitrobenzene-d5	80.000	82.515	-3.1	133	0.00
24	Nitrobenzene	40.000	40.808	-2.0	133	0.00
25	Isophorone	40.000	40.050	-0.1	133	0.00
26 C	2-Nitrophenol	40.000	43.921	-9.8	160	0.00
27	2,4-Dimethylphenol	40.000	40.155	-0.4	133	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.519	1.2	133	0.00
29 C	2,4-Dichlorophenol	40.000	41.824	-4.6	134	0.00
30	1,2,4-Trichlorobenzene	40.000	40.269	-0.7	134	0.00
31	Naphthalene	40.000	39.196	2.0	132	0.00
32	Benzoic acid	40.000	42.382	-6.0	154	0.02
33	4-Chloroaniline	40.000	39.799	0.5	129	0.00
34 C	Hexachlorobutadiene	40.000	41.302	-3.3	138	0.00
35	Caprolactam	40.000	39.792	0.5	130	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.633	-4.1	135	0.00
37	2-Methylnaphthalene	40.000	39.691	0.8	132	0.00
38	1-Methylnaphthalene	40.000	39.302	1.7	133	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	133	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.065	-2.7	135	0.00
41 P	Hexachlorocyclopentadiene	40.000	47.038	-17.6	158	-0.01
42 S	2,4,6-Tribromophenol	80.000	90.534	-13.2	145	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.923	-9.8	138	0.00
44	2,4,5-Trichlorophenol	40.000	42.929	-7.3	137	0.00
45 S	2-Fluorobiphenyl	80.000	78.719	1.6	132	0.00
46	1,1'-Biphenyl	40.000	39.468	1.3	131	0.00
47	2-Chloronaphthalene	40.000	39.883	0.3	133	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071425\
 Data File : BP025121.D
 Acq On : 14 Jul 2025 15:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 16:32:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 03 16:05:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.340	-3.4	141	0.00
49	Acenaphthylene	40.000	40.432	-1.1	132	0.00
50	Dimethylphthalate	40.000	39.745	0.6	131	0.00
51	2,6-Dinitrotoluene	40.000	44.258	-10.6	136	0.00
52 C	Acenaphthene	40.000	39.270	1.8	132	0.00
53	3-Nitroaniline	40.000	44.664	-11.7	135	0.00
54 P	2,4-Dinitrophenol	40.000	44.033	-10.1	157	0.00
55	Dibenzofuran	40.000	39.396	1.5	132	0.00
56 P	4-Nitrophenol	40.000	42.357	-5.9	133	0.00
57	2,4-Dinitrotoluene	40.000	41.952	-4.9	142	0.00
58	Fluorene	40.000	39.747	0.6	131	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.672	-9.2	140	0.00
60	Diethylphthalate	40.000	40.657	-1.6	133	0.00
61	4-Chlorophenyl-phenylether	40.000	40.005	-0.0	134	0.00
62	4-Nitroaniline	40.000	43.967	-9.9	131	0.00
63	Azobenzene	40.000	39.068	2.3	129	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	132	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.532	-8.8	158	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.151	-0.4	132	0.00
67	4-Bromophenyl-phenylether	40.000	42.017	-5.0	139	0.00
68	Hexachlorobenzene	40.000	40.644	-1.6	136	0.00
69	Atrazine	40.000	35.680	10.8	115	0.00
70 C	Pentachlorophenol	40.000	44.288	-10.7	144	0.00
71	Phenanthrene	40.000	39.423	1.4	130	0.00
72	Anthracene	40.000	39.471	1.3	129	0.00
73	Carbazole	40.000	39.354	1.6	127	0.00
74	Di-n-butylphthalate	40.000	43.657	-9.1	135	0.00
75 C	Fluoranthene	40.000	40.052	-0.1	128	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	125	0.00
77	Benzydine	40.000	40.945	-2.4	79	0.00
78	Pyrene	40.000	40.388	-1.0	127	0.00
79 S	Terphenyl-d14	80.000	80.478	-0.6	125	0.00
80	Butylbenzylphthalate	40.000	43.786	-9.5	147	0.00
81	Benzo(a)anthracene	40.000	39.713	0.7	123	0.00
82	3,3'-Dichlorobenzidine	40.000	45.835	-14.6	134	-0.02
83	Chrysene	40.000	39.428	1.4	124	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.115	-17.8	136	-0.02
85 c	Di-n-octyl phthalate	40.000	43.571	-8.9	146	-0.03
86 I	Perylene-d12	20.000	20.000	0.0	126	-0.03
87	Indeno(1,2,3-cd)pyrene	40.000	41.298	-3.2	125	-0.04
88	Benzo(b)fluoranthene	40.000	40.529	-1.3	124	-0.02
89	Benzo(k)fluoranthene	40.000	39.000	2.5	124	-0.03
90 C	Benzo(a)pyrene	40.000	41.325	-3.3	125	-0.03
91	Dibenzo(a,h)anthracene	40.000	41.216	-3.0	125	-0.05
92	Benzo(g,h,i)perylene	40.000	40.657	-1.6	124	-0.05

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071425\
Data File : BP025121.D
Acq On : 14 Jul 2025 15:36
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jul 14 16:32:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0