

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071822\  
 Data File : BP010995.D  
 Acq On : 18 Jul 2022 16:26  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :

Quant Time: Jul 19 00:45:53 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP071422.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Jul 14 15:34:34 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.681	152	376451	20.000	ng/ul	0.00
20) Naphthalene-d8	10.475	136	1579913	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.345	164	895776	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.110	188	1797534	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.233	240	1763592	20.000	ng/ul	# 0.00
88) Perylene-d12	23.586	264	1848541	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	75812	7.260	ng/uL	0.00
4) Pyridine-d5	3.587	84	472491	17.506	ng/ul	0.00
7) Phenol-d5	6.858	99	516567	16.717	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	353876	17.809	ng/ul	0.00
11) 2-Chlorophenol-d4	7.216	132	445610	17.846	ng/ul	0.00
15) 4-Methylphenol-d8	8.399	113	417857	16.891	ng/ul	0.00
21) Nitrobenzene-d5	8.846	128	199984	17.369	ng/ul	0.00
24) 2-Nitrophenol-d4	9.563	143	189196	16.607	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.099	165	359552	17.257	ng/ul	0.00
31) 4-Chloroaniline-d4	10.622	131	571902	16.650	ng/ul	0.00
46) Dimethylphthalate-d6	13.763	166	1084535	17.714	ng/ul	0.00
49) Acenaphthylene-d8	14.034	160	1303780	18.049	ng/ul	0.00
54) 4-Nitrophenol-d4	14.557	143	196519	15.860	ng/ul	0.00
60) Fluorene-d10	15.345	176	923844	17.753	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.469	200	136138	15.305	ng/ul	0.00
73) Anthracene-d10	17.210	188	1368960	17.914	ng/ul	0.00
81) Pyrene-d10	19.469	212	1573548	16.989	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.433	264	1569264	17.675	ng/ul	-0.02
Target Compounds						
2) 1,4-Dioxane	3.211	88	79048	7.215	ng/uL#	84
5) Pyridine	3.605	79	494919	17.793	ng/ul#	78
6) Benzaldehyde	6.828	77	327940	21.848	ng/ul	98
8) Phenol	6.887	94	556720	16.949	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.116	93	462092	17.726	ng/ul	91
12) 2-Chlorophenol	7.246	128	464719	18.005	ng/ul	92
13) 2-Methylphenol	8.134	108	410952	16.918	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.222	45	647912	18.172	ng/ul#	98
16) Acetophenone	8.510	105	665345	17.766	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.499	70	335880	17.875	ng/ul	91
18) 4-Methylphenol	8.463	108	441317	17.016	ng/ul	99
19) Hexachloroethane	8.752	117	186271	18.237	ng/ul#	77
22) Nitrobenzene	8.887	77	489157	18.066	ng/ul	93
23) Isophorone	9.416	82	887425	17.512	ng/ul	98
25) 2-Nitrophenol	9.593	139	218366	17.076	ng/ul#	84
26) 2,4-Dimethylphenol	9.663	107	479203	17.720	ng/ul	96
27) Bis(2-Chloroethoxy)met...	9.904	93	580031	17.296	ng/ul	98
29) 2,4-Dichlorophenol	10.128	162	372740	17.284	ng/ul	99
30) Naphthalene	10.528	128	1451365	18.027	ng/ul	100
32) 4-Chloroaniline	10.646	127	579679	17.062	ng/ul	98
33) Hexachlorobutadiene	10.810	225	221362	17.634	ng/ul	96
34) Caprolactam	11.434	113	150861	19.961	ng/ul	91
35) 4-Chloro-3-methylphenol	11.781	107	412745	16.936	ng/ul	95
36) 2-Methylnaphthalene	12.146	142	920894	17.644	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071822\  
 Data File : BP010995.D  
 Acq On : 18 Jul 2022 16:26  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :

Quant Time: Jul 19 00:45:53 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP071422.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Jul 14 15:34:34 2022  
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.369	142	925223	17.685	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.516	216	412990	17.481	ng/ul#	98
40) Hexachlorocyclopentadiene	12.498	237	248034	16.950	ng/ul	95
41) 2,4,6-Trichlorophenol	12.769	196	257542	16.553	ng/ul	97
42) 2,4,5-Trichlorophenol	12.834	196	284834	17.130	ng/ul	97
44) 2-Chloronaphthalene	13.210	162	892216	17.772	ng/ul	99
47) Dimethylphthalate	13.810	163	1081969	17.514	ng/ul	98
48) 2,6-Dinitrotoluene	13.928	165	192146	17.180	ng/ul	98
50) Acenaphthylene	14.063	152	1473382	18.178	ng/ul	98
51) 3-Nitroaniline	14.257	138	222873	16.586	ng/ul	94
52) Acenaphthene	14.404	153	946701	17.631	ng/ul	98
53) 2,4-Dinitrophenol	14.463	184	113837	17.397	ng/ul	89
55) 4-Nitrophenol	14.569	109	181426	19.124	ng/ul#	76
56) Dibenzofuran	14.745	168	1324652	17.878	ng/ul	98
57) 2,4-Dinitrotoluene	14.716	165	285470	17.894	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	14.975	232	219514	16.858	ng/ul#	86
59) Diethylphthalate	15.187	149	1097606	17.301	ng/ul	98
61) Fluorene	15.398	166	1054807	17.740	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.398	204	484433	17.697	ng/ul	94
63) 4-Nitroaniline	15.428	138	221548	17.940	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.481	198	141194	15.473	ng/ul	98
67) N-Nitrosodiphenylamine	15.616	169	901723	17.561	ng/ul	99
68) 4-Bromophenyl-phenylether	16.304	248	279525	17.242	ng/ul	98
69) Hexachlorobenzene	16.404	284	326446	17.350	ng/ul#	92
70) Atrazine	16.586	200	305571	17.203	ng/ul	97
71) Pentachlorophenol	16.757	266	204874	17.680	ng/ul	99
72) Phenanthrene	17.151	178	1655008	17.672	ng/ul	99
74) Anthracene	17.245	178	1682017	18.056	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.134	216	430177	16.717	ng/uL	98
76) Pentachlorobenzene	14.663	250	393404	17.440	ng/uL	99
77) Carbazole	17.522	167	1559041	17.881	ng/ul	100
78) Di-n-butylphthalate	18.098	149	1810436	16.612	ng/ul	99
80) Fluoranthene	19.139	202	1917081	17.697	ng/ul#	87
82) Pyrene	19.498	202	1980479	17.719	ng/ul#	86
83) Butylbenzylphthalate	20.392	149	812846	16.302	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.157	252	570326	18.264	ng/ul#	98
85) Benzo(a)anthracene	21.216	228	1911246	17.795	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.157	149	1235036	16.601	ng/ul#	95
87) Chrysene	21.269	228	1877786	17.806	ng/ul	100
89) Di-n-octyl phthalate	22.074	149	2072227	16.715	ng/ul	100
90) Benzo(b)fluoranthene	22.869	252	1998290	17.507	ng/ul#	95
91) Benzo(k)fluoranthene	22.916	252	1911486	17.829	ng/ul#	93
93) Benzo(a)pyrene	23.480	252	1947337	17.765	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.027	276	2272593	17.840	ng/ul#	88
95) Dibenzo(a,h)anthracene	26.051	278	1957571	17.759	ng/ul#	90
96) Benzo(g,h,i)perylene	26.774	276	1943903	17.917	ng/ul#	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP071822\  
 Data File : BP010995.D  
 Acq On : 18 Jul 2022 16:26  
 Operator : CG/JU  
 Sample : SSTCCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :


Quant Time: Jul 19 00:45:53 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP071422.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Jul 14 15:34:34 2022  
 Response via : Initial Calibration








#2  
 1,4-Dioxane  
 Concen: 7.215 ng/uL  
 RT: 3.211 min Scan# 21  
 Delta R.T. -0.005 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion: 88 Resp: 79048  
 Ion Ratio Lower Upper  
 88 100  
 43 28.6 35.8 53.6#  
 58 74.4 68.1 102.1



#5  
 Pyridine  
 Concen: 17.793 ng/uL  
 RT: 3.605 min Scan# 88  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion: 79 Resp: 494919  
 Ion Ratio Lower Upper  
 79 100  
 52 80.5 83.8 125.8#  
 51 37.7 41.0 61.4#








#6  
Benzaldehyde  
Concen: 21.848 ng/u1  
RT: 6.828 min Scan# 61  
Delta R.T. -0.006 min  
Lab File: BP010995.D  
Acq: 18 Jul 2022 16:26

Instrument :  
BNA\_P  
ClientSampleId :







Tgt Ion: 77 Resp: 327940  
Ion Ratio Lower Upper  
77 100  
105 96.9 76.0 114.0  
106 92.7 73.2 109.8



#8  
Phenol  
Concen: 16.949 ng/u1  
RT: 6.887 min Scan# 646  
Delta R.T. -0.006 min  
Lab File: BP010995.D  
Acq: 18 Jul 2022 16:26


Tgt Ion: 94 Resp: 556720  
Ion Ratio Lower Upper  
94 100  
65 27.6 27.4 41.2  
66 35.4 37.4 56.2#








#10  
 Bis(2-Chloroethyl)ether  
 Concen: 17.726 ng/ul  
 RT: 7.116 min Scan# 685  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :




Tgt Ion: 93 Resp: 462092

Ion	Ratio	Lower	Upper
93	100		
63	79.5	72.4	108.6
95	32.1	27.3	40.9






#12  
 2-Chlorophenol  
 Concen: 18.005 ng/ul  
 RT: 7.246 min Scan# 707  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26



Tgt Ion: 128 Resp: 464719

Ion	Ratio	Lower	Upper
128	100		
64	46.7	44.8	67.2
130	31.9	25.8	38.6









#13  
 2-Methylphenol  
 Concen: 16.918 ng/ul  
 RT: 8.134 min Scan# 81  
 Delta R.T. 0.000 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Instrument :  
 BNA\_P  
 ClientSampleId :


Tgt Ion:108 Resp: 410952  
 Ion Ratio Lower Upper  
 108 100  
 107 86.7 71.4 107.0



#14  
 2,2'-oxybis(1-Chloropropane)  
 Concen: 18.172 ng/ul  
 RT: 8.222 min Scan# 873  
 Delta R.T. 0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion: 45 Resp: 647912  
 Ion Ratio Lower Upper  
 45 100  
 77 14.1 11.2 16.8  
 79 10.4 6.9 10.3#








#16  
 Acetophenone  
 Concen: 17.766 ng/ul  
 RT: 8.510 min Scan# 911  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :




Tgt Ion: 105 Resp: 665345

Ion	Ratio	Lower	Upper
105	100		
77	75.7	67.5	101.3
51	27.0	29.6	44.4





#17  
 N-Nitroso-di-n-propylamine  
 Concen: 17.875 ng/ul  
 RT: 8.499 min Scan# 920  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Tgt Ion: 70 Resp: 335880

Ion	Ratio	Lower	Upper
70	100		
42	58.5	53.8	80.8
101	10.0	7.8	11.6
130	21.8	15.4	23.2











#18  
 4-Methylphenol  
 Concen: 17.016 ng/ul  
 RT: 8.463 min Scan# 914  
 Delta R.T. 0.000 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:108 Resp: 441317  
 Ion Ratio Lower Upper  
 108 100  
 107 111.1 88.3 132.5



#19  
 Hexachloroethane  
 Concen: 18.237 ng/ul  
 RT: 8.752 min Scan# 963  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:117 Resp: 186271  
 Ion Ratio Lower Upper  
 117 100  
 201 83.2 85.7 128.5#  
 199 52.4 57.4 86.0#








#22  
 Nitrobenzene  
 Concen: 18.066 ng/u1  
 RT: 8.887 min Scan# 986  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Instrument :  
 BNA\_P  
 ClientSampleId :





Tgt Ion: 77 Resp: 489157  
 Ion Ratio Lower Upper  
 77 100  
 123 44.5 31.5 47.3  
 65 12.9 11.4 17.2




#23  
 Isophorone  
 Concen: 17.512 ng/u1  
 RT: 9.416 min Scan# 1076  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26



Tgt Ion: 82 Resp: 887425  
 Ion Ratio Lower Upper  
 82 100  
 95 7.3 6.6 10.0  
 138 17.7 13.8 20.6









#25  
 2-Nitrophenol  
 Concen: 17.076 ng/ul  
 RT: 9.593 min Scan# 1106  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Instrument :  
 BNA\_P  
 ClientSampleId :


Tgt Ion:139 Resp: 218366  
 Ion Ratio Lower Upper  
 139 100  
 65 54.3 54.2 81.4  
 109 29.8 30.6 45.8#



#26  
 2,4-Dimethylphenol  
 Concen: 17.720 ng/ul  
 RT: 9.663 min Scan# 1118  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:107 Resp: 479203  
 Ion Ratio Lower Upper  
 107 100  
 121 51.4 39.1 58.7  
 122 85.8 66.1 99.1








#27  
 Bis(2-Chloroethoxy)methane  
 Concen: 17.296 ng/ul  
 RT: 9.904 min Scan# 1159  
 Delta R.T. 0.000 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument : BNA\_P  
 ClientSampleId :



Tgt Ion: 93 Resp: 580031




Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.6	39.8
123	11.5	7.8	11.8




#29  
 2,4-Dichlorophenol  
 Concen: 17.284 ng/ul  
 RT: 10.128 min Scan# 1197  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion: 162 Resp: 372740


Ion	Ratio	Lower	Upper
162	100		
164	64.5	52.5	78.7
98	38.8	31.4	47.0








#30  
 Naphthalene  
 Concen: 18.027 ng/uI  
 RT: 10.528 min Scan# 1265  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:128 Resp: 1451365  
 Ion Ratio Lower Upper  
 128 100  
 129 10.8 8.7 13.1  
 127 12.9 10.2 15.4



#32  
 4-Chloroaniline  
 Concen: 17.062 ng/uI  
 RT: 10.646 min Scan# 1285  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:127 Resp: 579679  
 Ion Ratio Lower Upper  
 127 100  
 129 31.7 26.2 39.2








#33  
 Hexachlorobutadiene  
 Concen: 17.634 ng/uI  
 RT: 10.810 min Scan# 1111  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:225 Resp: 221362  
 Ion Ratio Lower Upper  
 225 100  
 223 64.1 53.7 80.5  
 227 65.3 49.4 74.2



#34  
 Caprolactam  
 Concen: 19.961 ng/uI  
 RT: 11.434 min Scan# 1419  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:113 Resp: 150861  
 Ion Ratio Lower Upper  
 113 100  
 55 173.8 153.8 230.8  
 56 132.4 110.2 165.4








#35  
 4-Chloro-3-methylphenol  
 Concen: 16.936 ng/ul  
 RT: 11.781 min Scan# 1478  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Instrument :  
 BNA\_P  
 ClientSampleId :





Tgt Ion:107 Resp: 412745  
 Ion Ratio Lower Upper  
 107 100  
 144 25.5 18.6 27.8  
 142 78.6 59.6 89.4




#36  
 2-Methylnaphthalene  
 Concen: 17.644 ng/ul  
 RT: 12.146 min Scan# 1540  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Tgt Ion:142 Resp: 920894  
 Ion Ratio Lower Upper  
 142 100  
 141 87.6 70.1 105.1








#37  
 1-Methylnaphthalene  
 Concen: 17.685 ng/ul  
 RT: 12.369 min Scan# 11  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Instrument :  
 BNA\_P  
 ClientSampleId :





Tgt Ion:142 Resp: 925223  
 Ion Ratio Lower Upper  
 142 100  
 141 92.6 73.3 109.9  
 116 3.8 4.0 6.0#




#39  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 17.481 ng/ul  
 RT: 12.516 min Scan# 1603  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26



Tgt Ion:216 Resp: 412990  
 Ion Ratio Lower Upper  
 216 100  
 214 77.7 63.0 94.6  
 179 22.5 18.1 27.1  
 108 20.4 13.6 20.4#











#40  
 Hexachlorocyclopentadiene  
 Concen: 16.950 ng/ul  
 RT: 12.498 min Scan# 1600  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument : BNA\_P  
 ClientSampleId :



Tgt Ion: 237 Resp: 248034




Ion	Ratio	Lower	Upper
237	100		
235	61.6	52.8	79.2
272	12.6	11.5	17.3




#41  
 2,4,6-Trichlorophenol  
 Concen: 16.553 ng/ul  
 RT: 12.769 min Scan# 1646  
 Delta R.T. 0.000 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion: 196 Resp: 257542


Ion	Ratio	Lower	Upper
196	100		
198	94.6	78.2	117.4
200	30.6	25.0	37.6








#42  
 2,4,5-Trichlorophenol  
 Concen: 17.130 ng/ul  
 RT: 12.834 min Scan# 10  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:196 Resp: 284834  
 Ion Ratio Lower Upper  
 196 100  
 198 92.3 77.0 115.4  
 200 30.0 24.8 37.2



#44  
 2-Chloronaphthalene  
 Concen: 17.772 ng/ul  
 RT: 13.210 min Scan# 1721  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:162 Resp: 892216  
 Ion Ratio Lower Upper  
 162 100  
 164 32.1 25.5 38.3  
 127 41.1 31.9 47.9








#47  
 Dimethylphthalate  
 Concen: 17.514 ng/ul  
 RT: 13.810 min Scan# 1823  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:163 Resp: 1081969  
 Ion Ratio Lower Upper  
 163 100  
 194 3.7 3.4 5.2  
 164 10.1 7.6 11.4



#48  
 2,6-Dinitrotoluene  
 Concen: 17.180 ng/ul  
 RT: 13.928 min Scan# 1843  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:165 Resp: 192146  
 Ion Ratio Lower Upper  
 165 100  
 89 66.7 52.9 79.3  
 121 26.6 19.4 29.2








#50  
 Acenaphthylene  
 Concen: 18.178 ng/uI  
 RT: 14.063 min Scan# 1866  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:152 Resp: 1473382  
 Ion Ratio Lower Upper  
 152 100  
 151 19.7 16.5 24.7  
 153 13.3 10.1 15.1



#51  
 3-Nitroaniline  
 Concen: 16.586 ng/uI  
 RT: 14.257 min Scan# 1899  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:138 Resp: 222873  
 Ion Ratio Lower Upper  
 138 100  
 108 13.2 11.0 16.4  
 92 123.4 104.6 156.8








#52  
 Acenaphthene  
 Concen: 17.631 ng/ul  
 RT: 14.404 min Scan# 1924  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:153 Resp: 946701  
 Ion Ratio Lower Upper  
 153 100  
 152 48.0 36.2 54.2  
 154 89.0 70.8 106.2



#53  
 2,4-Dinitrophenol  
 Concen: 17.397 ng/ul  
 RT: 14.463 min Scan# 1934  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:184 Resp: 113837  
 Ion Ratio Lower Upper  
 184 100  
 63 79.9 73.8 110.6  
 154 64.2 57.2 85.8









#55  
 4-Nitrophenol  
 Concen: 19.124 ng/ul  
 RT: 14.569 min Scan# 1953  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Instrument :  
 BNA\_P  
 ClientSampleId :


Tgt Ion:109 Resp: 181426  
 Ion Ratio Lower Upper  
 109 100  
 139 136.4 75.6 113.4#  
 65 146.6 107.6 161.4



#56  
 Dibenzofuran  
 Concen: 17.878 ng/ul  
 RT: 14.745 min Scan# 1982  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:168 Resp: 1324652  
 Ion Ratio Lower Upper  
 168 100  
 139 39.6 32.8 49.2








#57  
 2,4-Dinitrotoluene  
 Concen: 17.894 ng/ul  
 RT: 14.716 min Scan# 1977  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :



Tgt Ion:165 Resp: 285470




Ion	Ratio	Lower	Upper
165	100		
63	49.2	47.0	70.6
182	2.6	2.2	3.2




#58  
 2,3,4,6-Tetrachlorophenol  
 Concen: 16.858 ng/ul  
 RT: 14.975 min Scan# 2021  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:232 Resp: 219514


Ion	Ratio	Lower	Upper
232	100		
131	56.4	35.8	53.8#
130	3.0	1.8	2.8#
166	34.6	23.7	35.5








#59  
 Diethylphthalate  
 Concen: 17.301 ng/ul  
 RT: 15.187 min Scan# 2057  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :






Tgt Ion:149 Resp: 1097606  
 Ion Ratio Lower Upper  
 149 100  
 177 21.1 17.6 26.4  
 150 12.6 10.3 15.5




#61  
 Fluorene  
 Concen: 17.740 ng/ul  
 RT: 15.398 min Scan# 2093  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:166 Resp: 1054807  
 Ion Ratio Lower Upper  
 166 100  
 165 97.9 77.4 116.2  
 167 13.4 10.8 16.2











#62  
 4-Chlorophenyl-phenylether  
 Concen: 17.697 ng/ul  
 RT: 15.398 min Scan# 2093  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Instrument : BNA\_P  
 ClientSampleId :


Tgt Ion:204 Resp: 484433  
 Ion Ratio Lower Upper  
 204 100  
 206 32.7 26.0 39.0  
 141 68.5 49.3 73.9



#63  
 4-Nitroaniline  
 Concen: 17.940 ng/ul  
 RT: 15.428 min Scan# 2098  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:138 Resp: 221548  
 Ion Ratio Lower Upper  
 138 100  
 92 53.5 46.2 69.2  
 108 72.7 85.8 128.8#








#66  
 4,6-Dinitro-2-methylphenol  
 Concen: 15.473 ng/ul  
 RT: 15.481 min Scan# 2107  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:198 Resp: 141194  
 Ion Ratio Lower Upper  
 198 100  
 182 6.1 4.3 6.5  
 77 30.8 23.7 35.5



#67  
 N-Nitrosodiphenylamine  
 Concen: 17.561 ng/ul  
 RT: 15.616 min Scan# 2130  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:169 Resp: 901723  
 Ion Ratio Lower Upper  
 169 100  
 168 65.9 52.6 78.8  
 167 36.1 27.8 41.6









#68  
 4-Bromophenyl-phenylether  
 Concen: 17.242 ng/ul  
 RT: 16.304 min Scan# 2247  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :

Tgt Ion:248 Resp: 279525




Ion	Ratio	Lower	Upper
248	100		
250	100.0	80.8	121.2
141	83.7	63.9	95.9




#69  
 Hexachlorobenzene  
 Concen: 17.350 ng/ul  
 RT: 16.404 min Scan# 2264  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:284 Resp: 326446


Ion	Ratio	Lower	Upper
284	100		
142	43.3	27.8	41.6#
249	32.6	25.4	38.2








#70  
 Atrazine  
 Concen: 17.203 ng/ul  
 RT: 16.586 min Scan# 21  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Instrument :  
 BNA\_P  
 ClientSampleId :





Tgt Ion:200 Resp: 305571  
 Ion Ratio Lower Upper  
 200 100  
 173 28.9 20.7 31.1  
 215 47.4 38.6 57.8




#71  
 Pentachlorophenol  
 Concen: 17.680 ng/ul  
 RT: 16.757 min Scan# 2324  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Tgt Ion:266 Resp: 204874  
 Ion Ratio Lower Upper  
 266 100  
 264 63.4 51.4 77.2  
 268 67.0 53.5 80.3








#72  
 Phenanthrene  
 Concen: 17.672 ng/uI  
 RT: 17.151 min Scan# 21  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:178 Resp: 1655008  
 Ion Ratio Lower Upper  
 178 100  
 179 15.5 12.7 19.1  
 176 19.1 14.6 22.0



#74  
 Anthracene  
 Concen: 18.056 ng/uI  
 RT: 17.245 min Scan# 2407  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:178 Resp: 1682017  
 Ion Ratio Lower Upper  
 178 100  
 179 15.5 12.2 18.2  
 176 18.8 14.5 21.7








#75  
 1,2,3,4-Tetrachlorobenzene  
 Concen: 16.717 ng/uL  
 RT: 13.134 min Scan# 1708  
 Delta R.T. 0.000 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Instrument : BNA\_P  
 ClientSampleId :





Tgt Ion:216 Resp: 430177  
 Ion Ratio Lower Upper  
 216 100  
 214 79.1 65.0 97.4  
 218 47.8 39.1 58.7




#76  
 Pentachlorobenzene  
 Concen: 17.440 ng/uL  
 RT: 14.663 min Scan# 1968  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Tgt Ion:250 Resp: 393404  
 Ion Ratio Lower Upper  
 250 100  
 248 62.7 49.8 74.6  
 252 63.6 51.2 76.8








#77  
 Carbazole  
 Concen: 17.881 ng/ul  
 RT: 17.522 min Scan# 2454  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:167 Resp: 1559041  
 Ion Ratio Lower Upper  
 167 100  
 166 21.2 16.9 25.3  
 139 13.6 10.6 16.0



#78  
 Di-n-butylphthalate  
 Concen: 16.612 ng/ul  
 RT: 18.098 min Scan# 2552  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:149 Resp: 1810436  
 Ion Ratio Lower Upper  
 149 100  
 150 9.1 7.1 10.7  
 104 5.1 4.6 6.8








#80  
 Fluoranthene  
 Concen: 17.697 ng/uI  
 RT: 19.139 min Scan# 2729  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :






Tgt Ion:202 Resp: 1917081  
 Ion Ratio Lower Upper  
 202 100  
 101 14.4 7.1 10.7#  
 100 10.9 5.7 8.5#




#82  
 Pyrene  
 Concen: 17.719 ng/uI  
 RT: 19.498 min Scan# 2790  
 Delta R.T. -0.006 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:202 Resp: 1980479  
 Ion Ratio Lower Upper  
 202 100  
 101 15.4 7.9 11.9#  
 100 13.0 6.7 10.1#











#83  
 Butylbenzylphthalate  
 Concen: 16.302 ng/ul  
 RT: 20.392 min Scan# 2942  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26




Instrument : BNA\_P  
 ClientSampleId :


Tgt Ion	Resp	Lower	Upper
149	100		
91	75.2	60.0	90.0
206	18.9	17.4	26.2



#84  
 3,3'-Dichlorobenzidine  
 Concen: 18.264 ng/ul  
 RT: 21.157 min Scan# 3072  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion	Resp	Lower	Upper
252	100		
254	63.7	51.0	76.4
126	15.8	8.6	12.8








#85  
 Benzo(a)anthracene  
 Concen: 17.795 ng/ul  
 RT: 21.216 min Scan# 3082  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:228 Resp: 1911246  
 Ion Ratio Lower Upper  
 228 100  
 229 19.8 15.0 22.4  
 226 26.6 20.6 31.0



#86  
 Bis(2-ethylhexyl)phthalate  
 Concen: 16.601 ng/ul  
 RT: 21.157 min Scan# 3072  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:149 Resp: 1235036  
 Ion Ratio Lower Upper  
 149 100  
 167 25.5 22.6 33.8  
 279 3.0 4.2 6.4#








#87  
 Chrysene  
 Concen: 17.806 ng/ul  
 RT: 21.269 min Scan# 3091  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:228 Resp: 1877786  
 Ion Ratio Lower Upper  
 228 100  
 226 29.6 23.6 35.4  
 229 19.5 15.6 23.4



#89  
 Di-n-octyl phthalate  
 Concen: 16.715 ng/ul  
 RT: 22.074 min Scan# 3228  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:149 Resp: 2072227








#90  
 Benzo(b)fluoranthene  
 Concen: 17.507 ng/ul  
 RT: 22.869 min Scan# 3363  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument : BNA\_P  
 ClientSampleId :







Tgt Ion:252 Resp: 1998290  
 Ion Ratio Lower Upper  
 252 100  
 253 21.6 17.8 26.6  
 125 13.4 6.2 9.2#



#91  
 Benzo(k)fluoranthene  
 Concen: 17.829 ng/ul  
 RT: 22.916 min Scan# 3371  
 Delta R.T. -0.012 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:252 Resp: 1911486  
 Ion Ratio Lower Upper  
 252 100  
 253 22.4 17.0 25.4  
 125 13.9 6.2 9.4#








#93  
 Benzo(a)pyrene  
 Concen: 17.765 ng/u1  
 RT: 23.480 min Scan# 3468  
 Delta R.T. -0.018 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :







Tgt Ion:252 Resp: 1947337  
 Ion Ratio Lower Upper  
 252 100  
 253 21.8 17.4 26.0  
 125 14.8 6.6 10.0#



#94  
 Indeno(1,2,3-cd)pyrene  
 Concen: 17.840 ng/u1  
 RT: 26.027 min Scan# 3900  
 Delta R.T. -0.018 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26


Tgt Ion:276 Resp: 2272593  
 Ion Ratio Lower Upper  
 276 100  
 138 30.1 14.6 22.0#  
 277 25.2 19.6 29.4








#95  
 Dibenzo(a,h)anthracene  
 Concen: 17.759 ng/ul  
 RT: 26.051 min Scan# 3904  
 Delta R.T. -0.018 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Instrument :  
 BNA\_P  
 ClientSampleId :



Tgt Ion:278 Resp: 1957571  
 Ion Ratio Lower Upper  
 278 100  
 139 22.3 9.4 14.2#  
 279 24.2 18.9 28.3



#96  
 Benzo(g,h,i)perylene  
 Concen: 17.917 ng/ul  
 RT: 26.774 min Scan# 4027  
 Delta R.T. -0.018 min  
 Lab File: BP010995.D  
 Acq: 18 Jul 2022 16:26

Tgt Ion:276 Resp: 1943903  
 Ion Ratio Lower Upper  
 276 100  
 138 28.4 12.7 19.1#  
 277 24.0 18.2 27.4

