

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071822\
 Data File : BP011010.D
 Acq On : 19 Jul 2022 01:02
 Operator : CG/JU
 Sample : N3710-09MSD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 H0B06MSD

Quant Time: Jul 19 01:49:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 14 15:34:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.681	152	382093	20.000	ng/ul	0.00
20) Naphthalene-d8	10.475	136	1449731	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.340	164	761882	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.110	188	1482576	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.228	240	1440662	20.000	ng/ul	#-0.01
88) Perylene-d12	23.580	264	1590026	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	44633	4.211	ng/uL	0.00
4) Pyridine-d5	3.593	84	423598	15.463	ng/ul	0.00
7) Phenol-d5	6.858	99	265708	8.472	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	716749	35.538	ng/ul	0.00
11) 2-Chlorophenol-d4	7.211	132	721056	28.451	ng/ul	-0.01
15) 4-Methylphenol-d8	8.399	113	478386	19.052	ng/ul	0.00
21) Nitrobenzene-d5	8.840	128	426006	40.323	ng/ul	0.00
24) 2-Nitrophenol-d4	9.563	143	400087	38.271	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.099	165	688666	36.022	ng/ul	0.00
31) 4-Chloroaniline-d4	10.622	131	803110	25.481	ng/ul	0.00
46) Dimethylphthalate-d6	13.763	166	2431774	46.700	ng/ul	0.00
49) Acenaphthylene-d8	14.034	160	2751823	44.790	ng/ul	0.00
54) 4-Nitrophenol-d4	14.563	143	112884	10.711	ng/ul	0.00
60) Fluorene-d10	15.340	176	1988013	44.915	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.469	200	342235	46.650	ng/ul	0.00
73) Anthracene-d10	17.210	188	3114669	49.417	ng/ul	0.00
81) Pyrene-d10	19.463	212	3613505	47.757	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.433	264	3877757	50.778	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.217	88	94343	8.484	ng/uL#	88
5) Pyridine	3.611	79	457025	16.188	ng/ul#	77
6) Benzaldehyde	6.828	77	764319	50.169	ng/ul	99
8) Phenol	6.887	94	297335	8.918	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.117	93	930546	35.169	ng/ul	90
12) 2-Chlorophenol	7.246	128	738754	28.200	ng/ul	93
13) 2-Methylphenol	8.128	108	543575	22.047	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.222	45	1280850	35.393	ng/ul#	97
16) Acetophenone	8.511	105	1335415	35.133	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.499	70	662001	34.710	ng/ul	92
18) 4-Methylphenol	8.464	108	509896	19.370	ng/ul	99
19) Hexachloroethane	8.752	117	331419	31.968	ng/ul#	81
22) Nitrobenzene	8.887	77	999671	40.236	ng/ul#	90
23) Isophorone	9.416	82	1848819	39.760	ng/ul#	97
25) 2-Nitrophenol	9.593	139	432721	36.876	ng/ul#	83
26) 2,4-Dimethylphenol	9.663	107	826186	33.294	ng/ul	95
27) Bis(2-Chloroethoxy)met...	9.899	93	1159644	37.684	ng/ul	98
29) 2,4-Dichlorophenol	10.122	162	689199	34.829	ng/ul	98
30) Naphthalene	10.522	128	2717691	36.787	ng/ul	99
32) 4-Chloroaniline	10.646	127	1022307	32.792	ng/ul	100
33) Hexachlorobutadiene	10.805	225	391118	33.955	ng/ul	96
34) Caprolactam	11.458	113	52162	7.522	ng/ul	97
35) 4-Chloro-3-methylphenol	11.775	107	784338	35.073	ng/ul	95
36) 2-Methylnaphthalene	12.146	142	1763203	36.815	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071822\
 Data File : BP011010.D
 Acq On : 19 Jul 2022 01:02
 Operator : CG/JU
 Sample : N3710-09MSD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 H0B06MSD

Quant Time: Jul 19 01:49:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jul 14 15:34:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.369	142	1758032	36.621	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.516	216	793696	39.498	ng/ul#	98
40) Hexachlorocyclopentadiene	12.493	237	394290	31.681	ng/ul	96
41) 2,4,6-Trichlorophenol	12.763	196	559013	42.244	ng/ul	99
42) 2,4,5-Trichlorophenol	12.834	196	617406	43.656	ng/ul	99
43) 1,1'-Biphenyl	13.169	154	2289234	40.270	ng/ul#	97
44) 2-Chloronaphthalene	13.210	162	1751114	41.011	ng/ul	98
45) 2-Nitroaniline	13.422	65	620876	49.908	ng/ul#	85
47) Dimethylphthalate	13.810	163	2431041	46.268	ng/ul	98
48) 2,6-Dinitrotoluene	13.928	165	490167	51.528	ng/ul	96
50) Acenaphthylene	14.063	152	2861365	41.506	ng/ul	99
51) 3-Nitroaniline	14.257	138	534329	46.753	ng/ul	91
52) Acenaphthene	14.404	153	1973813	43.219	ng/ul	99
53) 2,4-Dinitrophenol	14.463	184	225983	40.606	ng/ul#	84
55) 4-Nitrophenol	14.575	109	89966	11.150	ng/ul#	74
56) Dibenzofuran	14.746	168	2692559	42.727	ng/ul	97
57) 2,4-Dinitrotoluene	14.716	165	709502	52.289	ng/ul	86
58) 2,3,4,6-Tetrachlorophenol	14.975	232	520678	47.013	ng/ul	90
59) Diethylphthalate	15.187	149	2592049	48.037	ng/ul	98
61) Fluorene	15.398	166	2263573	44.758	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.398	204	1014367	43.568	ng/ul	95
63) 4-Nitroaniline	15.428	138	585940	55.785	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.481	198	343977	45.704	ng/ul	99
67) N-Nitrosodiphenylamine	15.616	169	1997988	47.177	ng/ul	98
68) 4-Bromophenyl-phenylether	16.298	248	651216	48.702	ng/ul	95
69) Hexachlorobenzene	16.404	284	726150	46.791	ng/ul	94
70) Atrazine	16.593	200	697768	47.628	ng/ul	98
71) Pentachlorophenol	16.757	266	435396	45.555	ng/ul	97
72) Phenanthrene	17.151	178	3723606	48.208	ng/ul	99
74) Anthracene	17.245	178	3719593	48.411	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.128	216	836777	39.426	ng/uL	96
76) Pentachlorobenzene	14.657	250	811624	43.624	ng/uL	99
77) Carbazole	17.522	167	3626095	50.424	ng/ul	99
78) Di-n-butylphthalate	18.092	149	4703609	52.326	ng/ul	99
80) Fluoranthene	19.134	202	4320094	48.818	ng/ul#	85
82) Pyrene	19.492	202	4426892	48.486	ng/ul#	83
83) Butylbenzylphthalate	20.386	149	2304133	56.570	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.157	252	378882	14.853	ng/ul#	97
85) Benzo(a)anthracene	21.216	228	4392955	50.069	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.157	149	3411288	56.132	ng/ul#	96
87) Chrysene	21.269	228	4163185	48.325	ng/ul	99
89) Di-n-octyl phthalate	22.069	149	5895193	55.283	ng/ul	100
90) Benzo(b)fluoranthene	22.869	252	4889256	49.800	ng/ul#	95
91) Benzo(k)fluoranthene	22.916	252	4562682	49.477	ng/ul#	94
93) Benzo(a)pyrene	23.480	252	4222394	44.783	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.021	276	5788577	52.829	ng/ul#	88
95) Dibenzo(a,h)anthracene	26.045	278	4805642	50.685	ng/ul#	90
96) Benzo(g,h,i)perylene	26.774	276	4775419	51.171	ng/ul#	87





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.484 ng/uL
RT: 3.217 min Scan# 21
Delta R.T. 0.001 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion: 88 Resp: 94343
Ion Ratio Lower Upper
88 100
43 31.6 35.8 53.6#
58 78.2 68.1 102.1



#5
Pyridine
Concen: 16.188 ng/uL
RT: 3.611 min Scan# 89
Delta R.T. 0.000 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion: 79 Resp: 457025
Ion Ratio Lower Upper
79 100
52 79.7 83.8 125.8#
51 37.8 41.0 61.4#









#6
Benzaldehyde
Concen: 50.169 ng/ul
RT: 6.828 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion: 77 Resp: 764319
Ion Ratio Lower Upper
77 100
105 96.8 76.0 114.0
106 92.6 73.2 109.8



#8
Phenol
Concen: 8.918 ng/ul
RT: 6.887 min Scan# 646
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion: 94 Resp: 297335
Ion Ratio Lower Upper
94 100
65 27.6 27.4 41.2
66 35.6 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 35.169 ng/ul
RT: 7.117 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion: 93 Resp: 930546
Ion Ratio Lower Upper
93 100
63 78.0 72.4 108.6
95 32.7 27.3 40.9



#12
2-Chlorophenol
Concen: 28.200 ng/ul
RT: 7.246 min Scan# 707
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:128 Resp: 738754
Ion Ratio Lower Upper
128 100
64 48.7 44.8 67.2
130 32.5 25.8 38.6





#13

2-Methylphenol

Concen: 22.047 ng/ul

RT: 8.128 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Instrument :

BNA_P

ClientSampleId :





H0B06MSD

Tgt Ion:108 Resp: 543575

Ion Ratio Lower Upper

108 100

107 90.8 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 35.393 ng/ul

RT: 8.222 min Scan# 873

Delta R.T. 0.006 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02




Tgt Ion: 45 Resp: 1280850


Ion Ratio Lower Upper

45 100

77 14.2 11.2 16.8

79 11.1 6.9 10.3#









#16
Acetophenone
Concen: 35.133 ng/ul
RT: 8.511 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:105 Resp: 1335415
Ion Ratio Lower Upper
105 100
77 75.5 67.5 101.3
51 26.8 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 34.710 ng/ul
RT: 8.499 min Scan# 920
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion: 70 Resp: 662001
Ion Ratio Lower Upper
70 100
42 59.0 53.8 80.8
101 10.1 7.8 11.6
130 21.3 15.4 23.2









#18
4-Methylphenol
Concen: 19.370 ng/ul
RT: 8.464 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:108 Resp: 509896
Ion Ratio Lower Upper
108 100
107 109.3 88.3 132.5



#19
Hexachloroethane
Concen: 31.968 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:117 Resp: 331419
Ion Ratio Lower Upper
117 100
201 87.2 85.7 128.5
199 55.7 57.4 86.0#









#22
Nitrobenzene
Concen: 40.236 ng/ul
RT: 8.887 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion: 77 Resp: 999671
Ion Ratio Lower Upper
77 100
123 47.3 31.5 47.3#
65 13.3 11.4 17.2



#23
Isophorone
Concen: 39.760 ng/ul
RT: 9.416 min Scan# 1076
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion: 82 Resp: 1848819
Ion Ratio Lower Upper
82 100
95 6.6 6.6 10.0#
138 18.0 13.8 20.6









#25
2-Nitrophenol
Concen: 36.876 ng/ul
RT: 9.593 min Scan# 1106
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:139 Resp: 432721
Ion Ratio Lower Upper
139 100
65 53.2 54.2 81.4#
109 29.6 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 33.294 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:107 Resp: 826186
Ion Ratio Lower Upper
107 100
121 51.9 39.1 58.7
122 86.8 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 37.684 ng/ul

RT: 9.899 min Scan# 1158

Delta R.T. -0.006 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Instrument :





BNA_P

ClientSampleId :

H0B06MSD

Tgt Ion: 93 Resp: 1159644

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.6	39.8
123	11.3	7.8	11.8



#29

2,4-Dichlorophenol

Concen: 34.829 ng/ul

RT: 10.122 min Scan# 1196




Delta R.T. -0.011 min


Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Tgt Ion:162 Resp: 689199

Ion	Ratio	Lower	Upper
162	100		
164	63.8	52.5	78.7
98	39.8	31.4	47.0









#30
Naphthalene
Concen: 36.787 ng/ul
RT: 10.522 min Scan# 1265
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:128 Resp: 2717691
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 32.792 ng/ul
RT: 10.646 min Scan# 1285
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:127 Resp: 1022307
Ion Ratio Lower Upper
127 100
129 32.4 26.2 39.2









#33
Hexachlorobutadiene
Concen: 33.955 ng/ul
RT: 10.805 min Scan# 11
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:225 Resp: 391118
Ion Ratio Lower Upper
225 100
223 61.7 53.7 80.5
227 62.1 49.4 74.2



#34
Caprolactam
Concen: 7.522 ng/ul
RT: 11.458 min Scan# 1423
Delta R.T. 0.012 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:113 Resp: 52162
Ion Ratio Lower Upper
113 100
55 186.8 153.8 230.8
56 140.1 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 35.073 ng/ul
RT: 11.775 min Scan# 1477
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:107 Resp: 784338
Ion Ratio Lower Upper
107 100
144 25.3 18.6 27.8
142 78.5 59.6 89.4



#36
2-Methylnaphthalene
Concen: 36.815 ng/ul
RT: 12.146 min Scan# 1540
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:142 Resp: 1763203
Ion Ratio Lower Upper
142 100
141 88.1 70.1 105.1









#37
1-Methylnaphthalene
Concen: 36.621 ng/ul
RT: 12.369 min Scan# 1578
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:142 Resp: 1758032
Ion Ratio Lower Upper
142 100
141 92.4 73.3 109.9
116 3.7 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.498 ng/ul
RT: 12.516 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:216 Resp: 793696
Ion Ratio Lower Upper
216 100
214 77.7 63.0 94.6
179 22.5 18.1 27.1
108 20.4 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 31.681 ng/ul
RT: 12.493 min Scan# 1
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:237 Resp: 394290
Ion Ratio Lower Upper
237 100
235 63.3 52.8 79.2
272 12.5 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 42.244 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:196 Resp: 559013
Ion Ratio Lower Upper
196 100
198 96.3 78.2 117.4
200 30.6 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 43.656 ng/ul

RT: 12.834 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Instrument :

BNA_P

ClientSampleId :

H0B06MSD





Tgt Ion:196 Resp: 617406

Ion Ratio Lower Upper

196 100

198 95.9 77.0 115.4

200 30.1 24.8 37.2



#43

1,1'-Biphenyl

Concen: 40.270 ng/ul

RT: 13.169 min Scan# 1714

Delta R.T. -0.011 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02




Tgt Ion:154 Resp: 2289234


Ion Ratio Lower Upper

154 100

153 40.6 31.9 47.9

76 14.8 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 41.011 ng/ul
RT: 13.210 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:162 Resp: 1751114
Ion Ratio Lower Upper
162 100
164 32.5 25.5 38.3
127 41.5 31.9 47.9



#45
2-Nitroaniline
Concen: 49.908 ng/ul
RT: 13.422 min Scan# 1757
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion: 65 Resp: 620876
Ion Ratio Lower Upper
65 100
92 68.0 49.0 73.6
138 100.1 66.4 99.6#









#47
Dimethylphthalate
Concen: 46.268 ng/ul
RT: 13.810 min Scan# 1823
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:163 Resp: 2431041
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.2 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 51.528 ng/ul
RT: 13.928 min Scan# 1843
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:165 Resp: 490167
Ion Ratio Lower Upper
165 100
89 62.2 52.9 79.3
121 22.9 19.4 29.2









#50
Acenaphthylene
Concen: 41.506 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:152 Resp: 2861365
Ion Ratio Lower Upper
152 100
151 20.0 16.5 24.7
153 13.0 10.1 15.1



#51
3-Nitroaniline
Concen: 46.753 ng/ul
RT: 14.257 min Scan# 1899
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:138 Resp: 534329
Ion Ratio Lower Upper
138 100
108 13.1 11.0 16.4
92 119.8 104.6 156.8









#52
Acenaphthene
Concen: 43.219 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:153 Resp: 1973813
Ion Ratio Lower Upper
153 100
152 47.1 36.2 54.2
154 89.1 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 40.606 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:184 Resp: 225983
Ion Ratio Lower Upper
184 100
63 71.3 73.8 110.6#
154 65.1 57.2 85.8









#55
4-Nitrophenol
Concen: 11.150 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:109 Resp: 89966
Ion Ratio Lower Upper
109 100
139 139.6 75.6 113.4#
65 148.8 107.6 161.4



#56
Dibenzofuran
Concen: 42.727 ng/ul
RT: 14.746 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:168 Resp: 2692559
Ion Ratio Lower Upper
168 100
139 39.2 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 52.289 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:165 Resp: 709502
Ion Ratio Lower Upper
165 100
63 48.1 47.0 70.6
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 47.013 ng/ul
RT: 14.975 min Scan# 2021
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:232 Resp: 520678
Ion Ratio Lower Upper
232 100
131 52.9 35.8 53.8
130 2.7 1.8 2.8
166 34.0 23.7 35.5









#59
Diethylphthalate
Concen: 48.037 ng/ul
RT: 15.187 min Scan# 2057
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:149 Resp: 2592049
Ion Ratio Lower Upper
149 100
177 21.1 17.6 26.4
150 12.4 10.3 15.5



#61
Fluorene
Concen: 44.758 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:166 Resp: 2263573
Ion Ratio Lower Upper
166 100
165 97.5 77.4 116.2
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 43.568 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:204 Resp: 1014367
Ion Ratio Lower Upper
204 100
206 31.9 26.0 39.0
141 67.0 49.3 73.9



#63
4-Nitroaniline
Concen: 55.785 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:138 Resp: 585940
Ion Ratio Lower Upper
138 100
92 54.8 46.2 69.2
108 76.9 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 45.704 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:198 Resp: 343977
Ion Ratio Lower Upper
198 100
182 4.7 4.3 6.5
77 28.9 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 47.177 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:169 Resp: 1997988
Ion Ratio Lower Upper
169 100
168 67.0 52.6 78.8
167 36.4 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 48.702 ng/ul
RT: 16.298 min Scan# 2246
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:248 Resp: 651216
Ion Ratio Lower Upper
248 100
250 97.9 80.8 121.2
141 85.6 63.9 95.9



#69
Hexachlorobenzene
Concen: 46.791 ng/ul
RT: 16.404 min Scan# 2264
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:284 Resp: 726150
Ion Ratio Lower Upper
284 100
142 40.7 27.8 41.6
249 31.0 25.4 38.2





#70

Atrazine

Concen: 47.628 ng/ul

RT: 16.593 min Scan# 2295

Delta R.T. 0.000 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Instrument :

BNA_P

ClientSampleId :

H0B06MSD





Tgt Ion:200 Resp: 697768

Ion Ratio Lower Upper

200 100

173 27.4 20.7 31.1

215 47.5 38.6 57.8



#71

Pentachlorophenol

Concen: 45.555 ng/ul

RT: 16.757 min Scan# 2324

Delta R.T. -0.006 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02




Tgt Ion:266 Resp: 435396


Ion Ratio Lower Upper

266 100

264 62.8 51.4 77.2

268 63.0 53.5 80.3









#72
Phenanthrene
Concen: 48.208 ng/ul
RT: 17.151 min Scan# 21
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:178 Resp: 3723606
Ion Ratio Lower Upper
178 100
179 15.6 12.7 19.1
176 19.2 14.6 22.0



#74
Anthracene
Concen: 48.411 ng/ul
RT: 17.245 min Scan# 2407
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:178 Resp: 3719593
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 39.426 ng/uL
RT: 13.128 min Scan# 1707
Delta R.T. -0.006 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:216 Resp: 836777
Ion Ratio Lower Upper
216 100
214 77.5 65.0 97.4
218 47.2 39.1 58.7



#76
Pentachlorobenzene
Concen: 43.624 ng/uL
RT: 14.657 min Scan# 1967
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:250 Resp: 811624
Ion Ratio Lower Upper
250 100
248 63.4 49.8 74.6
252 63.3 51.2 76.8









#77
 Carbazole
 Concen: 50.424 ng/ul
 RT: 17.522 min Scan# 2454
 Delta R.T. -0.011 min
 Lab File: BP011010.D
 Acq: 19 Jul 2022 01:02




Instrument :
 BNA_P
 ClientSampleId :
 H0B06MSD


Tgt Ion:167 Resp: 3626095
 Ion Ratio Lower Upper
 167 100
 166 21.8 16.9 25.3
 139 13.6 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 52.326 ng/ul
 RT: 18.092 min Scan# 2551
 Delta R.T. -0.011 min
 Lab File: BP011010.D
 Acq: 19 Jul 2022 01:02

Tgt Ion:149 Resp: 4703609
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 5.1 4.6 6.8









#80
Fluoranthene
Concen: 48.818 ng/ul
RT: 19.134 min Scan# 21
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:202 Resp: 4320094
Ion Ratio Lower Upper
202 100
101 15.1 7.1 10.7#
100 11.5 5.7 8.5#



#82
Pyrene
Concen: 48.486 ng/ul
RT: 19.492 min Scan# 2789
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:202 Resp: 4426892
Ion Ratio Lower Upper
202 100
101 16.8 7.9 11.9#
100 13.5 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 56.570 ng/ul
RT: 20.386 min Scan# 2941
Delta R.T. -0.017 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:149 Resp: 2304133
Ion Ratio Lower Upper
149 100
91 74.2 60.0 90.0
206 18.4 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 14.853 ng/ul
RT: 21.157 min Scan# 3072
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:252 Resp: 378882
Ion Ratio Lower Upper
252 100
254 65.0 51.0 76.4
126 14.6 8.6 12.8#





#85

Benzo(a)anthracene

Concen: 50.069 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.011 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Instrument :

BNA_P

ClientSampleId :

H0B06MSD





Tgt Ion:228 Resp: 4392955

Ion Ratio Lower Upper

228 100

229 20.0 15.0 22.4

226 26.9 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 56.132 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. -0.011 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02




Tgt Ion:149 Resp: 3411288


Ion Ratio Lower Upper

149 100

167 26.4 22.6 33.8

279 3.8 4.2 6.4#





#87

Chrysene

Concen: 48.325 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. -0.011 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Instrument :

BNA_P

ClientSampleId :

H0B06MSD





Tgt Ion:228 Resp: 4163185

Ion Ratio Lower Upper

228 100

226 29.9 23.6 35.4

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 55.283 ng/ul




RT: 22.069 min Scan# 3227


Delta R.T. -0.017 min

Lab File: BP011010.D

Acq: 19 Jul 2022 01:02

Tgt Ion:149 Resp: 5895193









#90
Benzo(b)fluoranthene
Concen: 49.800 ng/ul
RT: 22.869 min Scan# 3363
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:252 Resp: 4889256
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 13.4 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 49.477 ng/ul
RT: 22.916 min Scan# 3371
Delta R.T. -0.011 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:252 Resp: 4562682
Ion Ratio Lower Upper
252 100
253 22.0 17.0 25.4
125 13.6 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 44.783 ng/ul
RT: 23.480 min Scan# 3467
Delta R.T. -0.017 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02




Instrument :
BNA_P
ClientSampleId :
H0B06MSD


Tgt Ion:252 Resp: 4222394
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 14.8 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 52.829 ng/ul
RT: 26.021 min Scan# 3899
Delta R.T. -0.023 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:276 Resp: 5788577
Ion Ratio Lower Upper
276 100
138 30.2 14.6 22.0#
277 24.9 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 50.685 ng/ul
RT: 26.045 min Scan# 3904
Delta R.T. -0.023 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Instrument :
BNA_P
ClientSampleId :
H0B06MSD

Tgt Ion:278 Resp: 4805642
Ion Ratio Lower Upper
278 100
139 21.8 9.4 14.2#
279 24.4 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 51.171 ng/ul
RT: 26.774 min Scan# 4027
Delta R.T. -0.017 min
Lab File: BP011010.D
Acq: 19 Jul 2022 01:02

Tgt Ion:276 Resp: 4775419
Ion Ratio Lower Upper
276 100
138 28.4 12.7 19.1#
277 23.7 18.2 27.4

