

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071822\
 Data File : BP011014.D
 Acq On : 19 Jul 2022 03:53
 Operator : CG/JU
 Sample : PB146262BL
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK262

Quant Time: Jul 19 04:36:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 19 03:25:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	485112	20.000	ng/u1	0.00
20) Naphthalene-d8	10.469	136	1974342	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.340	164	1027657	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.104	188	1963949	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.228	240	1972220	20.000	ng/u1	# 0.00
88) Perylene-d12	23.580	264	2126971	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	86399	6.421	ng/uL	0.00
4) Pyridine-d5	3.587	84	927508	26.667	ng/u1	0.00
7) Phenol-d5	6.858	99	1205687	30.278	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	810357	31.647	ng/u1	0.00
11) 2-Chlorophenol-d4	7.211	132	1037956	32.258	ng/u1	0.00
15) 4-Methylphenol-d8	8.393	113	934846	29.325	ng/u1	0.00
21) Nitrobenzene-d5	8.840	128	461689	32.088	ng/u1	0.00
24) 2-Nitrophenol-d4	9.558	143	451654	31.724	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.093	165	760471	29.208	ng/u1	0.00
31) 4-Chloroaniline-d4	10.616	131	1183183	27.565	ng/u1	0.00
46) Dimethylphthalate-d6	13.757	166	2151622	30.634	ng/u1	0.00
49) Acenaphthylene-d8	14.028	160	2657539	32.069	ng/u1	0.00
54) 4-Nitrophenol-d4	14.557	143	329541	23.183	ng/u1	0.00
60) Fluorene-d10	15.340	176	1817361	30.441	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.463	200	226579	23.315	ng/u1	0.00
73) Anthracene-d10	17.204	188	2716728	32.539	ng/u1	0.00
81) Pyrene-d10	19.463	212	3089950	29.831	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.427	264	3290294	32.208	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071822\
Data File : BP011014.D
Acq On : 19 Jul 2022 03:53
Operator : CG/JU
Sample : PB146262BL
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK262

Quant Time: Jul 19 04:36:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071422.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 19 03:25:54 2022
Response via : Initial Calibration

