

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071923\
 Data File : BP016330.D
 Acq On : 19 Jul 2023 18:45
 Operator : MA/JU
 Sample : 03472-10
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A46Q4

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/20/2023
 Supervised By :mohammad ahmed 07/20/2023

Quant Time: Jul 19 23:09:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 19 14:47:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	111054	20.000	ng/u1	0.00
20) Naphthalene-d8	10.775	136	513372	20.000	ng/u1	0.02
38) Acenaphthene-d10	14.604	164	353210	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.369	188	848080	20.000	ng/u1	0.00
79) Chrysene-d12	21.469	240	968928	20.000	ng/u1	0.00
88) Perylene-d12	23.951	264	1102944	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	10502	3.346	ng/uL	0.00
4) Pyridine-d5	3.740	84	98926	12.639	ng/u1	0.02
7) Phenol-d5	7.158	99	174096	17.356	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	7.281	67	131024	19.317	ng/u1	0.00
11) 2-Chlorophenol-d4	7.487	132	144411	19.943	ng/u1	0.01
15) 4-Methylphenol-d8	8.763	113	50834m	6.291	ng/u1	0.08
21) Nitrobenzene-d5	9.169	128	69873	19.728	ng/u1	0.04
24) 2-Nitrophenol-d4	9.899	143	84966	21.049	ng/u1	0.04
28) 2,4-Dichlorophenol-d3	10.510	165	135443m	19.376	ng/u1	0.09
31) 4-Chloroaniline-d4	11.087	131	47504m	4.240	ng/u1	0.15
46) Dimethylphthalate-d6	14.028	166	650118	24.129	ng/u1	0.02
49) Acenaphthylene-d8	14.304	160	658322	22.295	ng/u1	0.01
54) 4-Nitrophenol-d4	14.987	143	72396m	13.166	ng/u1	0.06
60) Fluorene-d10	15.616	176	554174	24.576	ng/u1	0.02
65) 4,6-Dinitro-2-methylph...	15.816	200	69759m	14.249	ng/u1	0.04
73) Anthracene-d10	17.481	188	657636	17.728	ng/u1	0.01
81) Pyrene-d10	19.704	212	1216230	24.190	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.786	264	859922	16.235	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071923\
Data File : BP016330.D
Acq On : 19 Jul 2023 18:45
Operator : MA/JU
Sample : 03472-10
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
A46Q4

Quant Time: Jul 19 23:09:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 19 14:47:57 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/20/2023
Supervised By :mohammad ahmed 07/20/2023

