

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072523\
 Data File : BP016346.D
 Acq On : 25 Jul 2023 13:34
 Operator : MA/JU
 Sample : SST04005
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040607

Quant Time: Jul 25 23:29:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 25 16:07:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.102	152	342347	20.000	ng/u1	0.00
20) Naphthalene-d8	10.937	136	1492580	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.742	164	900490	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.507	188	1918212	20.000	ng/u1	0.00
79) Chrysene-d12	21.601	240	1724929	20.000	ng/u1	0.00
88) Perylene-d12	24.201	264	1977645	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.425	96	142805	16.448	ng/uL	0.00
4) Pyridine-d5	3.855	84	1021056	41.449	ng/u1	0.00
7) Phenol-d5	7.231	99	1198700	41.842	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.437	67	724457	42.120	ng/u1	0.00
11) 2-Chlorophenol-d4	7.619	132	948148	42.422	ng/u1	0.00
15) 4-Methylphenol-d8	8.796	113	971753	42.178	ng/u1	0.00
21) Nitrobenzene-d5	9.296	128	418777	45.564	ng/u1	0.00
24) 2-Nitrophenol-d4	10.013	143	368125	48.564	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.537	165	912580	43.434	ng/u1	0.00
31) 4-Chloroaniline-d4	11.084	131	1367041	41.823	ng/u1	0.00
46) Dimethylphthalate-d6	14.160	166	2721112	41.289	ng/u1	0.00
49) Acenaphthylene-d8	14.442	160	3244023	41.665	ng/u1	0.00
54) 4-Nitrophenol-d4	14.931	143	477064	43.202	ng/u1	0.00
60) Fluorene-d10	15.737	176	2301457	41.049	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.854	200	244719	38.733	ng/u1	0.00
73) Anthracene-d10	17.607	188	3530465	41.550	ng/u1	0.00
81) Pyrene-d10	19.830	212	4010198	40.944	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.036	264	4024052	42.174	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.467	88	151114	16.413	ng/uL	98
5) Pyridine	3.872	79	1045856	41.729	ng/u1	97
6) Benzaldehyde	7.249	77	733409	46.925	ng/u1	96
8) Phenol	7.255	94	1249407	41.937	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.531	93	1022165	41.882	ng/u1	99
12) 2-Chlorophenol	7.649	128	963550	41.931	ng/u1	99
13) 2-Methylphenol	8.531	108	961450	42.087	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.619	45	814515	27.571	ng/u1	99
16) Acetophenone	8.949	105	1559352	42.888	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.925	70	794722	41.974	ng/u1	99
18) 4-Methylphenol	8.866	108	1027366	42.092	ng/u1	96
19) Hexachloroethane	9.178	117	396375	41.229	ng/u1	100
22) Nitrobenzene	9.343	77	1086881	44.172	ng/u1	97
23) Isophorone	9.854	82	2292391	41.362	ng/u1	99
25) 2-Nitrophenol	10.043	139	457963	48.419	ng/u1	98
26) 2,4-Dimethylphenol	10.078	107	1114281	41.666	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.349	93	1370695	41.020	ng/u1	100
29) 2,4-Dichlorophenol	10.566	162	909788	42.833	ng/u1	98
30) Naphthalene	10.990	128	3166846	40.966	ng/u1	100
32) 4-Chloroaniline	11.107	127	1357547	42.365	ng/u1	99
33) Hexachlorobutadiene	11.225	225	568922	40.754	ng/u1	99
34) Caprolactam	11.901	113	181140	23.107	ng/u1	95
35) 4-Chloro-3-methylphenol	12.196	107	1009162	42.923	ng/u1	99
36) 2-Methylnaphthalene	12.584	142	2163185	41.060	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072523\
 Data File : BP016346.D
 Acq On : 25 Jul 2023 13:34
 Operator : MA/JU
 Sample : SST04005
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST040607


Quant Time: Jul 25 23:29:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 25 16:07:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.801	142	2170446	41.205	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.925	216	1100968	41.384	ng/ul	99
40) Hexachlorocyclopentadiene	12.884	237	661196	34.815	ng/ul	98
41) 2,4,6-Trichlorophenol	13.172	196	684050	43.641	ng/ul	100
42) 2,4,5-Trichlorophenol	13.237	196	737302	43.953	ng/ul	99
43) 1,1'-Biphenyl	13.584	154	2786465	40.981	ng/ul	99
44) 2-Chloronaphthalene	13.631	162	2193890	41.245	ng/ul	99
45) 2-Nitroaniline	13.848	65	539648	48.488	ng/ul	96
47) Dimethylphthalate	14.207	163	2725074	41.002	ng/ul	100
48) 2,6-Dinitrotoluene	14.337	165	468291	48.895	ng/ul	98
50) Acenaphthylene	14.472	152	3643361	41.074	ng/ul	99
51) 3-Nitroaniline	14.672	138	498632	44.502	ng/ul#	98
52) Acenaphthene	14.807	153	2348407	41.058	ng/ul	100
53) 2,4-Dinitrophenol	14.866	184	126433	30.866	ng/ul	95
55) 4-Nitrophenol	14.942	109	384702	43.219	ng/ul	95
56) Dibenzofuran	15.142	168	3184821	40.674	ng/ul	99
57) 2,4-Dinitrotoluene	15.113	165	679513	49.764	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.354	232	597472	44.647	ng/ul	99
59) Diethylphthalate	15.566	149	2759393	41.222	ng/ul	100
61) Fluorene	15.795	166	2598851	40.863	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.789	204	1274652	41.111	ng/ul	99
63) 4-Nitroaniline	15.837	138	481478	46.862	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.866	198	308399	40.829	ng/ul	100
67) N-Nitrosodiphenylamine	16.007	169	2217077	41.222	ng/ul	99
68) 4-Bromophenyl-phenylether	16.689	248	744859	40.945	ng/ul	98
69) Hexachlorobenzene	16.766	284	900534	41.281	ng/ul	99
70) Atrazine	16.960	200	833177	42.459	ng/ul	99
71) Pentachlorophenol	17.125	266	508426	41.359	ng/ul	99
72) Phenanthrene	17.548	178	4105742	41.133	ng/ul	99
74) Anthracene	17.642	178	4158992	40.904	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.537	216	1149675	41.521	ng/ul	99
76) Pentachlorobenzene	15.037	250	1010458	41.297	ng/ul	99
77) Carbazole	17.919	167	3794374	43.355	ng/ul	99
78) Di-n-butylphthalate	18.442	149	4708451	42.245	ng/ul	100
80) Fluoranthene	19.507	202	4754097	40.493	ng/ul	99
82) Pyrene	19.860	202	4899381	40.392	ng/ul	99
83) Butylbenzylphthalate	20.730	149	2039901	44.422	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.513	252	1665775	43.938	ng/ul	99
85) Benzo(a)anthracene	21.583	228	4846645	41.483	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.483	149	3069995	43.827	ng/ul	98
87) Chrysene	21.636	228	4511800	41.701	ng/ul	100
89) Di-n-octyl phthalate	22.489	149	5226963	42.047	ng/ul	100
90) Benzo(b)fluoranthene	23.395	252	4981855	42.064	ng/ul	99
91) Benzo(k)fluoranthene	23.448	252	4809084	40.803	ng/ul	99
93) Benzo(a)pyrene	24.089	252	4639012	41.590	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.989	276	5600418	43.482	ng/ul	98
95) Dibenzo(a,h)anthracene	27.030	278	4626257	43.428	ng/ul	99
96) Benzo(g,h,i)perylene	27.859	276	4382352	42.977	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 23:29:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 25 16:07:21 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 16.413 ng/uL
RT: 3.467 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion: 88 Resp: 151114
Ion Ratio Lower Upper
88 100
43 25.5 21.8 32.8
58 65.8 53.1 79.7



#5
Pyridine
Concen: 41.729 ng/uL
RT: 3.872 min Scan# 159
Delta R.T. -0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion: 79 Resp: 1045856
Ion Ratio Lower Upper
79 100
52 66.1 51.0 76.4
51 30.7 25.4 38.2









#6
Benzaldehyde
Concen: 46.925 ng/ul
RT: 7.249 min Scan# 733
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion: 77 Resp: 733409
Ion Ratio Lower Upper
77 100
105 96.0 80.2 120.2
106 92.2 76.5 114.7



#8
Phenol
Concen: 41.937 ng/ul
RT: 7.255 min Scan# 734
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion: 94 Resp: 1249407
Ion Ratio Lower Upper
94 100
65 26.4 21.4 32.2
66 35.6 28.2 42.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 41.882 ng/ul
 RT: 7.531 min Scan# 781
 Delta R.T. -0.000 min
 Lab File: BP016346.D
 Acq: 25 Jul 2023 13:34

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040607

Tgt Ion: 93 Resp: 1022165




Ion	Ratio	Lower	Upper
93	100		
63	69.9	57.0	85.4
95	32.4	25.8	38.6




#12
 2-Chlorophenol
 Concen: 41.931 ng/ul
 RT: 7.649 min Scan# 801
 Delta R.T. -0.000 min
 Lab File: BP016346.D
 Acq: 25 Jul 2023 13:34

Tgt Ion: 128 Resp: 963550

Ion	Ratio	Lower	Upper
128	100		
64	44.5	36.4	54.6
130	32.7	25.6	38.4









#13
2-Methylphenol
Concen: 42.087 ng/ul
RT: 8.531 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:108 Resp: 961450
Ion Ratio Lower Upper
108 100
107 88.5 71.3 106.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.571 ng/ul
RT: 8.619 min Scan# 966
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion: 45 Resp: 814515
Ion Ratio Lower Upper
45 100
77 16.7 13.7 20.5
79 12.4 9.6 14.4









#16
Acetophenone
Concen: 42.888 ng/ul
RT: 8.949 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:105 Resp: 1559352
Ion Ratio Lower Upper
105 100
77 76.7 61.2 91.8
51 23.8 19.1 28.7



#17
N-Nitroso-di-n-propylamine
Concen: 41.974 ng/ul
RT: 8.925 min Scan# 1018
Delta R.T. 0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion: 70 Resp: 794722
Ion Ratio Lower Upper
70 100
42 49.8 40.3 60.5
101 10.2 7.8 11.8
130 21.3 17.1 25.7









#18
4-Methylphenol
Concen: 42.092 ng/ul
RT: 8.866 min Scan# 1007
Delta R.T. 0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:108 Resp: 1027366
Ion Ratio Lower Upper
108 100
107 109.9 91.6 137.4



#19
Hexachloroethane
Concen: 41.229 ng/ul
RT: 9.178 min Scan# 1061
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:117 Resp: 396375
Ion Ratio Lower Upper
117 100
201 93.9 75.0 112.6
199 59.0 47.4 71.0









#22
Nitrobenzene
Concen: 44.172 ng/ul
RT: 9.343 min Scan# 1088
Delta R.T. 0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion: 77 Resp: 1086881
Ion Ratio Lower Upper
77 100
123 46.1 35.3 52.9
65 13.2 10.2 15.2



#23
Isophorone
Concen: 41.362 ng/ul
RT: 9.854 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion: 82 Resp: 2292391
Ion Ratio Lower Upper
82 100
95 7.2 5.6 8.4
138 17.1 14.0 21.0









#25
 2-Nitrophenol
 Concen: 48.419 ng/ul
 RT: 10.043 min Scan# 1208
 Delta R.T. -0.000 min
 Lab File: BP016346.D
 Acq: 25 Jul 2023 13:34




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040607


Tgt Ion:	139	Resp:	457963
Ion Ratio	Lower	Upper	
139	100		
65	50.3	38.7	58.1
109	31.4	25.3	37.9



#26
 2,4-Dimethylphenol
 Concen: 41.666 ng/ul
 RT: 10.078 min Scan# 1214
 Delta R.T. -0.000 min
 Lab File: BP016346.D
 Acq: 25 Jul 2023 13:34

Tgt Ion:	107	Resp:	1114281
Ion Ratio	Lower	Upper	
107	100		
121	49.8	39.8	59.6
122	86.2	69.0	103.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 41.020 ng/ul
 RT: 10.349 min Scan# 1259
 Delta R.T. 0.006 min
 Lab File: BP016346.D
 Acq: 25 Jul 2023 13:34

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040607

Tgt Ion: 93 Resp: 1370695




Ion	Ratio	Lower	Upper
93	100		
95	32.5	25.9	38.9
123	11.2	8.8	13.2




#29
 2,4-Dichlorophenol
 Concen: 42.833 ng/ul
 RT: 10.566 min Scan# 1297
 Delta R.T. 0.006 min
 Lab File: BP016346.D
 Acq: 25 Jul 2023 13:34

Tgt Ion: 162 Resp: 909788

Ion	Ratio	Lower	Upper
162	100		
164	65.5	51.4	77.0
98	36.7	30.6	46.0









#30
Naphthalene
Concen: 40.966 ng/ul
RT: 10.990 min Scan# 1369
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:128 Resp: 3166846
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 42.365 ng/ul
RT: 11.107 min Scan# 1389
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:127 Resp: 1357547
Ion Ratio Lower Upper
127 100
129 31.2 25.2 37.8









#33
Hexachlorobutadiene
Concen: 40.754 ng/ul
RT: 11.225 min Scan# 1409
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:225 Resp: 568922
Ion Ratio Lower Upper
225 100
223 63.2 50.1 75.1
227 62.7 51.5 77.3



#34
Caprolactam
Concen: 23.107 ng/ul
RT: 11.901 min Scan# 1524
Delta R.T. 0.012 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:113 Resp: 181140
Ion Ratio Lower Upper
113 100
55 146.6 122.2 183.4
56 111.7 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 42.923 ng/ul
RT: 12.196 min Scan# 1573
Delta R.T. 0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:107 Resp: 1009162
Ion Ratio Lower Upper
107 100
144 25.7 20.2 30.2
142 78.8 62.6 93.8



#36
2-Methylnaphthalene
Concen: 41.060 ng/ul
RT: 12.584 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:142 Resp: 2163185
Ion Ratio Lower Upper
142 100
141 88.3 71.9 107.9





#37

1-Methylnaphthalene

Concen: 41.205 ng/ul

RT: 12.801 min Scan# 1677

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:142 Resp: 2170446

Ion Ratio Lower Upper

142 100

141 91.9 73.2 109.8

116 3.7 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.384 ng/ul

RT: 12.925 min Scan# 1698

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Tgt Ion:216 Resp: 1100968




Ion Ratio Lower Upper


216 100

214 78.4 63.6 95.4

179 21.1 17.4 26.0

108 18.2 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 34.815 ng/ul

RT: 12.884 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:237 Resp: 661196

Ion Ratio Lower Upper

237 100

235 62.6 51.2 76.8

272 12.7 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 43.641 ng/ul

RT: 13.172 min Scan# 1740

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:196 Resp: 684050


Ion Ratio Lower Upper

196 100

198 97.2 77.7 116.5

200 31.1 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 43.953 ng/ul

RT: 13.237 min Scan# 1751

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:196 Resp: 737302

Ion Ratio Lower Upper

196 100

198 95.1 77.0 115.4

200 31.2 24.8 37.2



#43

1,1'-Biphenyl

Concen: 40.981 ng/ul

RT: 13.584 min Scan# 1810

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:154 Resp: 2786465


Ion Ratio Lower Upper

154 100

153 40.0 32.3 48.5

76 13.6 10.7 16.1









#44
2-Chloronaphthalene
Concen: 41.245 ng/ul
RT: 13.631 min Scan# 1818
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:162 Resp: 2193890
Ion Ratio Lower Upper
162 100
164 33.4 26.6 40.0
127 38.2 31.0 46.4



#45
2-Nitroaniline
Concen: 48.488 ng/ul
RT: 13.848 min Scan# 1855
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion: 65 Resp: 539648
Ion Ratio Lower Upper
65 100
92 72.1 56.8 85.2
138 110.2 83.9 125.9









#47
Dimethylphthalate
Concen: 41.002 ng/ul
RT: 14.207 min Scan# 1916
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:163 Resp: 2725074
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 48.895 ng/ul
RT: 14.337 min Scan# 1938
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:165 Resp: 468291
Ion Ratio Lower Upper
165 100
89 65.6 53.8 80.6
121 23.6 20.0 30.0





#50

Acenaphthylene

Concen: 41.074 ng/ul

RT: 14.472 min Scan# 1961

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:152 Resp: 3643361

Ion Ratio Lower Upper

152 100

151 20.4 16.0 24.0

153 13.2 10.6 16.0



#51

3-Nitroaniline

Concen: 44.502 ng/ul

RT: 14.672 min Scan# 1995

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:138 Resp: 498632


Ion Ratio Lower Upper

138 100

108 15.6 16.6 24.8#

92 117.7 94.3 141.5









#52
Acenaphthene
Concen: 41.058 ng/ul
RT: 14.807 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:153 Resp: 2348407
Ion Ratio Lower Upper
153 100
152 48.6 38.8 58.2
154 89.0 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 30.866 ng/ul
RT: 14.866 min Scan# 2028
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:184 Resp: 126433
Ion Ratio Lower Upper
184 100
63 72.6 60.0 90.0
154 63.3 55.2 82.8





#55

4-Nitrophenol

Concen: 43.219 ng/ul

RT: 14.942 min Scan# 2056

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:109 Resp: 384702

Ion Ratio Lower Upper

109 100

139 131.7 103.5 155.3

65 131.6 98.6 148.0



#56

Dibenzofuran

Concen: 40.674 ng/ul

RT: 15.142 min Scan# 2075

Delta R.T. -0.000 min

Lab File: BP016346.D




Acq: 25 Jul 2023 13:34


Tgt Ion:168 Resp: 3184821

Ion Ratio Lower Upper

168 100

139 37.1 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 49.764 ng/ul
RT: 15.113 min Scan# 2070
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:165 Resp: 679513
Ion Ratio Lower Upper
165 100
63 46.9 37.4 56.2
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 44.647 ng/ul
RT: 15.354 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:232 Resp: 597472
Ion Ratio Lower Upper
232 100
131 44.5 35.1 52.7
130 2.4 2.1 3.1
166 31.1 24.2 36.2









#59
Diethylphthalate
Concen: 41.222 ng/ul
RT: 15.566 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:149 Resp: 2759393
Ion Ratio Lower Upper
149 100
177 22.4 17.9 26.9
150 12.3 9.9 14.9



#61
Fluorene
Concen: 40.863 ng/ul
RT: 15.795 min Scan# 2186
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:166 Resp: 2598851
Ion Ratio Lower Upper
166 100
165 96.8 77.0 115.4
167 13.4 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 41.111 ng/ul
RT: 15.789 min Scan# 2185
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:204 Resp: 1274652
Ion Ratio Lower Upper
204 100
206 33.5 26.5 39.7
141 56.6 46.2 69.4



#63
4-Nitroaniline
Concen: 46.862 ng/ul
RT: 15.837 min Scan# 2193
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:138 Resp: 481478
Ion Ratio Lower Upper
138 100
92 55.3 45.3 67.9
108 91.5 76.4 114.6









#66
4,6-Dinitro-2-methylphenol
Concen: 40.829 ng/ul
RT: 15.866 min Scan# 2198
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:198 Resp: 308399
Ion Ratio Lower Upper
198 100
182 4.7 3.8 5.8
77 27.3 21.6 32.4



#67
N-Nitrosodiphenylamine
Concen: 41.222 ng/ul
RT: 16.007 min Scan# 2222
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:169 Resp: 2217077
Ion Ratio Lower Upper
169 100
168 67.0 54.0 81.0
167 37.1 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 40.945 ng/ul
RT: 16.689 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SSTD040607


Tgt Ion:248 Resp: 744859
Ion Ratio Lower Upper
248 100
250 98.7 78.1 117.1
141 74.4 61.2 91.8



#69
Hexachlorobenzene
Concen: 41.281 ng/ul
RT: 16.766 min Scan# 2351
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:284 Resp: 900534
Ion Ratio Lower Upper
284 100
142 36.7 29.7 44.5
249 30.4 24.8 37.2





#70

Atrazine

Concen: 42.459 ng/ul

RT: 16.960 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:200 Resp: 833177

Ion Ratio Lower Upper

200 100

173 26.9 21.7 32.5

215 49.1 38.5 57.7



#71

Pentachlorophenol

Concen: 41.359 ng/ul

RT: 17.125 min Scan# 2412

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:266 Resp: 508426


Ion Ratio Lower Upper

266 100

264 62.5 50.9 76.3

268 64.3 51.5 77.3





#72

Phenanthrene

Concen: 41.133 ng/ul

RT: 17.548 min Scan# 2485

Delta R.T. -0.006 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:178 Resp: 4105742

Ion Ratio Lower Upper

178 100

179 15.9 12.3 18.5

176 19.8 15.4 23.0



#74

Anthracene

Concen: 40.904 ng/ul

RT: 17.642 min Scan# 2500

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:178 Resp: 4158992


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.0 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 41.521 ng/uL

RT: 13.537 min Scan# 1802

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:216 Resp: 1149675

Ion Ratio Lower Upper

216 100

214 77.9 62.8 94.2

218 48.6 38.6 58.0



#76

Pentachlorobenzene

Concen: 41.297 ng/uL

RT: 15.037 min Scan# 2057

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:250 Resp: 1010458


Ion Ratio Lower Upper

250 100

248 62.1 50.1 75.1

252 64.2 51.0 76.6





#77

Carbazole

Concen: 43.355 ng/ul

RT: 17.919 min Scan# 2547

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:167 Resp: 3794374

Ion Ratio Lower Upper

167 100

166 22.1 17.3 25.9

139 13.2 10.4 15.6



#78

Di-n-butylphthalate

Concen: 42.245 ng/ul

RT: 18.442 min Scan# 2636

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:149 Resp: 4708451


Ion Ratio Lower Upper

149 100

150 9.3 7.5 11.3

104 5.3 4.3 6.5





#80

Fluoranthene

Concen: 40.493 ng/ul

RT: 19.507 min Scan# 2817

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:202 Resp: 4754097

Ion Ratio Lower Upper

202 100

101 11.3 9.1 13.7

100 8.6 7.1 10.7



#82

Pyrene

Concen: 40.392 ng/ul

RT: 19.860 min Scan# 2877

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:202 Resp: 4899381


Ion Ratio Lower Upper

202 100

101 13.9 11.2 16.8

100 11.2 9.4 14.0





#83

Butylbenzylphthalate

Concen: 44.422 ng/ul

RT: 20.730 min Scan# 3025

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:149 Resp: 2039901

Ion Ratio Lower Upper

149 100

91 70.3 58.5 87.7

206 21.1 17.0 25.4



#84

3,3'-Dichlorobenzidine

Concen: 43.938 ng/ul

RT: 21.513 min Scan# 3158

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:252 Resp: 1665775


Ion Ratio Lower Upper

252 100

254 64.6 51.4 77.0

126 12.3 10.3 15.5





#85

Benzo(a)anthracene

Concen: 41.483 ng/ul

RT: 21.583 min Scan# 3170

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:228 Resp: 4846645

Ion Ratio Lower Upper

228 100

229 19.8 15.9 23.9

226 27.8 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 43.827 ng/ul

RT: 21.483 min Scan# 3153

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:149 Resp: 3069995


Ion Ratio Lower Upper

149 100

167 28.0 21.7 32.5

279 4.4 3.0 4.4





#87

Chrysene

Concen: 41.701 ng/ul

RT: 21.636 min Scan# 3180

Delta R.T. -0.006 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:228 Resp: 4511800

Ion Ratio Lower Upper

228 100

226 30.5 24.5 36.7

229 19.9 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 42.047 ng/ul




RT: 22.489 min Scan# 3324


Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Tgt Ion:149 Resp: 5226963





#90

Benzo(b)fluoranthene

Concen: 42.064 ng/ul

RT: 23.395 min Scan# 3478

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:252 Resp: 4981855

Ion Ratio Lower Upper

252 100

253 22.0 17.9 26.9

125 10.4 8.8 13.2



#91

Benzo(k)fluoranthene

Concen: 40.803 ng/ul

RT: 23.448 min Scan# 3487

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:252 Resp: 4809084


Ion Ratio Lower Upper

252 100

253 22.6 17.7 26.5

125 10.7 9.0 13.6





#93

Benzo(a)pyrene

Concen: 41.590 ng/ul

RT: 24.089 min Scan# 3596

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SSTD040607





Tgt Ion:252 Resp: 4639012

Ion Ratio Lower Upper

252 100

253 22.3 18.0 27.0

125 11.8 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.482 ng/ul

RT: 26.989 min Scan# 4089

Delta R.T. -0.000 min

Lab File: BP016346.D

Acq: 25 Jul 2023 13:34




Tgt Ion:276 Resp: 5600418


Ion Ratio Lower Upper

276 100

138 21.8 18.1 27.1

277 24.5 19.0 28.4









#95
Dibenzo(a,h)anthracene
Concen: 43.428 ng/ul
RT: 27.030 min Scan# 4095
Delta R.T. 0.006 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Instrument :
BNA_P
ClientSampleId :
SSTD040607

Tgt Ion:278 Resp: 4626257
Ion Ratio Lower Upper
278 100
139 16.5 14.2 21.2
279 24.3 19.4 29.2



#96
Benzo(g,h,i)perylene
Concen: 42.977 ng/ul
RT: 27.859 min Scan# 4237
Delta R.T. -0.000 min
Lab File: BP016346.D
Acq: 25 Jul 2023 13:34

Tgt Ion:276 Resp: 4382352
Ion Ratio Lower Upper
276 100
138 23.0 17.8 26.6
277 23.9 19.0 28.4

