

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP072523\  
 Data File : BP016349.D  
 Acq On : 25 Jul 2023 16:33  
 Operator : MA/JU  
 Sample : SSTDICV020  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SICV610

Quant Time: Jul 25 23:57:54 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP072523.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue Jul 25 23:56:16 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| Internal Standards            |        |      |          |        |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.099  | 152  | 421813   | 20.000 | ng/ul  | 0.00     |
| 20) Naphthalene-d8            | 10.934 | 136  | 1831944  | 20.000 | ng/ul  | 0.00     |
| 38) Acenaphthene-d10          | 14.740 | 164  | 1162951  | 20.000 | ng/ul  | 0.00     |
| 64) Phenanthrene-d10          | 17.504 | 188  | 2591048  | 20.000 | ng/ul  | 0.00     |
| 79) Chrysene-d12              | 21.598 | 240  | 2417109  | 20.000 | ng/ul  | 0.00     |
| 88) Perylene-d12              | 24.198 | 264  | 2651476  | 20.000 | ng/ul  | 0.00     |
| System Monitoring Compounds   |        |      |          |        |        |          |
| 3) 1,4-Dioxane-d8             | 3.428  | 96   | 74872    | 6.999  | ng/uL  | 0.00     |
| 4) Pyridine-d5                | 3.852  | 84   | 587352   | 19.351 | ng/ul  | 0.00     |
| 7) Phenol-d5                  | 7.222  | 99   | 689043   | 19.521 | ng/ul  | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.434  | 67   | 440893   | 20.804 | ng/ul  | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.611  | 132  | 517745   | 18.801 | ng/ul  | 0.00     |
| 15) 4-Methylphenol-d8         | 8.793  | 113  | 530021   | 18.671 | ng/ul  | 0.00     |
| 21) Nitrobenzene-d5           | 9.293  | 128  | 237694   | 21.071 | ng/ul  | 0.00     |
| 24) 2-Nitrophenol-d4          | 10.010 | 143  | 211668   | 22.751 | ng/ul  | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.534 | 165  | 509913   | 19.773 | ng/ul  | 0.00     |
| 31) 4-Chloroaniline-d4        | 11.081 | 131  | 752231   | 18.750 | ng/ul  | 0.00     |
| 46) Dimethylphthalate-d6      | 14.157 | 166  | 1572125  | 18.471 | ng/ul  | 0.00     |
| 49) Acenaphthylene-d8         | 14.440 | 160  | 1824848  | 18.148 | ng/ul  | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.922 | 143  | 271885   | 19.065 | ng/ul  | 0.00     |
| 60) Fluorene-d10              | 15.734 | 176  | 1326272  | 18.317 | ng/ul  | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.845 | 200  | 144073   | 16.882 | ng/ul  | 0.00     |
| 73) Anthracene-d10            | 17.604 | 188  | 2082734  | 18.146 | ng/ul  | 0.00     |
| 81) Pyrene-d10                | 19.828 | 212  | 2364652  | 17.229 | ng/ul  | 0.00     |
| 92) Benzo(a)pyrene-d12        | 24.027 | 264  | 2259564  | 17.663 | ng/ul  | 0.00     |
| Target Compounds              |        |      |          |        |        |          |
|                               |        |      |          |        | Qvalue |          |
| 2) 1,4-Dioxane                | 3.464  | 88   | 78739    | 6.941  | ng/uL  | 99       |
| 5) Pyridine                   | 3.875  | 79   | 496047   | 16.063 | ng/ul  | 98       |
| 6) Benzaldehyde               | 7.246  | 77   | 454124   | 23.582 | ng/ul  | 99       |
| 8) Phenol                     | 7.252  | 94   | 647160   | 17.630 | ng/ul  | 99       |
| 10) Bis(2-Chloroethyl)ether   | 7.528  | 93   | 521199   | 17.332 | ng/ul  | 99       |
| 12) 2-Chlorophenol            | 7.646  | 128  | 502618   | 17.752 | ng/ul  | 99       |
| 13) 2-Methylphenol            | 8.522  | 108  | 489999   | 17.409 | ng/ul  | 99       |
| 14) 2,2'-oxybis(1-Chloropr... | 8.616  | 45   | 316305   | 8.693  | ng/ul  | 99       |
| 16) Acetophenone              | 8.946  | 105  | 868671   | 19.390 | ng/ul  | 99       |
| 17) N-Nitroso-di-n-propyla... | 8.916  | 70   | 418447   | 17.937 | ng/ul  | 99       |
| 18) 4-Methylphenol            | 8.858  | 108  | 528818   | 17.584 | ng/ul  | 96       |
| 19) Hexachloroethane          | 9.175  | 117  | 207253   | 17.496 | ng/ul  | 99       |
| 22) Nitrobenzene              | 9.334  | 77   | 569676   | 18.863 | ng/ul  | 98       |
| 23) Isophorone                | 9.846  | 82   | 1166694  | 17.151 | ng/ul  | 99       |
| 25) 2-Nitrophenol             | 10.040 | 139  | 239232   | 20.608 | ng/ul  | 98       |
| 26) 2,4-Dimethylphenol        | 10.075 | 107  | 513708   | 15.651 | ng/ul  | 99       |
| 27) Bis(2-Chloroethoxy)met... | 10.340 | 93   | 702446   | 17.128 | ng/ul  | 99       |
| 29) 2,4-Dichlorophenol        | 10.557 | 162  | 475860   | 18.253 | ng/ul  | 99       |
| 30) Naphthalene               | 10.981 | 128  | 1629677  | 17.176 | ng/ul  | 100      |
| 32) 4-Chloroaniline           | 11.105 | 127  | 706422   | 17.961 | ng/ul  | 99       |
| 33) Hexachlorobutadiene       | 11.222 | 225  | 290658   | 16.964 | ng/ul  | 99       |
| 34) Caprolactam               | 11.893 | 113  | 179839   | 19.366 | ng/ul  | 97       |
| 35) 4-Chloro-3-methylphenol   | 12.187 | 107  | 527741   | 18.288 | ng/ul  | 98       |
| 36) 2-Methylnaphthalene       | 12.581 | 142  | 1116374  | 17.265 | ng/ul  | 98       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP072523\  
 Data File : BP016349.D  
 Acq On : 25 Jul 2023 16:33  
 Operator : MA/JU  
 Sample : SSTDICV020  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SICV610


Quant Time: Jul 25 23:57:54 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP072523.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue Jul 25 23:56:16 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 37) 1-Methylnaphthalene       | 12.799 | 142  | 1093878  | 16.920 | ng/ul  | 100      |
| 39) 1,2,4,5-Tetrachloroben... | 12.922 | 216  | 607172   | 17.672 | ng/ul  | 99       |
| 40) Hexachlorocyclopentadiene | 12.881 | 237  | 417448   | 17.020 | ng/ul  | 98       |
| 41) 2,4,6-Trichlorophenol     | 13.169 | 196  | 363236   | 17.944 | ng/ul  | 97       |
| 42) 2,4,5-Trichlorophenol     | 13.234 | 196  | 400803   | 18.501 | ng/ul  | 99       |
| 43) 1,1'-Biphenyl             | 13.581 | 154  | 1563684  | 17.807 | ng/ul  | 99       |
| 44) 2-Chloronaphthalene       | 13.628 | 162  | 1151138  | 16.757 | ng/ul  | 99       |
| 45) 2-Nitroaniline            | 13.846 | 65   | 294552   | 20.493 | ng/ul  | 98       |
| 47) Dimethylphthalate         | 14.204 | 163  | 1453772  | 16.937 | ng/ul  | 99       |
| 48) 2,6-Dinitrotoluene        | 14.334 | 165  | 292469   | 23.645 | ng/ul  | 95       |
| 50) Acenaphthylene            | 14.469 | 152  | 1934343  | 16.885 | ng/ul  | 100      |
| 51) 3-Nitroaniline            | 14.669 | 138  | 320586   | 22.155 | ng/ul# | 96       |
| 52) Acenaphthene              | 14.804 | 153  | 1214532  | 16.442 | ng/ul  | 99       |
| 53) 2,4-Dinitrophenol         | 14.863 | 184  | 77497    | 14.649 | ng/ul  | 91       |
| 55) 4-Nitrophenol             | 14.940 | 109  | 214555   | 18.664 | ng/ul  | 98       |
| 56) Dibenzofuran              | 15.140 | 168  | 1749262  | 17.299 | ng/ul  | 99       |
| 57) 2,4-Dinitrotoluene        | 15.110 | 165  | 371205   | 21.050 | ng/ul  | 97       |
| 58) 2,3,4,6-Tetrachlorophenol | 15.345 | 232  | 342560   | 19.821 | ng/ul  | 96       |
| 59) Diethylphthalate          | 15.557 | 149  | 1475889  | 17.072 | ng/ul  | 99       |
| 61) Fluorene                  | 15.793 | 166  | 1414641  | 17.223 | ng/ul  | 98       |
| 62) 4-Chlorophenyl-phenyle... | 15.787 | 204  | 684657   | 17.098 | ng/ul  | 99       |
| 63) 4-Nitroaniline            | 15.834 | 138  | 315572   | 23.783 | ng/ul  | 94       |
| 66) 4,6-Dinitro-2-methylph... | 15.863 | 198  | 158362   | 15.521 | ng/ul  | 99       |
| 67) N-Nitrosodiphenylamine    | 16.004 | 169  | 1249918  | 17.205 | ng/ul  | 99       |
| 68) 4-Bromophenyl-phenylether | 16.687 | 248  | 409756   | 16.675 | ng/ul  | 99       |
| 69) Hexachlorobenzene         | 16.763 | 284  | 475772   | 16.146 | ng/ul  | 99       |
| 70) Atrazine                  | 16.957 | 200  | 458930   | 17.314 | ng/ul  | 98       |
| 71) Pentachlorophenol         | 17.122 | 266  | 254722   | 15.340 | ng/ul  | 99       |
| 72) Phenanthrene              | 17.551 | 178  | 2245961  | 16.658 | ng/ul  | 100      |
| 74) Anthracene                | 17.639 | 178  | 2263912  | 16.484 | ng/ul  | 100      |
| 75) 1,2,3,4-Tetrachloroben... | 13.534 | 216  | 592081   | 15.831 | ng/ul  | 99       |
| 76) Pentachlorobenzene        | 15.034 | 250  | 555459   | 16.806 | ng/ul  | 99       |
| 77) Carbazole                 | 17.910 | 167  | 2120298  | 17.936 | ng/ul  | 100      |
| 78) Di-n-butylphthalate       | 18.439 | 149  | 2602282  | 17.285 | ng/ul  | 99       |
| 80) Fluoranthene              | 19.498 | 202  | 2584748  | 15.711 | ng/ul  | 98       |
| 82) Pyrene                    | 19.857 | 202  | 2717641  | 15.989 | ng/ul  | 99       |
| 83) Butylbenzylphthalate      | 20.728 | 149  | 1103170  | 17.144 | ng/ul  | 99       |
| 84) 3,3'-Dichlorobenzidine    | 21.510 | 252  | 1040049  | 19.577 | ng/ul  | 99       |
| 85) Benzo(a)anthracene        | 21.575 | 228  | 2685524  | 16.403 | ng/ul  | 99       |
| 86) Bis(2-ethylhexyl)phtha... | 21.480 | 149  | 1692589  | 17.244 | ng/ul  | 99       |
| 87) Chrysene                  | 21.633 | 228  | 2455275  | 16.195 | ng/ul  | 99       |
| 89) Di-n-octyl phthalate      | 22.480 | 149  | 2831728  | 16.990 | ng/ul  | 100      |
| 90) Benzo(b)fluoranthene      | 23.386 | 252  | 2633465  | 16.585 | ng/ul  | 99       |
| 91) Benzo(k)fluoranthene      | 23.439 | 252  | 2473492  | 15.653 | ng/ul  | 99       |
| 93) Benzo(a)pyrene            | 24.080 | 252  | 2298905  | 15.373 | ng/ul  | 99       |
| 94) Indeno(1,2,3-cd)pyrene    | 26.980 | 276  | 2738882  | 15.861 | ng/ul  | 99       |
| 95) Dibenzo(a,h)anthracene    | 27.015 | 278  | 2295738  | 16.074 | ng/ul  | 99       |
| 96) Benzo(g,h,i)perylene      | 27.839 | 276  | 2103818  | 15.389 | ng/ul  | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 25 23:57:54 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP072523.MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Tue Jul 25 23:56:16 2023  
Response via : Initial Calibration









#2  
1,4-Dioxane  
Concen: 6.941 ng/uL  
RT: 3.464 min Scan# 30  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion: 88 Resp: 78739  
Ion Ratio Lower Upper  
88 100  
43 27.4 21.8 32.8  
58 65.6 53.1 79.7



#5  
Pyridine  
Concen: 16.063 ng/uL  
RT: 3.875 min Scan# 100  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion: 79 Resp: 496047  
Ion Ratio Lower Upper  
79 100  
52 66.0 51.0 76.4  
51 32.7 25.4 38.2









#6  
Benzaldehyde  
Concen: 23.582 ng/ul  
RT: 7.246 min Scan# 61  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion: 77 Resp: 454124  
Ion Ratio Lower Upper  
77 100  
105 98.3 80.2 120.2  
106 95.2 76.5 114.7



#8  
Phenol  
Concen: 17.630 ng/ul  
RT: 7.252 min Scan# 674  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion: 94 Resp: 647160  
Ion Ratio Lower Upper  
94 100  
65 27.4 21.4 32.2  
66 35.7 28.2 42.4









#10  
 Bis(2-Chloroethyl)ether  
 Concen: 17.332 ng/ul  
 RT: 7.528 min Scan# 71  
 Delta R.T. -0.003 min  
 Lab File: BP016349.D  
 Acq: 25 Jul 2023 16:33




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SICV610


Tgt Ion: 93 Resp: 521199  
 Ion Ratio Lower Upper  
 93 100  
 63 70.7 57.0 85.4  
 95 31.9 25.8 38.6



#12  
 2-Chlorophenol  
 Concen: 17.752 ng/ul  
 RT: 7.646 min Scan# 741  
 Delta R.T. -0.003 min  
 Lab File: BP016349.D  
 Acq: 25 Jul 2023 16:33

Tgt Ion:128 Resp: 502618  
 Ion Ratio Lower Upper  
 128 100  
 64 45.9 36.4 54.6  
 130 32.5 25.6 38.4





#13

2-Methylphenol

Concen: 17.409 ng/ul

RT: 8.522 min Scan# 890

Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :





SICV610

Tgt Ion:108 Resp: 489999

Ion Ratio Lower Upper

108 100

107 88.0 71.3 106.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 8.693 ng/ul

RT: 8.616 min Scan# 906

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion: 45 Resp: 316305


Ion Ratio Lower Upper

45 100

77 16.1 13.7 20.5

79 12.1 9.6 14.4









#16  
Acetophenone  
Concen: 19.390 ng/ul  
RT: 8.946 min Scan# 90  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610

Tgt Ion:105 Resp: 868671  
Ion Ratio Lower Upper  
105 100  
77 76.0 61.2 91.8  
51 24.2 19.1 28.7




#17  
N-Nitroso-di-n-propylamine  
Concen: 17.937 ng/ul  
RT: 8.916 min Scan# 957  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion: 70 Resp: 418447  
Ion Ratio Lower Upper  
70 100  
42 50.8 40.3 60.5  
101 9.9 7.8 11.8  
130 21.8 17.1 25.7







#18

4-Methylphenol

Concen: 17.584 ng/ul

RT: 8.858 min Scan# 947

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :





SICV610

Tgt Ion:108 Resp: 528818

Ion Ratio Lower Upper

108 100

107 110.3 91.6 137.4



#19

Hexachloroethane

Concen: 17.496 ng/ul

RT: 9.175 min Scan# 1001

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:117 Resp: 207253


Ion Ratio Lower Upper

117 100

201 93.6 75.0 112.6

199 58.2 47.4 71.0









#22  
Nitrobenzene  
Concen: 18.863 ng/ul  
RT: 9.334 min Scan# 1105  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion: 77 Resp: 569676  
Ion Ratio Lower Upper  
77 100  
123 46.0 35.3 52.9  
65 13.1 10.2 15.2



#23  
Isophorone  
Concen: 17.151 ng/ul  
RT: 9.846 min Scan# 1115  
Delta R.T. -0.009 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion: 82 Resp: 1166694  
Ion Ratio Lower Upper  
82 100  
95 7.2 5.6 8.4  
138 17.0 14.0 21.0









#25  
2-Nitrophenol  
Concen: 20.608 ng/ul  
RT: 10.040 min Scan# 1154  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:139 Resp: 239232  
Ion Ratio Lower Upper  
139 100  
65 49.9 38.7 58.1  
109 31.3 25.3 37.9



#26  
2,4-Dimethylphenol  
Concen: 15.651 ng/ul  
RT: 10.075 min Scan# 1154  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:107 Resp: 513708  
Ion Ratio Lower Upper  
107 100  
121 50.6 39.8 59.6  
122 86.7 69.0 103.4









#27  
 Bis(2-Chloroethoxy)methane  
 Concen: 17.128 ng/ul  
 RT: 10.340 min Scan# 1199  
 Delta R.T. -0.003 min  
 Lab File: BP016349.D  
 Acq: 25 Jul 2023 16:33




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SICV610


Tgt Ion: 93 Resp: 702446  
 Ion Ratio Lower Upper  
 93 100  
 95 32.6 25.9 38.9  
 123 11.3 8.8 13.2



#29  
 2,4-Dichlorophenol  
 Concen: 18.253 ng/ul  
 RT: 10.557 min Scan# 1236  
 Delta R.T. -0.003 min  
 Lab File: BP016349.D  
 Acq: 25 Jul 2023 16:33

Tgt Ion:162 Resp: 475860  
 Ion Ratio Lower Upper  
 162 100  
 164 64.4 51.4 77.0  
 98 36.8 30.6 46.0









#30  
Naphthalene  
Concen: 17.176 ng/ul  
RT: 10.981 min Scan# 1308  
Delta R.T. -0.009 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:128 Resp: 1629677  
Ion Ratio Lower Upper  
128 100  
129 11.0 8.7 13.1  
127 13.3 10.6 15.8



#32  
4-Chloroaniline  
Concen: 17.961 ng/ul  
RT: 11.105 min Scan# 1329  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:127 Resp: 706422  
Ion Ratio Lower Upper  
127 100  
129 32.2 25.2 37.8





#33

Hexachlorobutadiene

Concen: 16.964 ng/ul

RT: 11.222 min Scan# 1409

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:225 Resp: 290658

Ion Ratio Lower Upper

225 100

223 62.1 50.1 75.1

227 64.0 51.5 77.3



#34

Caprolactam

Concen: 19.366 ng/ul

RT: 11.893 min Scan# 1463

Delta R.T. 0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:113 Resp: 179839


Ion Ratio Lower Upper

113 100

55 154.3 122.2 183.4

56 111.9 93.8 140.8









#35  
4-Chloro-3-methylphenol  
Concen: 18.288 ng/ul  
RT: 12.187 min Scan# 1573  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:107 Resp: 527741  
Ion Ratio Lower Upper  
107 100  
144 26.2 20.2 30.2  
142 79.8 62.6 93.8



#36  
2-Methylnaphthalene  
Concen: 17.265 ng/ul  
RT: 12.581 min Scan# 1580  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:142 Resp: 1116374  
Ion Ratio Lower Upper  
142 100  
141 87.9 71.9 107.9









#37  
1-Methylnaphthalene  
Concen: 16.920 ng/ul  
RT: 12.799 min Scan# 1617  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610

Tgt Ion:142 Resp: 1093878  
Ion Ratio Lower Upper  
142 100  
141 91.9 73.2 109.8  
116 3.7 3.0 4.6




#39  
1,2,4,5-Tetrachlorobenzene  
Concen: 17.672 ng/ul  
RT: 12.922 min Scan# 1638  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:216 Resp: 607172  
Ion Ratio Lower Upper  
216 100  
214 78.7 63.6 95.4  
179 21.4 17.4 26.0  
108 18.7 14.6 21.8











#40  
Hexachlorocyclopentadiene  
Concen: 17.020 ng/ul  
RT: 12.881 min Scan# 1680  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:237 Resp: 417448  
Ion Ratio Lower Upper  
237 100  
235 62.3 51.2 76.8  
272 13.0 10.3 15.5



#41  
2,4,6-Trichlorophenol  
Concen: 17.944 ng/ul  
RT: 13.169 min Scan# 1680  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:196 Resp: 363236  
Ion Ratio Lower Upper  
196 100  
198 93.9 77.7 116.5  
200 30.2 25.0 37.6









#42  
2,4,5-Trichlorophenol  
Concen: 18.501 ng/ul  
RT: 13.234 min Scan# 1750  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:196 Resp: 400803  
Ion Ratio Lower Upper  
196 100  
198 95.1 77.0 115.4  
200 31.0 24.8 37.2



#43  
1,1'-Biphenyl  
Concen: 17.807 ng/ul  
RT: 13.581 min Scan# 1750  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:154 Resp: 1563684  
Ion Ratio Lower Upper  
154 100  
153 39.7 32.3 48.5  
76 13.1 10.7 16.1





#44

2-Chloronaphthalene

Concen: 16.757 ng/ul

RT: 13.628 min Scan# 1

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:162 Resp: 1151138

Ion Ratio Lower Upper

162 100

164 32.6 26.6 40.0

127 38.1 31.0 46.4



#45

2-Nitroaniline

Concen: 20.493 ng/ul

RT: 13.846 min Scan# 1795

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion: 65 Resp: 294552


Ion Ratio Lower Upper

65 100

92 71.7 56.8 85.2

138 108.2 83.9 125.9









#47  
Dimethylphthalate  
Concen: 16.937 ng/ul  
RT: 14.204 min Scan# 1856  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:163 Resp: 1453772  
Ion Ratio Lower Upper  
163 100  
194 3.9 3.3 4.9  
164 10.1 7.9 11.9



#48  
2,6-Dinitrotoluene  
Concen: 23.645 ng/ul  
RT: 14.334 min Scan# 1878  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:165 Resp: 292469  
Ion Ratio Lower Upper  
165 100  
89 63.5 53.8 80.6  
121 22.5 20.0 30.0









#50  
Acenaphthylene  
Concen: 16.885 ng/ul  
RT: 14.469 min Scan# 1961  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:152 Resp: 1934343  
Ion Ratio Lower Upper  
152 100  
151 20.0 16.0 24.0  
153 13.2 10.6 16.0



#51  
3-Nitroaniline  
Concen: 22.155 ng/ul  
RT: 14.669 min Scan# 1935  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:138 Resp: 320586  
Ion Ratio Lower Upper  
138 100  
108 13.3 16.6 24.8#  
92 119.4 94.3 141.5









#52  
Acenaphthene  
Concen: 16.442 ng/ul  
RT: 14.804 min Scan# 1958  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:153 Resp: 1214532  
Ion Ratio Lower Upper  
153 100  
152 47.7 38.8 58.2  
154 88.4 71.7 107.5



#53  
2,4-Dinitrophenol  
Concen: 14.649 ng/ul  
RT: 14.863 min Scan# 1968  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:184 Resp: 77497  
Ion Ratio Lower Upper  
184 100  
63 67.9 60.0 90.0  
154 60.8 55.2 82.8









#55  
4-Nitrophenol  
Concen: 18.664 ng/ul  
RT: 14.940 min Scan# 1981  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:109 Resp: 214555  
Ion Ratio Lower Upper  
109 100  
139 127.3 103.5 155.3  
65 121.8 98.6 148.0



#56  
Dibenzofuran  
Concen: 17.299 ng/ul  
RT: 15.140 min Scan# 2015  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:168 Resp: 1749262  
Ion Ratio Lower Upper  
168 100  
139 38.2 30.0 45.0









#57  
2,4-Dinitrotoluene  
Concen: 21.050 ng/ul  
RT: 15.110 min Scan# 2070  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610

Tgt Ion:165 Resp: 371205  
Ion Ratio Lower Upper  
165 100  
63 44.9 37.4 56.2  
182 2.6 2.2 3.4




#58  
2,3,4,6-Tetrachlorophenol  
Concen: 19.821 ng/ul  
RT: 15.345 min Scan# 2050  
Delta R.T. -0.009 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:232 Resp: 342560  
Ion Ratio Lower Upper  
232 100  
131 47.4 35.1 52.7  
130 2.7 2.1 3.1  
166 31.1 24.2 36.2











#59  
Diethylphthalate  
Concen: 17.072 ng/ul  
RT: 15.557 min Scan# 2086  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:149 Resp: 1475889  
Ion Ratio Lower Upper  
149 100  
177 21.9 17.9 26.9  
150 12.2 9.9 14.9



#61  
Fluorene  
Concen: 17.223 ng/ul  
RT: 15.793 min Scan# 2126  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:166 Resp: 1414641  
Ion Ratio Lower Upper  
166 100  
165 98.4 77.0 115.4  
167 13.7 10.6 15.8









#62  
4-Chlorophenyl-phenylether  
Concen: 17.098 ng/ul  
RT: 15.787 min Scan# 2185  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:204 Resp: 684657  
Ion Ratio Lower Upper  
204 100  
206 32.8 26.5 39.7  
141 57.3 46.2 69.4



#63  
4-Nitroaniline  
Concen: 23.783 ng/ul  
RT: 15.834 min Scan# 2133  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:138 Resp: 315572  
Ion Ratio Lower Upper  
138 100  
92 52.3 45.3 67.9  
108 90.2 76.4 114.6









#66  
4,6-Dinitro-2-methylphenol  
Concen: 15.521 ng/ul  
RT: 15.863 min Scan# 2198  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:198 Resp: 158362  
Ion Ratio Lower Upper  
198 100  
182 4.4 3.8 5.8  
77 26.3 21.6 32.4



#67  
N-Nitrosodiphenylamine  
Concen: 17.205 ng/ul  
RT: 16.004 min Scan# 2162  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:169 Resp: 1249918  
Ion Ratio Lower Upper  
169 100  
168 66.2 54.0 81.0  
167 36.3 29.4 44.2









#68  
4-Bromophenyl-phenylether  
Concen: 16.675 ng/ul  
RT: 16.687 min Scan# 21  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:248 Resp: 409756  
Ion Ratio Lower Upper  
248 100  
250 98.8 78.1 117.1  
141 75.9 61.2 91.8



#69  
Hexachlorobenzene  
Concen: 16.146 ng/ul  
RT: 16.763 min Scan# 2291  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:284 Resp: 475772  
Ion Ratio Lower Upper  
284 100  
142 38.3 29.7 44.5  
249 31.3 24.8 37.2





#70

Atrazine

Concen: 17.314 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:200 Resp: 458930

Ion Ratio Lower Upper

200 100

173 26.1 21.7 32.5

215 46.9 38.5 57.7



#71

Pentachlorophenol

Concen: 15.340 ng/ul

RT: 17.122 min Scan# 2352

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:266 Resp: 254722


Ion Ratio Lower Upper

266 100

264 65.1 50.9 76.3

268 64.0 51.5 77.3





#72

Phenanthrene

Concen: 16.658 ng/ul

RT: 17.551 min Scan# 2485

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:178 Resp: 2245961

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.4 15.4 23.0



#74

Anthracene

Concen: 16.484 ng/ul

RT: 17.639 min Scan# 2440

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:178 Resp: 2263912


Ion Ratio Lower Upper

178 100

179 15.7 12.3 18.5

176 18.9 15.1 22.7









#75  
1,2,3,4-Tetrachlorobenzene  
Concen: 15.831 ng/uL  
RT: 13.534 min Scan# 11  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:216 Resp: 592081  
Ion Ratio Lower Upper  
216 100  
214 78.8 62.8 94.2  
218 49.1 38.6 58.0



#76  
Pentachlorobenzene  
Concen: 16.806 ng/uL  
RT: 15.034 min Scan# 1997  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:250 Resp: 555459  
Ion Ratio Lower Upper  
250 100  
248 62.4 50.1 75.1  
252 64.5 51.0 76.6





#77

Carbazole

Concen: 17.936 ng/ul

RT: 17.910 min Scan# 24

Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:167 Resp: 2120298

Ion Ratio Lower Upper

167 100

166 21.6 17.3 25.9

139 12.9 10.4 15.6



#78

Di-n-butylphthalate

Concen: 17.285 ng/ul

RT: 18.439 min Scan# 2576

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:149 Resp: 2602282

Ion Ratio Lower Upper


149 100

150 9.1 7.5 11.3

104 5.2 4.3 6.5











#80  
Fluoranthene  
Concen: 15.711 ng/ul  
RT: 19.498 min Scan# 21  
Delta R.T. -0.009 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:202 Resp: 2584748  
Ion Ratio Lower Upper  
202 100  
101 12.3 9.1 13.7  
100 9.6 7.1 10.7



#82  
Pyrene  
Concen: 15.989 ng/ul  
RT: 19.857 min Scan# 2817  
Delta R.T. -0.003 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:202 Resp: 2717641  
Ion Ratio Lower Upper  
202 100  
101 13.5 11.2 16.8  
100 11.3 9.4 14.0





#83

Butylbenzylphthalate

Concen: 17.144 ng/ul

RT: 20.728 min Scan# 2965

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:149 Resp: 1103170

Ion Ratio Lower Upper

149 100

91 72.2 58.5 87.7

206 20.6 17.0 25.4



#84

3,3'-Dichlorobenzidine

Concen: 19.577 ng/ul

RT: 21.510 min Scan# 3098

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:252 Resp: 1040049


Ion Ratio Lower Upper

252 100

254 65.0 51.4 77.0

126 12.5 10.3 15.5





#85

Benzo(a)anthracene

Concen: 16.403 ng/ul

RT: 21.575 min Scan# 3170

Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:228 Resp: 2685524

Ion Ratio Lower Upper

228 100

229 20.0 15.9 23.9

226 26.9 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.244 ng/ul

RT: 21.480 min Scan# 3093

Delta R.T. -0.003 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:149 Resp: 1692589


Ion Ratio Lower Upper

149 100

167 27.3 21.7 32.5

279 4.3 3.0 4.4





#87

Chrysene

Concen: 16.195 ng/ul

RT: 21.633 min Scan# 31

Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:228 Resp: 2455275

Ion Ratio Lower Upper

228 100

226 30.5 24.5 36.7

229 20.0 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 16.990 ng/ul




RT: 22.480 min Scan# 3263


Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Tgt Ion:149 Resp: 2831728









#90  
Benzo(b)fluoranthene  
Concen: 16.585 ng/ul  
RT: 23.386 min Scan# 3478  
Delta R.T. -0.009 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33




Instrument :  
BNA\_P  
ClientSampleId :  
SICV610


Tgt Ion:252 Resp: 2633465  
Ion Ratio Lower Upper  
252 100  
253 22.1 17.9 26.9  
125 10.6 8.8 13.2



#91  
Benzo(k)fluoranthene  
Concen: 15.653 ng/ul  
RT: 23.439 min Scan# 3426  
Delta R.T. -0.009 min  
Lab File: BP016349.D  
Acq: 25 Jul 2023 16:33

Tgt Ion:252 Resp: 2473492  
Ion Ratio Lower Upper  
252 100  
253 21.5 17.7 26.5  
125 11.0 9.0 13.6





#93

Benzo(a)pyrene

Concen: 15.373 ng/ul

RT: 24.080 min Scan# 31

Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33

Instrument :

BNA\_P

ClientSampleId :

SICV610





Tgt Ion:252 Resp: 2298905

Ion Ratio Lower Upper

252 100

253 21.7 18.0 27.0

125 11.4 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 15.861 ng/ul

RT: 26.980 min Scan# 4028

Delta R.T. -0.009 min

Lab File: BP016349.D

Acq: 25 Jul 2023 16:33




Tgt Ion:276 Resp: 2738882


Ion Ratio Lower Upper

276 100

138 23.0 18.1 27.1

277 24.1 19.0 28.4









#95  
 Dibenzo(a,h)anthracene  
 Concen: 16.074 ng/ul  
 RT: 27.015 min Scan# 4095  
 Delta R.T. -0.009 min  
 Lab File: BP016349.D  
 Acq: 25 Jul 2023 16:33

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SICV610

Tgt Ion:278 Resp: 2295738  
 Ion Ratio Lower Upper  
 278 100  
 139 16.9 14.2 21.2  
 279 23.9 19.4 29.2



#96  
 Benzo(g,h,i)perylene  
 Concen: 15.389 ng/ul  
 RT: 27.839 min Scan# 4174  
 Delta R.T. -0.021 min  
 Lab File: BP016349.D  
 Acq: 25 Jul 2023 16:33

Tgt Ion:276 Resp: 2103818  
 Ion Ratio Lower Upper  
 276 100  
 138 21.9 17.8 26.6  
 277 24.2 19.0 28.4

