

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072523\
 Data File : BP016352.D
 Acq On : 25 Jul 2023 18:49
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020671

Quant Time: Jul 26 00:29:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 25 23:56:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.098	152	538586	20.000	ng/u1	0.00
20) Naphthalene-d8	10.934	136	2304811	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.739	164	1414049	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.504	188	3033460	20.000	ng/u1	0.00
79) Chrysene-d12	21.598	240	2549971	20.000	ng/u1	0.00
88) Perylene-d12	24.204	264	2481078	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	110397	8.082	ng/uL	0.00
4) Pyridine-d5	3.858	84	770626	19.885	ng/u1	0.00
7) Phenol-d5	7.228	99	891774	19.786	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	545867	20.173	ng/u1	0.00
11) 2-Chlorophenol-d4	7.616	132	703685	20.013	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	723718	19.967	ng/u1	0.00
21) Nitrobenzene-d5	9.293	128	325248	22.917	ng/u1	0.00
24) 2-Nitrophenol-d4	10.010	143	286583	24.484	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	677839	20.892	ng/u1	0.00
31) 4-Chloroaniline-d4	11.081	131	1029593	20.398	ng/u1	0.00
46) Dimethylphthalate-d6	14.151	166	2070403	20.006	ng/u1	0.00
49) Acenaphthylene-d8	14.439	160	2449617	20.035	ng/u1	0.00
54) 4-Nitrophenol-d4	14.922	143	357284	20.604	ng/u1	0.00
60) Fluorene-d10	15.733	176	1736899	19.728	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.845	200	190189	19.035	ng/u1	0.00
73) Anthracene-d10	17.604	188	2642717	19.667	ng/u1	0.00
81) Pyrene-d10	19.833	212	2926715	20.214	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.033	264	2351314	19.643	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.469	88	115978	8.007	ng/uL	97
5) Pyridine	3.875	79	792291	20.094	ng/u1	99
6) Benzaldehyde	7.246	77	566128	23.024	ng/u1	99
8) Phenol	7.251	94	934093	19.930	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.528	93	764854	19.920	ng/u1	100
12) 2-Chlorophenol	7.646	128	725687	20.073	ng/u1	100
13) 2-Methylphenol	8.528	108	711555	19.799	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.616	45	498126	10.722	ng/u1	99
16) Acetophenone	8.946	105	1181726	20.659	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.916	70	601837	20.205	ng/u1	98
18) 4-Methylphenol	8.857	108	762330	19.853	ng/u1	98
19) Hexachloroethane	9.181	117	302650	20.010	ng/u1	97
22) Nitrobenzene	9.334	77	828823	21.814	ng/u1	99
23) Isophorone	9.851	82	1725666	20.164	ng/u1	99
25) 2-Nitrophenol	10.040	139	349426	23.924	ng/u1	98
26) 2,4-Dimethylphenol	10.075	107	841291	20.372	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.340	93	1034624	20.051	ng/u1	100
29) 2,4-Dichlorophenol	10.557	162	677495	20.656	ng/u1	99
30) Naphthalene	10.987	128	2374592	19.892	ng/u1	100
32) 4-Chloroaniline	11.104	127	1003509	20.280	ng/u1	99
33) Hexachlorobutadiene	11.222	225	426780	19.798	ng/u1	99
34) Caprolactam	11.892	113	155674	13.324	ng/u1	98
35) 4-Chloro-3-methylphenol	12.192	107	751062	20.687	ng/u1	99
36) 2-Methylnaphthalene	12.581	142	1619364	19.905	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072523\
 Data File : BP016352.D
 Acq On : 25 Jul 2023 18:49
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020671


Quant Time: Jul 26 00:29:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 25 23:56:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.798	142	1648869	20.272	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.922	216	828505	19.832	ng/ul	99
40) Hexachlorocyclopentadiene	12.881	237	499208	16.739	ng/ul	97
41) 2,4,6-Trichlorophenol	13.169	196	519069	21.089	ng/ul	98
42) 2,4,5-Trichlorophenol	13.234	196	564581	21.433	ng/ul	99
43) 1,1'-Biphenyl	13.581	154	2099461	19.663	ng/ul	99
44) 2-Chloronaphthalene	13.628	162	1644459	19.688	ng/ul	99
45) 2-Nitroaniline	13.845	65	426163	24.384	ng/ul	96
47) Dimethylphthalate	14.198	163	2090192	20.028	ng/ul	99
48) 2,6-Dinitrotoluene	14.333	165	367955	24.466	ng/ul	95
50) Acenaphthylene	14.469	152	2764876	19.850	ng/ul	100
51) 3-Nitroaniline	14.669	138	389428	22.133	ng/ul#	98
52) Acenaphthene	14.804	153	1763717	19.637	ng/ul	100
53) 2,4-Dinitrophenol	14.863	184	103429	16.080	ng/ul	95
55) 4-Nitrophenol	14.939	109	289508	20.712	ng/ul	99
56) Dibenzofuran	15.139	168	2413212	19.627	ng/ul	100
57) 2,4-Dinitrotoluene	15.110	165	522802	24.382	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.345	232	450579	21.442	ng/ul	95
59) Diethylphthalate	15.557	149	2101284	19.990	ng/ul	99
61) Fluorene	15.792	166	1968501	19.710	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.786	204	974416	20.014	ng/ul	99
63) 4-Nitroaniline	15.833	138	365029	22.625	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.863	198	239842	20.079	ng/ul	98
67) N-Nitrosodiphenylamine	16.004	169	1685878	19.821	ng/ul	99
68) 4-Bromophenyl-phenylether	16.686	248	574866	19.983	ng/ul	98
69) Hexachlorobenzene	16.763	284	681084	19.743	ng/ul	100
70) Atrazine	16.957	200	625230	20.148	ng/ul	100
71) Pentachlorophenol	17.122	266	382904	19.697	ng/ul	99
72) Phenanthrene	17.551	178	3073010	19.468	ng/ul	100
74) Anthracene	17.639	178	3151681	19.601	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.534	216	869538	19.858	ng/ul	99
76) Pentachlorobenzene	15.033	250	769040	19.875	ng/ul	99
77) Carbazole	17.916	167	2811833	20.316	ng/ul	100
78) Di-n-butylphthalate	18.439	149	3550867	20.146	ng/ul	100
80) Fluoranthene	19.504	202	3469393	19.989	ng/ul	99
82) Pyrene	19.857	202	3593545	20.041	ng/ul	100
83) Butylbenzylphthalate	20.727	149	1436872	21.166	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.515	252	1038787	18.535	ng/ul	99
85) Benzo(a)anthracene	21.580	228	3361031	19.460	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.480	149	2104249	20.321	ng/ul	100
87) Chrysene	21.639	228	3118405	19.497	ng/ul	100
89) Di-n-octyl phthalate	22.486	149	3311593	21.234	ng/ul	100
90) Benzo(b)fluoranthene	23.392	252	3012355	20.274	ng/ul	100
91) Benzo(k)fluoranthene	23.445	252	2971233	20.094	ng/ul	99
93) Benzo(a)pyrene	24.086	252	2765759	19.765	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.986	276	2807386	17.374	ng/ul	98
95) Dibenzo(a,h)anthracene	27.021	278	2371929	17.748	ng/ul	99
96) Benzo(g,h,i)perylene	27.845	276	2148164	16.792	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 26 00:29:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP072523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 25 23:56:16 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.007 ng/uL
RT: 3.469 min Scan# 31
Delta R.T. 0.002 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion: 88 Resp: 115978
Ion Ratio Lower Upper
88 100
43 25.4 21.8 32.8
58 68.4 53.1 79.7



#5
Pyridine
Concen: 20.094 ng/uL
RT: 3.875 min Scan# 100
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion: 79 Resp: 792291
Ion Ratio Lower Upper
79 100
52 65.0 51.0 76.4
51 31.6 25.4 38.2









#6
Benzaldehyde
Concen: 23.024 ng/ul
RT: 7.246 min Scan# 61
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion: 77 Resp: 566128
Ion Ratio Lower Upper
77 100
105 97.8 80.2 120.2
106 95.2 76.5 114.7



#8
Phenol
Concen: 19.930 ng/ul
RT: 7.251 min Scan# 674
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion: 94 Resp: 934093
Ion Ratio Lower Upper
94 100
65 26.9 21.4 32.2
66 34.3 28.2 42.4









#10
Bis(2-Chloroethyl)ether
Concen: 19.920 ng/ul
RT: 7.528 min Scan# 71
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion: 93 Resp: 764854
Ion Ratio Lower Upper
93 100
63 70.9 57.0 85.4
95 32.2 25.8 38.6



#12
2-Chlorophenol
Concen: 20.073 ng/ul
RT: 7.646 min Scan# 741
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:128 Resp: 725687
Ion Ratio Lower Upper
128 100
64 45.5 36.4 54.6
130 32.6 25.6 38.4





#13

2-Methylphenol

Concen: 19.799 ng/ul

RT: 8.528 min Scan# 891

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :





SSTD020671

Tgt Ion:108 Resp: 711555

Ion Ratio Lower Upper

108 100

107 89.7 71.3 106.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 10.722 ng/ul

RT: 8.616 min Scan# 906

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion: 45 Resp: 498126


Ion Ratio Lower Upper

45 100

77 16.4 13.7 20.5

79 12.3 9.6 14.4









#16
Acetophenone
Concen: 20.659 ng/ul
RT: 8.946 min Scan# 910
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:105 Resp: 1181726
Ion Ratio Lower Upper
105 100
77 76.4 61.2 91.8
51 24.3 19.1 28.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.205 ng/ul
RT: 8.916 min Scan# 957
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion: 70 Resp: 601837
Ion Ratio Lower Upper
70 100
42 52.0 40.3 60.5
101 9.4 7.8 11.8
130 21.1 17.1 25.7









#18
4-Methylphenol
Concen: 19.853 ng/ul
RT: 8.857 min Scan# 947
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:108 Resp: 762330
Ion Ratio Lower Upper
108 100
107 111.9 91.6 137.4



#19
Hexachloroethane
Concen: 20.010 ng/ul
RT: 9.181 min Scan# 1002
Delta R.T. 0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:117 Resp: 302650
Ion Ratio Lower Upper
117 100
201 95.8 75.0 112.6
199 61.6 47.4 71.0









#22
Nitrobenzene
Concen: 21.814 ng/ul
RT: 9.334 min Scan# 1088
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion: 77 Resp: 828823
Ion Ratio Lower Upper
77 100
123 45.1 35.3 52.9
65 13.0 10.2 15.2



#23
Isophorone
Concen: 20.164 ng/ul
RT: 9.851 min Scan# 1116
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion: 82 Resp: 1725666
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 18.0 14.0 21.0









#25
2-Nitrophenol
Concen: 23.924 ng/ul
RT: 10.040 min Scan# 1148
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:139 Resp: 349426
Ion Ratio Lower Upper
139 100
65 50.5 38.7 58.1
109 30.8 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 20.372 ng/ul
RT: 10.075 min Scan# 1154
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:107 Resp: 841291
Ion Ratio Lower Upper
107 100
121 50.7 39.8 59.6
122 84.9 69.0 103.4









#27
Bis(2-Chloroethoxy)methane
Concen: 20.051 ng/ul
RT: 10.340 min Scan# 1199
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion: 93 Resp: 1034624
Ion Ratio Lower Upper
93 100
95 32.3 25.9 38.9
123 11.0 8.8 13.2



#29
2,4-Dichlorophenol
Concen: 20.656 ng/ul
RT: 10.557 min Scan# 1236
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:162 Resp: 677495
Ion Ratio Lower Upper
162 100
164 64.8 51.4 77.0
98 38.1 30.6 46.0









#30
Naphthalene
Concen: 19.892 ng/ul
RT: 10.987 min Scan# 1309
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:128 Resp: 2374592
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 20.280 ng/ul
RT: 11.104 min Scan# 1329
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:127 Resp: 1003509
Ion Ratio Lower Upper
127 100
129 32.2 25.2 37.8





#33

Hexachlorobutadiene

Concen: 19.798 ng/ul

RT: 11.222 min Scan# 1409

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:225 Resp: 426780

Ion Ratio Lower Upper

225 100

223 63.8 50.1 75.1

227 63.8 51.5 77.3



#34

Caprolactam

Concen: 13.324 ng/ul

RT: 11.892 min Scan# 1463

Delta R.T. 0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:113 Resp: 155674


Ion Ratio Lower Upper

113 100

55 155.3 122.2 183.4

56 115.1 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 20.687 ng/ul
RT: 12.192 min Scan# 1573
Delta R.T. 0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:107 Resp: 751062
Ion Ratio Lower Upper
107 100
144 25.8 20.2 30.2
142 79.3 62.6 93.8



#36
2-Methylnaphthalene
Concen: 19.905 ng/ul
RT: 12.581 min Scan# 1580
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:142 Resp: 1619364
Ion Ratio Lower Upper
142 100
141 90.0 71.9 107.9









#37
1-Methylnaphthalene
Concen: 20.272 ng/ul
RT: 12.798 min Scan# 1617
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:142 Resp: 1648869
Ion Ratio Lower Upper
142 100
141 92.8 73.2 109.8
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.832 ng/ul
RT: 12.922 min Scan# 1638
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:216 Resp: 828505
Ion Ratio Lower Upper
216 100
214 78.9 63.6 95.4
179 21.7 17.4 26.0
108 18.5 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 16.739 ng/ul

RT: 12.881 min Scan# 1

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:237 Resp: 499208

Ion Ratio Lower Upper

237 100

235 61.1 51.2 76.8

272 13.0 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 21.089 ng/ul

RT: 13.169 min Scan# 1680

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:196 Resp: 519069


Ion Ratio Lower Upper

196 100

198 95.3 77.7 116.5

200 29.9 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 21.433 ng/ul

RT: 13.234 min Scan# 1

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:196 Resp: 564581

Ion Ratio Lower Upper

196 100

198 96.7 77.0 115.4

200 30.1 24.8 37.2



#43

1,1'-Biphenyl

Concen: 19.663 ng/ul

RT: 13.581 min Scan# 1750

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:154 Resp: 2099461


Ion Ratio Lower Upper

154 100

153 39.7 32.3 48.5

76 12.8 10.7 16.1





#44

2-Chloronaphthalene

Concen: 19.688 ng/ul

RT: 13.628 min Scan# 1758

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:162 Resp: 1644459

Ion Ratio Lower Upper

162 100

164 32.6 26.6 40.0

127 39.3 31.0 46.4



#45

2-Nitroaniline

Concen: 24.384 ng/ul

RT: 13.845 min Scan# 1795

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion: 65 Resp: 426163


Ion Ratio Lower Upper

65 100

92 73.5 56.8 85.2

138 109.8 83.9 125.9









#47
Dimethylphthalate
Concen: 20.028 ng/ul
RT: 14.198 min Scan# 1875
Delta R.T. -0.009 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:163 Resp: 2090192
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.5 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 24.466 ng/ul
RT: 14.333 min Scan# 1878
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:165 Resp: 367955
Ion Ratio Lower Upper
165 100
89 63.1 53.8 80.6
121 22.6 20.0 30.0









#50
Acenaphthylene
Concen: 19.850 ng/ul
RT: 14.469 min Scan# 1901
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:152 Resp: 2764876
Ion Ratio Lower Upper
152 100
151 19.9 16.0 24.0
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 22.133 ng/ul
RT: 14.669 min Scan# 1935
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:138 Resp: 389428
Ion Ratio Lower Upper
138 100
108 14.7 16.6 24.8#
92 117.5 94.3 141.5





#52

Acenaphthene

Concen: 19.637 ng/ul

RT: 14.804 min Scan# 1968

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:153 Resp: 1763717

Ion Ratio Lower Upper

153 100

152 48.3 38.8 58.2

154 89.5 71.7 107.5



#53

2,4-Dinitrophenol

Concen: 16.080 ng/ul

RT: 14.863 min Scan# 1968

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:184 Resp: 103429


Ion Ratio Lower Upper

184 100

63 71.5 60.0 90.0

154 64.2 55.2 82.8









#55
4-Nitrophenol
Concen: 20.712 ng/ul
RT: 14.939 min Scan# 1981
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:109 Resp: 289508
Ion Ratio Lower Upper
109 100
139 129.5 103.5 155.3
65 126.2 98.6 148.0



#56
Dibenzofuran
Concen: 19.627 ng/ul
RT: 15.139 min Scan# 2015
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:168 Resp: 2413212
Ion Ratio Lower Upper
168 100
139 37.4 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 24.382 ng/ul
RT: 15.110 min Scan# 2070
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:165 Resp: 522802
Ion Ratio Lower Upper
165 100
63 45.4 37.4 56.2
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.442 ng/ul
RT: 15.345 min Scan# 2050
Delta R.T. -0.009 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:232 Resp: 450579
Ion Ratio Lower Upper
232 100
131 47.6 35.1 52.7
130 2.6 2.1 3.1
166 32.5 24.2 36.2





#59

Diethylphthalate

Concen: 19.990 ng/ul

RT: 15.557 min Scan# 2101

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:149 Resp: 2101284

Ion Ratio Lower Upper

149 100

177 21.9 17.9 26.9

150 12.5 9.9 14.9



#61

Fluorene

Concen: 19.710 ng/ul

RT: 15.792 min Scan# 2126

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:166 Resp: 1968501


Ion Ratio Lower Upper

166 100

165 98.2 77.0 115.4

167 13.3 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 20.014 ng/ul
RT: 15.786 min Scan# 2185
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:204 Resp: 974416
Ion Ratio Lower Upper
204 100
206 33.2 26.5 39.7
141 56.6 46.2 69.4



#63
4-Nitroaniline
Concen: 22.625 ng/ul
RT: 15.833 min Scan# 2133
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:138 Resp: 365029
Ion Ratio Lower Upper
138 100
92 52.3 45.3 67.9
108 86.8 76.4 114.6









#66
4,6-Dinitro-2-methylphenol
Concen: 20.079 ng/ul
RT: 15.863 min Scan# 2198
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:198 Resp: 239842
Ion Ratio Lower Upper
198 100
182 3.9 3.8 5.8
77 25.8 21.6 32.4



#67
N-Nitrosodiphenylamine
Concen: 19.821 ng/ul
RT: 16.004 min Scan# 2162
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:169 Resp: 1685878
Ion Ratio Lower Upper
169 100
168 66.7 54.0 81.0
167 36.4 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.983 ng/ul
RT: 16.686 min Scan# 2290
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:248 Resp: 574866
Ion Ratio Lower Upper
248 100
250 98.2 78.1 117.1
141 74.0 61.2 91.8



#69
Hexachlorobenzene
Concen: 19.743 ng/ul
RT: 16.763 min Scan# 2291
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:284 Resp: 681084
Ion Ratio Lower Upper
284 100
142 37.5 29.7 44.5
249 31.0 24.8 37.2





#70

Atrazine

Concen: 20.148 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:200 Resp: 625230

Ion Ratio Lower Upper

200 100

173 27.1 21.7 32.5

215 47.6 38.5 57.7



#71

Pentachlorophenol

Concen: 19.697 ng/ul

RT: 17.122 min Scan# 2352

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:266 Resp: 382904


Ion Ratio Lower Upper

266 100

264 63.3 50.9 76.3

268 63.1 51.5 77.3





#72

Phenanthrene

Concen: 19.468 ng/ul

RT: 17.551 min Scan# 2485

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:178 Resp: 3073010

Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.5 15.4 23.0



#74

Anthracene

Concen: 19.601 ng/ul

RT: 17.639 min Scan# 2440

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:178 Resp: 3151681


Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 18.7 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.858 ng/uL
RT: 13.534 min Scan# 11
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:216 Resp: 869538
Ion Ratio Lower Upper
216 100
214 77.7 62.8 94.2
218 48.4 38.6 58.0



#76
Pentachlorobenzene
Concen: 19.875 ng/uL
RT: 15.033 min Scan# 1997
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:250 Resp: 769040
Ion Ratio Lower Upper
250 100
248 62.4 50.1 75.1
252 65.0 51.0 76.6









#77
 Carbazole
 Concen: 20.316 ng/ul
 RT: 17.916 min Scan# 2487
 Delta R.T. -0.003 min
 Lab File: BP016352.D
 Acq: 25 Jul 2023 18:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020671


Tgt Ion:167 Resp: 2811833
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.3 25.9
 139 13.1 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 20.146 ng/ul
 RT: 18.439 min Scan# 2576
 Delta R.T. -0.003 min
 Lab File: BP016352.D
 Acq: 25 Jul 2023 18:49

Tgt Ion:149 Resp: 3550867
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.5 11.3
 104 5.4 4.3 6.5









#80
Fluoranthene
Concen: 19.989 ng/ul
RT: 19.504 min Scan# 21
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:202 Resp: 3469393
Ion Ratio Lower Upper
202 100
101 11.7 9.1 13.7
100 9.0 7.1 10.7



#82
Pyrene
Concen: 20.041 ng/ul
RT: 19.857 min Scan# 2817
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:202 Resp: 3593545
Ion Ratio Lower Upper
202 100
101 14.1 11.2 16.8
100 11.4 9.4 14.0





#83

Butylbenzylphthalate

Concen: 21.166 ng/ul

RT: 20.727 min Scan# 2099

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:149 Resp: 1436872

Ion Ratio Lower Upper

149 100

91 73.6 58.5 87.7

206 20.3 17.0 25.4



#84

3,3'-Dichlorobenzidine

Concen: 18.535 ng/ul

RT: 21.515 min Scan# 3099

Delta R.T. 0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:252 Resp: 1038787


Ion Ratio Lower Upper

252 100

254 64.2 51.4 77.0

126 11.2 10.3 15.5





#85

Benzo(a)anthracene

Concen: 19.460 ng/ul

RT: 21.580 min Scan# 3170

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:228 Resp: 3361031

Ion Ratio Lower Upper

228 100

229 19.7 15.9 23.9

226 27.4 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.321 ng/ul

RT: 21.480 min Scan# 3093

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:149 Resp: 2104249


Ion Ratio Lower Upper

149 100

167 27.0 21.7 32.5

279 4.0 3.0 4.4





#87

Chrysene

Concen: 19.497 ng/ul

RT: 21.639 min Scan# 31

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:228 Resp: 3118405

Ion Ratio Lower Upper

228 100

226 30.7 24.5 36.7

229 19.9 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 21.234 ng/ul




RT: 22.486 min Scan# 3264


Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Tgt Ion:149 Resp: 3311593









#90
Benzo(b)fluoranthene
Concen: 20.274 ng/ul
RT: 23.392 min Scan# 3478
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49




Instrument :
BNA_P
ClientSampleId :
SSTD020671


Tgt Ion:252 Resp: 3012355
Ion Ratio Lower Upper
252 100
253 22.2 17.9 26.9
125 10.9 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 20.094 ng/ul
RT: 23.445 min Scan# 3427
Delta R.T. -0.003 min
Lab File: BP016352.D
Acq: 25 Jul 2023 18:49

Tgt Ion:252 Resp: 2971233
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 10.5 9.0 13.6





#93

Benzo(a)pyrene

Concen: 19.765 ng/ul

RT: 24.086 min Scan# 3

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:252 Resp: 2765759

Ion Ratio Lower Upper

252 100

253 22.0 18.0 27.0

125 11.8 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.374 ng/ul

RT: 26.986 min Scan# 4029

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49




Tgt Ion:276 Resp: 2807386


Ion Ratio Lower Upper

276 100

138 23.1 18.1 27.1

277 24.7 19.0 28.4





#95

Dibenzo(a,h)anthracene

Concen: 17.748 ng/ul

RT: 27.021 min Scan# 4095

Delta R.T. -0.003 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Instrument :

BNA_P

ClientSampleId :

SSTD020671





Tgt Ion:278 Resp: 2371929

Ion Ratio Lower Upper

278 100

139 16.8 14.2 21.2

279 23.9 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 16.792 ng/ul

RT: 27.845 min Scan# 4175

Delta R.T. -0.015 min

Lab File: BP016352.D

Acq: 25 Jul 2023 18:49

Tgt Ion:276 Resp: 2148164

Ion Ratio Lower Upper

276 100

138 22.8 17.8 26.6

277 24.2 19.0 28.4

