

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072823\
 Data File : BP016411.D
 Acq On : 28 Jul 2023 08:03
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 28 08:40:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 01:48:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	42.391	-6.0	124	0.00
3	Pyridine	40.000	42.905	-7.3	122	0.00
4	n-Nitrosodimethylamine	40.000	43.132	-7.8	125	0.00
5 S	2-Fluorophenol	80.000	84.777	-6.0	119	0.00
6	Aniline	40.000	42.695	-6.7	119	0.00
7 S	Phenol-d6	80.000	85.006	-6.3	119	0.00
8	2-Chlorophenol	40.000	42.006	-5.0	116	0.00
9	Benzaldehyde	40.000	48.142	-20.4	125	0.00
10 C	Phenol	40.000	42.571	-6.4	119	0.00
11	bis(2-Chloroethyl)ether	40.000	42.180	-5.4	119	0.00
12	1,3-Dichlorobenzene	40.000	40.717	-1.8	115	0.00
13 C	1,4-Dichlorobenzene	40.000	40.746	-1.9	115	0.00
14	1,2-Dichlorobenzene	40.000	40.663	-1.7	114	0.00
15	Benzyl Alcohol	40.000	43.150	-7.9	119	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.692	-9.2	125	0.00
17	2-Methylphenol	40.000	42.317	-5.8	117	0.00
18	Hexachloroethane	40.000	41.599	-4.0	118	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.496	-6.2	119	0.00
20	3+4-Methylphenols	40.000	42.344	-5.9	116	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	41.032	-2.6	117	0.00
23 S	Nitrobenzene-d5	80.000	84.257	-5.3	119	0.00
24	Nitrobenzene	40.000	42.292	-5.7	120	0.00
25	Isophorone	40.000	41.602	-4.0	117	0.00
26 C	2-Nitrophenol	40.000	39.379	1.6	115	0.00
27	2,4-Dimethylphenol	40.000	40.851	-2.1	115	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.688	-4.2	118	0.00
29 C	2,4-Dichlorophenol	40.000	40.835	-2.1	111	0.00
30	1,2,4-Trichlorobenzene	40.000	39.173	2.1	108	0.00
31	Naphthalene	40.000	40.344	-0.9	114	0.00
32	Benzoic acid	40.000	39.018	2.5	116	0.01
33	4-Chloroaniline	40.000	40.668	-1.7	114	0.00
34 C	Hexachlorobutadiene	40.000	37.634	5.9	102	0.00
35	Caprolactam	40.000	40.432	-1.1	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.211	-3.0	114	0.00
37	2-Methylnaphthalene	40.000	40.110	-0.3	112	0.00
38	1-Methylnaphthalene	40.000	40.146	-0.4	112	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.126	2.2	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.485	-1.2	101	0.00
42 S	2,4,6-Tribromophenol	80.000	70.473	11.9	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.224	-3.1	104	0.00
44	2,4,5-Trichlorophenol	40.000	40.954	-2.4	105	0.00
45 S	2-Fluorobiphenyl	80.000	79.527	0.6	106	0.00
46	1,1'-Biphenyl	40.000	40.688	-1.7	109	0.00
47	2-Chloronaphthalene	40.000	40.730	-1.8	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072823\
 Data File : BP016411.D
 Acq On : 28 Jul 2023 08:03
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 28 08:40:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 01:48:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.162	-12.9	118	0.00
49	Acenaphthylene	40.000	41.045	-2.6	110	0.00
50	Dimethylphthalate	40.000	40.098	-0.2	108	0.00
51	2,6-Dinitrotoluene	40.000	41.955	-4.9	108	0.00
52 C	Acenaphthene	40.000	40.848	-2.1	109	0.00
53	3-Nitroaniline	40.000	42.831	-7.1	112	0.00
54 P	2,4-Dinitrophenol	40.000	37.884	5.3	110	0.00
55	Dibenzofuran	40.000	40.100	-0.3	108	0.00
56 P	4-Nitrophenol	40.000	42.667	-6.7	110	0.00
57	2,4-Dinitrotoluene	40.000	43.116	-7.8	108	0.00
58	Fluorene	40.000	39.743	0.6	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.091	2.3	99	0.00
60	Diethylphthalate	40.000	40.448	-1.1	109	0.00
61	4-Chlorophenyl-phenylether	40.000	38.721	3.2	101	0.00
62	4-Nitroaniline	40.000	42.293	-5.7	109	0.00
63	Azobenzene	40.000	42.108	-5.3	116	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.612	-1.5	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.810	-4.5	107	0.00
67	4-Bromophenyl-phenylether	40.000	39.912	0.2	97	0.00
68	Hexachlorobenzene	40.000	38.299	4.3	93	0.00
69	Atrazine	40.000	40.797	-2.0	102	0.00
70 C	Pentachlorophenol	40.000	38.880	2.8	96	0.00
71	Phenanthrene	40.000	40.516	-1.3	104	0.00
72	Anthracene	40.000	41.041	-2.6	104	0.00
73	Carbazole	40.000	40.993	-2.5	105	0.00
74	Di-n-butylphthalate	40.000	42.757	-6.9	107	0.00
75 C	Fluoranthene	40.000	39.550	1.1	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	91	0.00
77	Benzydine	40.000	53.696	-34.2#	111	0.00
78	Pyrene	40.000	44.366	-10.9	99	0.00
79 S	Terphenyl-d14	80.000	85.607	-7.0	93	0.00
80	Butylbenzylphthalate	40.000	46.991	-17.5	104	0.00
81	Benzo(a)anthracene	40.000	41.564	-3.9	93	0.00
82	3,3'-Dichlorobenzidine	40.000	41.576	-3.9	89	0.00
83	Chrysene	40.000	40.963	-2.4	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.581	-14.0	101	0.00
85 c	Di-n-octyl phthalate	40.000	43.386	-8.5	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	79	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.703	8.2	73	0.00
88	Benzo(b)fluoranthene	40.000	44.260	-10.6	84	0.00
89	Benzo(k)fluoranthene	40.000	43.400	-8.5	83	0.00
90 C	Benzo(a)pyrene	40.000	42.294	-5.7	81	0.00
91	Dibenzo(a,h)anthracene	40.000	37.289	6.8	73	0.00
92	Benzo(g,h,i)perylene	40.000	36.183	9.5	72	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072823\
Data File : BP016411.D
Acq On : 28 Jul 2023 08:03
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jul 28 08:40:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 01:48:25 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0