

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072823\
 Data File : BP016423.D
 Acq On : 28 Jul 2023 15:05
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 00:45:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 01:48:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	40.905	-2.3	96	0.00
3	Pyridine	40.000	42.367	-5.9	97	0.00
4	n-Nitrosodimethylamine	40.000	42.914	-7.3	99	0.00
5 S	2-Fluorophenol	80.000	83.441	-4.3	94	0.00
6	Aniline	40.000	42.589	-6.5	95	0.00
7 S	Phenol-d6	80.000	85.465	-6.8	95	0.00
8	2-Chlorophenol	40.000	41.744	-4.4	92	0.00
9	Benzaldehyde	40.000	48.473	-21.2	100	0.00
10 C	Phenol	40.000	42.848	-7.1	96	0.00
11	bis(2-Chloroethyl)ether	40.000	41.757	-4.4	94	0.00
12	1,3-Dichlorobenzene	40.000	40.361	-0.9	91	0.00
13 C	1,4-Dichlorobenzene	40.000	40.106	-0.3	90	0.00
14	1,2-Dichlorobenzene	40.000	40.444	-1.1	90	0.00
15	Benzyl Alcohol	40.000	43.467	-8.7	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.958	-7.4	98	0.00
17	2-Methylphenol	40.000	42.472	-6.2	94	0.00
18	Hexachloroethane	40.000	40.811	-2.0	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.540	-6.3	95	0.00
20	3+4-Methylphenols	40.000	42.857	-7.1	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	41.146	-2.9	94	0.00
23 S	Nitrobenzene-d5	80.000	85.377	-6.7	96	0.00
24	Nitrobenzene	40.000	42.414	-6.0	97	0.00
25	Isophorone	40.000	42.024	-5.1	95	0.00
26 C	2-Nitrophenol	40.000	40.209	-0.5	94	0.00
27	2,4-Dimethylphenol	40.000	41.015	-2.5	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.195	-3.0	94	0.00
29 C	2,4-Dichlorophenol	40.000	40.728	-1.8	88	0.00
30	1,2,4-Trichlorobenzene	40.000	38.662	3.3	86	0.00
31	Naphthalene	40.000	40.419	-1.0	92	0.00
32	Benzoic acid	40.000	40.627	-1.6	97	0.00
33	4-Chloroaniline	40.000	41.312	-3.3	93	0.00
34 C	Hexachlorobutadiene	40.000	36.941	7.6	80	0.00
35	Caprolactam	40.000	42.196	-5.5	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.589	-6.5	94	0.00
37	2-Methylnaphthalene	40.000	40.164	-0.4	90	0.00
38	1-Methylnaphthalene	40.000	40.294	-0.7	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.864	2.8	83	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.096	7.3	75	0.00
42 S	2,4,6-Tribromophenol	80.000	72.108	9.9	74	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.516	-3.8	86	0.00
44	2,4,5-Trichlorophenol	40.000	40.963	-2.4	85	0.00
45 S	2-Fluorobiphenyl	80.000	80.153	-0.2	87	0.00
46	1,1'-Biphenyl	40.000	40.717	-1.8	89	0.00
47	2-Chloronaphthalene	40.000	40.807	-2.0	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072823\
 Data File : BP016423.D
 Acq On : 28 Jul 2023 15:05
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 29 00:45:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 01:48:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.770	-16.9	99	0.00
49	Acenaphthylene	40.000	41.572	-3.9	91	0.00
50	Dimethylphthalate	40.000	40.815	-2.0	90	0.00
51	2,6-Dinitrotoluene	40.000	42.467	-6.2	89	0.00
52 C	Acenaphthene	40.000	41.020	-2.6	89	0.00
53	3-Nitroaniline	40.000	43.528	-8.8	93	0.00
54 P	2,4-Dinitrophenol	40.000	38.668	3.3	92	0.00
55	Dibenzofuran	40.000	40.426	-1.1	89	0.00
56 P	4-Nitrophenol	40.000	45.187	-13.0	95	0.00
57	2,4-Dinitrotoluene	40.000	43.809	-9.5	89	0.00
58	Fluorene	40.000	40.872	-2.2	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.814	0.5	82	0.00
60	Diethylphthalate	40.000	41.020	-2.6	90	0.00
61	4-Chlorophenyl-phenylether	40.000	39.406	1.5	84	0.00
62	4-Nitroaniline	40.000	44.414	-11.0	94	0.00
63	Azobenzene	40.000	42.917	-7.3	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.096	-0.2	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.307	-3.3	89	0.00
67	4-Bromophenyl-phenylether	40.000	38.642	3.4	80	0.00
68	Hexachlorobenzene	40.000	37.374	6.6	77	0.00
69	Atrazine	40.000	40.093	-0.2	85	0.00
70 C	Pentachlorophenol	40.000	39.229	1.9	82	0.00
71	Phenanthrene	40.000	40.764	-1.9	89	0.00
72	Anthracene	40.000	41.296	-3.2	89	0.00
73	Carbazole	40.000	41.484	-3.7	90	0.00
74	Di-n-butylphthalate	40.000	42.534	-6.3	90	0.00
75 C	Fluoranthene	40.000	40.782	-2.0	86	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzydine	40.000	37.805	5.5	72	0.00
78	Pyrene	40.000	41.999	-5.0	86	0.00
79 S	Terphenyl-d14	80.000	82.992	-3.7	82	0.00
80	Butylbenzylphthalate	40.000	45.676	-14.2	92	0.00
81	Benzo(a)anthracene	40.000	40.655	-1.6	83	0.00
82	3,3'-Dichlorobenzidine	40.000	40.878	-2.2	80	0.00
83	Chrysene	40.000	41.003	-2.5	85	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.474	-11.2	91	0.00
85 c	Di-n-octyl phthalate	40.000	45.065	-12.7	92	0.00
86 I	Perylene-d12	20.000	20.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.154	-0.4	82	-0.01
88	Benzo(b)fluoranthene	40.000	41.804	-4.5	81	0.00
89	Benzo(k)fluoranthene	40.000	42.380	-6.0	83	0.00
90 C	Benzo(a)pyrene	40.000	41.985	-5.0	82	0.00
91	Dibenzo(a,h)anthracene	40.000	40.450	-1.1	82	0.00
92	Benzo(g,h,i)perylene	40.000	39.697	0.8	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072823\
Data File : BP016423.D
Acq On : 28 Jul 2023 15:05
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jul 29 00:45:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP072723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 01:48:25 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0