

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072924\
 Data File : BP021226.D
 Acq On : 30 Jul 2024 03:18
 Operator : MA/JU
 Sample : P3300-11
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E1G65

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/30/2024
 Supervised By :mohammad ahmed 09/04/2024

Quant Time: Jul 30 03:48:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 29 12:46:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	195529	20.000	ng/u1	0.00
20) Naphthalene-d8	10.393	136	772517	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.269	164	502485	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.081	188	1066256	20.000	ng/u1	-0.01
79) Chrysene-d12	21.533	240	1076749	20.000	ng/u1	0.00
88) Perylene-d12	24.839	264	1331491	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.176	96	18866	4.391	ng/uL	0.00
4) Pyridine-d5	3.611	84	39161	3.148	ng/u1	0.02
7) Phenol-d5	6.869	99	102993	6.461	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.987	67	168124	17.403	ng/u1	0.00
11) 2-Chlorophenol-d4	7.193	132	237627	18.093	ng/u1	0.01
15) 4-Methylphenol-d8	8.375	113	182467	13.782	ng/u1	0.01
21) Nitrobenzene-d5	8.799	128	99082	16.513	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	112071	16.319	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.063	165	212708	17.129	ng/u1	0.01
31) 4-Chloroaniline-d4	10.569	131	267437	16.256	ng/u1	0.01
46) Dimethylphthalate-d6	13.687	166	693500	18.985	ng/u1	0.00
49) Acenaphthylene-d8	13.963	160	736180	17.949	ng/u1	0.00
54) 4-Nitrophenol-d4	14.628	143	38994m	7.503	ng/u1	0.04
60) Fluorene-d10	15.293	176	614963	20.521	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.463	200	96582	14.971	ng/u1	0.00
73) Anthracene-d10	17.192	188	1004440	21.210	ng/u1	0.00
81) Pyrene-d10	19.581	212	1257352	18.888	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.604	264	1466334	22.329	ng/u1	0.00
Target Compounds						
78) Di-n-butylphthalate	18.087	149	84543	1.329	ng/u1	97
86) Bis(2-ethylhexyl)phtha...	21.416	149	52816	1.077	ng/u1#	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP072924\
Data File : BP021226.D
Acq On : 30 Jul 2024 03:18
Operator : MA/JU
Sample : P3300-11
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E1G65

Quant Time: Jul 30 03:48:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 29 12:46:18 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/30/2024
Supervised By :mohammad ahmed 09/04/2024

