

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/2/2023	BNA_P	<u>BP016463.D</u>
Compound Name	Response	Retention Time
DDT	2240361	20.874
DDD	65737	20.486
DDE	0	19.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
65737	2306098	2.85