

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080423\
 Data File : BP016498.D
 Acq On : 04 Aug 2023 12:22
 Operator : MA/JU
 Sample : SST01012
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010612

Quant Time: Aug 04 14:11:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 14:09:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	467247	20.000	ng/u1	0.00
20) Naphthalene-d8	10.899	136	1789766	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.716	164	941077	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.486	188	1812291	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1492885	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1629992	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	48759	4.115	ng/uL	0.00
4) Pyridine-d5	3.840	84	312387	9.291	ng/u1	0.00
7) Phenol-d5	7.205	99	338687	8.662	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	218876	9.324	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	274612	9.002	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	257898	8.202	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	118230	10.728	ng/u1	0.00
24) 2-Nitrophenol-d4	9.981	143	102075	11.230	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.504	165	227481	9.029	ng/u1	0.00
31) 4-Chloroaniline-d4	11.051	131	339945	8.673	ng/u1	0.00
46) Dimethylphthalate-d6	14.128	166	610742	8.867	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	746070	9.169	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	92302	7.998	ng/u1	0.00
60) Fluorene-d10	15.710	176	527167	8.997	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.828	200	51213	8.579	ng/u1	0.00
73) Anthracene-d10	17.581	188	750501	9.349	ng/u1	0.00
81) Pyrene-d10	19.816	212	792977	9.355	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.998	264	716812	9.115	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	51471	4.096	ng/uL	98
5) Pyridine	3.864	79	325709	9.522	ng/u1	99
6) Benzaldehyde	7.222	77	232189	10.885	ng/u1	98
8) Phenol	7.228	94	365080	8.979	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.499	93	306310	9.196	ng/u1	99
12) 2-Chlorophenol	7.616	128	289648	9.235	ng/u1	95
13) 2-Methylphenol	8.499	108	260669	8.361	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.581	45	369831	9.176	ng/u1	98
16) Acetophenone	8.916	105	434185	8.749	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.881	70	194444	7.525	ng/u1	98
18) 4-Methylphenol	8.834	108	277991	8.345	ng/u1	97
19) Hexachloroethane	9.146	117	123162	9.386	ng/u1	96
22) Nitrobenzene	9.304	77	309076	10.475	ng/u1	99
23) Isophorone	9.822	82	527558	7.938	ng/u1	100
25) 2-Nitrophenol	10.010	139	126265	11.133	ng/u1	95
26) 2,4-Dimethylphenol	10.046	107	294255	9.176	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.310	93	369286	9.216	ng/u1	99
29) 2,4-Dichlorophenol	10.528	162	235424	9.243	ng/u1	100
30) Naphthalene	10.951	128	904938	9.762	ng/u1	100
32) 4-Chloroaniline	11.075	127	342903	8.924	ng/u1	100
33) Hexachlorobutadiene	11.187	225	161845	9.669	ng/u1	98
34) Caprolactam	11.869	113	53476	5.894	ng/u1	95
35) 4-Chloro-3-methylphenol	12.163	107	224744	7.972	ng/u1	98
36) 2-Methylnaphthalene	12.551	142	566578	8.969	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080423\
 Data File : BP016498.D
 Acq On : 04 Aug 2023 12:22
 Operator : MA/JU
 Sample : SSTD01012
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD010612


Quant Time: Aug 04 14:11:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 14:09:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.769	142	567289	8.981	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.898	216	289371	10.408	ng/ul	98
40) Hexachlorocyclopentadiene	12.851	237	150957	7.606	ng/ul	98
41) 2,4,6-Trichlorophenol	13.145	196	154394	9.425	ng/ul	97
42) 2,4,5-Trichlorophenol	13.210	196	168544	9.614	ng/ul	98
43) 1,1'-Biphenyl	13.551	154	713307	10.038	ng/ul	100
44) 2-Chloronaphthalene	13.598	162	570501	10.263	ng/ul	99
45) 2-Nitroaniline	13.822	65	116228	9.993	ng/ul	99
47) Dimethylphthalate	14.175	163	617905	8.896	ng/ul	100
48) 2,6-Dinitrotoluene	14.310	165	97501	9.741	ng/ul	99
50) Acenaphthylene	14.445	152	882327	9.518	ng/ul	99
51) 3-Nitroaniline	14.645	138	109887	9.384	ng/ul#	95
52) Acenaphthene	14.781	153	582138	9.739	ng/ul	99
53) 2,4-Dinitrophenol	14.845	184	26970	6.300	ng/ul	94
55) 4-Nitrophenol	14.922	109	74530	8.012	ng/ul	98
56) Dibenzofuran	15.116	168	793195	9.693	ng/ul	98
57) 2,4-Dinitrotoluene	15.092	165	135021	9.462	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.328	232	126367	9.036	ng/ul	99
59) Diethylphthalate	15.528	149	580597	8.299	ng/ul	99
61) Fluorene	15.769	166	620445	9.335	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.757	204	304041	9.383	ng/ul	99
63) 4-Nitroaniline	15.810	138	99520	9.269	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.839	198	60955	8.542	ng/ul	97
67) N-Nitrosodiphenylamine	15.981	169	496110	9.763	ng/ul	99
68) 4-Bromophenyl-phenylether	16.663	248	170225	9.904	ng/ul	98
69) Hexachlorobenzene	16.745	284	200863	9.746	ng/ul	98
70) Atrazine	16.928	200	149738	8.077	ng/ul	98
71) Pentachlorophenol	17.104	266	89216	7.682	ng/ul	98
72) Phenanthrene	17.528	178	922049	9.777	ng/ul	99
74) Anthracene	17.616	178	920049	9.578	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.504	216	293903	11.235	ng/uL	100
76) Pentachlorobenzene	15.010	250	243496	10.533	ng/uL	99
77) Carbazole	17.898	167	786407	9.511	ng/ul	99
78) Di-n-butylphthalate	18.416	149	844103	8.016	ng/ul	99
80) Fluoranthene	19.486	202	957630	9.424	ng/ul	99
82) Pyrene	19.839	202	1030299	9.814	ng/ul	98
83) Butylbenzylphthalate	20.704	149	322385	8.112	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.498	252	278032	8.473	ng/ul	99
85) Benzo(a)anthracene	21.563	228	943616	9.332	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.451	149	456380	7.528	ng/ul	100
87) Chrysene	21.621	228	893319	9.540	ng/ul	99
89) Di-n-octyl phthalate	22.439	149	753236	7.351	ng/ul	100
90) Benzo(b)fluoranthene	23.363	252	902017	9.241	ng/ul	99
91) Benzo(k)fluoranthene	23.415	252	894736	9.211	ng/ul	99
93) Benzo(a)pyrene	24.057	252	851733	9.265	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.933	276	1060737	9.992	ng/ul	98
95) Dibenzo(a,h)anthracene	26.962	278	886536	10.097	ng/ul	98
96) Benzo(g,h,i)perylene	27.792	276	859056	10.222	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 04 14:11:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 04 14:09:57 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 4.096 ng/uL
 RT: 3.446 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP016498.D
 Acq: 04 Aug 2023 12:22




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010612


Tgt Ion: 88 Resp: 51471
 Ion Ratio Lower Upper
 88 100
 43 24.5 21.0 31.4
 58 67.3 55.0 82.4



#5
 Pyridine
 Concen: 9.522 ng/uL
 RT: 3.864 min Scan# 98
 Delta R.T. -0.000 min
 Lab File: BP016498.D
 Acq: 04 Aug 2023 12:22

Tgt Ion: 79 Resp: 325709
 Ion Ratio Lower Upper
 79 100
 52 63.1 50.2 75.2
 51 31.1 23.8 35.6









#6
Benzaldehyde
Concen: 10.885 ng/ul
RT: 7.222 min Scan# 60
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion: 77 Resp: 232189
Ion Ratio Lower Upper
77 100
105 98.5 79.6 119.4
106 96.0 74.9 112.3



#8
Phenol
Concen: 8.979 ng/ul
RT: 7.228 min Scan# 670
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion: 94 Resp: 365080
Ion Ratio Lower Upper
94 100
65 25.4 20.3 30.5
66 33.8 26.6 39.8









#10
Bis(2-Chloroethyl)ether
Concen: 9.196 ng/ul
RT: 7.499 min Scan# 716
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion: 93 Resp: 306310
Ion Ratio Lower Upper
93 100
63 71.0 57.0 85.4
95 32.1 26.2 39.2



#12
2-Chlorophenol
Concen: 9.235 ng/ul
RT: 7.616 min Scan# 736
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:128 Resp: 289648
Ion Ratio Lower Upper
128 100
64 46.5 34.9 52.3
130 33.3 24.6 36.8





#13

2-Methylphenol

Concen: 8.361 ng/ul

RT: 8.499 min Scan# 8

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :





SSTD010612

Tgt Ion:108 Resp: 260669

Ion Ratio Lower Upper

108 100

107 90.0 71.0 106.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.176 ng/ul

RT: 8.581 min Scan# 900

Delta R.T. -0.006 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion: 45 Resp: 369831


Ion Ratio Lower Upper

45 100

77 17.0 12.5 18.7

79 11.9 9.8 14.8









#16
Acetophenone
Concen: 8.749 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:105 Resp: 434185
Ion Ratio Lower Upper
105 100
77 73.5 59.6 89.4
51 22.6 18.5 27.7



#17
N-Nitroso-di-n-propylamine
Concen: 7.525 ng/ul
RT: 8.881 min Scan# 951
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion: 70 Resp: 194444
Ion Ratio Lower Upper
70 100
42 50.9 39.9 59.9
101 9.8 7.5 11.3
130 19.7 17.0 25.4





#18

4-Methylphenol

Concen: 8.345 ng/ul

RT: 8.834 min Scan# 943

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :





SSTD010612

Tgt Ion:108 Resp: 277991

Ion Ratio Lower Upper

108 100

107 108.2 89.1 133.7



#19

Hexachloroethane

Concen: 9.386 ng/ul

RT: 9.146 min Scan# 996

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:117 Resp: 123162


Ion Ratio Lower Upper

117 100

201 92.4 77.6 116.4

199 58.0 48.2 72.2









#22
Nitrobenzene
Concen: 10.475 ng/ul
RT: 9.304 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion: 77 Resp: 309076
Ion Ratio Lower Upper
77 100
123 44.3 35.3 52.9
65 13.2 10.2 15.4



#23
Isophorone
Concen: 7.938 ng/ul
RT: 9.822 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion: 82 Resp: 527558
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.1 13.5 20.3









#25
2-Nitrophenol
Concen: 11.133 ng/ul
RT: 10.010 min Scan# 1144
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:139 Resp: 126265
Ion Ratio Lower Upper
139 100
65 50.5 37.6 56.4
109 31.0 22.9 34.3



#26
2,4-Dimethylphenol
Concen: 9.176 ng/ul
RT: 10.046 min Scan# 1149
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:107 Resp: 294255
Ion Ratio Lower Upper
107 100
121 50.4 40.5 60.7
122 88.1 70.2 105.4









#27
Bis(2-Chloroethoxy)methane
Concen: 9.216 ng/ul
RT: 10.310 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion: 93 Resp: 369286
Ion Ratio Lower Upper
93 100
95 32.0 26.5 39.7
123 11.0 8.8 13.2



#29
2,4-Dichlorophenol
Concen: 9.243 ng/ul
RT: 10.528 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion: 162 Resp: 235424
Ion Ratio Lower Upper
162 100
164 66.7 53.4 80.0
98 38.0 29.8 44.8









#30
Naphthalene
Concen: 9.762 ng/ul
RT: 10.951 min Scan# 1303
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:128 Resp: 904938
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 8.924 ng/ul
RT: 11.075 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:127 Resp: 342903
Ion Ratio Lower Upper
127 100
129 32.1 25.8 38.6





#33

Hexachlorobutadiene

Concen: 9.669 ng/ul

RT: 11.187 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:225 Resp: 161845

Ion Ratio Lower Upper

225 100

223 61.7 51.4 77.2

227 64.1 50.4 75.6



#34

Caprolactam

Concen: 5.894 ng/ul

RT: 11.869 min Scan# 1459

Delta R.T. -0.006 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:113 Resp: 53476


Ion Ratio Lower Upper

113 100

55 149.9 127.0 190.4

56 113.1 93.6 140.4









#35
4-Chloro-3-methylphenol
Concen: 7.972 ng/ul
RT: 12.163 min Scan# 1509
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:107 Resp: 224744
Ion Ratio Lower Upper
107 100
144 25.7 20.5 30.7
142 79.2 65.2 97.8



#36
2-Methylnaphthalene
Concen: 8.969 ng/ul
RT: 12.551 min Scan# 1575
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:142 Resp: 566578
Ion Ratio Lower Upper
142 100
141 88.7 70.9 106.3









#37
1-Methylnaphthalene
Concen: 8.981 ng/ul
RT: 12.769 min Scan# 1612
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:142 Resp: 567289
Ion Ratio Lower Upper
142 100
141 90.1 73.5 110.3
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 10.408 ng/ul
RT: 12.898 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:216 Resp: 289371
Ion Ratio Lower Upper
216 100
214 78.1 63.2 94.8
179 20.6 17.4 26.2
108 17.0 14.8 22.2





#40

Hexachlorocyclopentadiene

Concen: 7.606 ng/ul

RT: 12.851 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:237 Resp: 150957

Ion Ratio Lower Upper

237 100

235 60.9 50.3 75.5

272 12.3 10.2 15.4



#41

2,4,6-Trichlorophenol

Concen: 9.425 ng/ul

RT: 13.145 min Scan# 1676

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:196 Resp: 154394


Ion Ratio Lower Upper

196 100

198 98.9 77.1 115.7

200 33.3 25.0 37.4





#42

2,4,5-Trichlorophenol

Concen: 9.614 ng/ul

RT: 13.210 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:196 Resp: 168544

Ion Ratio Lower Upper

196 100

198 94.1 76.8 115.2

200 32.0 24.8 37.2



#43

1,1'-Biphenyl

Concen: 10.038 ng/ul

RT: 13.551 min Scan# 1745

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:154 Resp: 713307


Ion Ratio Lower Upper

154 100

153 39.9 31.9 47.9

76 13.7 11.2 16.8









#44
2-Chloronaphthalene
Concen: 10.263 ng/ul
RT: 13.598 min Scan# 1753
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:162 Resp: 570501
Ion Ratio Lower Upper
162 100
164 32.3 26.6 39.8
127 38.7 31.1 46.7



#45
2-Nitroaniline
Concen: 9.993 ng/ul
RT: 13.822 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion: 65 Resp: 116228
Ion Ratio Lower Upper
65 100
92 71.8 57.1 85.7
138 106.5 86.8 130.2









#47
Dimethylphthalate
Concen: 8.896 ng/ul
RT: 14.175 min Scan# 1851
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:163 Resp: 617905
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 9.741 ng/ul
RT: 14.310 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:165 Resp: 97501
Ion Ratio Lower Upper
165 100
89 62.8 50.1 75.1
121 23.5 17.9 26.9









#50
Acenaphthylene
Concen: 9.518 ng/ul
RT: 14.445 min Scan# 1897
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:152 Resp: 882327
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 13.2 10.7 16.1



#51
3-Nitroaniline
Concen: 9.384 ng/ul
RT: 14.645 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:138 Resp: 109887
Ion Ratio Lower Upper
138 100
108 14.7 9.7 14.5#
92 120.8 92.9 139.3









#52
Acenaphthene
Concen: 9.739 ng/ul
RT: 14.781 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:153 Resp: 582138
Ion Ratio Lower Upper
153 100
152 48.1 37.8 56.8
154 88.3 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 6.300 ng/ul
RT: 14.845 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:184 Resp: 26970
Ion Ratio Lower Upper
184 100
63 70.9 51.4 77.0
154 61.1 50.6 75.8









#55
4-Nitrophenol
Concen: 8.012 ng/ul
RT: 14.922 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:109 Resp: 74530
Ion Ratio Lower Upper
109 100
139 133.5 108.4 162.6
65 130.2 101.2 151.8



#56
Dibenzofuran
Concen: 9.693 ng/ul
RT: 15.116 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:168 Resp: 793195
Ion Ratio Lower Upper
168 100
139 36.1 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 9.462 ng/ul
RT: 15.092 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:165 Resp: 135021
Ion Ratio Lower Upper
165 100
63 44.3 33.5 50.3
182 2.5 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.036 ng/ul
RT: 15.328 min Scan# 2047
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:232 Resp: 126367
Ion Ratio Lower Upper
232 100
131 45.6 37.0 55.4
130 2.9 1.9 2.9
166 30.8 25.1 37.7









#59
Diethylphthalate
Concen: 8.299 ng/ul
RT: 15.528 min Scan# 2081
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:149 Resp: 580597
Ion Ratio Lower Upper
149 100
177 21.7 17.7 26.5
150 12.0 10.0 15.0



#61
Fluorene
Concen: 9.335 ng/ul
RT: 15.769 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:166 Resp: 620445
Ion Ratio Lower Upper
166 100
165 97.8 78.5 117.7
167 13.5 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 9.383 ng/ul
RT: 15.757 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:204 Resp: 304041
Ion Ratio Lower Upper
204 100
206 32.5 26.1 39.1
141 61.7 48.6 72.8



#63
4-Nitroaniline
Concen: 9.269 ng/ul
RT: 15.810 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:138 Resp: 99520
Ion Ratio Lower Upper
138 100
92 57.7 43.4 65.0
108 84.7 70.2 105.4









#66
4,6-Dinitro-2-methylphenol
Concen: 8.542 ng/ul
RT: 15.839 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:198 Resp: 60955
Ion Ratio Lower Upper
198 100
182 6.2 4.2 6.4
77 31.6 23.8 35.8



#67
N-Nitrosodiphenylamine
Concen: 9.763 ng/ul
RT: 15.981 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:169 Resp: 496110
Ion Ratio Lower Upper
169 100
168 66.6 53.6 80.4
167 35.6 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 9.904 ng/ul
RT: 16.663 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:248 Resp: 170225
Ion Ratio Lower Upper
248 100
250 98.4 78.1 117.1
141 75.1 57.5 86.3



#69
Hexachlorobenzene
Concen: 9.746 ng/ul
RT: 16.745 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:284 Resp: 200863
Ion Ratio Lower Upper
284 100
142 35.8 26.9 40.3
249 30.4 24.2 36.4





#70

Atrazine

Concen: 8.077 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:200 Resp: 149738

Ion Ratio Lower Upper

200 100

173 27.0 21.0 31.4

215 45.6 38.1 57.1



#71

Pentachlorophenol

Concen: 7.682 ng/ul

RT: 17.104 min Scan# 2349

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:266 Resp: 89216


Ion Ratio Lower Upper

266 100

264 63.3 51.4 77.0

268 66.4 51.5 77.3









#72
Phenanthrene
Concen: 9.777 ng/ul
RT: 17.528 min Scan# 2421
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:178 Resp: 922049
Ion Ratio Lower Upper
178 100
179 15.5 12.7 19.1
176 19.2 15.6 23.4



#74
Anthracene
Concen: 9.578 ng/ul
RT: 17.616 min Scan# 2436
Delta R.T. -0.006 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:178 Resp: 920049
Ion Ratio Lower Upper
178 100
179 15.8 12.5 18.7
176 18.2 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 11.235 ng/uL

RT: 13.504 min Scan# 1737

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:216 Resp: 293903

Ion Ratio Lower Upper

216 100

214 78.7 63.0 94.6

218 46.8 37.8 56.8



#76

Pentachlorobenzene

Concen: 10.533 ng/uL

RT: 15.010 min Scan# 1993

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:250 Resp: 243496


Ion Ratio Lower Upper

250 100

248 63.0 49.6 74.4

252 63.3 50.7 76.1





#77

Carbazole

Concen: 9.511 ng/ul

RT: 17.898 min Scan# 2484

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:167 Resp: 786407

Ion Ratio Lower Upper

167 100

166 20.9 17.2 25.8

139 12.5 10.5 15.7



#78

Di-n-butylphthalate

Concen: 8.016 ng/ul

RT: 18.416 min Scan# 2572

Delta R.T. 0.006 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:149 Resp: 844103


Ion Ratio Lower Upper

149 100

150 8.7 7.4 11.2

104 5.1 4.2 6.4









#80
Fluoranthene
Concen: 9.424 ng/ul
RT: 19.486 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:202 Resp: 957630
Ion Ratio Lower Upper
202 100
101 11.5 9.5 14.3
100 9.1 7.3 10.9



#82
Pyrene
Concen: 9.814 ng/ul
RT: 19.839 min Scan# 2814
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:202 Resp: 1030299
Ion Ratio Lower Upper
202 100
101 14.2 12.1 18.1
100 11.6 9.8 14.6





#83

Butylbenzylphthalate

Concen: 8.112 ng/ul

RT: 20.704 min Scan# 2961

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:149 Resp: 322385

Ion Ratio Lower Upper

149 100

91 75.2 58.1 87.1

206 20.1 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 8.473 ng/ul

RT: 21.498 min Scan# 3096

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:252 Resp: 278032


Ion Ratio Lower Upper

252 100

254 63.3 51.0 76.4

126 12.9 9.5 14.3





#85

Benzo(a)anthracene

Concen: 9.332 ng/ul

RT: 21.563 min Scan# 3107

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:228 Resp: 943616

Ion Ratio Lower Upper

228 100

229 19.9 15.7 23.5

226 26.7 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 7.528 ng/ul

RT: 21.451 min Scan# 3088

Delta R.T. 0.006 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:149 Resp: 456380


Ion Ratio Lower Upper

149 100

167 26.8 21.4 32.2

279 4.1 3.0 4.6









#87
Chrysene
Concen: 9.540 ng/ul
RT: 21.621 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:228 Resp: 893319
Ion Ratio Lower Upper
228 100
226 29.6 24.1 36.1
229 19.7 15.4 23.2



#89
Di-n-octyl phthalate
Concen: 7.351 ng/ul
RT: 22.439 min Scan# 3256
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:149 Resp: 753236









#90
Benzo(b)fluoranthene
Concen: 9.241 ng/ul
RT: 23.363 min Scan# 3413
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22




Instrument :
BNA_P
ClientSampleId :
SSTD010612


Tgt Ion:252 Resp: 902017
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 11.5 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 9.211 ng/ul
RT: 23.415 min Scan# 3422
Delta R.T. -0.000 min
Lab File: BP016498.D
Acq: 04 Aug 2023 12:22

Tgt Ion:252 Resp: 894736
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 11.7 9.4 14.0





#93

Benzo(a)pyrene

Concen: 9.265 ng/ul

RT: 24.057 min Scan# 3531

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:252 Resp: 851733

Ion Ratio Lower Upper

252 100

253 21.4 17.8 26.8

125 11.6 9.7 14.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.992 ng/ul

RT: 26.933 min Scan# 4020

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22




Tgt Ion:276 Resp: 1060737


Ion Ratio Lower Upper

276 100

138 24.6 18.7 28.1

277 24.2 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 10.097 ng/ul

RT: 26.962 min Scan# 4025

Delta R.T. -0.000 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Instrument :

BNA_P

ClientSampleId :

SSTD010612





Tgt Ion:278 Resp: 886536

Ion Ratio Lower Upper

278 100

139 16.9 14.4 21.6

279 24.5 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 10.222 ng/ul

RT: 27.792 min Scan# 4166

Delta R.T. -0.006 min

Lab File: BP016498.D

Acq: 04 Aug 2023 12:22

Tgt Ion:276 Resp: 859056

Ion Ratio Lower Upper

276 100

138 22.5 18.1 27.1

277 23.9 19.0 28.4

