

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080423\
 Data File : BP016505.D
 Acq On : 04 Aug 2023 16:20
 Operator : MA/JU
 Sample : 03820-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC-BLANK

Quant Time: Aug 04 16:51:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 15:15:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	335065	20.000	ng/ul	0.00
20) Naphthalene-d8	10.899	136	1270604	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.716	164	648526	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.481	188	1319257	20.000	ng/ul	0.00
79) Chrysene-d12	21.580	240	1131724	20.000	ng/ul	0.00
88) Perylene-d12	24.169	264	1293468	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	138	0.015	ng/uL	0.00
4) Pyridine-d5	3.852	84	1033	0.043	ng/ul	0.01
7) Phenol-d5	7.205	99	163	0.006	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.399	67	121	0.007	ng/ul	0.00
11) 2-Chlorophenol-d4	7.581	132	63	0.003	ng/ul	0.00
15) 4-Methylphenol-d8	8.781	113	112	0.006	ng/ul	0.01
21) Nitrobenzene-d5	9.263	128	49	0.005	ng/ul	0.00
24) 2-Nitrophenol-d4	9.975	143	11	0.001	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	150	0.009	ng/ul	0.00
31) 4-Chloroaniline-d4	11.046	131	115	0.004	ng/ul	0.00
46) Dimethylphthalate-d6	14.134	166	343	0.008	ng/ul	0.00
49) Acenaphthylene-d8	14.416	160	332	0.006	ng/ul	0.00
54) 4-Nitrophenol-d4	14.922	143	36	0.004	ng/ul	0.01
60) Fluorene-d10	15.710	176	1068	0.028	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	29	0.005	ng/ul	-0.02
73) Anthracene-d10	17.481	188	1319257	23.163	ng/ul	-0.11
81) Pyrene-d10	19.816	212	1858	0.030	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.992	264	3649	0.059	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080423\
Data File : BP016505.D
Acq On : 04 Aug 2023 16:20
Operator : MA/JU
Sample : 03820-05
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: Aug 04 16:51:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 04 15:15:50 2023
Response via : Initial Calibration

