

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080423\
 Data File : BP016507.D
 Acq On : 04 Aug 2023 17:29
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020675

Quant Time: Aug 04 23:50:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 15:15:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	499961	20.000	ng/u1	0.00
20) Naphthalene-d8	10.899	136	1964178	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.716	164	1045877	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.481	188	2081184	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1813256	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	2061955	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	90454	6.562	ng/uL	0.00
4) Pyridine-d5	3.840	84	692142	19.323	ng/u1	0.00
7) Phenol-d5	7.205	99	770698	19.711	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	509289	21.144	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	606667	19.211	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	565396	18.753	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	266088	18.727	ng/u1	0.00
24) 2-Nitrophenol-d4	9.981	143	252228	19.028	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.504	165	519876	19.274	ng/u1	0.00
31) 4-Chloroaniline-d4	11.052	131	743293	18.692	ng/u1	0.00
46) Dimethylphthalate-d6	14.128	166	1282071	17.727	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	1633028	18.751	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	228028	16.745	ng/u1	0.00
60) Fluorene-d10	15.710	176	1130284	18.200	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.828	200	142177	15.516	ng/u1	0.00
73) Anthracene-d10	17.581	188	1633524	18.181	ng/u1	0.00
81) Pyrene-d10	19.810	212	1801482	18.120	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.998	264	1768009	17.906	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	94618	6.407	ng/uL	99
5) Pyridine	3.864	79	577634	15.703	ng/u1	97
6) Benzaldehyde	7.222	77	512997	25.001	ng/u1	99
8) Phenol	7.234	94	740428	18.114	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.499	93	598742	17.680	ng/u1	99
12) 2-Chlorophenol	7.622	128	591072	18.005	ng/u1	98
13) 2-Methylphenol	8.499	108	529721	17.614	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.581	45	415887	10.255	ng/u1	99
16) Acetophenone	8.916	105	898549	19.244	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.887	70	398715	18.135	ng/u1	99
18) 4-Methylphenol	8.834	108	563027	17.605	ng/u1	100
19) Hexachloroethane	9.146	117	241555	17.304	ng/u1	98
22) Nitrobenzene	9.305	77	630154	17.596	ng/u1	100
23) Isophorone	9.822	82	1055363	16.959	ng/u1	99
25) 2-Nitrophenol	10.010	139	275312	17.874	ng/u1	95
26) 2,4-Dimethylphenol	10.046	107	524178	15.797	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.310	93	728028	17.228	ng/u1	98
29) 2,4-Dichlorophenol	10.534	162	493301	17.853	ng/u1	98
30) Naphthalene	10.951	128	1736065	17.058	ng/u1	100
32) 4-Chloroaniline	11.075	127	694161	17.731	ng/u1	100
33) Hexachlorobutadiene	11.187	225	310029	16.743	ng/u1	98
34) Caprolactam	11.881	113	133048	16.119	ng/u1	98
35) 4-Chloro-3-methylphenol	12.163	107	482167	17.723	ng/u1	98
36) 2-Methylnaphthalene	12.551	142	1102668	17.124	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080423\
 Data File : BP016507.D
 Acq On : 04 Aug 2023 17:29
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020675

Quant Time: Aug 04 23:50:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 15:15:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.769	142	1074459	16.737	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.898	216	589443	17.804	ng/ul	99
40) Hexachlorocyclopentadiene	12.851	237	386397	19.829	ng/ul	99
41) 2,4,6-Trichlorophenol	13.146	196	330467	17.452	ng/ul	98
42) 2,4,5-Trichlorophenol	13.210	196	358271	17.427	ng/ul	99
43) 1,1'-Biphenyl	13.551	154	1478719	18.292	ng/ul	99
44) 2-Chloronaphthalene	13.598	162	1086338	16.889	ng/ul	100
45) 2-Nitroaniline	13.822	65	271031	17.913	ng/ul	99
47) Dimethylphthalate	14.175	163	1210548	16.531	ng/ul	100
48) 2,6-Dinitrotoluene	14.310	165	252412	19.600	ng/ul	99
50) Acenaphthylene	14.445	152	1736557	17.164	ng/ul	100
51) 3-Nitroaniline	14.645	138	272993	19.271	ng/ul	95
52) Acenaphthene	14.781	153	1090864	16.404	ng/ul	99
53) 2,4-Dinitrophenol	14.845	184	80743	13.096	ng/ul	97
55) 4-Nitrophenol	14.922	109	170194	16.806	ng/ul	96
56) Dibenzofuran	15.116	168	1569412	17.404	ng/ul	99
57) 2,4-Dinitrotoluene	15.092	165	311194	17.133	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.328	232	289068	18.187	ng/ul	100
59) Diethylphthalate	15.528	149	1156611	16.460	ng/ul	100
61) Fluorene	15.763	166	1199626	16.830	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.757	204	578377	16.484	ng/ul	98
63) 4-Nitroaniline	15.816	138	256368	20.513	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.839	198	143011	13.893	ng/ul	99
67) N-Nitrosodiphenylamine	15.981	169	1015100	17.375	ng/ul	99
68) 4-Bromophenyl-phenylether	16.663	248	332242	16.271	ng/ul	98
69) Hexachlorobenzene	16.745	284	387801	16.199	ng/ul	99
70) Atrazine	16.928	200	329051	16.504	ng/ul	97
71) Pentachlorophenol	17.104	266	200135	14.370	ng/ul	99
72) Phenanthrene	17.528	178	1813853	16.714	ng/ul	99
74) Anthracene	17.616	178	1825116	16.734	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.510	216	561163	16.902	ng/ul	99
76) Pentachlorobenzene	15.010	250	491546	17.454	ng/ul	99
77) Carbazole	17.892	167	1633271	17.266	ng/ul	99
78) Di-n-butylphthalate	18.410	149	1829127	16.689	ng/ul	100
80) Fluoranthene	19.486	202	1959010	16.524	ng/ul	99
82) Pyrene	19.839	202	2080676	16.608	ng/ul	100
83) Butylbenzylphthalate	20.704	149	743486	16.157	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.498	252	716258	18.475	ng/ul	100
85) Benzo(a)anthracene	21.563	228	1985195	16.514	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.445	149	1102781	16.743	ng/ul	99
87) Chrysene	21.616	228	1854128	16.365	ng/ul	99
89) Di-n-octyl phthalate	22.433	149	1802825	15.701	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	1980785	16.279	ng/ul	100
91) Benzo(k)fluoranthene	23.416	252	1896177	15.486	ng/ul	99
93) Benzo(a)pyrene	24.051	252	1786198	15.478	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.927	276	2383486	16.721	ng/ul	99
95) Dibenzo(a,h)anthracene	26.957	278	1982448	16.661	ng/ul	99
96) Benzo(g,h,i)perylene	27.786	276	1910859	16.596	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
 1,4-Dioxane
 Concen: 6.407 ng/uL
 RT: 3.446 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BP016507.D
 Acq: 04 Aug 2023 17:29

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020675

Tgt Ion: 88 Resp: 94618




Ion	Ratio	Lower	Upper
88	100		
43	27.4	21.0	31.4
58	68.8	55.0	82.4




#5
 Pyridine
 Concen: 15.703 ng/uL
 RT: 3.864 min Scan# 98
 Delta R.T. 0.000 min
 Lab File: BP016507.D
 Acq: 04 Aug 2023 17:29

Tgt Ion: 79 Resp: 577634

Ion	Ratio	Lower	Upper
79	100		
52	65.8	50.2	75.2
51	30.6	23.8	35.6









#6
Benzaldehyde
Concen: 25.001 ng/ul
RT: 7.222 min Scan# 609
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion: 77 Resp: 512997
Ion Ratio Lower Upper
77 100
105 98.9 79.6 119.4
106 95.6 74.9 112.3



#8
Phenol
Concen: 18.114 ng/ul
RT: 7.234 min Scan# 671
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion: 94 Resp: 740428
Ion Ratio Lower Upper
94 100
65 26.0 20.3 30.5
66 34.1 26.6 39.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 17.680 ng/ul
 RT: 7.499 min Scan# 716
 Delta R.T. 0.000 min
 Lab File: BP016507.D
 Acq: 04 Aug 2023 17:29




Instrument :
 BNA_P
 ClientSampleId :
 SST020675


Tgt Ion: 93 Resp: 598742
 Ion Ratio Lower Upper
 93 100
 63 72.3 57.0 85.4
 95 33.3 26.2 39.2



#12
 2-Chlorophenol
 Concen: 18.005 ng/ul
 RT: 7.622 min Scan# 737
 Delta R.T. 0.006 min
 Lab File: BP016507.D
 Acq: 04 Aug 2023 17:29

Tgt Ion:128 Resp: 591072
 Ion Ratio Lower Upper
 128 100
 64 43.9 34.9 52.3
 130 32.4 24.6 36.8





#13

2-Methylphenol

Concen: 17.614 ng/ul

RT: 8.499 min Scan# 886

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :





SSTD020675

Tgt Ion:108 Resp: 529721

Ion Ratio Lower Upper

108 100

107 89.7 71.0 106.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 10.255 ng/ul

RT: 8.581 min Scan# 900

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion: 45 Resp: 415887


Ion Ratio Lower Upper

45 100

77 16.2 12.5 18.7

79 11.9 9.8 14.8









#16
Acetophenone
Concen: 19.244 ng/ul
RT: 8.916 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:105 Resp: 898549
Ion Ratio Lower Upper
105 100
77 76.3 59.6 89.4
51 23.2 18.5 27.7



#17
N-Nitroso-di-n-propylamine
Concen: 18.135 ng/ul
RT: 8.887 min Scan# 952
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion: 70 Resp: 398715
Ion Ratio Lower Upper
70 100
42 50.5 39.9 59.9
101 9.8 7.5 11.3
130 21.8 17.0 25.4









#18
4-Methylphenol
Concen: 17.605 ng/ul
RT: 8.834 min Scan# 943
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:108 Resp: 563027
Ion Ratio Lower Upper
108 100
107 110.9 89.1 133.7



#19
Hexachloroethane
Concen: 17.304 ng/ul
RT: 9.146 min Scan# 996
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:117 Resp: 241555
Ion Ratio Lower Upper
117 100
201 94.1 77.6 116.4
199 60.5 48.2 72.2









#22
Nitrobenzene
Concen: 17.596 ng/ul
RT: 9.305 min Scan# 1023
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion: 77 Resp: 630154
Ion Ratio Lower Upper
77 100
123 44.1 35.3 52.9
65 13.0 10.2 15.4



#23
Isophorone
Concen: 16.959 ng/ul
RT: 9.822 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion: 82 Resp: 1055363
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 17.1 13.5 20.3









#25
2-Nitrophenol
Concen: 17.874 ng/ul
RT: 10.010 min Scan# 1144
Delta R.T. -0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:139 Resp: 275312
Ion Ratio Lower Upper
139 100
65 51.1 37.6 56.4
109 30.1 22.9 34.3



#26
2,4-Dimethylphenol
Concen: 15.797 ng/ul
RT: 10.046 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:107 Resp: 524178
Ion Ratio Lower Upper
107 100
121 50.3 40.5 60.7
122 85.1 70.2 105.4









#27
Bis(2-Chloroethoxy)methane
Concen: 17.228 ng/ul
RT: 10.310 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion: 93 Resp: 728028
Ion Ratio Lower Upper
93 100
95 32.0 26.5 39.7
123 11.2 8.8 13.2



#29
2,4-Dichlorophenol
Concen: 17.853 ng/ul
RT: 10.534 min Scan# 1232
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion: 162 Resp: 493301
Ion Ratio Lower Upper
162 100
164 64.9 53.4 80.0
98 35.9 29.8 44.8









#30
Naphthalene
Concen: 17.058 ng/ul
RT: 10.951 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:128 Resp: 1736065
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 17.731 ng/ul
RT: 11.075 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:127 Resp: 694161
Ion Ratio Lower Upper
127 100
129 32.2 25.8 38.6









#33
Hexachlorobutadiene
Concen: 16.743 ng/ul
RT: 11.187 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:225 Resp: 310029
Ion Ratio Lower Upper
225 100
223 62.6 51.4 77.2
227 64.1 50.4 75.6



#34
Caprolactam
Concen: 16.119 ng/ul
RT: 11.881 min Scan# 1461
Delta R.T. 0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:113 Resp: 133048
Ion Ratio Lower Upper
113 100
55 161.2 127.0 190.4
56 119.3 93.6 140.4









#35
4-Chloro-3-methylphenol
Concen: 17.723 ng/ul
RT: 12.163 min Scan# 1510
Delta R.T. -0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:107 Resp: 482167
Ion Ratio Lower Upper
107 100
144 25.7 20.5 30.7
142 79.0 65.2 97.8



#36
2-Methylnaphthalene
Concen: 17.124 ng/ul
RT: 12.551 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:142 Resp: 1102668
Ion Ratio Lower Upper
142 100
141 88.1 70.9 106.3









#37
1-Methylnaphthalene
Concen: 16.737 ng/ul
RT: 12.769 min Scan# 1612
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:142 Resp: 1074459
Ion Ratio Lower Upper
142 100
141 91.8 73.5 110.3
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.804 ng/ul
RT: 12.898 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:216 Resp: 589443
Ion Ratio Lower Upper
216 100
214 79.0 63.2 94.8
179 21.0 17.4 26.2
108 17.3 14.8 22.2





#40

Hexachlorocyclopentadiene

Concen: 19.829 ng/ul

RT: 12.851 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:237 Resp: 386397

Ion Ratio Lower Upper

237 100

235 62.5 50.3 75.5

272 13.7 10.2 15.4



#41

2,4,6-Trichlorophenol

Concen: 17.452 ng/ul

RT: 13.146 min Scan# 1676

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:196 Resp: 330467


Ion Ratio Lower Upper

196 100

198 98.0 77.1 115.7

200 32.0 25.0 37.4





#42

2,4,5-Trichlorophenol

Concen: 17.427 ng/ul

RT: 13.210 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:196 Resp: 358271

Ion Ratio Lower Upper

196 100

198 95.3 76.8 115.2

200 30.9 24.8 37.2



#43

1,1'-Biphenyl

Concen: 18.292 ng/ul

RT: 13.551 min Scan# 1745

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:154 Resp: 1478719


Ion Ratio Lower Upper

154 100

153 40.1 31.9 47.9

76 14.6 11.2 16.8





#44

2-Chloronaphthalene

Concen: 16.889 ng/ul

RT: 13.598 min Scan# 1753

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:162 Resp: 1086338

Ion Ratio Lower Upper

162 100

164 33.0 26.6 39.8

127 38.8 31.1 46.7



#45

2-Nitroaniline

Concen: 17.913 ng/ul

RT: 13.822 min Scan# 1791

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion: 65 Resp: 271031


Ion Ratio Lower Upper

65 100

92 72.3 57.1 85.7

138 107.6 86.8 130.2









#47
Dimethylphthalate
Concen: 16.531 ng/ul
RT: 14.175 min Scan# 1851
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:163 Resp: 1210548
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 19.600 ng/ul
RT: 14.310 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:165 Resp: 252412
Ion Ratio Lower Upper
165 100
89 61.8 50.1 75.1
121 21.8 17.9 26.9









#50
Acenaphthylene
Concen: 17.164 ng/ul
RT: 14.445 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:152 Resp: 1736557
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.2 10.7 16.1



#51
3-Nitroaniline
Concen: 19.271 ng/ul
RT: 14.645 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:138 Resp: 272993
Ion Ratio Lower Upper
138 100
108 11.2 9.7 14.5
92 122.2 92.9 139.3





#52

Acenaphthene

Concen: 16.404 ng/ul

RT: 14.781 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNP_P

ClientSampleId :

SSTD020675





Tgt Ion:153 Resp: 1090864

Ion Ratio Lower Upper

153 100

152 48.7 37.8 56.8

154 88.6 70.2 105.4



#53

2,4-Dinitrophenol

Concen: 13.096 ng/ul

RT: 14.845 min Scan# 1965

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:184 Resp: 80743


Ion Ratio Lower Upper

184 100

63 66.2 51.4 77.0

154 61.0 50.6 75.8









#55
4-Nitrophenol
Concen: 16.806 ng/ul
RT: 14.922 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:109 Resp: 170194
Ion Ratio Lower Upper
109 100
139 132.7 108.4 162.6
65 132.7 101.2 151.8



#56
Dibenzofuran
Concen: 17.404 ng/ul
RT: 15.116 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:168 Resp: 1569412
Ion Ratio Lower Upper
168 100
139 36.3 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 17.133 ng/ul
RT: 15.092 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:165 Resp: 311194
Ion Ratio Lower Upper
165 100
63 42.8 33.5 50.3
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.187 ng/ul
RT: 15.328 min Scan# 2047
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:232 Resp: 289068
Ion Ratio Lower Upper
232 100
131 46.3 37.0 55.4
130 2.2 1.9 2.9
166 31.3 25.1 37.7





#59

Diethylphthalate

Concen: 16.460 ng/ul

RT: 15.528 min Scan# 2081

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:149 Resp: 1156611

Ion Ratio Lower Upper

149 100

177 22.1 17.7 26.5

150 12.5 10.0 15.0



#61

Fluorene

Concen: 16.830 ng/ul

RT: 15.763 min Scan# 2121

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:166 Resp: 1199626


Ion Ratio Lower Upper

166 100

165 97.5 78.5 117.7

167 13.5 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 16.484 ng/ul
RT: 15.757 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:204 Resp: 578377
Ion Ratio Lower Upper
204 100
206 33.5 26.1 39.1
141 59.2 48.6 72.8



#63
4-Nitroaniline
Concen: 20.513 ng/ul
RT: 15.816 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:138 Resp: 256368
Ion Ratio Lower Upper
138 100
92 51.4 43.4 65.0
108 85.3 70.2 105.4









#66
4,6-Dinitro-2-methylphenol
Concen: 13.893 ng/ul
RT: 15.839 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:198 Resp: 143011
Ion Ratio Lower Upper
198 100
182 4.5 4.2 6.4
77 29.5 23.8 35.8



#67
N-Nitrosodiphenylamine
Concen: 17.375 ng/ul
RT: 15.981 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:169 Resp: 1015100
Ion Ratio Lower Upper
169 100
168 68.0 53.6 80.4
167 35.8 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 16.271 ng/ul
RT: 16.663 min Scan# 211
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:248 Resp: 332242
Ion Ratio Lower Upper
248 100
250 98.2 78.1 117.1
141 75.0 57.5 86.3



#69
Hexachlorobenzene
Concen: 16.199 ng/ul
RT: 16.745 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:284 Resp: 387801
Ion Ratio Lower Upper
284 100
142 33.9 26.9 40.3
249 29.9 24.2 36.4





#70

Atrazine

Concen: 16.504 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:200 Resp: 329051

Ion Ratio Lower Upper

200 100

173 27.2 21.0 31.4

215 45.4 38.1 57.1



#71

Pentachlorophenol

Concen: 14.370 ng/ul

RT: 17.104 min Scan# 2349

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:266 Resp: 200135


Ion Ratio Lower Upper

266 100

264 62.4 51.4 77.0

268 64.6 51.5 77.3





#72

Phenanthrene

Concen: 16.714 ng/ul

RT: 17.528 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:178 Resp: 1813853

Ion Ratio Lower Upper

178 100

179 15.1 12.7 19.1

176 18.9 15.6 23.4



#74

Anthracene

Concen: 16.734 ng/ul

RT: 17.616 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:178 Resp: 1825116


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.4 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 16.902 ng/uL

RT: 13.510 min Scan# 1738

Delta R.T. 0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:216 Resp: 561163

Ion Ratio Lower Upper

216 100

214 79.8 63.0 94.6

218 48.4 37.8 56.8



#76

Pentachlorobenzene

Concen: 17.454 ng/uL

RT: 15.010 min Scan# 1993

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:250 Resp: 491546


Ion Ratio Lower Upper

250 100

248 62.6 49.6 74.4

252 64.1 50.7 76.1









#77
Carbazole
Concen: 17.266 ng/ul
RT: 17.892 min Scan# 2484
Delta R.T. -0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020675


Tgt Ion:167 Resp: 1633271
Ion Ratio Lower Upper
167 100
166 21.6 17.2 25.8
139 12.6 10.5 15.7



#78
Di-n-butylphthalate
Concen: 16.689 ng/ul
RT: 18.410 min Scan# 2571
Delta R.T. 0.000 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:149 Resp: 1829127
Ion Ratio Lower Upper
149 100
150 9.4 7.4 11.2
104 5.4 4.2 6.4





#80

Fluoranthene

Concen: 16.524 ng/ul

RT: 19.486 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:202 Resp: 1959010

Ion Ratio Lower Upper

202 100

101 12.2 9.5 14.3

100 9.0 7.3 10.9



#82

Pyrene

Concen: 16.608 ng/ul

RT: 19.839 min Scan# 2814

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:202 Resp: 2080676


Ion Ratio Lower Upper

202 100

101 14.9 12.1 18.1

100 12.3 9.8 14.6





#83

Butylbenzylphthalate

Concen: 16.157 ng/ul

RT: 20.704 min Scan# 2961

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:149 Resp: 743486

Ion Ratio Lower Upper

149 100

91 72.6 58.1 87.1

206 20.9 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 18.475 ng/ul

RT: 21.498 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:252 Resp: 716258


Ion Ratio Lower Upper

252 100

254 63.7 51.0 76.4

126 11.2 9.5 14.3





#85

Benzo(a)anthracene

Concen: 16.514 ng/ul

RT: 21.563 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:228 Resp: 1985195

Ion Ratio Lower Upper

228 100

229 19.7 15.7 23.5

226 27.1 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.743 ng/ul

RT: 21.445 min Scan# 3087

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:149 Resp: 1102781


Ion Ratio Lower Upper

149 100

167 27.5 21.4 32.2

279 4.1 3.0 4.6









#87
 Chrysene
 Concen: 16.365 ng/ul
 RT: 21.616 min Scan# 3116
 Delta R.T. -0.006 min
 Lab File: BP016507.D
 Acq: 04 Aug 2023 17:29

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020675




Tgt Ion:228 Resp: 1854128


Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.1	36.1
229	19.5	15.4	23.2



#89
 Di-n-octyl phthalate
 Concen: 15.701 ng/ul
 RT: 22.433 min Scan# 3255
 Delta R.T. -0.006 min
 Lab File: BP016507.D
 Acq: 04 Aug 2023 17:29

Tgt Ion:149 Resp: 1802825





#90

Benzo(b)fluoranthene

Concen: 16.279 ng/ul

RT: 23.357 min Scan# 3413

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:252 Resp: 1980785

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 11.1 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 15.486 ng/ul

RT: 23.416 min Scan# 3422

Delta R.T. 0.000 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:252 Resp: 1896177


Ion Ratio Lower Upper

252 100

253 22.3 17.6 26.4

125 11.1 9.4 14.0





#93

Benzo(a)pyrene

Concen: 15.478 ng/ul

RT: 24.051 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020675





Tgt Ion:252 Resp: 1786198

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8

125 12.3 9.7 14.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.721 ng/ul

RT: 26.927 min Scan# 4019

Delta R.T. -0.006 min

Lab File: BP016507.D

Acq: 04 Aug 2023 17:29




Tgt Ion:276 Resp: 2383486


Ion Ratio Lower Upper

276 100

138 24.0 18.7 28.1

277 24.5 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 16.661 ng/ul
RT: 26.957 min Scan# 4025
Delta R.T. -0.006 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Instrument :
BNA_P
ClientSampleId :
SSTD020675

Tgt Ion:278 Resp: 1982448
Ion Ratio Lower Upper
278 100
139 17.8 14.4 21.6
279 23.6 19.2 28.8



#96
Benzo(g,h,i)perylene
Concen: 16.596 ng/ul
RT: 27.786 min Scan# 4165
Delta R.T. -0.012 min
Lab File: BP016507.D
Acq: 04 Aug 2023 17:29

Tgt Ion:276 Resp: 1910859
Ion Ratio Lower Upper
276 100
138 24.6 18.1 27.1
277 24.0 19.0 28.4

