

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080522\
 Data File : BP011287.D
 Acq On : 05 Aug 2022 20:07
 Operator : CG/JU
 Sample : N4026-09
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 05 23:03:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 02 23:52:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.599	152	90792	20.000	ng/u1	0.00
20) Naphthalene-d8	10.387	136	337247	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.269	164	175100	20.000	ng/u1	# 0.00
64) Phenanthrene-d10	17.039	188	327008	20.000	ng/u1	0.00
79) Chrysene-d12	21.163	240	488537	20.000	ng/u1	-0.01
88) Perylene-d12	23.474	264	542017	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.087	96	3174	2.827	ng/uL	0.00
4) Pyridine-d5	3.511	84	12503	4.128	ng/u1	0.00
7) Phenol-d5	6.781	99	22959	3.034	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.946	67	157907	28.922	ng/u1	0.00
11) 2-Chlorophenol-d4	7.128	132	92891	15.062	ng/u1	-0.01
15) 4-Methylphenol-d8	8.316	113	45509	7.310	ng/u1	-0.01
21) Nitrobenzene-d5	8.763	128	114368	47.148	ng/u1	0.00
24) 2-Nitrophenol-d4	9.481	143	67563	29.714	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.016	165	92107	21.243	ng/u1	0.00
31) 4-Chloroaniline-d4	10.540	131	41194	6.141	ng/u1	0.00
46) Dimethylphthalate-d6	13.692	166	321517	26.810	ng/u1	0.00
49) Acenaphthylene-d8	13.963	160	513372	34.023	ng/u1	0.00
54) 4-Nitrophenol-d4	14.504	143	4262	2.090	ng/u1	0.00
60) Fluorene-d10	15.275	176	449338	44.034	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	52006	34.250	ng/u1	-0.01
73) Anthracene-d10	17.139	188	900658	62.669	ng/u1	0.00
81) Pyrene-d10	19.398	212	1028963	41.657	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.321	264	1123878	41.610	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080522\
Data File : BP011287.D
Acq On : 05 Aug 2022 20:07
Operator : CG/JU
Sample : N4026-09
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Aug 05 23:03:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080222.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 02 23:52:37 2022
Response via : Initial Calibration

