

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080524\
 Data File : BP021370.D
 Acq On : 05 Aug 2024 13:03
 Operator : CG/JU
 Sample : P3300-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E1GD3

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/06/2024
 Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 06 03:58:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 29 12:46:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	145762	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.375	136	593020	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.257	164	382830	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.075	188	839227	20.000	ng/u1	-0.02
79) Chrysene-d12	21.533	240	840018	20.000	ng/u1	0.00
88) Perylene-d12	24.821	264	965479	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	14500	4.527	ng/uL	-0.02
4) Pyridine-d5	3.599	84	73580m	7.935	ng/u1	0.00
7) Phenol-d5	6.869	99	92383m	7.774	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.969	67	209990	29.158	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.175	132	259538	26.508	ng/u1	0.00
15) 4-Methylphenol-d8	8.363	113	184294	18.672	ng/u1	0.00
21) Nitrobenzene-d5	8.787	128	131765	28.607	ng/u1	0.00
24) 2-Nitrophenol-d4	9.493	143	141623	26.864	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.052	165	222623	23.353	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	274910	21.768	ng/u1	0.00
46) Dimethylphthalate-d6	13.669	166	804972	28.925	ng/u1	-0.01
49) Acenaphthylene-d8	13.945	160	919786	29.434	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.704	143	24172m	6.105	ng/u1	0.12
60) Fluorene-d10	15.269	176	713065	31.232	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.475	200	84594	16.660	ng/u1	0.01
73) Anthracene-d10	17.181	188	1167279	31.316	ng/u1	-0.01
81) Pyrene-d10	19.575	212	1455830	28.033	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.598	264	1573135	33.037	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080524\
Data File : BP021370.D
Acq On : 05 Aug 2024 13:03
Operator : CG/JU
Sample : P3300-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E1GD3

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/06/2024
Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 06 03:58:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 29 12:46:18 2024
Response via : Initial Calibration

