

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080723\
 Data File : BP016527.D
 Acq On : 07 Aug 2023 20:16
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020678

Quant Time: Aug 08 02:26:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 15:15:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.069	152	435153	20.000	ng/u1	0.00
20) Naphthalene-d8	10.899	136	1648042	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.716	164	853976	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.481	188	1654099	20.000	ng/u1	0.00
79) Chrysene-d12	21.575	240	1378330	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	1511573	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	100969	8.415	ng/uL	0.00
4) Pyridine-d5	3.840	84	647274	20.762	ng/u1	0.00
7) Phenol-d5	7.199	99	698644	20.530	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.399	67	438088	20.897	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	577475	21.010	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	531357	20.249	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	253065	21.227	ng/u1	0.00
24) 2-Nitrophenol-d4	9.981	143	243481	21.891	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.504	165	482482	21.319	ng/u1	0.00
31) 4-Chloroaniline-d4	11.052	131	698248	20.928	ng/u1	0.00
46) Dimethylphthalate-d6	14.128	166	1226687	20.772	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	1526598	21.468	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	214081	19.253	ng/u1	0.00
60) Fluorene-d10	15.710	176	1056341	20.832	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	133433	18.322	ng/u1	0.00
73) Anthracene-d10	17.581	188	1518486	21.264	ng/u1	0.00
81) Pyrene-d10	19.810	212	1624243	21.493	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.992	264	1540469	21.282	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.440	88	106364	8.274	ng/uL	96
5) Pyridine	3.858	79	664004	20.739	ng/u1	97
6) Benzaldehyde	7.216	77	440191	24.647	ng/u1	99
8) Phenol	7.228	94	740276	20.808	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.493	93	615362	20.877	ng/u1	98
12) 2-Chlorophenol	7.616	128	591488	20.702	ng/u1	98
13) 2-Methylphenol	8.499	108	531086	20.289	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.581	45	475724	13.477	ng/u1	99
16) Acetophenone	8.916	105	848621	20.882	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.887	70	398310	20.815	ng/u1	98
18) 4-Methylphenol	8.834	108	566119	20.338	ng/u1	100
19) Hexachloroethane	9.146	117	249414	20.527	ng/u1	99
22) Nitrobenzene	9.305	77	636572	21.185	ng/u1	100
23) Isophorone	9.822	82	1093458	20.942	ng/u1	99
25) 2-Nitrophenol	10.010	139	278560	21.554	ng/u1	97
26) 2,4-Dimethylphenol	10.046	107	584479	20.993	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.310	93	741197	20.904	ng/u1	100
29) 2,4-Dichlorophenol	10.528	162	488821	21.084	ng/u1	97
30) Naphthalene	10.952	128	1781503	20.862	ng/u1	100
32) 4-Chloroaniline	11.075	127	684620	20.842	ng/u1	99
33) Hexachlorobutadiene	11.187	225	326156	20.992	ng/u1	99
34) Caprolactam	11.881	113	125667	18.146	ng/u1	93
35) 4-Chloro-3-methylphenol	12.163	107	479211	20.993	ng/u1	95
36) 2-Methylnaphthalene	12.546	142	1120942	20.747	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080723\
 Data File : BP016527.D
 Acq On : 07 Aug 2023 20:16
 Operator : MA/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020678


Quant Time: Aug 08 02:26:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 04 15:15:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.769	142	1125497	20.895	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.893	216	571793	21.152	ng/ul	99
40) Hexachlorocyclopentadiene	12.851	237	332662	20.908	ng/ul	99
41) 2,4,6-Trichlorophenol	13.146	196	339320	21.946	ng/ul	98
42) 2,4,5-Trichlorophenol	13.210	196	356567	21.241	ng/ul	100
43) 1,1'-Biphenyl	13.551	154	1398475	21.186	ng/ul	99
44) 2-Chloronaphthalene	13.598	162	1107974	21.096	ng/ul	99
45) 2-Nitroaniline	13.822	65	261770	21.188	ng/ul	99
47) Dimethylphthalate	14.175	163	1239856	20.736	ng/ul	99
48) 2,6-Dinitrotoluene	14.310	165	222553	21.164	ng/ul	98
50) Acenaphthylene	14.445	152	1761423	21.322	ng/ul	99
51) 3-Nitroaniline	14.645	138	237570	20.539	ng/ul	98
52) Acenaphthene	14.781	153	1134827	20.900	ng/ul	98
53) 2,4-Dinitrophenol	14.845	184	78772	15.647	ng/ul	98
55) 4-Nitrophenol	14.922	109	165316	19.993	ng/ul	96
56) Dibenzofuran	15.116	168	1539010	20.902	ng/ul	99
57) 2,4-Dinitrotoluene	15.087	165	309095	20.842	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.328	232	274086	21.119	ng/ul	96
59) Diethylphthalate	15.528	149	1184432	20.644	ng/ul	100
61) Fluorene	15.763	166	1216053	20.894	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.757	204	592622	20.686	ng/ul	97
63) 4-Nitroaniline	15.810	138	216496	21.215	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.839	198	154470	18.880	ng/ul	94
67) N-Nitrosodiphenylamine	15.975	169	992617	21.377	ng/ul	100
68) 4-Bromophenyl-phenylether	16.657	248	345024	21.260	ng/ul	97
69) Hexachlorobenzene	16.739	284	402362	21.147	ng/ul	98
70) Atrazine	16.928	200	322556	20.356	ng/ul	98
71) Pentachlorophenol	17.104	266	218628	19.751	ng/ul	99
72) Phenanthrene	17.522	178	1792259	20.779	ng/ul	100
74) Anthracene	17.616	178	1824915	21.052	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.504	216	582698	22.082	ng/uL	99
76) Pentachlorobenzene	15.004	250	488451	21.823	ng/uL	99
77) Carbazole	17.892	167	1591999	21.175	ng/ul	99
78) Di-n-butylphthalate	18.410	149	1807491	20.750	ng/ul	99
80) Fluoranthene	19.480	202	1934014	21.460	ng/ul	98
82) Pyrene	19.839	202	2044269	21.466	ng/ul	98
83) Butylbenzylphthalate	20.698	149	731044	20.900	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.492	252	596394	20.237	ng/ul	98
85) Benzo(a)anthracene	21.557	228	1917258	20.981	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.439	149	1032466	20.622	ng/ul	99
87) Chrysene	21.616	228	1811303	21.031	ng/ul	99
89) Di-n-octyl phthalate	22.433	149	1732783	20.586	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	1905313	21.360	ng/ul	99
91) Benzo(k)fluoranthene	23.410	252	1869164	20.824	ng/ul	100
93) Benzo(a)pyrene	24.051	252	1793716	21.203	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.921	276	2233519	21.374	ng/ul	99
95) Dibenzo(a,h)anthracene	26.951	278	1864374	21.374	ng/ul	99
96) Benzo(g,h,i)perylene	27.786	276	1793086	21.243	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 08 02:26:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 04 15:15:50 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.274 ng/uL
RT: 3.440 min Scan# 20
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion: 88 Resp: 106364
Ion Ratio Lower Upper
88 100
43 23.7 21.0 31.4
58 65.7 55.0 82.4



#5
Pyridine
Concen: 20.739 ng/uL
RT: 3.858 min Scan# 97
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion: 79 Resp: 664004
Ion Ratio Lower Upper
79 100
52 65.7 50.2 75.2
51 30.7 23.8 35.6









#6
Benzaldehyde
Concen: 24.647 ng/ul
RT: 7.216 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion: 77 Resp: 440191
Ion Ratio Lower Upper
77 100
105 98.7 79.6 119.4
106 95.1 74.9 112.3



#8
Phenol
Concen: 20.808 ng/ul
RT: 7.228 min Scan# 670
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion: 94 Resp: 740276
Ion Ratio Lower Upper
94 100
65 26.1 20.3 30.5
66 33.3 26.6 39.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.877 ng/ul
RT: 7.493 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion: 93 Resp: 615362
Ion Ratio Lower Upper
93 100
63 69.8 57.0 85.4
95 32.0 26.2 39.2



#12
2-Chlorophenol
Concen: 20.702 ng/ul
RT: 7.616 min Scan# 736
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:128 Resp: 591488
Ion Ratio Lower Upper
128 100
64 44.8 34.9 52.3
130 32.2 24.6 36.8





#13

2-Methylphenol

Concen: 20.289 ng/ul

RT: 8.499 min Scan# 886

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :





SSTD020678

Tgt Ion:108 Resp: 531086

Ion Ratio Lower Upper

108 100

107 88.0 71.0 106.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.477 ng/ul

RT: 8.581 min Scan# 900

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion: 45 Resp: 475724


Ion Ratio Lower Upper

45 100

77 16.5 12.5 18.7

79 12.1 9.8 14.8









#16
Acetophenone
Concen: 20.882 ng/ul
RT: 8.916 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:105 Resp: 848621
Ion Ratio Lower Upper
105 100
77 75.4 59.6 89.4
51 22.6 18.5 27.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.815 ng/ul
RT: 8.887 min Scan# 952
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion: 70 Resp: 398310
Ion Ratio Lower Upper
70 100
42 48.6 39.9 59.9
101 9.8 7.5 11.3
130 22.2 17.0 25.4









#18
4-Methylphenol
Concen: 20.338 ng/ul
RT: 8.834 min Scan# 943
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:108 Resp: 566119
Ion Ratio Lower Upper
108 100
107 110.9 89.1 133.7



#19
Hexachloroethane
Concen: 20.527 ng/ul
RT: 9.146 min Scan# 996
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:117 Resp: 249414
Ion Ratio Lower Upper
117 100
201 98.0 77.6 116.4
199 59.6 48.2 72.2









#22
Nitrobenzene
Concen: 21.185 ng/ul
RT: 9.305 min Scan# 1023
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion: 77 Resp: 636572
Ion Ratio Lower Upper
77 100
123 44.4 35.3 52.9
65 12.7 10.2 15.4



#23
Isophorone
Concen: 20.942 ng/ul
RT: 9.822 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion: 82 Resp: 1093458
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 17.6 13.5 20.3









#25
2-Nitrophenol
Concen: 21.554 ng/ul
RT: 10.010 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:139 Resp: 278560
Ion Ratio Lower Upper
139 100
65 48.8 37.6 56.4
109 30.2 22.9 34.3



#26
2,4-Dimethylphenol
Concen: 20.993 ng/ul
RT: 10.046 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:107 Resp: 584479
Ion Ratio Lower Upper
107 100
121 51.0 40.5 60.7
122 88.0 70.2 105.4









#27
Bis(2-Chloroethoxy)methane
Concen: 20.904 ng/ul
RT: 10.310 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion: 93 Resp: 741197
Ion Ratio Lower Upper
93 100
95 33.0 26.5 39.7
123 10.8 8.8 13.2



#29
2,4-Dichlorophenol
Concen: 21.084 ng/ul
RT: 10.528 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion: 162 Resp: 488821
Ion Ratio Lower Upper
162 100
164 63.9 53.4 80.0
98 36.5 29.8 44.8









#30
Naphthalene
Concen: 20.862 ng/ul
RT: 10.952 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:128 Resp: 1781503
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 12.7 10.3 15.5



#32
4-Chloroaniline
Concen: 20.842 ng/ul
RT: 11.075 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:127 Resp: 684620
Ion Ratio Lower Upper
127 100
129 31.6 25.8 38.6





#33

Hexachlorobutadiene

Concen: 20.992 ng/ul

RT: 11.187 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:225 Resp: 326156

Ion Ratio Lower Upper

225 100

223 62.9 51.4 77.2

227 63.9 50.4 75.6



#34

Caprolactam

Concen: 18.146 ng/ul

RT: 11.881 min Scan# 1461

Delta R.T. 0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:113 Resp: 125667


Ion Ratio Lower Upper

113 100

55 148.5 127.0 190.4

56 111.1 93.6 140.4









#35
4-Chloro-3-methylphenol
Concen: 20.993 ng/ul
RT: 12.163 min Scan# 1509
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:107 Resp: 479211
Ion Ratio Lower Upper
107 100
144 24.9 20.5 30.7
142 76.1 65.2 97.8



#36
2-Methylnaphthalene
Concen: 20.747 ng/ul
RT: 12.546 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:142 Resp: 1120942
Ion Ratio Lower Upper
142 100
141 89.0 70.9 106.3





#37

1-Methylnaphthalene

Concen: 20.895 ng/ul

RT: 12.769 min Scan# 1612

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:142 Resp: 1125497

Ion Ratio Lower Upper

142 100

141 92.2 73.5 110.3

116 3.7 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.152 ng/ul

RT: 12.893 min Scan# 1633

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Tgt Ion:216 Resp: 571793




Ion Ratio Lower Upper


216 100

214 80.0 63.2 94.8

179 21.7 17.4 26.2

108 18.4 14.8 22.2









#40
Hexachlorocyclopentadiene
Concen: 20.908 ng/ul
RT: 12.851 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:237 Resp: 332662
Ion Ratio Lower Upper
237 100
235 62.5 50.3 75.5
272 14.4 10.2 15.4



#41
2,4,6-Trichlorophenol
Concen: 21.946 ng/ul
RT: 13.146 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:196 Resp: 339320
Ion Ratio Lower Upper
196 100
198 94.6 77.1 115.7
200 30.9 25.0 37.4





#42

2,4,5-Trichlorophenol

Concen: 21.241 ng/ul

RT: 13.210 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:196 Resp: 356567

Ion Ratio Lower Upper

196 100

198 95.8 76.8 115.2

200 31.7 24.8 37.2



#43

1,1'-Biphenyl

Concen: 21.186 ng/ul

RT: 13.551 min Scan# 1745

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:154 Resp: 1398475


Ion Ratio Lower Upper

154 100

153 40.1 31.9 47.9

76 13.6 11.2 16.8









#44
2-Chloronaphthalene
Concen: 21.096 ng/ul
RT: 13.598 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:162 Resp: 1107974
Ion Ratio Lower Upper
162 100
164 32.1 26.6 39.8
127 39.0 31.1 46.7



#45
2-Nitroaniline
Concen: 21.188 ng/ul
RT: 13.822 min Scan# 1791
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion: 65 Resp: 261770
Ion Ratio Lower Upper
65 100
92 73.5 57.1 85.7
138 108.9 86.8 130.2









#47
Dimethylphthalate
Concen: 20.736 ng/ul
RT: 14.175 min Scan# 1851
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:163 Resp: 1239856
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 9.7 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.164 ng/ul
RT: 14.310 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:165 Resp: 222553
Ion Ratio Lower Upper
165 100
89 60.5 50.1 75.1
121 22.4 17.9 26.9





#50

Acenaphthylene

Concen: 21.322 ng/ul

RT: 14.445 min Scan# 1897

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:152 Resp: 1761423

Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.8

153 13.1 10.7 16.1



#51

3-Nitroaniline

Concen: 20.539 ng/ul

RT: 14.645 min Scan# 1931

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:138 Resp: 237570


Ion Ratio Lower Upper

138 100

108 12.6 9.7 14.5

92 118.9 92.9 139.3









#52
Acenaphthene
Concen: 20.900 ng/ul
RT: 14.781 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:153 Resp: 1134827
Ion Ratio Lower Upper
153 100
152 48.7 37.8 56.8
154 89.6 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 15.647 ng/ul
RT: 14.845 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:184 Resp: 78772
Ion Ratio Lower Upper
184 100
63 65.5 51.4 77.0
154 62.0 50.6 75.8









#55
4-Nitrophenol
Concen: 19.993 ng/ul
RT: 14.922 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:109 Resp: 165316
Ion Ratio Lower Upper
109 100
139 139.9 108.4 162.6
65 131.8 101.2 151.8



#56
Dibenzofuran
Concen: 20.902 ng/ul
RT: 15.116 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:168 Resp: 1539010
Ion Ratio Lower Upper
168 100
139 36.6 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 20.842 ng/ul
RT: 15.087 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:165 Resp: 309095
Ion Ratio Lower Upper
165 100
63 46.0 33.5 50.3
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.119 ng/ul
RT: 15.328 min Scan# 2047
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:232 Resp: 274086
Ion Ratio Lower Upper
232 100
131 43.3 37.0 55.4
130 2.3 1.9 2.9
166 29.9 25.1 37.7









#59
Diethylphthalate
Concen: 20.644 ng/ul
RT: 15.528 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:149 Resp: 1184432
Ion Ratio Lower Upper
149 100
177 22.1 17.7 26.5
150 12.4 10.0 15.0



#61
Fluorene
Concen: 20.894 ng/ul
RT: 15.763 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:166 Resp: 1216053
Ion Ratio Lower Upper
166 100
165 97.7 78.5 117.7
167 13.4 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 20.686 ng/ul
RT: 15.757 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:204 Resp: 592622
Ion Ratio Lower Upper
204 100
206 33.2 26.1 39.1
141 57.4 48.6 72.8



#63
4-Nitroaniline
Concen: 21.215 ng/ul
RT: 15.810 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:138 Resp: 216496
Ion Ratio Lower Upper
138 100
92 55.2 43.4 65.0
108 84.0 70.2 105.4









#66
4,6-Dinitro-2-methylphenol
Concen: 18.880 ng/ul
RT: 15.839 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:198 Resp: 154470
Ion Ratio Lower Upper
198 100
182 5.0 4.2 6.4
77 26.4 23.8 35.8



#67
N-Nitrosodiphenylamine
Concen: 21.377 ng/ul
RT: 15.975 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:169 Resp: 992617
Ion Ratio Lower Upper
169 100
168 67.0 53.6 80.4
167 36.7 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 21.260 ng/ul
RT: 16.657 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:248 Resp: 345024
Ion Ratio Lower Upper
248 100
250 96.4 78.1 117.1
141 76.0 57.5 86.3



#69
Hexachlorobenzene
Concen: 21.147 ng/ul
RT: 16.739 min Scan# 2287
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:284 Resp: 402362
Ion Ratio Lower Upper
284 100
142 35.7 26.9 40.3
249 30.9 24.2 36.4





#70

Atrazine

Concen: 20.356 ng/ul

RT: 16.928 min Scan# 2319

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:200 Resp: 322556

Ion Ratio Lower Upper

200 100

173 26.7 21.0 31.4

215 49.2 38.1 57.1



#71

Pentachlorophenol

Concen: 19.751 ng/ul

RT: 17.104 min Scan# 2349

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:266 Resp: 218628


Ion Ratio Lower Upper

266 100

264 63.2 51.4 77.0

268 64.9 51.5 77.3





#72

Phenanthrene

Concen: 20.779 ng/ul

RT: 17.522 min Scan# 2421

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:178 Resp: 1792259

Ion Ratio Lower Upper

178 100

179 15.7 12.7 19.1

176 19.3 15.6 23.4



#74

Anthracene

Concen: 21.052 ng/ul

RT: 17.616 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:178 Resp: 1824915


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.2 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.082 ng/uL

RT: 13.504 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:216 Resp: 582698

Ion Ratio Lower Upper

216 100

214 78.9 63.0 94.6

218 48.8 37.8 56.8



#76

Pentachlorobenzene

Concen: 21.823 ng/uL

RT: 15.004 min Scan# 1992

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:250 Resp: 488451


Ion Ratio Lower Upper

250 100

248 61.9 49.6 74.4

252 62.1 50.7 76.1









#77
 Carbazole
 Concen: 21.175 ng/ul
 RT: 17.892 min Scan# 2483
 Delta R.T. -0.006 min
 Lab File: BP016527.D
 Acq: 07 Aug 2023 20:16




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020678


Tgt Ion:167 Resp: 1591999
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.2 25.8
 139 12.7 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 20.750 ng/ul
 RT: 18.410 min Scan# 2571
 Delta R.T. 0.000 min
 Lab File: BP016527.D
 Acq: 07 Aug 2023 20:16

Tgt Ion:149 Resp: 1807491
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 5.0 4.2 6.4









#80
Fluoranthene
Concen: 21.460 ng/ul
RT: 19.480 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16




Instrument :
BNA_P
ClientSampleId :
SSTD020678


Tgt Ion:202 Resp: 1934014
Ion Ratio Lower Upper
202 100
101 12.8 9.5 14.3
100 9.5 7.3 10.9



#82
Pyrene
Concen: 21.466 ng/ul
RT: 19.839 min Scan# 2814
Delta R.T. 0.000 min
Lab File: BP016527.D
Acq: 07 Aug 2023 20:16

Tgt Ion:202 Resp: 2044269
Ion Ratio Lower Upper
202 100
101 14.3 12.1 18.1
100 11.6 9.8 14.6





#83

Butylbenzylphthalate

Concen: 20.900 ng/ul

RT: 20.698 min Scan# 2961

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:149 Resp: 731044

Ion Ratio Lower Upper

149 100

91 75.2 58.1 87.1

206 20.7 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 20.237 ng/ul

RT: 21.492 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:252 Resp: 596394


Ion Ratio Lower Upper

252 100

254 65.2 51.0 76.4

126 12.4 9.5 14.3





#85

Benzo(a)anthracene

Concen: 20.981 ng/ul

RT: 21.557 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:228 Resp: 1917258

Ion Ratio Lower Upper

228 100

229 19.9 15.7 23.5

226 27.2 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.622 ng/ul

RT: 21.439 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:149 Resp: 1032466


Ion Ratio Lower Upper

149 100

167 27.1 21.4 32.2

279 3.7 3.0 4.6





#87

Chrysene

Concen: 21.031 ng/ul

RT: 21.616 min Scan# 3116

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:228 Resp: 1811303

Ion Ratio Lower Upper

228 100

226 29.9 24.1 36.1

229 19.7 15.4 23.2



#89

Di-n-octyl phthalate

Concen: 20.586 ng/ul




RT: 22.433 min Scan# 3255


Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Tgt Ion:149 Resp: 1732783





#90

Benzo(b)fluoranthene

Concen: 21.360 ng/ul

RT: 23.357 min Scan# 3413

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:252 Resp: 1905313

Ion Ratio Lower Upper

252 100

253 22.2 17.4 26.0

125 10.9 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 20.824 ng/ul

RT: 23.410 min Scan# 3421

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:252 Resp: 1869164


Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 11.6 9.4 14.0





#93

Benzo(a)pyrene

Concen: 21.203 ng/ul

RT: 24.051 min Scan# 3530

Delta R.T. -0.006 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:252 Resp: 1793716

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 11.8 9.7 14.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.374 ng/ul

RT: 26.921 min Scan# 4018

Delta R.T. -0.012 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16




Tgt Ion:276 Resp: 2233519


Ion Ratio Lower Upper

276 100

138 24.3 18.7 28.1

277 24.1 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 21.374 ng/ul

RT: 26.951 min Scan# 4024

Delta R.T. -0.012 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Instrument :

BNA_P

ClientSampleId :

SSTD020678





Tgt Ion:278 Resp: 1864374

Ion Ratio Lower Upper

278 100

139 17.3 14.4 21.6

279 24.5 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 21.243 ng/ul

RT: 27.786 min Scan# 4165

Delta R.T. -0.012 min

Lab File: BP016527.D

Acq: 07 Aug 2023 20:16

Tgt Ion:276 Resp: 1793086

Ion Ratio Lower Upper

276 100

138 23.1 18.1 27.1

277 24.6 19.0 28.4

