

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
 Data File : BP006627.D
 Acq On : 10 Aug 2021 17:17
 Operator : CG/JU
 Sample : M3208-11
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C00F1

Quant Time: Aug 10 17:48:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 10 15:09:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	128743	20.000	ng/ul	0.00
20) Naphthalene-d8	10.505	136	560295	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.357	164	328628	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.116	188	658484	20.000	ng/ul	0.00
79) Chrysene-d12	21.216	240	646474	20.000	ng/ul	0.00
88) Perylene-d12	23.527	264	690657	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	11671	3.066	ng/uL	0.00
4) Pyridine-d5	3.646	84	60967	5.572	ng/ul	0.01
7) Phenol-d5	6.899	99	192300	14.858	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.064	67	130589	16.418	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	153308	15.905	ng/ul	0.00
15) 4-Methylphenol-d8	8.434	113	114530	11.077	ng/ul	0.00
21) Nitrobenzene-d5	8.875	128	80448	16.915	ng/ul	0.00
24) 2-Nitrophenol-d4	9.593	143	81318	16.340	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	130392	15.508	ng/ul	0.00
31) 4-Chloroaniline-d4	10.652	131	160057	11.712	ng/ul	0.00
46) Dimethylphthalate-d6	13.781	166	464101	18.415	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	556432	18.000	ng/ul	0.00
54) 4-Nitrophenol-d4	14.569	143	69332	13.170	ng/ul	0.00
60) Fluorene-d10	15.357	176	376558	18.371	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	52386	12.036	ng/ul	0.00
73) Anthracene-d10	17.216	188	550906	17.682	ng/ul	0.00
81) Pyrene-d10	19.457	212	644163	18.974	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.380	264	592243	15.902	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
Data File : BP006627.D
Acq On : 10 Aug 2021 17:17
Operator : CG/JU
Sample : M3208-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C00F1

Quant Time: Aug 10 17:48:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 10 15:09:30 2021
Response via : Initial Calibration

