

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
 Data File : BP006631.D
 Acq On : 10 Aug 2021 19:34
 Operator : CG/JU
 Sample : M3208-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C00E2

Quant Time: Aug 11 01:16:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 10 15:09:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	146271	20.000	ng/ul	0.00
20) Naphthalene-d8	10.504	136	639160	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.357	164	372758	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.110	188	734500	20.000	ng/ul	0.00
79) Chrysene-d12	21.215	240	712745	20.000	ng/ul	0.00
88) Perylene-d12	23.527	264	752897	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	13909	3.216	ng/uL	0.00
4) Pyridine-d5	3.640	84	131482	10.576	ng/ul	0.00
7) Phenol-d5	6.898	99	250404	17.029	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	164786	18.235	ng/ul	0.00
11) 2-Chlorophenol-d4	7.251	132	198539	18.129	ng/ul	0.00
15) 4-Methylphenol-d8	8.434	113	205839	17.522	ng/ul	0.00
21) Nitrobenzene-d5	8.875	128	101976	18.796	ng/ul	0.00
24) 2-Nitrophenol-d4	9.592	143	106577	18.773	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	169428	17.664	ng/ul	0.00
31) 4-Chloroaniline-d4	10.651	131	269228	17.269	ng/ul	0.00
46) Dimethylphthalate-d6	13.780	166	595886	20.845	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	701016	19.992	ng/ul	0.00
54) 4-Nitrophenol-d4	14.569	143	88616	14.841	ng/ul	0.00
60) Fluorene-d10	15.357	176	476445	20.493	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.480	200	64527	13.291	ng/ul	0.00
73) Anthracene-d10	17.210	188	715301	20.583	ng/ul	0.00
81) Pyrene-d10	19.457	212	788314	21.062	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.374	264	764256	18.824	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
Data File : BP006631.D
Acq On : 10 Aug 2021 19:34
Operator : CG/JU
Sample : M3208-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C00E2

Quant Time: Aug 11 01:16:20 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 10 15:09:30 2021
Response via : Initial Calibration

