

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
 Data File : BP006650.D
 Acq On : 11 Aug 2021 11:31
 Operator : CG/JU
 Sample : PB138246BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS246

Manual Integrations
 APPROVED

mohammad
 8/11/2021 9:29:30 PM

Quant Time: Aug 11 11:45:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 10 15:09:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	150426	20.000	ng/u1	0.00
20) Naphthalene-d8	10.505	136	658040	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.357	164	376857	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.110	188	741360	20.000	ng/u1	0.00
79) Chrysene-d12	21.216	240	705553	20.000	ng/u1	0.00
88) Perylene-d12	23.533	264	729292	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	31628	7.112	ng/uL	0.00
4) Pyridine-d5	3.628	84	455616	35.637	ng/u1	0.00
7) Phenol-d5	6.899	99	568218	37.575	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.064	67	351675	37.841	ng/u1	0.00
11) 2-Chlorophenol-d4	7.252	132	438394	38.925	ng/u1	0.00
15) 4-Methylphenol-d8	8.434	113	446884	36.990	ng/u1	0.00
21) Nitrobenzene-d5	8.875	128	221277	39.614	ng/u1	0.00
24) 2-Nitrophenol-d4	9.593	143	230686	39.468	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	381475	38.630	ng/u1	0.00
31) 4-Chloroaniline-d4	10.646	131	580329	36.156	ng/u1	0.00
46) Dimethylphthalate-d6	13.781	166	1149556	39.776	ng/u1	0.00
49) Acenaphthylene-d8	14.045	160	1432538	40.410	ng/u1	0.00
54) 4-Nitrophenol-d4	14.581	143	244578	40.514	ng/u1	0.00
60) Fluorene-d10	15.357	176	937052	39.866	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.487	200	180362	36.806	ng/u1	0.00
73) Anthracene-d10	17.216	188	1413348	40.292	ng/u1	0.00
81) Pyrene-d10	19.457	212	1577427	42.574	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.380	264	1532199	38.960	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	67153	14.861	ng/uL	98
5) Pyridine	3.652	79	473367	35.766	ng/u1	97
6) Benzaldehyde	6.869	77	382917	51.500	ng/u1	98
8) Phenol	6.928	94	594349	38.484	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.158	93	457290	38.415	ng/u1	99
12) 2-Chlorophenol	7.281	128	456794	40.222	ng/u1	98
13) 2-Methylphenol	8.169	108	430509	37.887	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.246	45	531678	38.226	ng/u1	99
16) Acetophenone	8.546	105	708240	37.240	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.534	70	390773	38.200	ng/u1	98
18) 4-Methylphenol	8.499	108	463274	37.501	ng/u1	95
19) Hexachloroethane	8.781	117	197550	39.973	ng/u1	98
22) Nitrobenzene	8.916	77	588714	40.257	ng/u1	99
23) Isophorone	9.446	82	1054739	38.677	ng/u1	100
25) 2-Nitrophenol	9.622	139	248873	40.371	ng/u1	98
26) 2,4-Dimethylphenol	9.693	107	459868	34.724	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.934	93	613619	38.817	ng/u1	100
29) 2,4-Dichlorophenol	10.157	162	384221	40.105	ng/u1	99
30) Naphthalene	10.552	128	1413967	39.592	ng/u1	99
32) 4-Chloroaniline	10.675	127	577049	36.638	ng/u1	99
33) Hexachlorobutadiene	10.834	225	226849	40.160	ng/u1	99
34) Caprolactam	11.469	113	149537m	38.987	ng/u1	
35) 4-Chloro-3-methylphenol	11.804	107	494502	40.350	ng/u1	97
36) 2-Methylnaphthalene	12.169	142	960444	39.137	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
 Data File : BP006650.D
 Acq On : 11 Aug 2021 11:31
 Operator : CG/JU
 Sample : PB138246BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS246

Manual Integrations
 APPROVED

mohammad
 8/11/2021 9:29:30 PM

Quant Time: Aug 11 11:45:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 10 15:09:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.387	142	935568	38.308	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.540	216	411971	40.564	ng/ul	98
40) Hexachlorocyclopentadiene	12.516	237	189154	28.389	ng/ul	98
41) 2,4,6-Trichlorophenol	12.787	196	285928	40.995	ng/ul	100
42) 2,4,5-Trichlorophenol	12.857	196	312933	42.142	ng/ul	99
43) 1,1'-Biphenyl	13.193	154	1177503	40.192	ng/ul	99
44) 2-Chloronaphthalene	13.228	162	886291	39.875	ng/ul	100
45) 2-Nitroaniline	13.446	65	344631	42.997	ng/ul	100
47) Dimethylphthalate	13.828	163	1153301	40.345	ng/ul	100
48) 2,6-Dinitrotoluene	13.945	165	233848	42.384	ng/ul	91
50) Acenaphthylene	14.081	152	1386592	40.012	ng/ul	99
51) 3-Nitroaniline	14.275	138	279122	44.176	ng/ul	96
52) Acenaphthene	14.422	153	986647	39.930	ng/ul	99
53) 2,4-Dinitrophenol	14.487	184	129501	36.289	ng/ul	98
55) 4-Nitrophenol	14.593	109	230000	42.487	ng/ul	97
56) Dibenzofuran	14.757	168	1335437	40.420	ng/ul	96
57) 2,4-Dinitrotoluene	14.740	165	349866	43.103	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.987	232	253042	41.050	ng/ul	94
59) Diethylphthalate	15.198	149	1257570	41.388	ng/ul	99
61) Fluorene	15.410	166	1112088	40.369	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.410	204	501699	40.229	ng/ul	98
63) 4-Nitroaniline	15.445	138	300234	50.744	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.504	198	184737	37.712	ng/ul	93
67) N-Nitrosodiphenylamine	15.628	169	935080	41.063	ng/ul	99
68) 4-Bromophenyl-phenylether	16.310	248	299028	40.897	ng/ul	97
69) Hexachlorobenzene	16.416	284	338715	40.991	ng/ul	98
70) Atrazine	16.592	200	324281	40.142	ng/ul	99
71) Pentachlorophenol	16.763	266	215735	39.859	ng/ul	99
72) Phenanthrene	17.157	178	1717115	40.999	ng/ul	99
74) Anthracene	17.251	178	1717139	40.375	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.151	216	410897	40.759	ng/uL	98
76) Pentachlorobenzene	14.675	250	405149	40.790	ng/uL	98
77) Carbazole	17.528	167	1625426	40.919	ng/ul	100
78) Di-n-butylphthalate	18.092	149	2126883	41.411	ng/ul	100
80) Fluoranthene	19.133	202	1941930	41.700	ng/ul	100
82) Pyrene	19.486	202	1988187	41.334	ng/ul	98
83) Butylbenzylphthalate	20.375	149	983415	41.355	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.139	252	606711	37.683	ng/ul	100
85) Benzo(a)anthracene	21.198	228	1880716	40.924	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.139	149	1503715	41.037	ng/ul	100
87) Chrysene	21.251	228	1808133	40.807	ng/ul	100
89) Di-n-octyl phthalate	22.045	149	2571597	40.793	ng/ul	100
90) Benzo(b)fluoranthene	22.827	252	1963673	41.449	ng/ul	99
91) Benzo(k)fluoranthene	22.874	252	1891576	41.240	ng/ul	99
93) Benzo(a)pyrene	23.433	252	1753687	40.862	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.921	276	2286062	40.738	ng/ul	98
95) Dibenzo(a,h)anthracene	25.945	278	1919468	40.722	ng/ul	99
96) Benzo(g,h,i)perylene	26.662	276	1898831	40.754	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081021\
 Data File : BP006650.D
 Acq On : 11 Aug 2021 11:31
 Operator : CG/JU
 Sample : PB138246BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS246

Manual Integrations
 APPROVED

mohammad
 8/11/2021 9:29:30 PM

Quant Time: Aug 11 11:45:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP081021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 10 15:09:30 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 14.861 ng/uL
RT: 3.258 min Scan# 29
Delta R.T. 0.001 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion: 88 Resp: 67153
Ion Ratio Lower Upper
88 100
43 25.2 19.5 29.3
58 64.8 53.1 79.7

Instrument :
BNA_P
ClientSampled :
SLCS246




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#5
Pyridine
Concen: 35.766 ng/uL
RT: 3.652 min Scan# 96
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion: 79 Resp: 473367
Ion Ratio Lower Upper
79 100
52 56.8 48.1 72.1
51 28.3 23.1 34.7





#6
Benzaldehyde
Concen: 51.500 ng/ul
RT: 6.869 min Scan# 643
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246



Tgt Ion: 77 Resp: 382917
Ion Ratio Lower Upper
77 100
105 92.9 72.6 109.0
106 87.2 68.7 103.1




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#8
Phenol
Concen: 38.484 ng/ul
RT: 6.928 min Scan# 653
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion: 94 Resp: 594349
Ion Ratio Lower Upper
94 100
65 29.6 23.0 34.6
66 39.3 31.2 46.8





#10
Bis(2-Chloroethyl)ether
Concen: 38.415 ng/ul
RT: 7.158 min Scan# 692
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Instrument :
BNA_P
ClientSampleId :
SLCS246






Tgt Ion: 93 Resp: 457290


Ion	Ratio	Lower	Upper
93	100		
63	75.9	61.1	91.7
95	32.1	26.1	39.1

Manual Integrations
APPROVED

mohammad
8/11/2021 9:29:30 PM






#12
2-Chlorophenol
Concen: 40.222 ng/ul
RT: 7.281 min Scan# 713
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31



Tgt Ion: 128 Resp: 456794


Ion	Ratio	Lower	Upper
128	100		
64	50.7	40.2	60.4
130	31.0	26.5	39.7





#13
2-Methylphenol
Concen: 37.887 ng/ul
RT: 8.169 min Scan# 864
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:108 Resp: 430509
Ion Ratio Lower Upper
108 100
107 91.6 74.2 111.2

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#14
2,2'-oxybis(1-Chloropropane)
Concen: 38.226 ng/ul
RT: 8.246 min Scan# 877
Delta R.T. -0.012 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion: 45 Resp: 531678
Ion Ratio Lower Upper
45 100
77 18.1 14.5 21.7
79 13.3 11.6 17.4





#16
Acetophenone
Concen: 37.240 ng/ul
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:105 Resp: 708240
Ion Ratio Lower Upper
105 100
77 83.9 67.0 100.6
51 24.0 19.4 29.0

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#17
N-Nitroso-di-n-propylamine
Concen: 38.200 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion: 70 Resp: 390773
Ion Ratio Lower Upper
70 100
42 51.6 40.3 60.5
101 9.5 7.0 10.4
130 18.3 15.2 22.8





#18
4-Methylphenol
Concen: 37.501 ng/ul
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:108 Resp: 463274
Ion Ratio Lower Upper
108 100
107 111.0 93.0 139.4

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#19
Hexachloroethane
Concen: 39.973 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:117 Resp: 197550
Ion Ratio Lower Upper
117 100
201 82.0 67.7 101.5
199 51.7 40.7 61.1





#22
Nitrobenzene
Concen: 40.257 ng/ul
RT: 8.916 min Scan# 991
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion: 77 Resp: 588714
Ion Ratio Lower Upper
77 100
123 39.7 32.5 48.7
65 13.5 10.6 15.8

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#23
Isophorone
Concen: 38.677 ng/ul
RT: 9.446 min Scan# 1081
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion: 82 Resp: 1054739
Ion Ratio Lower Upper
82 100
95 7.3 5.9 8.9
138 15.0 12.2 18.2









#25
2-Nitrophenol
Concen: 40.371 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31





Tgt Ion:139 Resp: 248873
Ion Ratio Lower Upper
139 100
65 60.4 46.6 70.0
109 34.8 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 34.724 ng/ul
RT: 9.693 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:107 Resp: 459868
Ion Ratio Lower Upper
107 100
121 48.1 37.8 56.8
122 81.4 64.3 96.5



Instrument :
BNA_P
ClientSampleId :
SLCS246


Manual Integrations
APPROVED

mohammad
8/11/2021 9:29:30 PM



#27
Bis(2-Chloroethoxy)methane
Concen: 38.817 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246



Tgt Ion: 93 Resp: 613619
Ion Ratio Lower Upper
93 100
95 31.5 25.4 38.0
123 10.5 8.3 12.5




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#29
2,4-Dichlorophenol
Concen: 40.105 ng/ul
RT: 10.157 min Scan# 1202
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion: 162 Resp: 384221
Ion Ratio Lower Upper
162 100
164 65.0 52.9 79.3
98 43.1 34.5 51.7





#30
Naphthalene
Concen: 39.592 ng/ul
RT: 10.552 min Scan# 1269
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:128 Resp: 1413967
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.4
127 13.3 10.7 16.1

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#32
4-Chloroaniline
Concen: 36.638 ng/ul
RT: 10.675 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:127 Resp: 577049
Ion Ratio Lower Upper
127 100
129 32.4 25.7 38.5







#33
Hexachlorobutadiene
Concen: 40.160 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:225 Resp: 226849
Ion Ratio Lower Upper
225 100
223 64.1 50.3 75.5
227 61.2 48.6 72.8




Instrument :
BNA_P
ClientSampleId :
SLCS246

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#34
Caprolactam
Concen: 38.987 ng/ul m
RT: 11.469 min Scan# 1425
Delta R.T. 0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31



Tgt Ion:113 Resp: 149537
Ion Ratio Lower Upper
113 100
55 161.2 126.5 189.7
56 124.6 95.0 142.4









#35
4-Chloro-3-methylphenol
Concen: 40.350 ng/ul
RT: 11.804 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:107 Resp: 494502
Ion Ratio Lower Upper
107 100
144 23.7 18.0 27.0
142 72.9 56.3 84.5

Instrument :
BNA_P
ClientSampleId :
SLCS246




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM




#36
2-Methylnaphthalene
Concen: 39.137 ng/ul
RT: 12.169 min Scan# 1544
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:142 Resp: 960444
Ion Ratio Lower Upper
142 100
141 89.9 71.2 106.8







#37
1-Methylnaphthalene
Concen: 38.308 ng/ul
RT: 12.387 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:142 Resp: 935568
Ion Ratio Lower Upper
142 100
141 93.9 73.8 110.6
116 4.4 3.1 4.7




Instrument :
BNA_P
ClientSampleId :
SLCS246

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.564 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:216 Resp: 411971
Ion Ratio Lower Upper
216 100
214 76.8 60.2 90.2
179 23.9 18.1 27.1
108 22.7 16.6 24.8





#40
Hexachlorocyclopentadiene
Concen: 28.389 ng/ul
RT: 12.516 min Scan# 1603
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:	237	Resp:	189154
Ion	Ratio	Lower	Upper
237	100		
235	63.0	49.6	74.4
272	12.3	10.7	16.1

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#41
2,4,6-Trichlorophenol
Concen: 40.995 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:	196	Resp:	285928
Ion	Ratio	Lower	Upper
196	100		
198	94.4	75.8	113.6
200	29.8	24.0	36.0





#42
2,4,5-Trichlorophenol
Concen: 42.142 ng/ul
RT: 12.857 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246




Tgt Ion:196 Resp: 312933
Ion Ratio Lower Upper
196 100
198 96.3 75.7 113.5
200 31.7 25.4 38.2

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#43
1,1'-Biphenyl
Concen: 40.192 ng/ul
RT: 13.193 min Scan# 1718
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:154 Resp: 1177503
Ion Ratio Lower Upper
154 100
153 40.4 31.8 47.6
76 15.0 12.6 19.0





#44
2-Chloronaphthalene
Concen: 39.875 ng/ul
RT: 13.228 min Scan# 1724
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246



Tgt Ion:162 Resp: 886291
Ion Ratio Lower Upper
162 100
164 32.4 25.9 38.9
127 45.7 36.5 54.7




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#45
2-Nitroaniline
Concen: 42.997 ng/ul
RT: 13.446 min Scan# 1761
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion: 65 Resp: 344631
Ion Ratio Lower Upper
65 100
92 68.1 54.2 81.4
138 92.3 73.6 110.4





#47
Dimethylphthalate
Concen: 40.345 ng/ul
RT: 13.828 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:163 Resp: 1153301
Ion Ratio Lower Upper
163 100
194 3.6 2.8 4.2
164 10.2 8.1 12.1

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#48
2,6-Dinitrotoluene
Concen: 42.384 ng/ul
RT: 13.945 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:165 Resp: 233848
Ion Ratio Lower Upper
165 100
89 77.5 54.9 82.3
121 26.7 19.4 29.0





#50
Acenaphthylene
Concen: 40.012 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246




Tgt Ion:152 Resp: 1386592
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 12.7 10.4 15.6

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#51
3-Nitroaniline
Concen: 44.176 ng/ul
RT: 14.275 min Scan# 1902
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:138 Resp: 279122
Ion Ratio Lower Upper
138 100
108 11.8 10.1 15.1
92 129.9 108.4 162.6





#52
Acenaphthene
Concen: 39.930 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:153 Resp: 986647
Ion Ratio Lower Upper
153 100
152 47.8 38.4 57.6
154 86.2 69.4 104.2

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#53
2,4-Dinitrophenol
Concen: 36.289 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:184 Resp: 129501
Ion Ratio Lower Upper
184 100
63 81.7 66.5 99.7
154 72.3 56.6 84.8





#55
4-Nitrophenol
Concen: 42.487 ng/ul
RT: 14.593 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246



Tgt Ion:109 Resp: 230000
Ion Ratio Lower Upper
109 100
139 115.4 95.6 143.4
65 134.2 110.2 165.4




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#56
Dibenzofuran
Concen: 40.420 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion:168 Resp: 1335437
Ion Ratio Lower Upper
168 100
139 45.2 34.4 51.6





#57
2,4-Dinitrotoluene
Concen: 43.103 ng/ul
RT: 14.740 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246



Tgt Ion:165 Resp: 349866
Ion Ratio Lower Upper
165 100
63 54.4 42.6 64.0
182 2.6 2.2 3.2




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.050 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion:232 Resp: 253042
Ion Ratio Lower Upper
232 100
131 64.9 48.2 72.4
130 3.7 2.6 3.8
166 37.7 27.5 41.3





#59
Diethylphthalate
Concen: 41.388 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:149 Resp: 1257570
Ion Ratio Lower Upper
149 100
177 19.6 15.7 23.5
150 12.9 9.8 14.8

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#61
Fluorene
Concen: 40.369 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:166 Resp: 1112088
Ion Ratio Lower Upper
166 100
165 99.4 78.7 118.1
167 13.7 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 40.229 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246




Tgt Ion:204 Resp: 501699
Ion Ratio Lower Upper
204 100
206 32.9 26.6 40.0
141 75.4 58.3 87.5

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#63
4-Nitroaniline
Concen: 50.744 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:138 Resp: 300234
Ion Ratio Lower Upper
138 100
92 59.7 48.2 72.2
108 96.5 73.0 109.4





#66
4,6-Dinitro-2-methylphenol
Concen: 37.712 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. 0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246




Tgt Ion:198 Resp: 184737
Ion Ratio Lower Upper
198 100
182 4.7 4.1 6.1
77 32.2 29.3 43.9

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#67
N-Nitrosodiphenylamine
Concen: 41.063 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:169 Resp: 935080
Ion Ratio Lower Upper
169 100
168 68.5 54.5 81.7
167 37.3 29.2 43.8





#68
4-Bromophenyl-phenylether
Concen: 40.897 ng/ul
RT: 16.310 min Scan# 2248
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246



Tgt Ion:248 Resp: 299028
Ion Ratio Lower Upper
248 100
250 97.3 79.8 119.6
141 92.1 76.9 115.3




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#69
Hexachlorobenzene
Concen: 40.991 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion:284 Resp: 338715
Ion Ratio Lower Upper
284 100
142 44.8 37.0 55.4
249 33.1 27.0 40.6





#70
Atrazine
Concen: 40.142 ng/ul
RT: 16.592 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246




Tgt Ion:200 Resp: 324281
Ion Ratio Lower Upper
200 100
173 27.6 22.5 33.7
215 44.6 36.1 54.1

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#71
Pentachlorophenol
Concen: 39.859 ng/ul
RT: 16.763 min Scan# 2325
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31



Tgt Ion:266 Resp: 215735
Ion Ratio Lower Upper
266 100
264 63.1 49.7 74.5
268 64.4 51.1 76.7









#72
Phenanthrene
Concen: 40.999 ng/ul
RT: 17.157 min Scan# 2392
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Instrument :
BNA_P
ClientSampleId :
SLCS246

Tgt Ion:178 Resp: 1717115
Ion Ratio Lower Upper
178 100
179 15.0 12.5 18.7
176 19.2 15.4 23.2




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#74
Anthracene
Concen: 40.375 ng/ul
RT: 17.251 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion:178 Resp: 1717139
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 18.9 14.6 21.8





#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.759 ng/uL
RT: 13.151 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:216 Resp: 410897
Ion Ratio Lower Upper
216 100
214 77.8 63.6 95.4
218 47.5 39.4 59.0

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#76
Pentachlorobenzene
Concen: 40.790 ng/uL
RT: 14.675 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:250 Resp: 405149
Ion Ratio Lower Upper
250 100
248 63.8 50.6 76.0
252 63.0 52.6 78.8





#77
 Carbazole
 Concen: 40.919 ng/ul
 RT: 17.528 min Scan# 2455
 Delta R.T. -0.000 min
 Lab File: BP006650.D
 Acq: 11 Aug 2021 11:31




Instrument :
 BNA_P
 ClientSampleId :
 SLCS246



Tgt Ion:167 Resp: 1625426
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.4 26.0
 139 14.7 11.8 17.8




Manual Integrations
 APPROVED


mohammad
 8/11/2021 9:29:30 PM



#78
 Di-n-butylphthalate
 Concen: 41.411 ng/ul
 RT: 18.092 min Scan# 2551
 Delta R.T. -0.000 min
 Lab File: BP006650.D
 Acq: 11 Aug 2021 11:31


Tgt Ion:149 Resp: 2126883
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 6.1 4.9 7.3





#80
Fluoranthene
Concen: 41.700 ng/ul
RT: 19.133 min Scan# 2728
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Instrument :
BNA_P
ClientSampleId :
SLCS246






Tgt Ion:202 Resp: 1941930

Ion	Ratio	Lower	Upper
202	100		
101	14.0	11.0	16.6
100	10.9	8.7	13.1

Manual Integrations
APPROVED




mohammad
8/11/2021 9:29:30 PM




#82
Pyrene
Concen: 41.334 ng/ul
RT: 19.486 min Scan# 2788
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:202 Resp: 1988187


Ion	Ratio	Lower	Upper
202	100		
101	15.9	12.1	18.1
100	13.9	10.6	16.0





#83
Butylbenzylphthalate
Concen: 41.355 ng/ul
RT: 20.375 min Scan# 2939
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246



Tgt Ion:149 Resp: 983415
Ion Ratio Lower Upper
149 100
91 84.2 63.0 94.6
206 17.2 14.9 22.3




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#84
3,3'-Dichlorobenzidine
Concen: 37.683 ng/ul
RT: 21.139 min Scan# 3069
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:252 Resp: 606711
Ion Ratio Lower Upper
252 100
254 63.8 51.2 76.8
126 14.6 11.9 17.9









#85
Benzo(a)anthracene
Concen: 40.924 ng/ul
RT: 21.198 min Scan# 3079
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:228 Resp: 1880716
Ion Ratio Lower Upper
228 100
229 20.0 15.3 22.9
226 27.4 21.2 31.8

Instrument :
BNA_P
ClientSampled :
SLCS246




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#86
Bis(2-ethylhexyl)phthalate
Concen: 41.037 ng/ul
RT: 21.139 min Scan# 3069
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31


Tgt Ion:149 Resp: 1503715
Ion Ratio Lower Upper
149 100
167 24.9 20.0 30.0
279 3.3 2.9 4.3





#87
Chrysene
Concen: 40.807 ng/ul
RT: 21.251 min Scan# 3088
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Instrument :
BNA_P
ClientSampleId :
SLCS246






Tgt Ion:228 Resp: 1808133

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.2
229	19.9	15.8	23.6




Manual Integrations
APPROVED


mohammad
8/11/2021 9:29:30 PM



#89
Di-n-octyl phthalate
Concen: 40.793 ng/ul
RT: 22.045 min Scan# 3223
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:149 Resp: 2571597





#90
Benzo(b)fluoranthene
Concen: 41.449 ng/ul
RT: 22.827 min Scan# 3356
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31





Tgt Ion:252 Resp: 1963673

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.8	26.6
125	11.8	10.3	15.5

Instrument :
BNA_P
ClientSampleId :
SLCS246

Manual Integrations
APPROVED




mohammad
8/11/2021 9:29:30 PM




#91
Benzo(k)fluoranthene
Concen: 41.240 ng/ul
RT: 22.874 min Scan# 3364
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31

Tgt Ion:252 Resp: 1891576


Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.3	25.9
125	12.5	10.4	15.6





#93
Benzo(a)pyrene
Concen: 40.862 ng/ul
RT: 23.433 min Scan# 3459
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampled :
SLCS246




Tgt Ion:252 Resp: 1753687
Ion Ratio Lower Upper
252 100
253 21.4 17.6 26.4
125 13.6 10.6 16.0

Manual Integrations
APPROVED



mohammad
8/11/2021 9:29:30 PM




#94
Indeno(1,2,3-cd)pyrene
Concen: 40.738 ng/ul
RT: 25.921 min Scan# 3882
Delta R.T. -0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Tgt Ion:276 Resp: 2286062
Ion Ratio Lower Upper
276 100
138 26.7 23.0 34.4
277 24.8 20.0 30.0





#95
Dibenzo(a,h)anthracene
Concen: 40.722 ng/ul
RT: 25.945 min Scan# 3886
Delta R.T. -0.000 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31




Instrument :
BNA_P
ClientSampleId :
SLCS246




Tgt Ion:278 Resp: 1919468
Ion Ratio Lower Upper
278 100
139 18.7 14.4 21.6
279 24.2 19.1 28.7

Manual Integrations
APPROVED

mohammad
8/11/2021 9:29:30 PM



#96
Benzo(g,h,i)perylene
Concen: 40.754 ng/ul
RT: 26.662 min Scan# 4008
Delta R.T. 0.006 min
Lab File: BP006650.D
Acq: 11 Aug 2021 11:31



Tgt Ion:276 Resp: 1898831
Ion Ratio Lower Upper
276 100
138 25.6 21.5 32.3
277 23.4 19.0 28.6

