

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081423\
 Data File : BP016559.D
 Acq On : 14 Aug 2023 13:51
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV630

Quant Time: Aug 14 17:41:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 15:11:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	301207	20.000	ng/u1	0.00
20) Naphthalene-d8	10.893	136	1376317	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	836025	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.481	188	1806520	20.000	ng/u1	0.00
79) Chrysene-d12	21.574	240	1757934	20.000	ng/u1	0.00
88) Perylene-d12	24.163	264	1992424	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	65468	8.259	ng/uL	0.00
4) Pyridine-d5	3.828	84	492884	20.437	ng/u1	0.00
7) Phenol-d5	7.199	99	569757	20.518	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.393	67	375740	20.544	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	427201	21.419	ng/u1	0.00
15) 4-Methylphenol-d8	8.763	113	458995	20.788	ng/u1	0.00
21) Nitrobenzene-d5	9.257	128	213083	22.003	ng/u1	0.00
24) 2-Nitrophenol-d4	9.975	143	208148	22.630	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.499	165	402571	22.240	ng/u1	0.00
31) 4-Chloroaniline-d4	11.046	131	652336	21.004	ng/u1	0.00
46) Dimethylphthalate-d6	14.122	166	1271484	21.027	ng/u1	0.00
49) Acenaphthylene-d8	14.410	160	1482672	21.724	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	257305	20.888	ng/u1	0.00
60) Fluorene-d10	15.704	176	1080657	20.963	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	174365	20.289	ng/u1	0.00
73) Anthracene-d10	17.575	188	1683354	21.542	ng/u1	0.00
81) Pyrene-d10	19.810	212	1975219	21.444	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.992	264	2031945	21.280	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.434	88	70985	8.144	ng/uL	97
5) Pyridine	3.852	79	514824	20.575	ng/u1	98
6) Benzaldehyde	7.210	77	372023	24.303	ng/u1	96
8) Phenol	7.222	94	609423	20.533	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.493	93	493249	20.574	ng/u1	99
12) 2-Chlorophenol	7.610	128	443003	21.109	ng/u1	98
13) 2-Methylphenol	8.493	108	441718	20.489	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.575	45	440319	13.180	ng/u1	98
16) Acetophenone	8.910	105	750710	21.054	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.881	70	400624	20.950	ng/u1	99
18) 4-Methylphenol	8.828	108	487432	20.572	ng/u1	96
19) Hexachloroethane	9.140	117	181108	20.656	ng/u1	99
22) Nitrobenzene	9.299	77	583349	21.518	ng/u1	98
23) Isophorone	9.816	82	1103344	21.548	ng/u1	100
25) 2-Nitrophenol	10.004	139	238681	22.699	ng/u1	100
26) 2,4-Dimethylphenol	10.040	107	529722	21.369	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.304	93	680428	20.730	ng/u1	99
29) 2,4-Dichlorophenol	10.522	162	405208	21.758	ng/u1	99
30) Naphthalene	10.946	128	1495642	20.799	ng/u1	100
32) 4-Chloroaniline	11.069	127	645573	20.991	ng/u1	99
33) Hexachlorobutadiene	11.181	225	238985	21.288	ng/u1	99
34) Caprolactam	11.875	113	150903	20.630	ng/u1	94
35) 4-Chloro-3-methylphenol	12.163	107	485391	21.323	ng/u1	99
36) 2-Methylnaphthalene	12.540	142	997429	20.707	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081423\
 Data File : BP016559.D
 Acq On : 14 Aug 2023 13:51
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Quant Time: Aug 14 17:41:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 15:11:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	1011689	20.859	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.893	216	476905	21.327	ng/ul	99
40) Hexachlorocyclopentadiene	12.845	237	272122	20.001	ng/ul	99
41) 2,4,6-Trichlorophenol	13.140	196	308011	22.133	ng/ul	99
42) 2,4,5-Trichlorophenol	13.204	196	334741	22.190	ng/ul	99
43) 1,1'-Biphenyl	13.545	154	1305516	21.008	ng/ul	99
44) 2-Chloronaphthalene	13.593	162	1010439	20.898	ng/ul	99
45) 2-Nitroaniline	13.822	65	326174	22.311	ng/ul	96
47) Dimethylphthalate	14.169	163	1287907	20.831	ng/ul	99
48) 2,6-Dinitrotoluene	14.310	165	243583	22.746	ng/ul	97
50) Acenaphthylene	14.440	152	1720643	21.407	ng/ul	100
51) 3-Nitroaniline	14.645	138	267006	22.237	ng/ul	95
52) Acenaphthene	14.775	153	1123709	20.884	ng/ul	98
53) 2,4-Dinitrophenol	14.839	184	105968	17.775	ng/ul	99
55) 4-Nitrophenol	14.922	109	211424	21.112	ng/ul	99
56) Dibenzofuran	15.110	168	1512814	20.607	ng/ul	99
57) 2,4-Dinitrotoluene	15.087	165	366473	22.853	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.322	232	278836	22.062	ng/ul	98
59) Diethylphthalate	15.522	149	1322353	20.919	ng/ul	99
61) Fluorene	15.763	166	1259373	20.922	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.751	204	591264	20.925	ng/ul	99
63) 4-Nitroaniline	15.810	138	268380	23.356	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.839	198	200855	21.078	ng/ul	95
67) N-Nitrosodiphenylamine	15.975	169	1069372	21.838	ng/ul	99
68) 4-Bromophenyl-phenylether	16.657	248	349521	21.669	ng/ul	97
69) Hexachlorobenzene	16.739	284	398677	21.154	ng/ul	98
70) Atrazine	16.928	200	374778	21.273	ng/ul	99
71) Pentachlorophenol	17.098	266	235718	20.314	ng/ul	98
72) Phenanthrene	17.522	178	2012516	21.145	ng/ul	99
74) Anthracene	17.616	178	2033093	21.267	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.498	216	505651	22.025	ng/ul	96
76) Pentachlorobenzene	15.004	250	448319	21.764	ng/ul	99
77) Carbazole	17.892	167	1900745	21.575	ng/ul	100
78) Di-n-butylphthalate	18.404	149	2264476	22.204	ng/ul	100
80) Fluoranthene	19.480	202	2336450	21.253	ng/ul	100
82) Pyrene	19.839	202	2466047	21.243	ng/ul	97
83) Butylbenzylphthalate	20.698	149	1039709	22.232	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.492	252	830577	21.697	ng/ul	100
85) Benzo(a)anthracene	21.557	228	2516204	21.298	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.439	149	1515209	22.325	ng/ul	98
87) Chrysene	21.616	228	2306877	20.866	ng/ul	98
89) Di-n-octyl phthalate	22.427	149	2637279	21.215	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	2469885	20.819	ng/ul	99
91) Benzo(k)fluoranthene	23.410	252	2491282	21.168	ng/ul	99
93) Benzo(a)pyrene	24.045	252	2368364	21.163	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.909	276	2731776	20.683	ng/ul	98
95) Dibenzo(a,h)anthracene	26.945	278	2264368	20.601	ng/ul	100
96) Benzo(g,h,i)perylene	27.780	276	2160901	20.556	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 14 17:41:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 14 15:11:15 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.144 ng/uL
 RT: 3.434 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP016559.D
 Acq: 14 Aug 2023 13:51




Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Tgt Ion: 88 Resp: 70985
 Ion Ratio Lower Upper
 88 100
 43 28.2 23.5 35.3
 58 68.6 57.1 85.7



#5
 Pyridine
 Concen: 20.575 ng/uL
 RT: 3.852 min Scan# 96
 Delta R.T. -0.000 min
 Lab File: BP016559.D
 Acq: 14 Aug 2023 13:51

Tgt Ion: 79 Resp: 514824
 Ion Ratio Lower Upper
 79 100
 52 69.0 56.6 84.8
 51 32.8 26.9 40.3









#6
Benzaldehyde
Concen: 24.303 ng/ul
RT: 7.210 min Scan# 667
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion: 77 Resp: 372023
Ion Ratio Lower Upper
77 100
105 93.0 71.8 107.6
106 91.2 70.2 105.2



#8
Phenol
Concen: 20.533 ng/ul
RT: 7.222 min Scan# 669
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion: 94 Resp: 609423
Ion Ratio Lower Upper
94 100
65 27.7 22.3 33.5
66 37.5 28.7 43.1









#10
Bis(2-Chloroethyl)ether
Concen: 20.574 ng/ul
RT: 7.493 min Scan# 715
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion: 93 Resp: 493249
Ion Ratio Lower Upper
93 100
63 75.7 60.1 90.1
95 32.1 25.4 38.2



#12
2-Chlorophenol
Concen: 21.109 ng/ul
RT: 7.610 min Scan# 735
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:128 Resp: 443003
Ion Ratio Lower Upper
128 100
64 50.2 41.4 62.0
130 31.6 26.0 39.0





#13

2-Methylphenol

Concen: 20.489 ng/ul

RT: 8.493 min Scan# 885

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :





SICV630

Tgt Ion:108 Resp: 441718

Ion Ratio Lower Upper

108 100

107 90.1 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.180 ng/ul

RT: 8.575 min Scan# 899

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion: 45 Resp: 440319


Ion Ratio Lower Upper

45 100

77 15.5 12.1 18.1

79 11.3 8.2 12.4









#16
Acetophenone
Concen: 21.054 ng/ul
RT: 8.910 min Scan# 910
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:105 Resp: 750710
Ion Ratio Lower Upper
105 100
77 82.5 64.4 96.6
51 26.5 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.950 ng/ul
RT: 8.881 min Scan# 951
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion: 70 Resp: 400624
Ion Ratio Lower Upper
70 100
42 52.9 43.0 64.4
101 8.9 6.8 10.2
130 18.8 14.2 21.4









#18
4-Methylphenol
Concen: 20.572 ng/ul
RT: 8.828 min Scan# 942
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:108 Resp: 487432
Ion Ratio Lower Upper
108 100
107 112.8 87.3 130.9



#19
Hexachloroethane
Concen: 20.656 ng/ul
RT: 9.140 min Scan# 995
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:117 Resp: 181108
Ion Ratio Lower Upper
117 100
201 89.1 71.8 107.6
199 57.7 45.2 67.8









#22
Nitrobenzene
Concen: 21.518 ng/ul
RT: 9.299 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion: 77 Resp: 583349
Ion Ratio Lower Upper
77 100
123 41.8 32.5 48.7
65 13.0 10.3 15.5



#23
Isophorone
Concen: 21.548 ng/ul
RT: 9.816 min Scan# 1110
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion: 82 Resp: 1103344
Ion Ratio Lower Upper
82 100
95 6.7 5.3 7.9
138 15.8 12.7 19.1









#25
2-Nitrophenol
Concen: 22.699 ng/ul
RT: 10.004 min Scan# 1142
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:139 Resp: 238681
Ion Ratio Lower Upper
139 100
65 58.0 46.4 69.6
109 31.9 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 21.369 ng/ul
RT: 10.040 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:107 Resp: 529722
Ion Ratio Lower Upper
107 100
121 48.8 39.1 58.7
122 83.4 66.9 100.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.730 ng/ul
 RT: 10.304 min Scan# 1193
 Delta R.T. -0.000 min
 Lab File: BP016559.D
 Acq: 14 Aug 2023 13:51




Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Tgt Ion: 93 Resp: 680428
 Ion Ratio Lower Upper
 93 100
 95 32.6 25.8 38.8
 123 10.7 8.7 13.1



#29
 2,4-Dichlorophenol
 Concen: 21.758 ng/ul
 RT: 10.522 min Scan# 1230
 Delta R.T. -0.000 min
 Lab File: BP016559.D
 Acq: 14 Aug 2023 13:51

Tgt Ion: 162 Resp: 405208
 Ion Ratio Lower Upper
 162 100
 164 66.7 52.9 79.3
 98 41.2 32.7 49.1









#30
Naphthalene
Concen: 20.799 ng/ul
RT: 10.946 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:128 Resp: 1495642
Ion Ratio Lower Upper
128 100
129 10.5 8.6 13.0
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 20.991 ng/ul
RT: 11.069 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:127 Resp: 645573
Ion Ratio Lower Upper
127 100
129 32.6 25.6 38.4





#33

Hexachlorobutadiene

Concen: 21.288 ng/ul

RT: 11.181 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:225 Resp: 238985

Ion Ratio Lower Upper

225 100

223 63.9 50.2 75.2

227 63.3 50.8 76.2



#34

Caprolactam

Concen: 20.630 ng/ul

RT: 11.875 min Scan# 1460

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:113 Resp: 150903


Ion Ratio Lower Upper

113 100

55 177.1 134.8 202.2

56 132.2 101.5 152.3









#35
4-Chloro-3-methylphenol
Concen: 21.323 ng/ul
RT: 12.163 min Scan# 1508
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:107 Resp: 485391
Ion Ratio Lower Upper
107 100
144 24.3 19.0 28.6
142 75.0 60.2 90.4



#36
2-Methylnaphthalene
Concen: 20.707 ng/ul
RT: 12.540 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:142 Resp: 997429
Ion Ratio Lower Upper
142 100
141 88.1 71.4 107.0









#37
1-Methylnaphthalene
Concen: 20.859 ng/ul
RT: 12.763 min Scan# 1611
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:142 Resp: 1011689
Ion Ratio Lower Upper
142 100
141 92.1 74.5 111.7
116 4.1 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.327 ng/ul
RT: 12.893 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:216 Resp: 476905
Ion Ratio Lower Upper
216 100
214 78.7 63.8 95.6
179 22.4 17.9 26.9
108 19.3 16.3 24.5





#40

Hexachlorocyclopentadiene

Concen: 20.001 ng/ul

RT: 12.845 min Scan# 1625

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:237 Resp: 272122

Ion Ratio Lower Upper

237 100

235 61.5 49.7 74.5

272 13.2 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 22.133 ng/ul

RT: 13.140 min Scan# 1675

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:196 Resp: 308011


Ion Ratio Lower Upper

196 100

198 97.3 77.0 115.4

200 32.0 24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 22.190 ng/ul
RT: 13.204 min Scan# 1686
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:196 Resp: 334741
Ion Ratio Lower Upper
196 100
198 97.1 76.7 115.1
200 30.5 24.2 36.2



#43
1,1'-Biphenyl
Concen: 21.008 ng/ul
RT: 13.545 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:154 Resp: 1305516
Ion Ratio Lower Upper
154 100
153 40.2 31.4 47.2
76 15.6 13.0 19.4









#44
2-Chloronaphthalene
Concen: 20.898 ng/ul
RT: 13.593 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:162 Resp: 1010439
Ion Ratio Lower Upper
162 100
164 32.3 25.8 38.6
127 41.6 32.9 49.3



#45
2-Nitroaniline
Concen: 22.311 ng/ul
RT: 13.822 min Scan# 1791
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion: 65 Resp: 326174
Ion Ratio Lower Upper
65 100
92 68.9 54.5 81.7
138 96.2 72.1 108.1





#47

Dimethylphthalate

Concen: 20.831 ng/ul

RT: 14.169 min Scan# 1850

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:163 Resp: 1287907

Ion Ratio Lower Upper

163 100

194 4.1 3.0 4.6

164 10.1 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 22.746 ng/ul

RT: 14.310 min Scan# 1874

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:165 Resp: 243583


Ion Ratio Lower Upper

165 100

89 64.1 53.2 79.8

121 22.4 18.8 28.2









#50
Acenaphthylene
Concen: 21.407 ng/ul
RT: 14.440 min Scan# 1896
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:152 Resp: 1720643
Ion Ratio Lower Upper
152 100
151 19.8 15.7 23.5
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 22.237 ng/ul
RT: 14.645 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:138 Resp: 267006
Ion Ratio Lower Upper
138 100
108 11.8 10.9 16.3
92 125.4 105.0 157.6









#52
Acenaphthene
Concen: 20.884 ng/ul
RT: 14.775 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:153 Resp: 1123709
Ion Ratio Lower Upper
153 100
152 47.9 38.7 58.1
154 87.2 72.0 108.0



#53
2,4-Dinitrophenol
Concen: 17.775 ng/ul
RT: 14.839 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:184 Resp: 105968
Ion Ratio Lower Upper
184 100
63 80.2 65.0 97.6
154 69.7 56.2 84.2









#55
4-Nitrophenol
Concen: 21.112 ng/ul
RT: 14.922 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:109 Resp: 211424
Ion Ratio Lower Upper
109 100
139 129.3 101.1 151.7
65 136.3 108.9 163.3



#56
Dibenzofuran
Concen: 20.607 ng/ul
RT: 15.110 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:168 Resp: 1512814
Ion Ratio Lower Upper
168 100
139 38.4 31.4 47.0









#57
2,4-Dinitrotoluene
Concen: 22.853 ng/ul
RT: 15.087 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:165 Resp: 366473
Ion Ratio Lower Upper
165 100
63 50.3 40.4 60.6
182 2.5 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.062 ng/ul
RT: 15.322 min Scan# 2046
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:232 Resp: 278836
Ion Ratio Lower Upper
232 100
131 51.6 42.9 64.3
130 3.0 2.2 3.4
166 33.7 26.7 40.1









#59
Diethylphthalate
Concen: 20.919 ng/ul
RT: 15.522 min Scan# 2080
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:149 Resp: 1322353
Ion Ratio Lower Upper
149 100
177 21.4 16.5 24.7
150 12.6 10.0 15.0



#61
Fluorene
Concen: 20.922 ng/ul
RT: 15.763 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:166 Resp: 1259373
Ion Ratio Lower Upper
166 100
165 96.5 78.3 117.5
167 13.2 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.925 ng/ul
RT: 15.751 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:204 Resp: 591264
Ion Ratio Lower Upper
204 100
206 32.8 25.5 38.3
141 65.1 52.2 78.2



#63
4-Nitroaniline
Concen: 23.356 ng/ul
RT: 15.810 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:138 Resp: 268380
Ion Ratio Lower Upper
138 100
92 58.4 46.7 70.1
108 93.3 73.7 110.5









#66
4,6-Dinitro-2-methylphenol
Concen: 21.078 ng/ul
RT: 15.839 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:198 Resp: 200855
Ion Ratio Lower Upper
198 100
182 4.5 3.9 5.9
77 29.4 25.8 38.6



#67
N-Nitrosodiphenylamine
Concen: 21.838 ng/ul
RT: 15.975 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:169 Resp: 1069372
Ion Ratio Lower Upper
169 100
168 66.6 53.9 80.9
167 36.6 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 21.669 ng/ul
RT: 16.657 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:248 Resp: 349521
Ion Ratio Lower Upper
248 100
250 96.0 76.9 115.3
141 79.3 67.6 101.4



#69
Hexachlorobenzene
Concen: 21.154 ng/ul
RT: 16.739 min Scan# 2287
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:284 Resp: 398677
Ion Ratio Lower Upper
284 100
142 40.1 31.9 47.9
249 33.3 24.7 37.1





#70

Atrazine

Concen: 21.273 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:200 Resp: 374778

Ion Ratio Lower Upper

200 100

173 26.6 22.2 33.2

215 47.2 38.4 57.6



#71

Pentachlorophenol

Concen: 20.314 ng/ul

RT: 17.098 min Scan# 2348

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:266 Resp: 235718


Ion Ratio Lower Upper

266 100

264 63.3 50.7 76.1

268 63.7 53.0 79.6





#72

Phenanthrene

Concen: 21.145 ng/ul

RT: 17.522 min Scan# 2420

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:178 Resp: 2012516

Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 19.2 15.5 23.3



#74

Anthracene

Concen: 21.267 ng/ul

RT: 17.616 min Scan# 2436

Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:178 Resp: 2033093


Ion Ratio Lower Upper

178 100

179 15.2 12.5 18.7

176 18.5 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.025 ng/uL

RT: 13.498 min Scan# 1736

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:216 Resp: 505651

Ion Ratio Lower Upper

216 100

214 76.7 65.4 98.2

218 47.5 39.3 58.9



#76

Pentachlorobenzene

Concen: 21.764 ng/uL

RT: 15.004 min Scan# 1992

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:250 Resp: 448319


Ion Ratio Lower Upper

250 100

248 63.4 50.2 75.2

252 63.1 51.0 76.4









#77
 Carbazole
 Concen: 21.575 ng/ul
 RT: 17.892 min Scan# 2483
 Delta R.T. -0.000 min
 Lab File: BP016559.D
 Acq: 14 Aug 2023 13:51




Instrument :
 BNA_P
 ClientSampleId :
 SICV630


Tgt Ion:167 Resp: 1900745
 Ion Ratio Lower Upper
 167 100
 166 21.0 16.8 25.2
 139 13.3 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 22.204 ng/ul
 RT: 18.404 min Scan# 2570
 Delta R.T. -0.000 min
 Lab File: BP016559.D
 Acq: 14 Aug 2023 13:51

Tgt Ion:149 Resp: 2264476
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 5.5 4.6 6.8









#80
Fluoranthene
Concen: 21.253 ng/ul
RT: 19.480 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:202 Resp: 2336450
Ion Ratio Lower Upper
202 100
101 13.2 10.6 16.0
100 10.2 8.1 12.1



#82
Pyrene
Concen: 21.243 ng/ul
RT: 19.839 min Scan# 2814
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:202 Resp: 2466047
Ion Ratio Lower Upper
202 100
101 15.4 13.4 20.2
100 12.4 11.0 16.4









#83
Butylbenzylphthalate
Concen: 22.232 ng/ul
RT: 20.698 min Scan# 2959
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:149 Resp: 1039709
Ion Ratio Lower Upper
149 100
91 77.5 67.8 101.6
206 19.5 14.4 21.6



#84
3,3'-Dichlorobenzidine
Concen: 21.697 ng/ul
RT: 21.492 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:252 Resp: 830577
Ion Ratio Lower Upper
252 100
254 63.9 51.3 76.9
126 13.4 10.2 15.4





#85

Benzo(a)anthracene

Concen: 21.298 ng/ul

RT: 21.557 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:228 Resp: 2516204

Ion Ratio Lower Upper

228 100

229 20.0 15.8 23.6

226 27.2 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.325 ng/ul

RT: 21.439 min Scan# 3086

Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:149 Resp: 1515209


Ion Ratio Lower Upper

149 100

167 26.9 20.9 31.3

279 3.7 2.6 4.0





#87

Chrysene

Concen: 20.866 ng/ul

RT: 21.616 min Scan# 3115

Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:228 Resp: 2306877

Ion Ratio Lower Upper

228 100

226 30.1 24.6 36.8

229 19.5 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 21.215 ng/ul




RT: 22.427 min Scan# 3254


Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Tgt Ion:149 Resp: 2637279









#90
Benzo(b)fluoranthene
Concen: 20.819 ng/ul
RT: 23.357 min Scan# 3411
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51




Instrument :
BNA_P
ClientSampleId :
SICV630


Tgt Ion:252 Resp: 2469885
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 11.3 10.0 15.0



#91
Benzo(k)fluoranthene
Concen: 21.168 ng/ul
RT: 23.410 min Scan# 3421
Delta R.T. 0.006 min
Lab File: BP016559.D
Acq: 14 Aug 2023 13:51

Tgt Ion:252 Resp: 2491282
Ion Ratio Lower Upper
252 100
253 22.0 18.0 27.0
125 11.8 9.8 14.8





#93

Benzo(a)pyrene

Concen: 21.163 ng/ul

RT: 24.045 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:252 Resp: 2368364

Ion Ratio Lower Upper

252 100

253 21.6 17.7 26.5

125 12.9 10.6 15.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.683 ng/ul

RT: 26.909 min Scan# 4016

Delta R.T. -0.000 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51




Tgt Ion:276 Resp: 2731776


Ion Ratio Lower Upper

276 100

138 25.1 21.7 32.5

277 24.4 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 20.601 ng/ul

RT: 26.945 min Scan# 4021

Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Instrument :

BNA_P

ClientSampleId :

SICV630





Tgt Ion:278 Resp: 2264368

Ion Ratio Lower Upper

278 100

139 19.4 15.5 23.3

279 23.9 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 20.556 ng/ul

RT: 27.780 min Scan# 4164

Delta R.T. 0.006 min

Lab File: BP016559.D

Acq: 14 Aug 2023 13:51

Tgt Ion:276 Resp: 2160901

Ion Ratio Lower Upper

276 100

138 25.0 21.5 32.3

277 23.6 19.3 28.9

