

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081423\
 Data File : BP016560.D
 Acq On : 14 Aug 2023 14:25
 Operator : MA/JU
 Sample : PB154773BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK773

Quant Time: Aug 14 17:42:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 15:11:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	440699	20.000	ng/ul	0.00
20) Naphthalene-d8	10.893	136	2037572	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.710	164	1266095	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.475	188	2850199	20.000	ng/ul	0.00
79) Chrysene-d12	21.569	240	2733952	20.000	ng/ul	0.00
88) Perylene-d12	24.157	264	2951957	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	42360	3.652	ng/uL	0.00
4) Pyridine-d5	3.828	84	633883	17.964	ng/ul	0.00
7) Phenol-d5	7.199	99	751734	18.502	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.393	67	487483	18.217	ng/ul	0.00
11) 2-Chlorophenol-d4	7.575	132	538374	18.449	ng/ul	0.00
15) 4-Methylphenol-d8	8.757	113	568530	17.598	ng/ul	0.00
21) Nitrobenzene-d5	9.257	128	271025	18.904	ng/ul	0.00
24) 2-Nitrophenol-d4	9.975	143	264519	19.426	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.499	165	474962	17.724	ng/ul	0.00
31) 4-Chloroaniline-d4	11.046	131	803870	17.483	ng/ul	0.00
46) Dimethylphthalate-d6	14.122	166	1669692	18.233	ng/ul	0.00
49) Acenaphthylene-d8	14.410	160	1872777	18.119	ng/ul	0.00
54) 4-Nitrophenol-d4	14.904	143	277000	14.848	ng/ul	0.00
60) Fluorene-d10	15.704	176	1439587	18.440	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	176465	13.014	ng/ul	0.00
73) Anthracene-d10	17.575	188	2314253	18.771	ng/ul	0.00
81) Pyrene-d10	19.804	212	2727047	19.037	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.986	264	2735887	19.339	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.434	88	103	0.008	ng/ul#	47
5) Pyridine	3.858	79	621	0.017	ng/ul#	1
6) Benzaldehyde	7.205	77	1561	0.070	ng/ul#	19
8) Phenol	7.228	94	258	0.006	ng/ul#	1
10) Bis(2-Chloroethyl)ether	7.487	93	231	0.007	ng/ul#	78
12) 2-Chlorophenol	7.610	128	305	0.010	ng/ul#	11
13) 2-Methylphenol	8.593	108	62	0.002	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.563	45	246	0.005	ng/ul#	1
16) Acetophenone	8.916	105	278	0.005	ng/ul#	1
17) N-Nitroso-di-n-propyla...	8.875	70	184	0.007	ng/ul#	31
18) 4-Methylphenol	8.828	108	268	0.008	ng/ul#	60
19) Hexachloroethane	9.140	117	68	0.005	ng/ul#	51
22) Nitrobenzene	9.299	77	506	0.013	ng/ul#	64
23) Isophorone	9.810	82	425	0.006	ng/ul#	78
25) 2-Nitrophenol	9.999	139	137	0.009	ng/ul#	1
26) 2,4-Dimethylphenol	10.040	107	238	0.006	ng/ul#	54
27) Bis(2-Chloroethoxy)met...	10.416	93	108	0.002	ng/ul	91
29) 2,4-Dichlorophenol	10.528	162	97	0.004	ng/ul#	1
30) Naphthalene	10.951	128	427	0.004	ng/ul#	76
32) 4-Chloroaniline	11.063	127	917	0.020	ng/ul#	1
33) Hexachlorobutadiene	11.157	225	29	0.002	ng/ul	85
34) Caprolactam	11.728	113	51	0.005	ng/ul	94
35) 4-Chloro-3-methylphenol	12.157	107	190	0.006	ng/ul#	71
36) 2-Methylnaphthalene	12.551	142	144	0.002	ng/ul	91


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081423\
 Data File : BP016560.D
 Acq On : 14 Aug 2023 14:25
 Operator : MA/JU
 Sample : PB154773BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK773

Quant Time: Aug 14 17:42:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 15:11:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	269	0.004	ng/ul#	91
39) 1,2,4,5-Tetrachloroben...	12.898	216	88	0.003	ng/ul#	25
40) Hexachlorocyclopentadiene	12.840	237	5	0.000	ng/ul	90
41) 2,4,6-Trichlorophenol	13.145	196	138	0.007	ng/ul	94
42) 2,4,5-Trichlorophenol	13.145	196	138	0.006	ng/ul	94
43) 1,1'-Biphenyl	13.545	154	436	0.005	ng/ul#	74
44) 2-Chloronaphthalene	13.592	162	353	0.005	ng/ul#	82
45) 2-Nitroaniline	13.822	65	170	0.008	ng/ul#	50
47) Dimethylphthalate	14.169	163	1052	0.011	ng/ul#	88
48) 2,6-Dinitrotoluene	14.310	165	231	0.014	ng/ul#	56
50) Acenaphthylene	14.434	152	870	0.007	ng/ul#	53
51) 3-Nitroaniline	14.645	138	104	0.006	ng/ul#	1
52) Acenaphthene	14.781	153	456	0.006	ng/ul#	84
53) 2,4-Dinitrophenol	14.839	184	11	0.001	ng/ul#	1
55) 4-Nitrophenol	14.922	109	492	0.032	ng/ul#	1
56) Dibenzofuran	15.057	168	33	0.000	ng/ul	92
57) 2,4-Dinitrotoluene	15.092	165	212	0.009	ng/ul#	87
58) 2,3,4,6-Tetrachlorophenol	15.322	232	147	0.008	ng/ul#	80
59) Diethylphthalate	15.522	149	1185	0.012	ng/ul#	92
61) Fluorene	15.763	166	672	0.007	ng/ul#	89
62) 4-Chlorophenyl-phenyle...	15.745	204	353	0.008	ng/ul#	77
63) 4-Nitroaniline	15.822	138	1347	0.077	ng/ul#	1
66) 4,6-Dinitro-2-methylph...	15.822	198	820	0.055	ng/ul#	1
67) N-Nitrosodiphenylamine	15.969	169	879	0.011	ng/ul	93
68) 4-Bromophenyl-phenylether	16.651	248	326	0.013	ng/ul#	82
69) Hexachlorobenzene	16.733	284	297	0.010	ng/ul#	80
70) Atrazine	16.922	200	118	0.004	ng/ul#	43
71) Pentachlorophenol	17.092	266	139	0.008	ng/ul#	79
72) Phenanthrene	17.516	178	2459	0.016	ng/ul#	86
74) Anthracene	17.616	178	2132	0.014	ng/ul#	86
75) 1,2,3,4-Tetrachloroben...	13.504	216	82	0.002	ng/uL	87
76) Pentachlorobenzene	14.998	250	114	0.004	ng/ul#	54
77) Carbazole	17.886	167	1042	0.007	ng/ul#	74
78) Di-n-butylphthalate	18.404	149	2631	0.016	ng/ul#	93
80) Fluoranthene	19.480	202	2353	0.014	ng/ul#	44
82) Pyrene	19.833	202	2643	0.015	ng/ul#	76
83) Butylbenzylphthalate	20.686	149	618	0.008	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.492	252	1469	0.025	ng/ul#	82
85) Benzo(a)anthracene	21.569	228	10513	0.057	ng/ul#	76
86) Bis(2-ethylhexyl)phtha...	21.433	149	2478	0.023	ng/ul#	90
87) Chrysene	21.610	228	2877	0.017	ng/ul#	86
89) Di-n-octyl phthalate	22.427	149	2530	0.014	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	4147	0.024	ng/ul#	32
91) Benzo(k)fluoranthene	23.398	252	3426	0.020	ng/ul#	37
93) Benzo(a)pyrene	24.033	252	3210	0.019	ng/ul#	4
94) Indeno(1,2,3-cd)pyrene	26.886	276	534	0.003	ng/ul	100
95) Dibenzo(a,h)anthracene	26.951	278	1724	0.011	ng/ul#	73
96) Benzo(g,h,i)perylene	27.756	276	494	0.003	ng/ul#	59





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 0.008 ng/uL
RT: 3.434 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion: 88 Resp: 103
Ion Ratio Lower Upper
88 100
43 53.5 23.5 35.3#
58 24.4 57.1 85.7#



#5
Pyridine
Concen: 0.017 ng/uL
RT: 3.858 min Scan# 97
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion: 79 Resp: 621
Ion Ratio Lower Upper
79 100
52 3568.9 56.6 84.8#
51 347.1 26.9 40.3#









#6
Benzaldehyde
Concen: 0.070 ng/ul
RT: 7.205 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion: 77 Resp: 1561
Ion Ratio Lower Upper
77 100
105 18.1 71.8 107.6#
106 8.0 70.2 105.2#



#8
Phenol
Concen: 0.006 ng/ul
RT: 7.228 min Scan# 670
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion: 94 Resp: 258
Ion Ratio Lower Upper
94 100
65 62.3 22.3 33.5#
66 158.5 28.7 43.1#









#10
Bis(2-Chloroethyl)ether
Concen: 0.007 ng/ul
RT: 7.487 min Scan# 715
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion: 93 Resp: 231
Ion Ratio Lower Upper
93 100
63 62.9 60.1 90.1
95 55.0 25.4 38.2#



#12
2-Chlorophenol
Concen: 0.010 ng/ul
RT: 7.610 min Scan# 735
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion: 128 Resp: 305
Ion Ratio Lower Upper
128 100
64 152.8 41.4 62.0#
130 32.5 26.0 39.0





#13

2-Methylphenol

Concen: 0.002 ng/ul

RT: 8.593 min Scan# 902

Delta R.T. 0.100 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :





SBLK773

Tgt Ion:108 Resp: 62

Ion Ratio Lower Upper

108 100

107 84.2 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.005 ng/ul

RT: 8.563 min Scan# 897

Delta R.T. -0.012 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion: 45 Resp: 246


Ion Ratio Lower Upper

45 100

77 86.0 12.1 18.1#

79 52.1 8.2 12.4#









#16
Acetophenone
Concen: 0.005 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:105 Resp: 278
Ion Ratio Lower Upper
105 100
77 143.8 64.4 96.6#
51 158.4 21.2 31.8#



#17
N-Nitroso-di-n-propylamine
Concen: 0.007 ng/ul
RT: 8.875 min Scan# 950
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion: 70 Resp: 184
Ion Ratio Lower Upper
70 100
42 121.5 43.0 64.4#
101 7.9 6.8 10.2
130 7.3 14.2 21.4#









#18
4-Methylphenol
Concen: 0.008 ng/ul
RT: 8.828 min Scan# 942
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:108 Resp: 268
Ion Ratio Lower Upper
108 100
107 66.8 87.3 130.9#



#19
Hexachloroethane
Concen: 0.005 ng/ul
RT: 9.140 min Scan# 995
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:117 Resp: 68
Ion Ratio Lower Upper
117 100
201 51.1 71.8 107.6#
199 11.1 45.2 67.8#









#22
Nitrobenzene
Concen: 0.013 ng/ul
RT: 9.299 min Scan# 1109
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion: 77 Resp: 506
Ion Ratio Lower Upper
77 100
123 22.4 32.5 48.7#
65 35.0 10.3 15.5#



#23
Isophorone
Concen: 0.006 ng/ul
RT: 9.810 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion: 82 Resp: 425
Ion Ratio Lower Upper
82 100
95 3.0 5.3 7.9#
138 4.6 12.7 19.1#









#25
2-Nitrophenol
Concen: 0.009 ng/ul
RT: 9.999 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:139 Resp: 137
Ion Ratio Lower Upper
139 100
65 409.6 46.4 69.6#
109 79.0 25.3 37.9#



#26
2,4-Dimethylphenol
Concen: 0.006 ng/ul
RT: 10.040 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:107 Resp: 238
Ion Ratio Lower Upper
107 100
121 24.7 39.1 58.7#
122 36.6 66.9 100.3#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 0.002 ng/ul
 RT: 10.416 min Scan# 1193
 Delta R.T. 0.112 min
 Lab File: BP016560.D
 Acq: 14 Aug 2023 14:25




Instrument :
 BNA_P
 ClientSampleId :
 SBLK773


Tgt Ion: 93 Resp: 108
 Ion Ratio Lower Upper
 93 100
 95 38.1 25.8 38.8
 123 12.4 8.7 13.1



#29
 2,4-Dichlorophenol
 Concen: 0.004 ng/ul
 RT: 10.528 min Scan# 1231
 Delta R.T. 0.006 min
 Lab File: BP016560.D
 Acq: 14 Aug 2023 14:25

Tgt Ion: 162 Resp: 97
 Ion Ratio Lower Upper
 162 100
 164 1659.4 52.9 79.3#
 98 121.8 32.7 49.1#









#30
Naphthalene
Concen: 0.004 ng/ul
RT: 10.951 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:	128	Resp:	427
Ion Ratio	Lower	Upper	
128	100		
129	28.4	8.6	13.0#
127	11.0	10.3	15.5



#32
4-Chloroaniline
Concen: 0.020 ng/ul
RT: 11.063 min Scan# 1322
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:	127	Resp:	917
Ion Ratio	Lower	Upper	
127	100		
129	136.0	25.6	38.4#









#33
Hexachlorobutadiene
Concen: 0.002 ng/ul
RT: 11.157 min Scan# 111
Delta R.T. -0.024 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:225 Resp: 29
Ion Ratio Lower Upper
225 100
223 51.6 50.2 75.2
227 51.6 50.8 76.2



#34
Caprolactam
Concen: 0.005 ng/ul
RT: 11.728 min Scan# 1435
Delta R.T. -0.147 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:113 Resp: 51
Ion Ratio Lower Upper
113 100
55 161.2 134.8 202.2
56 134.3 101.5 152.3









#35
4-Chloro-3-methylphenol
Concen: 0.006 ng/ul
RT: 12.157 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:107 Resp: 190
Ion Ratio Lower Upper
107 100
144 20.8 19.0 28.6
142 44.6 60.2 90.4#



#36
2-Methylnaphthalene
Concen: 0.002 ng/ul
RT: 12.551 min Scan# 1575
Delta R.T. 0.012 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:142 Resp: 144
Ion Ratio Lower Upper
142 100
141 80.9 71.4 107.0









#37
1-Methylnaphthalene
Concen: 0.004 ng/ul
RT: 12.757 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:142 Resp: 269
Ion Ratio Lower Upper
142 100
141 87.6 74.5 111.7
116 30.2 3.0 4.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 0.003 ng/ul
RT: 12.898 min Scan# 1634
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:216 Resp: 88
Ion Ratio Lower Upper
216 100
214 40.0 63.8 95.6#
179 113.3 17.9 26.9#
108 50.0 16.3 24.5#









#40
Hexachlorocyclopentadiene
Concen: 0.000 ng/ul
RT: 12.840 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:237	Resp:	5
Ion Ratio	Lower	Upper
237	100	
235	52.9	49.7 74.5
272	11.8	9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 0.007 ng/ul
RT: 13.145 min Scan# 1676
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:196	Resp:	138
Ion Ratio	Lower	Upper
196	100	
198	89.2	77.0 115.4
200	31.9	24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 0.006 ng/ul
RT: 13.145 min Scan# 1676
Delta R.T. -0.059 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:196 Resp: 138
Ion Ratio Lower Upper
196 100
198 89.2 76.7 115.1
200 31.9 24.2 36.2



#43
1,1'-Biphenyl
Concen: 0.005 ng/ul
RT: 13.545 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:154 Resp: 436
Ion Ratio Lower Upper
154 100
153 60.2 31.4 47.2#
76 18.2 13.0 19.4









#44
2-Chloronaphthalene
Concen: 0.005 ng/ul
RT: 13.592 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:	162	Resp:	353
Ion Ratio	Lower	Upper	
162	100		
164	26.4	25.8	38.6
127	56.0	32.9	49.3#



#45
2-Nitroaniline
Concen: 0.008 ng/ul
RT: 13.822 min Scan# 1791
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:	65	Resp:	170
Ion Ratio	Lower	Upper	
65	100		
92	38.9	54.5	81.7#
138	33.8	72.1	108.1#





#47

Dimethylphthalate

Concen: 0.011 ng/ul

RT: 14.169 min Scan# 1850

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:163 Resp: 1052

Ion Ratio Lower Upper

163 100

194 13.4 3.0 4.6#

164 7.9 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 0.014 ng/ul

RT: 14.310 min Scan# 1874

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:165 Resp: 231


Ion Ratio Lower Upper

165 100

89 33.2 53.2 79.8#

121 47.2 18.8 28.2#









#50
Acenaphthylene
Concen: 0.007 ng/ul
RT: 14.434 min Scan# 1896
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:152 Resp: 870
Ion Ratio Lower Upper
152 100
151 39.3 15.7 23.5#
153 34.0 10.6 15.8#



#51
3-Nitroaniline
Concen: 0.006 ng/ul
RT: 14.645 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:138 Resp: 104
Ion Ratio Lower Upper
138 100
108 124.2 10.9 16.3#
92 274.7 105.0 157.6#









#52
Acenaphthene
Concen: 0.006 ng/ul
RT: 14.781 min Scan# 1953
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:153 Resp: 456
Ion Ratio Lower Upper
153 100
152 47.0 38.7 58.1
154 67.8 72.0 108.0#



#53
2,4-Dinitrophenol
Concen: 0.001 ng/ul
RT: 14.839 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:184 Resp: 11
Ion Ratio Lower Upper
184 100
63 676.5 65.0 97.6#
154 147.1 56.2 84.2#









#55
4-Nitrophenol
Concen: 0.032 ng/ul
RT: 14.922 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:109 Resp: 492
Ion Ratio Lower Upper
109 100
139 64.3 101.1 151.7#
65 479.7 108.9 163.3#



#56
Dibenzofuran
Concen: 0.000 ng/ul
RT: 15.057 min Scan# 2001
Delta R.T. -0.053 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:168 Resp: 33
Ion Ratio Lower Upper
168 100
139 34.4 31.4 47.0









#57
2,4-Dinitrotoluene
Concen: 0.009 ng/ul
RT: 15.092 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:165 Resp: 212
Ion Ratio Lower Upper
165 100
63 58.7 40.4 60.6
182 17.5 1.9 2.9#



#58
2,3,4,6-Tetrachlorophenol
Concen: 0.008 ng/ul
RT: 15.322 min Scan# 2046
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:232 Resp: 147
Ion Ratio Lower Upper
232 100
131 74.5 42.9 64.3#
130 0.0 2.2 3.4#
166 29.8 26.7 40.1









#59
Diethylphthalate
Concen: 0.012 ng/ul
RT: 15.522 min Scan# 2080
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:149 Resp: 1185
Ion Ratio Lower Upper
149 100
177 19.0 16.5 24.7
150 19.1 10.0 15.0#



#61
Fluorene
Concen: 0.007 ng/ul
RT: 15.763 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:166 Resp: 672
Ion Ratio Lower Upper
166 100
165 88.0 78.3 117.5
167 7.0 10.8 16.2#









#62
4-Chlorophenyl-phenylether
Concen: 0.008 ng/ul
RT: 15.745 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:204 Resp: 353
Ion Ratio Lower Upper
204 100
206 40.7 25.5 38.3#
141 86.1 52.2 78.2#



#63
4-Nitroaniline
Concen: 0.077 ng/ul
RT: 15.822 min Scan# 2131
Delta R.T. 0.012 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:138 Resp: 1347
Ion Ratio Lower Upper
138 100
92 266.7 46.7 70.1#
108 4451.5 73.7 110.5#









#66
4,6-Dinitro-2-methylphenol
Concen: 0.055 ng/ul
RT: 15.822 min Scan# 2131
Delta R.T. -0.018 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:198 Resp: 820
Ion Ratio Lower Upper
198 100
182 385.8 3.9 5.9#
77 1551.5 25.8 38.6#



#67
N-Nitrosodiphenylamine
Concen: 0.011 ng/ul
RT: 15.969 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:169 Resp: 879
Ion Ratio Lower Upper
169 100
168 72.2 53.9 80.9
167 41.7 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 0.013 ng/ul
RT: 16.651 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:248	Resp:	326
Ion Ratio	Lower	Upper
248	100	
250	85.6	76.9 115.3
141	60.2	67.6 101.4#



#69
Hexachlorobenzene
Concen: 0.010 ng/ul
RT: 16.733 min Scan# 2286
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:284	Resp:	297
Ion Ratio	Lower	Upper
284	100	
142	18.2	31.9 47.9#
249	30.3	24.7 37.1





#70

Atrazine

Concen: 0.004 ng/ul

RT: 16.922 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:200 Resp: 118

Ion Ratio Lower Upper

200 100

173 54.9 22.2 33.2#

215 8.0 38.4 57.6#



#71

Pentachlorophenol

Concen: 0.008 ng/ul

RT: 17.092 min Scan# 2347

Delta R.T. -0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:266 Resp: 139


Ion Ratio Lower Upper

266 100

264 43.2 50.7 76.1#

268 53.8 53.0 79.6





#72

Phenanthrene

Concen: 0.016 ng/ul

RT: 17.516 min Scan# 2419

Delta R.T. -0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:178 Resp: 2459

Ion Ratio Lower Upper

178 100

179 18.1 12.4 18.6

176 28.7 15.5 23.3#



#74

Anthracene

Concen: 0.014 ng/ul

RT: 17.616 min Scan# 2436

Delta R.T. 0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:178 Resp: 2132


Ion Ratio Lower Upper

178 100

179 20.1 12.5 18.7#

176 25.9 14.9 22.3#





#75

1,2,3,4-Tetrachlorobenzene

Concen: 0.002 ng/uL

RT: 13.504 min Scan# 1737

Delta R.T. 0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:216 Resp: 82

Ion Ratio Lower Upper

216 100

214 93.7 65.4 98.2

218 40.0 39.3 58.9



#76

Pentachlorobenzene

Concen: 0.004 ng/uL

RT: 14.998 min Scan# 1991

Delta R.T. -0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:250 Resp: 114


Ion Ratio Lower Upper

250 100

248 15.2 50.2 75.2#

252 88.4 51.0 76.4#









#77
Carbazole
Concen: 0.007 ng/ul
RT: 17.886 min Scan# 2483
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:167 Resp: 1042
Ion Ratio Lower Upper
167 100
166 38.6 16.8 25.2#
139 16.7 10.6 16.0#



#78
Di-n-butylphthalate
Concen: 0.016 ng/ul
RT: 18.404 min Scan# 2570
Delta R.T. -0.000 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:149 Resp: 2631
Ion Ratio Lower Upper
149 100
150 7.1 7.3 10.9#
104 9.0 4.6 6.8#





#80

Fluoranthene

Concen: 0.014 ng/ul

RT: 19.480 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:202 Resp: 2353

Ion Ratio Lower Upper

202 100

101 38.6 10.6 16.0#

100 27.2 8.1 12.1#



#82

Pyrene

Concen: 0.015 ng/ul

RT: 19.833 min Scan# 2813

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:202 Resp: 2643


Ion Ratio Lower Upper

202 100

101 26.7 13.4 20.2#

100 24.1 11.0 16.4#





#83

Butylbenzylphthalate

Concen: 0.008 ng/ul

RT: 20.686 min Scan# 2958

Delta R.T. -0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:149 Resp: 618

Ion Ratio Lower Upper

149 100

91 64.8 67.8 101.6#

206 23.7 14.4 21.6#



#84

3,3'-Dichlorobenzidine

Concen: 0.025 ng/ul

RT: 21.492 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:252 Resp: 1469


Ion Ratio Lower Upper

252 100

254 50.3 51.3 76.9#

126 21.3 10.2 15.4#





#85

Benzo(a)anthracene

Concen: 0.057 ng/ul

RT: 21.569 min Scan# 3106

Delta R.T. 0.012 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:228 Resp: 10513

Ion Ratio Lower Upper

228 100

229 24.6 15.8 23.6#

226 10.1 21.8 32.6#



#86

Bis(2-ethylhexyl)phthalate

Concen: 0.023 ng/ul

RT: 21.433 min Scan# 3085

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25




Tgt Ion:149 Resp: 2478


Ion Ratio Lower Upper

149 100

167 21.2 20.9 31.3

279 6.3 2.6 4.0#





#87

Chrysene

Concen: 0.017 ng/ul

RT: 21.610 min Scan# 3115

Delta R.T. -0.000 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:228 Resp: 2877

Ion Ratio Lower Upper

228 100

226 41.6 24.6 36.8#

229 18.2 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 0.014 ng/ul




RT: 22.427 min Scan# 3254


Delta R.T. 0.006 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Tgt Ion:149 Resp: 2530









#90
Benzo(b)fluoranthene
Concen: 0.024 ng/ul
RT: 23.345 min Scan# 3410
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:252 Resp: 4147
Ion Ratio Lower Upper
252 100
253 45.6 17.4 26.2#
125 51.8 10.0 15.0#



#91
Benzo(k)fluoranthene
Concen: 0.020 ng/ul
RT: 23.398 min Scan# 3419
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:252 Resp: 3426
Ion Ratio Lower Upper
252 100
253 46.1 18.0 27.0#
125 47.4 9.8 14.8#









#93
Benzo(a)pyrene
Concen: 0.019 ng/ul
RT: 24.033 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25




Instrument :
BNA_P
ClientSampleId :
SBLK773


Tgt Ion:252 Resp: 3210
Ion Ratio Lower Upper
252 100
253 41.8 17.7 26.5#
125 88.4 10.6 15.8#



#94
Indeno(1,2,3-cd)pyrene
Concen: 0.003 ng/ul
RT: 26.886 min Scan# 4012
Delta R.T. -0.024 min
Lab File: BP016560.D
Acq: 14 Aug 2023 14:25

Tgt Ion:276 Resp: 534
Ion Ratio Lower Upper
276 100
138 27.3 21.7 32.5
277 24.4 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 0.011 ng/ul

RT: 26.951 min Scan# 4021

Delta R.T. 0.012 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Instrument :

BNA_P

ClientSampleId :

SBLK773





Tgt Ion:278 Resp: 1724

Ion Ratio Lower Upper

278 100

139 27.8 15.5 23.3#

279 40.3 19.1 28.7#



#96

Benzo(g,h,i)perylene

Concen: 0.003 ng/ul

RT: 27.756 min Scan# 4160

Delta R.T. -0.018 min

Lab File: BP016560.D

Acq: 14 Aug 2023 14:25

Tgt Ion:276 Resp: 494

Ion Ratio Lower Upper

276 100

138 57.2 21.5 32.3#

277 34.8 19.3 28.9#

