

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081523\
 Data File : BP016594.D
 Acq On : 15 Aug 2023 15:36
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020684

Quant Time: Aug 15 16:06:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 18:26:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.058	152	243849	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	1179245	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.704	164	760974	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.475	188	1798207	20.000	ng/u1	0.00
79) Chrysene-d12	21.569	240	1887386	20.000	ng/u1	0.00
88) Perylene-d12	24.157	264	2080666	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	52067	8.114	ng/uL	0.00
4) Pyridine-d5	3.828	84	417353	21.376	ng/u1	0.00
7) Phenol-d5	7.193	99	485121	21.579	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.393	67	305625	20.641	ng/u1	0.00
11) 2-Chlorophenol-d4	7.575	132	358652	22.212	ng/u1	0.00
15) 4-Methylphenol-d8	8.758	113	396426	22.177	ng/u1	0.00
21) Nitrobenzene-d5	9.252	128	181238	21.842	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	182052	23.101	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.493	165	354036	22.828	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	578596	21.743	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	1132376	20.573	ng/u1	0.00
49) Acenaphthylene-d8	14.404	160	1343128	21.620	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	259330	23.128	ng/u1	0.00
60) Fluorene-d10	15.698	176	994499	21.194	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	153177	17.906	ng/u1	0.00
73) Anthracene-d10	17.575	188	1656330	21.295	ng/u1	0.00
81) Pyrene-d10	19.804	212	2031390	20.541	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	2100505	21.065	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	59010	8.363	ng/uL	98
5) Pyridine	3.852	79	431484	21.300	ng/u1	98
6) Benzaldehyde	7.211	77	316660	25.552	ng/u1	99
8) Phenol	7.222	94	518839	21.593	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.487	93	403817	20.805	ng/u1	99
12) 2-Chlorophenol	7.611	128	369419	21.744	ng/u1	99
13) 2-Methylphenol	8.493	108	378716	21.699	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.569	45	553550	20.463	ng/u1	99
16) Acetophenone	8.905	105	632096	21.897	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.875	70	334179	21.586	ng/u1	99
18) 4-Methylphenol	8.822	108	418244	21.804	ng/u1	94
19) Hexachloroethane	9.134	117	146780	20.679	ng/u1	97
22) Nitrobenzene	9.299	77	494681	21.297	ng/u1	98
23) Isophorone	9.810	82	931067	21.222	ng/u1	99
25) 2-Nitrophenol	9.999	139	204423	22.690	ng/u1	98
26) 2,4-Dimethylphenol	10.040	107	454426	21.395	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.299	93	566023	20.127	ng/u1	100
29) 2,4-Dichlorophenol	10.522	162	356731	22.356	ng/u1	99
30) Naphthalene	10.940	128	1273675	20.672	ng/u1	100
32) 4-Chloroaniline	11.069	127	575547	21.837	ng/u1	99
33) Hexachlorobutadiene	11.175	225	198259	20.611	ng/u1	100
34) Caprolactam	11.869	113	143113	23.053	ng/u1	99
35) 4-Chloro-3-methylphenol	12.157	107	442070	22.665	ng/u1	97
36) 2-Methylnaphthalene	12.540	142	866317	20.991	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081523\
 Data File : BP016594.D
 Acq On : 15 Aug 2023 15:36
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020684

Quant Time: Aug 15 16:06:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 18:26:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	879388	21.162	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.887	216	418651	20.569	ng/ul	99
40) Hexachlorocyclopentadiene	12.840	237	196518	15.868	ng/ul	98
41) 2,4,6-Trichlorophenol	13.134	196	283814	22.405	ng/ul	98
42) 2,4,5-Trichlorophenol	13.204	196	312024	22.724	ng/ul	98
43) 1,1'-Biphenyl	13.546	154	1138103	20.121	ng/ul	98
44) 2-Chloronaphthalene	13.593	162	908100	20.634	ng/ul	99
45) 2-Nitroaniline	13.816	65	306120	23.005	ng/ul	95
47) Dimethylphthalate	14.163	163	1152923	20.487	ng/ul	99
48) 2,6-Dinitrotoluene	14.304	165	220319	22.603	ng/ul	99
50) Acenaphthylene	14.434	152	1553248	21.230	ng/ul	100
51) 3-Nitroaniline	14.640	138	264182	24.172	ng/ul	98
52) Acenaphthene	14.769	153	1007378	20.569	ng/ul	99
53) 2,4-Dinitrophenol	14.840	184	87237	16.076	ng/ul	94
55) 4-Nitrophenol	14.922	109	211412	23.193	ng/ul	98
56) Dibenzofuran	15.104	168	1393014	20.847	ng/ul	99
57) 2,4-Dinitrotoluene	15.081	165	340562	23.331	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.316	232	273428	23.767	ng/ul	98
59) Diethylphthalate	15.516	149	1186105	20.614	ng/ul	99
61) Fluorene	15.757	166	1163908	21.243	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.745	204	538372	20.932	ng/ul	99
63) 4-Nitroaniline	15.804	138	280486	26.817	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.834	198	168070	17.719	ng/ul	100
67) N-Nitrosodiphenylamine	15.969	169	992240	20.356	ng/ul	99
68) 4-Bromophenyl-phenylether	16.651	248	327045	20.370	ng/ul	100
69) Hexachlorobenzene	16.734	284	381026	20.310	ng/ul	99
70) Atrazine	16.922	200	365582	20.847	ng/ul	99
71) Pentachlorophenol	17.092	266	253377	21.937	ng/ul	98
72) Phenanthrene	17.516	178	1958686	20.675	ng/ul	100
74) Anthracene	17.610	178	2007007	21.091	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.498	216	445406	19.491	ng/ul	95
76) Pentachlorobenzene	14.998	250	406753	19.837	ng/ul	99
77) Carbazole	17.886	167	1925558	21.958	ng/ul	99
78) Di-n-butylphthalate	18.398	149	2222997	21.898	ng/ul	99
80) Fluoranthene	19.475	202	2389524	20.245	ng/ul	98
82) Pyrene	19.833	202	2534926	20.338	ng/ul	98
83) Butylbenzylphthalate	20.692	149	1110254	22.112	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.492	252	851570	20.720	ng/ul	99
85) Benzo(a)anthracene	21.551	228	2646879	20.867	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.433	149	1620903	22.244	ng/ul	98
87) Chrysene	21.610	228	2463479	20.754	ng/ul	99
89) Di-n-octyl phthalate	22.422	149	2869910	22.105	ng/ul	100
90) Benzo(b)fluoranthene	23.351	252	2566426	20.716	ng/ul	99
91) Benzo(k)fluoranthene	23.404	252	2633501	21.427	ng/ul	99
93) Benzo(a)pyrene	24.039	252	2435127	20.837	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.910	276	2665394	19.324	ng/ul	100
95) Dibenzo(a,h)anthracene	26.939	278	2236378	19.483	ng/ul	98
96) Benzo(g,h,i)perylene	27.768	276	2102624	19.154	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 8.363 ng/uL
 RT: 3.428 min Scan# 24
 Delta R.T. -0.006 min
 Lab File: BP016594.D
 Acq: 15 Aug 2023 15:36




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020684


Tgt Ion:	88	Resp:	59010
Ion	Ratio	Lower	Upper
88	100		
43	27.1	23.5	35.3
58	72.2	57.1	85.7



#5
 Pyridine
 Concen: 21.300 ng/uL
 RT: 3.852 min Scan# 96
 Delta R.T. -0.000 min
 Lab File: BP016594.D
 Acq: 15 Aug 2023 15:36

Tgt Ion:	79	Resp:	431484
Ion	Ratio	Lower	Upper
79	100		
52	69.0	56.6	84.8
51	32.6	26.9	40.3









#6
Benzaldehyde
Concen: 25.552 ng/ul
RT: 7.211 min Scan# 60
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion: 77 Resp: 316660
Ion Ratio Lower Upper
77 100
105 89.7 71.8 107.6
106 86.4 70.2 105.2



#8
Phenol
Concen: 21.593 ng/ul
RT: 7.222 min Scan# 669
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion: 94 Resp: 518839
Ion Ratio Lower Upper
94 100
65 29.0 22.3 33.5
66 36.5 28.7 43.1









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.805 ng/ul
 RT: 7.487 min Scan# 714
 Delta R.T. -0.006 min
 Lab File: BP016594.D
 Acq: 15 Aug 2023 15:36




Instrument :
 BNA_P
 ClientSampleId :
 SST020684


Tgt Ion: 93 Resp: 403817
 Ion Ratio Lower Upper
 93 100
 63 76.3 60.1 90.1
 95 31.7 25.4 38.2



#12
 2-Chlorophenol
 Concen: 21.744 ng/ul
 RT: 7.611 min Scan# 735
 Delta R.T. 0.000 min
 Lab File: BP016594.D
 Acq: 15 Aug 2023 15:36

Tgt Ion:128 Resp: 369419
 Ion Ratio Lower Upper
 128 100
 64 50.3 41.4 62.0
 130 32.7 26.0 39.0









#13
2-Methylphenol
Concen: 21.699 ng/ul
RT: 8.493 min Scan# 885
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:108 Resp: 378716
Ion Ratio Lower Upper
108 100
107 88.5 70.5 105.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.463 ng/ul
RT: 8.569 min Scan# 898
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion: 45 Resp: 553550
Ion Ratio Lower Upper
45 100
77 15.3 12.1 18.1
79 10.7 8.2 12.4









#16
Acetophenone
Concen: 21.897 ng/ul
RT: 8.905 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:105 Resp: 632096
Ion Ratio Lower Upper
105 100
77 80.0 64.4 96.6
51 26.8 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 21.586 ng/ul
RT: 8.875 min Scan# 950
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion: 70 Resp: 334179
Ion Ratio Lower Upper
70 100
42 53.1 43.0 64.4
101 9.4 6.8 10.2
130 18.9 14.2 21.4









#18
4-Methylphenol
Concen: 21.804 ng/ul
RT: 8.822 min Scan# 941
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:108 Resp: 418244
Ion Ratio Lower Upper
108 100
107 114.9 87.3 130.9



#19
Hexachloroethane
Concen: 20.679 ng/ul
RT: 9.134 min Scan# 994
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:117 Resp: 146780
Ion Ratio Lower Upper
117 100
201 86.5 71.8 107.6
199 54.7 45.2 67.8









#22
Nitrobenzene
Concen: 21.297 ng/ul
RT: 9.299 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion: 77 Resp: 494681
Ion Ratio Lower Upper
77 100
123 41.8 32.5 48.7
65 12.6 10.3 15.5



#23
Isophorone
Concen: 21.222 ng/ul
RT: 9.810 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion: 82 Resp: 931067
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 15.6 12.7 19.1









#25
2-Nitrophenol
Concen: 22.690 ng/ul
RT: 9.999 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:139 Resp: 204423
Ion Ratio Lower Upper
139 100
65 59.1 46.4 69.6
109 33.1 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 21.395 ng/ul
RT: 10.040 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:107 Resp: 454426
Ion Ratio Lower Upper
107 100
121 49.4 39.1 58.7
122 83.4 66.9 100.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.127 ng/ul

RT: 10.299 min Scan# 1192

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion: 93 Resp: 566023

Ion Ratio Lower Upper

93 100

95 32.4 25.8 38.8

123 10.7 8.7 13.1



#29

2,4-Dichlorophenol

Concen: 22.356 ng/ul

RT: 10.522 min Scan# 1230

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:162 Resp: 356731


Ion Ratio Lower Upper

162 100

164 65.0 52.9 79.3

98 39.9 32.7 49.1









#30
Naphthalene
Concen: 20.672 ng/ul
RT: 10.940 min Scan# 1301
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:128 Resp: 1273675
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 21.837 ng/ul
RT: 11.069 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:127 Resp: 575547
Ion Ratio Lower Upper
127 100
129 32.7 25.6 38.4





#33

Hexachlorobutadiene

Concen: 20.611 ng/ul

RT: 11.175 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:225 Resp: 198259

Ion Ratio Lower Upper

225 100

223 62.5 50.2 75.2

227 63.4 50.8 76.2



#34

Caprolactam

Concen: 23.053 ng/ul

RT: 11.869 min Scan# 1459

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:113 Resp: 143113


Ion Ratio Lower Upper

113 100

55 168.6 134.8 202.2

56 124.3 101.5 152.3









#35
4-Chloro-3-methylphenol
Concen: 22.665 ng/ul
RT: 12.157 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:107 Resp: 442070
Ion Ratio Lower Upper
107 100
144 24.6 19.0 28.6
142 78.0 60.2 90.4



#36
2-Methylnaphthalene
Concen: 20.991 ng/ul
RT: 12.540 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:142 Resp: 866317
Ion Ratio Lower Upper
142 100
141 89.1 71.4 107.0





#37

1-Methylnaphthalene

Concen: 21.162 ng/ul

RT: 12.757 min Scan# 1610

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:142 Resp: 879388

Ion Ratio Lower Upper

142 100

141 92.3 74.5 111.7

116 4.1 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.569 ng/ul

RT: 12.887 min Scan# 1632

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Tgt Ion:216 Resp: 418651




Ion Ratio Lower Upper


216 100

214 78.3 63.8 95.6

179 22.1 17.9 26.9

108 19.9 16.3 24.5









#40
Hexachlorocyclopentadiene
Concen: 15.868 ng/ul
RT: 12.840 min Scan# 1624
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:237 Resp: 196518
Ion Ratio Lower Upper
237 100
235 63.5 49.7 74.5
272 12.7 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 22.405 ng/ul
RT: 13.134 min Scan# 1674
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:196 Resp: 283814
Ion Ratio Lower Upper
196 100
198 94.3 77.0 115.4
200 29.8 24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 22.724 ng/ul
RT: 13.204 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:196 Resp: 312024
Ion Ratio Lower Upper
196 100
198 98.1 76.7 115.1
200 31.0 24.2 36.2



#43
1,1'-Biphenyl
Concen: 20.121 ng/ul
RT: 13.546 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:154 Resp: 1138103
Ion Ratio Lower Upper
154 100
153 39.9 31.4 47.2
76 14.4 13.0 19.4





#44

2-Chloronaphthalene

Concen: 20.634 ng/ul

RT: 13.593 min Scan# 1752

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:162 Resp: 908100

Ion Ratio Lower Upper

162 100

164 32.2 25.8 38.6

127 39.8 32.9 49.3



#45

2-Nitroaniline

Concen: 23.005 ng/ul

RT: 13.816 min Scan# 1790

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion: 65 Resp: 306120


Ion Ratio Lower Upper

65 100

92 69.8 54.5 81.7

138 96.2 72.1 108.1









#47
Dimethylphthalate
Concen: 20.487 ng/ul
RT: 14.163 min Scan# 1850
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:163 Resp: 1152923
Ion Ratio Lower Upper
163 100
194 3.7 3.0 4.6
164 10.3 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 22.603 ng/ul
RT: 14.304 min Scan# 1873
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:165 Resp: 220319
Ion Ratio Lower Upper
165 100
89 65.5 53.2 79.8
121 23.8 18.8 28.2





#50

Acenaphthylene

Concen: 21.230 ng/ul

RT: 14.434 min Scan# 1895

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:152 Resp: 1553248

Ion Ratio Lower Upper

152 100

151 19.9 15.7 23.5

153 13.2 10.6 15.8



#51

3-Nitroaniline

Concen: 24.172 ng/ul

RT: 14.640 min Scan# 1930

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:138 Resp: 264182


Ion Ratio Lower Upper

138 100

108 12.6 10.9 16.3

92 128.7 105.0 157.6









#52
Acenaphthene
Concen: 20.569 ng/ul
RT: 14.769 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:153 Resp: 1007378
Ion Ratio Lower Upper
153 100
152 48.3 38.7 58.1
154 88.2 72.0 108.0



#53
2,4-Dinitrophenol
Concen: 16.076 ng/ul
RT: 14.840 min Scan# 1964
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:184 Resp: 87237
Ion Ratio Lower Upper
184 100
63 74.8 65.0 97.6
154 66.7 56.2 84.2









#55
4-Nitrophenol
Concen: 23.193 ng/ul
RT: 14.922 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:109 Resp: 211412
Ion Ratio Lower Upper
109 100
139 129.5 101.1 151.7
65 133.6 108.9 163.3



#56
Dibenzofuran
Concen: 20.847 ng/ul
RT: 15.104 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:168 Resp: 1393014
Ion Ratio Lower Upper
168 100
139 38.5 31.4 47.0





#57

2,4-Dinitrotoluene

Concen: 23.331 ng/ul

RT: 15.081 min Scan# 2006

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:165 Resp: 340562

Ion Ratio Lower Upper

165 100

63 53.0 40.4 60.6

182 2.6 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.767 ng/ul

RT: 15.316 min Scan# 2045

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Tgt Ion:232 Resp: 273428




Ion Ratio Lower Upper


232 100

131 53.6 42.9 64.3

130 2.7 2.2 3.4

166 36.3 26.7 40.1









#59
Diethylphthalate
Concen: 20.614 ng/ul
RT: 15.516 min Scan# 2080
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:149 Resp: 1186105
Ion Ratio Lower Upper
149 100
177 21.2 16.5 24.7
150 12.5 10.0 15.0



#61
Fluorene
Concen: 21.243 ng/ul
RT: 15.757 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:166 Resp: 1163908
Ion Ratio Lower Upper
166 100
165 96.5 78.3 117.5
167 13.3 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.932 ng/ul
RT: 15.745 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:204 Resp: 538372
Ion Ratio Lower Upper
204 100
206 33.1 25.5 38.3
141 65.8 52.2 78.2



#63
4-Nitroaniline
Concen: 26.817 ng/ul
RT: 15.804 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:138 Resp: 280486
Ion Ratio Lower Upper
138 100
92 57.0 46.7 70.1
108 90.3 73.7 110.5









#66
4,6-Dinitro-2-methylphenol
Concen: 17.719 ng/ul
RT: 15.834 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:198 Resp: 168070
Ion Ratio Lower Upper
198 100
182 5.3 3.9 5.9
77 32.2 25.8 38.6



#67
N-Nitrosodiphenylamine
Concen: 20.356 ng/ul
RT: 15.969 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:169 Resp: 992240
Ion Ratio Lower Upper
169 100
168 65.7 53.9 80.9
167 36.3 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 20.370 ng/ul
RT: 16.651 min Scan# 2272
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:248 Resp: 327045
Ion Ratio Lower Upper
248 100
250 96.7 76.9 115.3
141 84.6 67.6 101.4



#69
Hexachlorobenzene
Concen: 20.310 ng/ul
RT: 16.734 min Scan# 2286
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:284 Resp: 381026
Ion Ratio Lower Upper
284 100
142 41.0 31.9 47.9
249 30.4 24.7 37.1





#70

Atrazine

Concen: 20.847 ng/ul

RT: 16.922 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:200 Resp: 365582

Ion Ratio Lower Upper

200 100

173 27.3 22.2 33.2

215 47.6 38.4 57.6



#71

Pentachlorophenol

Concen: 21.937 ng/ul

RT: 17.092 min Scan# 2347

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:266 Resp: 253377


Ion Ratio Lower Upper

266 100

264 63.5 50.7 76.1

268 63.5 53.0 79.6





#72

Phenanthrene

Concen: 20.675 ng/ul

RT: 17.516 min Scan# 2419

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:178 Resp: 1958686

Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 19.5 15.5 23.3



#74

Anthracene

Concen: 21.091 ng/ul

RT: 17.610 min Scan# 2435

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:178 Resp: 2007007


Ion Ratio Lower Upper

178 100

179 14.9 12.5 18.7

176 18.4 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.491 ng/uL

RT: 13.498 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:216 Resp: 445406

Ion Ratio Lower Upper

216 100

214 76.8 65.4 98.2

218 47.0 39.3 58.9



#76

Pentachlorobenzene

Concen: 19.837 ng/uL

RT: 14.998 min Scan# 1991

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:250 Resp: 406753


Ion Ratio Lower Upper

250 100

248 64.3 50.2 75.2

252 63.6 51.0 76.4





#77

Carbazole

Concen: 21.958 ng/ul

RT: 17.886 min Scan# 2483

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:167 Resp: 1925558

Ion Ratio Lower Upper

167 100

166 21.8 16.8 25.2

139 13.2 10.6 16.0



#78

Di-n-butylphthalate

Concen: 21.898 ng/ul

RT: 18.398 min Scan# 2569

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:149 Resp: 2222997


Ion Ratio Lower Upper

149 100

150 9.4 7.3 10.9

104 5.8 4.6 6.8









#80
Fluoranthene
Concen: 20.245 ng/ul
RT: 19.475 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:202 Resp: 2389524
Ion Ratio Lower Upper
202 100
101 14.0 10.6 16.0
100 10.7 8.1 12.1



#82
Pyrene
Concen: 20.338 ng/ul
RT: 19.833 min Scan# 2813
Delta R.T. 0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:202 Resp: 2534926
Ion Ratio Lower Upper
202 100
101 15.6 13.4 20.2
100 13.1 11.0 16.4





#83

Butylbenzylphthalate

Concen: 22.112 ng/ul

RT: 20.692 min Scan# 2959

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:149 Resp: 1110254

Ion Ratio Lower Upper

149 100

91 79.4 67.8 101.6

206 19.6 14.4 21.6



#84

3,3'-Dichlorobenzidine

Concen: 20.720 ng/ul

RT: 21.492 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36




Tgt Ion:252 Resp: 851570


Ion Ratio Lower Upper

252 100

254 64.5 51.3 76.9

126 12.2 10.2 15.4









#85
Benzo(a)anthracene
Concen: 20.867 ng/ul
RT: 21.551 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:228 Resp: 2646879
Ion Ratio Lower Upper
228 100
229 20.0 15.8 23.6
226 27.5 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.244 ng/ul
RT: 21.433 min Scan# 3085
Delta R.T. 0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:149 Resp: 1620903
Ion Ratio Lower Upper
149 100
167 27.2 20.9 31.3
279 3.7 2.6 4.0





#87

Chrysene

Concen: 20.754 ng/ul

RT: 21.610 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:228 Resp: 2463479

Ion Ratio Lower Upper

228 100

226 30.6 24.6 36.8

229 19.8 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 22.105 ng/ul




RT: 22.422 min Scan# 3253


Delta R.T. 0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Tgt Ion:149 Resp: 2869910









#90
Benzo(b)fluoranthene
Concen: 20.716 ng/ul
RT: 23.351 min Scan# 3411
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:252 Resp: 2566426
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 11.6 10.0 15.0



#91
Benzo(k)fluoranthene
Concen: 21.427 ng/ul
RT: 23.404 min Scan# 3420
Delta R.T. 0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:252 Resp: 2633501
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 12.0 9.8 14.8









#93
Benzo(a)pyrene
Concen: 20.837 ng/ul
RT: 24.039 min Scan# 3529
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36




Instrument :
BNA_P
ClientSampleId :
SSTD020684


Tgt Ion:252 Resp: 2435127
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 13.5 10.6 15.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.324 ng/ul
RT: 26.910 min Scan# 4016
Delta R.T. -0.000 min
Lab File: BP016594.D
Acq: 15 Aug 2023 15:36

Tgt Ion:276 Resp: 2665394
Ion Ratio Lower Upper
276 100
138 26.6 21.7 32.5
277 24.4 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 19.483 ng/ul

RT: 26.939 min Scan# 4022

Delta R.T. -0.000 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Instrument :

BNA_P

ClientSampleId :

SSTD020684





Tgt Ion:278 Resp: 2236378

Ion Ratio Lower Upper

278 100

139 18.5 15.5 23.3

279 23.2 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 19.154 ng/ul

RT: 27.768 min Scan# 4162

Delta R.T. -0.006 min

Lab File: BP016594.D

Acq: 15 Aug 2023 15:36

Tgt Ion:276 Resp: 2102624

Ion Ratio Lower Upper

276 100

138 25.2 21.5 32.3

277 23.5 19.3 28.9

