

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081623\
 Data File : BP016612.D
 Acq On : 16 Aug 2023 20:23
 Operator : MA/JU
 Sample : 03967-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A48L9

Quant Time: Aug 17 00:48:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 14 18:26:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.057	152	191544	20.000	ng/ul	0.00
20) Naphthalene-d8	10.887	136	903863	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.704	164	594618	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.469	188	1385978	20.000	ng/ul	0.00
79) Chrysene-d12	21.563	240	1421417	20.000	ng/ul	-0.01
88) Perylene-d12	24.145	264	1531168	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	41785	8.289	ng/uL	0.00
4) Pyridine-d5	3.828	84	512585	33.423	ng/ul	0.00
7) Phenol-d5	7.193	99	630368	35.697	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.387	67	400568	34.440	ng/ul	0.00
11) 2-Chlorophenol-d4	7.575	132	439722	34.669	ng/ul	0.00
15) 4-Methylphenol-d8	8.751	113	514441	36.638	ng/ul	-0.01
21) Nitrobenzene-d5	9.251	128	232848	36.612	ng/ul	0.00
24) 2-Nitrophenol-d4	9.969	143	243500	40.311	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.493	165	450379	37.887	ng/ul	0.00
31) 4-Chloroaniline-d4	11.040	131	732994	35.938	ng/ul	0.00
46) Dimethylphthalate-d6	14.116	166	1727453	40.166	ng/ul	0.00
49) Acenaphthylene-d8	14.404	160	1928846	39.734	ng/ul	0.00
54) 4-Nitrophenol-d4	14.904	143	315473	36.007	ng/ul	0.00
60) Fluorene-d10	15.698	176	1533021	41.811	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.816	200	180110	27.316	ng/ul	0.00
73) Anthracene-d10	17.569	188	2453092	40.919	ng/ul	0.00
81) Pyrene-d10	19.798	212	3064632	41.147	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.974	264	3045150	41.499	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081623\
Data File : BP016612.D
Acq On : 16 Aug 2023 20:23
Operator : MA/JU
Sample : 03967-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
A48L9

Quant Time: Aug 17 00:48:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 14 18:26:05 2023
Response via : Initial Calibration

