

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081823\
 Data File : BP016641.D
 Acq On : 18 Aug 2023 15:14
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020691

Quant Time: Aug 18 22:31:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 17 01:38:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	389796	20.000	ng/ul	0.00
20) Naphthalene-d8	10.881	136	1764410	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.704	164	1087692	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.469	188	2385043	20.000	ng/ul	0.00
79) Chrysene-d12	21.569	240	2344597	20.000	ng/ul	0.00
88) Perylene-d12	24.157	264	2732185	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	75643	7.374	ng/uL	0.00
4) Pyridine-d5	3.828	84	533211	17.085	ng/ul	0.00
7) Phenol-d5	7.187	99	618148	17.201	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.387	67	402421	17.002	ng/ul	0.00
11) 2-Chlorophenol-d4	7.569	132	464328	17.989	ng/ul	0.00
15) 4-Methylphenol-d8	8.752	113	506201	17.715	ng/ul	0.00
21) Nitrobenzene-d5	9.246	128	234516	18.890	ng/ul	0.00
24) 2-Nitrophenol-d4	9.963	143	241551	20.485	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	448263	19.317	ng/ul	0.00
31) 4-Chloroaniline-d4	11.040	131	712803	17.903	ng/ul	0.00
46) Dimethylphthalate-d6	14.110	166	1409658	17.918	ng/ul	0.00
49) Acenaphthylene-d8	14.404	160	1651984	18.604	ng/ul	0.00
54) 4-Nitrophenol-d4	14.904	143	299068	18.661	ng/ul	0.00
60) Fluorene-d10	15.698	176	1200721	17.903	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.816	200	197566	17.412	ng/ul	0.00
73) Anthracene-d10	17.569	188	1875593	18.180	ng/ul	0.00
81) Pyrene-d10	19.804	212	2202046	17.924	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.980	264	2333580	17.822	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	80050	7.097	ng/uL	97
5) Pyridine	3.846	79	561610	17.344	ng/ul	100
6) Benzaldehyde	7.205	77	409498	20.672	ng/ul	97
8) Phenol	7.216	94	657561	17.120	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.481	93	528694	17.040	ng/ul	96
12) 2-Chlorophenol	7.605	128	477769	17.592	ng/ul	97
13) 2-Methylphenol	8.487	108	488153	17.497	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.563	45	733245	16.957	ng/ul	99
16) Acetophenone	8.899	105	811181	17.580	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.869	70	449351	18.158	ng/ul	98
18) 4-Methylphenol	8.822	108	532736	17.374	ng/ul	99
19) Hexachloroethane	9.128	117	196957	17.358	ng/ul	96
22) Nitrobenzene	9.293	77	637745	18.350	ng/ul	97
23) Isophorone	9.805	82	1244849	18.964	ng/ul	99
25) 2-Nitrophenol	9.993	139	268122	19.890	ng/ul	99
26) 2,4-Dimethylphenol	10.034	107	578094	18.191	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.293	93	747221	17.758	ng/ul	100
29) 2,4-Dichlorophenol	10.516	162	453637	19.000	ng/ul	99
30) Naphthalene	10.934	128	1605199	17.413	ng/ul	100
32) 4-Chloroaniline	11.063	127	700644	17.767	ng/ul	99
33) Hexachlorobutadiene	11.169	225	258102	17.934	ng/ul	100
34) Caprolactam	11.869	113	177757	19.137	ng/ul	99
35) 4-Chloro-3-methylphenol	12.151	107	553068	18.952	ng/ul	99
36) 2-Methylnaphthalene	12.534	142	1093453	17.708	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081823\
 Data File : BP016641.D
 Acq On : 18 Aug 2023 15:14
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020691

Quant Time: Aug 18 22:31:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 17 01:38:34 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.751	142	1107846	17.818	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.881	216	523238	17.985	ng/ul	99
40) Hexachlorocyclopentadiene	12.834	237	291488	16.467	ng/ul	100
41) 2,4,6-Trichlorophenol	13.134	196	356828	19.708	ng/ul	99
42) 2,4,5-Trichlorophenol	13.198	196	387983	19.769	ng/ul	99
43) 1,1'-Biphenyl	13.540	154	1431237	17.702	ng/ul	98
44) 2-Chloronaphthalene	13.587	162	1105777	17.579	ng/ul	99
45) 2-Nitroaniline	13.810	65	379012	19.927	ng/ul	96
47) Dimethylphthalate	14.157	163	1432878	17.814	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	278142	19.964	ng/ul	99
50) Acenaphthylene	14.434	152	1900591	18.174	ng/ul	99
51) 3-Nitroaniline	14.640	138	314474	20.130	ng/ul	95
52) Acenaphthene	14.769	153	1227394	17.533	ng/ul	98
53) 2,4-Dinitrophenol	14.834	184	121017	15.603	ng/ul	97
55) 4-Nitrophenol	14.916	109	238787	18.327	ng/ul	97
56) Dibenzofuran	15.104	168	1665686	17.440	ng/ul	99
57) 2,4-Dinitrotoluene	15.081	165	411911	19.743	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.316	232	323876	19.696	ng/ul	97
59) Diethylphthalate	15.516	149	1470158	17.876	ng/ul	99
61) Fluorene	15.751	166	1385207	17.688	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.745	204	646708	17.591	ng/ul	97
63) 4-Nitroaniline	15.804	138	311579	20.841	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.834	198	220291	17.510	ng/ul	97
67) N-Nitrosodiphenylamine	15.969	169	1170613	18.107	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	382780	17.975	ng/ul	98
69) Hexachlorobenzene	16.728	284	443370	17.819	ng/ul	97
70) Atrazine	16.922	200	429831	18.480	ng/ul	99
71) Pentachlorophenol	17.092	266	292697	19.106	ng/ul	99
72) Phenanthrene	17.516	178	2197163	17.486	ng/ul	99
74) Anthracene	17.604	178	2249291	17.821	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.493	216	553302	18.255	ng/ul	98
76) Pentachlorobenzene	14.998	250	494460	18.181	ng/ul	99
77) Carbazole	17.886	167	2119620	18.224	ng/ul	99
78) Di-n-butylphthalate	18.398	149	2601402	19.321	ng/ul	99
80) Fluoranthene	19.475	202	2615138	17.835	ng/ul	99
82) Pyrene	19.833	202	2729406	17.628	ng/ul	96
83) Butylbenzylphthalate	20.686	149	1253201	20.092	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.486	252	950118	18.610	ng/ul	99
85) Benzo(a)anthracene	21.551	228	2807691	17.818	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.427	149	1842760	20.357	ng/ul	99
87) Chrysene	21.610	228	2603387	17.656	ng/ul	99
89) Di-n-octyl phthalate	22.416	149	3273928	19.204	ng/ul	100
90) Benzo(b)fluoranthene	23.351	252	2838229	17.447	ng/ul	99
91) Benzo(k)fluoranthene	23.404	252	2840089	17.598	ng/ul	99
93) Benzo(a)pyrene	24.039	252	2680387	17.466	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.904	276	2956273	16.322	ng/ul	99
95) Dibenzo(a,h)anthracene	26.933	278	2458758	16.313	ng/ul	99
96) Benzo(g,h,i)perylene	27.768	276	2295973	15.928	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081823\
 Data File : BP016641.D
 Acq On : 18 Aug 2023 15:14
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020691

Quant Time: Aug 18 22:31:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 17 01:38:34 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.097 ng/uL
 RT: 3.428 min Scan# 24
 Delta R.T. -0.000 min
 Lab File: BP016641.D
 Acq: 18 Aug 2023 15:14




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020691


Tgt Ion: 88 Resp: 80050
 Ion Ratio Lower Upper
 88 100
 43 30.1 23.5 35.3
 58 68.2 57.1 85.7



#5
 Pyridine
 Concen: 17.344 ng/uL
 RT: 3.846 min Scan# 95
 Delta R.T. -0.006 min
 Lab File: BP016641.D
 Acq: 18 Aug 2023 15:14

Tgt Ion: 79 Resp: 561610
 Ion Ratio Lower Upper
 79 100
 52 70.3 56.6 84.8
 51 33.5 26.9 40.3









#6
Benzaldehyde
Concen: 20.672 ng/ul
RT: 7.205 min Scan# 666
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion: 77 Resp: 409498
Ion Ratio Lower Upper
77 100
105 93.2 71.8 107.6
106 90.3 70.2 105.2



#8
Phenol
Concen: 17.120 ng/ul
RT: 7.216 min Scan# 668
Delta R.T. -0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion: 94 Resp: 657561
Ion Ratio Lower Upper
94 100
65 28.0 22.3 33.5
66 37.0 28.7 43.1









#10
Bis(2-Chloroethyl)ether
Concen: 17.040 ng/ul
RT: 7.481 min Scan# 713
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion: 93 Resp: 528694
Ion Ratio Lower Upper
93 100
63 78.6 60.1 90.1
95 33.1 25.4 38.2



#12
2-Chlorophenol
Concen: 17.592 ng/ul
RT: 7.605 min Scan# 734
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:128 Resp: 477769
Ion Ratio Lower Upper
128 100
64 49.0 41.4 62.0
130 32.1 26.0 39.0





#13

2-Methylphenol

Concen: 17.497 ng/ul

RT: 8.487 min Scan# 884

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :





SSTD020691

Tgt Ion:108 Resp: 488153

Ion Ratio Lower Upper

108 100

107 89.4 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.957 ng/ul

RT: 8.563 min Scan# 897

Delta R.T. -0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion: 45 Resp: 733245


Ion Ratio Lower Upper

45 100

77 15.2 12.1 18.1

79 10.6 8.2 12.4









#16
Acetophenone
Concen: 17.580 ng/ul
RT: 8.899 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:105 Resp: 811181
Ion Ratio Lower Upper
105 100
77 81.3 64.4 96.6
51 26.6 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 18.158 ng/ul
RT: 8.869 min Scan# 949
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion: 70 Resp: 449351
Ion Ratio Lower Upper
70 100
42 55.0 43.0 64.4
101 9.6 6.8 10.2
130 19.0 14.2 21.4









#18
4-Methylphenol
Concen: 17.374 ng/ul
RT: 8.822 min Scan# 941
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:108 Resp: 532736
Ion Ratio Lower Upper
108 100
107 109.8 87.3 130.9



#19
Hexachloroethane
Concen: 17.358 ng/ul
RT: 9.128 min Scan# 993
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:117 Resp: 196957
Ion Ratio Lower Upper
117 100
201 86.9 71.8 107.6
199 53.0 45.2 67.8









#22
Nitrobenzene
Concen: 18.350 ng/ul
RT: 9.293 min Scan# 1021
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion: 77 Resp: 637745
Ion Ratio Lower Upper
77 100
123 42.7 32.5 48.7
65 13.3 10.3 15.5



#23
Isophorone
Concen: 18.964 ng/ul
RT: 9.805 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion: 82 Resp: 1244849
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 15.7 12.7 19.1









#25
2-Nitrophenol
Concen: 19.890 ng/ul
RT: 9.993 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:139 Resp: 268122
Ion Ratio Lower Upper
139 100
65 57.2 46.4 69.6
109 31.9 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 18.191 ng/ul
RT: 10.034 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:107 Resp: 578094
Ion Ratio Lower Upper
107 100
121 48.6 39.1 58.7
122 83.6 66.9 100.3









#27
Bis(2-Chloroethoxy)methane
Concen: 17.758 ng/ul
RT: 10.293 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion: 93 Resp: 747221
Ion Ratio Lower Upper
93 100
95 32.5 25.8 38.8
123 10.8 8.7 13.1



#29
2,4-Dichlorophenol
Concen: 19.000 ng/ul
RT: 10.516 min Scan# 1229
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:162 Resp: 453637
Ion Ratio Lower Upper
162 100
164 66.1 52.9 79.3
98 39.5 32.7 49.1









#30
Naphthalene
Concen: 17.413 ng/ul
RT: 10.934 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:128 Resp: 1605199
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 17.767 ng/ul
RT: 11.063 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:127 Resp: 700644
Ion Ratio Lower Upper
127 100
129 32.4 25.6 38.4





#33

Hexachlorobutadiene

Concen: 17.934 ng/ul

RT: 11.169 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:225 Resp: 258102

Ion Ratio Lower Upper

225 100

223 62.8 50.2 75.2

227 63.9 50.8 76.2



#34

Caprolactam

Concen: 19.137 ng/ul

RT: 11.869 min Scan# 1459

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:113 Resp: 177757


Ion Ratio Lower Upper

113 100

55 170.7 134.8 202.2

56 126.6 101.5 152.3









#35
4-Chloro-3-methylphenol
Concen: 18.952 ng/ul
RT: 12.151 min Scan# 1507
Delta R.T. -0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:107 Resp: 553068
Ion Ratio Lower Upper
107 100
144 24.4 19.0 28.6
142 74.5 60.2 90.4



#36
2-Methylnaphthalene
Concen: 17.708 ng/ul
RT: 12.534 min Scan# 1572
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:142 Resp: 1093453
Ion Ratio Lower Upper
142 100
141 89.0 71.4 107.0





#37

1-Methylnaphthalene

Concen: 17.818 ng/ul

RT: 12.751 min Scan# 1609

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:142 Resp: 1107846

Ion Ratio Lower Upper

142 100

141 92.4 74.5 111.7

116 4.1 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.985 ng/ul

RT: 12.881 min Scan# 1631

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Tgt Ion:216 Resp: 523238




Ion Ratio Lower Upper


216 100

214 78.8 63.8 95.6

179 21.6 17.9 26.9

108 20.6 16.3 24.5





#40

Hexachlorocyclopentadiene

Concen: 16.467 ng/ul

RT: 12.834 min Scan# 1623

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:237 Resp: 291488

Ion Ratio Lower Upper

237 100

235 62.0 49.7 74.5

272 12.1 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 19.708 ng/ul

RT: 13.134 min Scan# 1674

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:196 Resp: 356828


Ion Ratio Lower Upper

196 100

198 94.8 77.0 115.4

200 30.5 24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 19.769 ng/ul
RT: 13.198 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:196 Resp: 387983
Ion Ratio Lower Upper
196 100
198 95.8 76.7 115.1
200 31.2 24.2 36.2



#43
1,1'-Biphenyl
Concen: 17.702 ng/ul
RT: 13.540 min Scan# 1743
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:154 Resp: 1431237
Ion Ratio Lower Upper
154 100
153 40.1 31.4 47.2
76 15.1 13.0 19.4









#44
2-Chloronaphthalene
Concen: 17.579 ng/ul
RT: 13.587 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:162 Resp: 1105777
Ion Ratio Lower Upper
162 100
164 32.3 25.8 38.6
127 40.6 32.9 49.3



#45
2-Nitroaniline
Concen: 19.927 ng/ul
RT: 13.810 min Scan# 1789
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion: 65 Resp: 379012
Ion Ratio Lower Upper
65 100
92 70.1 54.5 81.7
138 95.1 72.1 108.1





#47

Dimethylphthalate

Concen: 17.814 ng/ul

RT: 14.157 min Scan# 1848

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:163 Resp: 1432878

Ion Ratio Lower Upper

163 100

194 3.6 3.0 4.6

164 10.3 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 19.964 ng/ul

RT: 14.298 min Scan# 1872

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:165 Resp: 278142


Ion Ratio Lower Upper

165 100

89 67.5 53.2 79.8

121 23.5 18.8 28.2









#50
Acenaphthylene
Concen: 18.174 ng/ul
RT: 14.434 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:152 Resp: 1900591
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.0 10.6 15.8



#51
3-Nitroaniline
Concen: 20.130 ng/ul
RT: 14.640 min Scan# 1930
Delta R.T. 0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:138 Resp: 314474
Ion Ratio Lower Upper
138 100
108 11.4 10.9 16.3
92 125.1 105.0 157.6





#52

Acenaphthene

Concen: 17.533 ng/ul

RT: 14.769 min Scan# 1952

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:153 Resp: 1227394

Ion Ratio Lower Upper

153 100

152 47.9 38.7 58.1

154 87.6 72.0 108.0



#53

2,4-Dinitrophenol

Concen: 15.603 ng/ul

RT: 14.834 min Scan# 1963

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:184 Resp: 121017


Ion Ratio Lower Upper

184 100

63 80.8 65.0 97.6

154 65.7 56.2 84.2









#55
4-Nitrophenol
Concen: 18.327 ng/ul
RT: 14.916 min Scan# 1977
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:109 Resp: 238787
Ion Ratio Lower Upper
109 100
139 125.0 101.1 151.7
65 142.2 108.9 163.3



#56
Dibenzofuran
Concen: 17.440 ng/ul
RT: 15.104 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:168 Resp: 1665686
Ion Ratio Lower Upper
168 100
139 38.4 31.4 47.0









#57
2,4-Dinitrotoluene
Concen: 19.743 ng/ul
RT: 15.081 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:165 Resp: 411911
Ion Ratio Lower Upper
165 100
63 50.3 40.4 60.6
182 2.6 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.696 ng/ul
RT: 15.316 min Scan# 2045
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:232 Resp: 323876
Ion Ratio Lower Upper
232 100
131 50.8 42.9 64.3
130 2.8 2.2 3.4
166 32.0 26.7 40.1









#59
Diethylphthalate
Concen: 17.876 ng/ul
RT: 15.516 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:149 Resp: 1470158
Ion Ratio Lower Upper
149 100
177 21.2 16.5 24.7
150 12.5 10.0 15.0



#61
Fluorene
Concen: 17.688 ng/ul
RT: 15.751 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:166 Resp: 1385207
Ion Ratio Lower Upper
166 100
165 96.7 78.3 117.5
167 13.3 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 17.591 ng/ul
RT: 15.745 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:204 Resp: 646708
Ion Ratio Lower Upper
204 100
206 32.5 25.5 38.3
141 62.4 52.2 78.2



#63
4-Nitroaniline
Concen: 20.841 ng/ul
RT: 15.804 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:138 Resp: 311579
Ion Ratio Lower Upper
138 100
92 57.4 46.7 70.1
108 89.2 73.7 110.5









#66
4,6-Dinitro-2-methylphenol
Concen: 17.510 ng/ul
RT: 15.834 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:198 Resp: 220291
Ion Ratio Lower Upper
198 100
182 4.4 3.9 5.9
77 30.6 25.8 38.6



#67
N-Nitrosodiphenylamine
Concen: 18.107 ng/ul
RT: 15.969 min Scan# 2156
Delta R.T. 0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:169 Resp: 1170613
Ion Ratio Lower Upper
169 100
168 66.5 53.9 80.9
167 37.3 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 17.975 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:248 Resp: 382780
Ion Ratio Lower Upper
248 100
250 98.3 76.9 115.3
141 85.9 67.6 101.4



#69
Hexachlorobenzene
Concen: 17.819 ng/ul
RT: 16.728 min Scan# 2285
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:284 Resp: 443370
Ion Ratio Lower Upper
284 100
142 42.6 31.9 47.9
249 31.8 24.7 37.1





#70

Atrazine

Concen: 18.480 ng/ul

RT: 16.922 min Scan# 2318

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:200 Resp: 429831

Ion Ratio Lower Upper

200 100

173 27.2 22.2 33.2

215 47.6 38.4 57.6



#71

Pentachlorophenol

Concen: 19.106 ng/ul

RT: 17.092 min Scan# 2347

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:266 Resp: 292697


Ion Ratio Lower Upper

266 100

264 64.1 50.7 76.1

268 66.9 53.0 79.6





#72

Phenanthrene

Concen: 17.486 ng/ul

RT: 17.516 min Scan# 2419

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:178 Resp: 2197163

Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 19.5 15.5 23.3



#74

Anthracene

Concen: 17.821 ng/ul

RT: 17.604 min Scan# 2434

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:178 Resp: 2249291


Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 18.7 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.255 ng/uL

RT: 13.493 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:216 Resp: 553302

Ion Ratio Lower Upper

216 100

214 79.4 65.4 98.2

218 49.6 39.3 58.9



#76

Pentachlorobenzene

Concen: 18.181 ng/uL

RT: 14.998 min Scan# 1991

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:250 Resp: 494460


Ion Ratio Lower Upper

250 100

248 62.1 50.2 75.2

252 63.0 51.0 76.4





#77

Carbazole

Concen: 18.224 ng/ul

RT: 17.886 min Scan# 2482

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:167 Resp: 2119620

Ion Ratio Lower Upper

167 100

166 21.7 16.8 25.2

139 13.4 10.6 16.0



#78

Di-n-butylphthalate

Concen: 19.321 ng/ul

RT: 18.398 min Scan# 2569

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:149 Resp: 2601402


Ion Ratio Lower Upper

149 100

150 9.0 7.3 10.9

104 5.4 4.6 6.8





#80

Fluoranthene

Concen: 17.835 ng/ul

RT: 19.475 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:202 Resp: 2615138

Ion Ratio Lower Upper

202 100

101 13.0 10.6 16.0

100 10.2 8.1 12.1



#82

Pyrene

Concen: 17.628 ng/ul

RT: 19.833 min Scan# 2813

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:202 Resp: 2729406


Ion Ratio Lower Upper

202 100

101 14.8 13.4 20.2

100 12.3 11.0 16.4









#83
Butylbenzylphthalate
Concen: 20.092 ng/ul
RT: 20.686 min Scan# 2959
Delta R.T. -0.000 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:149 Resp: 1253201
Ion Ratio Lower Upper
149 100
91 82.0 67.8 101.6
206 18.7 14.4 21.6



#84
3,3'-Dichlorobenzidine
Concen: 18.610 ng/ul
RT: 21.486 min Scan# 3094
Delta R.T. 0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:252 Resp: 950118
Ion Ratio Lower Upper
252 100
254 63.5 51.3 76.9
126 13.2 10.2 15.4





#85

Benzo(a)anthracene

Concen: 17.818 ng/ul

RT: 21.551 min Scan# 3106

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:228 Resp: 2807691

Ion Ratio Lower Upper

228 100

229 19.6 15.8 23.6

226 27.2 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.357 ng/ul

RT: 21.427 min Scan# 3084

Delta R.T. 0.006 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:149 Resp: 1842760


Ion Ratio Lower Upper

149 100

167 26.4 20.9 31.3

279 3.6 2.6 4.0









#87
Chrysene
Concen: 17.656 ng/ul
RT: 21.610 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14




Instrument :
BNA_P
ClientSampleId :
SSTD020691


Tgt Ion:228 Resp: 2603387
Ion Ratio Lower Upper
228 100
226 30.1 24.6 36.8
229 19.5 16.4 24.6



#89
Di-n-octyl phthalate
Concen: 19.204 ng/ul
RT: 22.416 min Scan# 3252
Delta R.T. 0.006 min
Lab File: BP016641.D
Acq: 18 Aug 2023 15:14

Tgt Ion:149 Resp: 3273928





#90

Benzo(b)fluoranthene

Concen: 17.447 ng/ul

RT: 23.351 min Scan# 3411

Delta R.T. 0.012 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:252 Resp: 2838229

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.2

125 11.8 10.0 15.0



#91

Benzo(k)fluoranthene

Concen: 17.598 ng/ul

RT: 23.404 min Scan# 3420

Delta R.T. 0.012 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:252 Resp: 2840089


Ion Ratio Lower Upper

252 100

253 22.3 18.0 27.0

125 11.5 9.8 14.8





#93

Benzo(a)pyrene

Concen: 17.466 ng/ul

RT: 24.039 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:252 Resp: 2680387

Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 13.6 10.6 15.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.322 ng/ul

RT: 26.904 min Scan# 4015

Delta R.T. 0.018 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14




Tgt Ion:276 Resp: 2956273


Ion Ratio Lower Upper

276 100

138 26.1 21.7 32.5

277 24.3 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 16.313 ng/ul

RT: 26.933 min Scan# 4022

Delta R.T. 0.018 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Instrument :

BNA_P

ClientSampleId :

SSTD020691





Tgt Ion:278 Resp: 2458758

Ion Ratio Lower Upper

278 100

139 18.7 15.5 23.3

279 24.2 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 15.928 ng/ul

RT: 27.768 min Scan# 4162

Delta R.T. 0.024 min

Lab File: BP016641.D

Acq: 18 Aug 2023 15:14

Tgt Ion:276 Resp: 2295973

Ion Ratio Lower Upper

276 100

138 24.7 21.5 32.3

277 24.2 19.3 28.9

