

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081823\
 Data File : BP016663.D
 Acq On : 19 Aug 2023 04:52
 Operator : MA/JU
 Sample : 04049-05
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH9T5

Quant Time: Aug 19 05:23:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 17 01:38:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	361528	20.000	ng/u1	0.00
20) Naphthalene-d8	10.881	136	1554213	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.698	164	881716	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.463	188	1706230	20.000	ng/u1	0.00
79) Chrysene-d12	21.563	240	1065406	20.000	ng/u1	0.00
88) Perylene-d12	24.145	264	1205107	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	40763	4.284	ng/uL	0.00
4) Pyridine-d5	3.828	84	457626	15.809	ng/u1	0.00
7) Phenol-d5	7.187	99	292755	8.783	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	868643	39.569	ng/u1	0.00
11) 2-Chlorophenol-d4	7.569	132	756581	31.604	ng/u1	0.00
15) 4-Methylphenol-d8	8.752	113	521850	19.691	ng/u1	0.00
21) Nitrobenzene-d5	9.246	128	495628	45.321	ng/u1	0.00
24) 2-Nitrophenol-d4	9.963	143	494197	47.580	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	797341	39.008	ng/u1	0.00
31) 4-Chloroaniline-d4	11.034	131	1053523	30.039	ng/u1	0.00
46) Dimethylphthalate-d6	14.110	166	2737561	42.926	ng/u1	0.00
49) Acenaphthylene-d8	14.398	160	3156673	43.854	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	72806	5.604	ng/u1	0.00
60) Fluorene-d10	15.692	176	2293955	42.193	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.810	200	314328	38.724	ng/u1	0.00
73) Anthracene-d10	17.569	188	3325102	45.054	ng/u1	0.00
81) Pyrene-d10	19.798	212	3160506	56.615	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	2669404	46.221	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081823\
Data File : BP016663.D
Acq On : 19 Aug 2023 04:52
Operator : MA/JU
Sample : 04049-05
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH9T5

Quant Time: Aug 19 05:23:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 17 01:38:34 2023
Response via : Initial Calibration

