

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082323\
 Data File : BP016718.D
 Acq On : 23 Aug 2023 16:58
 Operator : MA/JU
 Sample : 04105-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BHAR7

Quant Time: Aug 23 18:39:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 17 01:38:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	326871	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.869	136	1493652	20.000	ng/ul	-0.02
38) Acenaphthene-d10	14.693	164	903936	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.463	188	1944329	20.000	ng/ul	0.00
79) Chrysene-d12	21.557	240	1395208	20.000	ng/ul	0.00
88) Perylene-d12	24.133	264	1251263	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	57517	6.686	ng/uL	0.00
4) Pyridine-d5	3.828	84	395050	15.095	ng/ul	0.00
7) Phenol-d5	7.181	99	287303	9.534	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.375	67	864531	43.557	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.558	132	740839	34.227	ng/ul	-0.01
15) 4-Methylphenol-d8	8.746	113	530017	22.120	ng/ul	-0.01
21) Nitrobenzene-d5	9.240	128	495211	47.119	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.952	143	489329	49.021	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.475	165	827791	42.139	ng/ul	-0.02
31) 4-Chloroaniline-d4	11.028	131	995564	29.538	ng/ul	-0.01
46) Dimethylphthalate-d6	14.104	166	2890442	44.209	ng/ul	0.00
49) Acenaphthylene-d8	14.393	160	3335762	45.203	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.904	143	78334	5.881	ng/ul	0.00
60) Fluorene-d10	15.687	176	2477203	44.444	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.810	200	378249	40.893	ng/ul	0.00
73) Anthracene-d10	17.563	188	3832640	45.571	ng/ul	0.00
81) Pyrene-d10	19.798	212	4014418	54.912	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.963	264	2825484	47.119	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082323\
Data File : BP016718.D
Acq On : 23 Aug 2023 16:58
Operator : MA/JU
Sample : 04105-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BHAR7

Quant Time: Aug 23 18:39:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 17 01:38:34 2023
Response via : Initial Calibration

