

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082423\
 Data File : BP016727.D
 Acq On : 24 Aug 2023 13:23
 Operator : MA/JU
 Sample : SSTD00531
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005631

Quant Time: Aug 25 00:53:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 25 00:53:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	336881	20.000	ng/u1	0.00
20) Naphthalene-d8	10.869	136	1422713	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.693	164	922486	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.457	188	2112329	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	2137340	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	2293532	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	15299	1.877	ng/uL	0.00
4) Pyridine-d5	3.823	84	109022	4.377	ng/u1	0.00
7) Phenol-d5	7.175	99	115280	3.848	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.375	67	81688	4.279	ng/u1	0.00
11) 2-Chlorophenol-d4	7.558	132	95943	4.260	ng/u1	0.00
15) 4-Methylphenol-d8	8.740	113	91338	3.735	ng/u1	0.00
21) Nitrobenzene-d5	9.234	128	43887	4.080	ng/u1	0.00
24) 2-Nitrophenol-d4	9.952	143	42064	3.710	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	87263	4.128	ng/u1	0.00
31) 4-Chloroaniline-d4	11.022	131	133427	4.116	ng/u1	0.00
46) Dimethylphthalate-d6	14.098	166	316266	4.546	ng/u1	0.00
49) Acenaphthylene-d8	14.387	160	339758	4.320	ng/u1	0.00
54) 4-Nitrophenol-d4	14.893	143	42066	2.925	ng/u1	0.00
60) Fluorene-d10	15.681	176	272403	4.650	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	33598	2.671	ng/u1	0.00
73) Anthracene-d10	17.557	188	427534	4.566	ng/u1	0.00
81) Pyrene-d10	19.792	212	510710	4.351	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	494824	4.307	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.423	88	17280	1.947	ng/uL	96
5) Pyridine	3.846	79	112684	4.392	ng/u1	97
6) Benzaldehyde	7.193	77	83515	5.098	ng/u1	97
8) Phenol	7.205	94	127491	4.067	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.469	93	110779	4.378	ng/u1	98
12) 2-Chlorophenol	7.587	128	101346	4.402	ng/u1	99
13) 2-Methylphenol	8.475	108	91300	3.882	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.552	45	108415	2.937	ng/u1	98
16) Acetophenone	8.887	105	164314	4.333	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.852	70	81787	4.093	ng/u1	96
18) 4-Methylphenol	8.805	108	99902	3.901	ng/u1	99
19) Hexachloroethane	9.110	117	43809	4.574	ng/u1	94
22) Nitrobenzene	9.275	77	120161	4.354	ng/u1	98
23) Isophorone	9.793	82	225159	4.201	ng/u1	99
25) 2-Nitrophenol	9.981	139	49292	3.932	ng/u1	98
26) 2,4-Dimethylphenol	10.016	107	109594	4.349	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.275	93	150031	4.537	ng/u1	97
29) 2,4-Dichlorophenol	10.505	162	90000	4.239	ng/u1	98
30) Naphthalene	10.922	128	353110	4.791	ng/u1	99
32) 4-Chloroaniline	11.052	127	136989	4.300	ng/u1	99
33) Hexachlorobutadiene	11.152	225	66142	5.057	ng/u1	99
34) Caprolactam	11.852	113	23266	2.946	ng/u1	97
35) 4-Chloro-3-methylphenol	12.140	107	93461	3.925	ng/u1	95
36) 2-Methylnaphthalene	12.522	142	235584	4.631	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082423\
 Data File : BP016727.D
 Acq On : 24 Aug 2023 13:23
 Operator : MA/JU
 Sample : SST00531
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005631


Quant Time: Aug 25 00:53:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 25 00:53:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.740	142	238547	4.650	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.869	216	125315	4.766	ng/ul	98
40) Hexachlorocyclopentadiene	12.816	237	65181	3.371	ng/ul	95
41) 2,4,6-Trichlorophenol	13.116	196	70122	3.994	ng/ul	99
42) 2,4,5-Trichlorophenol	13.187	196	77451	4.077	ng/ul	97
43) 1,1'-Biphenyl	13.522	154	323232	4.713	ng/ul	100
44) 2-Chloronaphthalene	13.575	162	252132	4.657	ng/ul	99
45) 2-Nitroaniline	13.798	65	58623	3.570	ng/ul	96
47) Dimethylphthalate	14.146	163	330366	4.689	ng/ul	99
48) 2,6-Dinitrotoluene	14.181	165	75	0.005	ng/ul	95
50) Acenaphthylene	14.416	152	413283	4.591	ng/ul	99
51) 3-Nitroaniline	14.628	138	52888	3.679	ng/ul#	98
52) Acenaphthene	14.751	153	284118	4.771	ng/ul	98
53) 2,4-Dinitrophenol	14.828	184	19875	1.981	ng/ul	97
55) 4-Nitrophenol	14.910	109	35819	3.298	ng/ul	98
56) Dibenzofuran	15.087	168	396765	4.856	ng/ul	99
57) 2,4-Dinitrotoluene	15.069	165	74358	3.659	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.304	232	69401	4.291	ng/ul	98
59) Diethylphthalate	15.498	149	326037	4.555	ng/ul	99
61) Fluorene	15.740	166	327052	4.785	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.734	204	160230	4.775	ng/ul	99
63) 4-Nitroaniline	15.792	138	49111	3.590	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.816	198	40518	2.988	ng/ul#	86
67) N-Nitrosodiphenylamine	15.957	169	267512	4.579	ng/ul	99
68) 4-Bromophenyl-phenylether	16.634	248	97341	4.806	ng/ul	95
69) Hexachlorobenzene	16.716	284	116567	5.141	ng/ul	96
70) Atrazine	16.904	200	92825	4.214	ng/ul	98
71) Pentachlorophenol	17.081	266	60388	3.588	ng/ul	99
72) Phenanthrene	17.498	178	536246	4.822	ng/ul	99
74) Anthracene	17.592	178	530555	4.718	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.481	216	129101	4.565	ng/ul	97
76) Pentachlorobenzene	14.981	250	123204	4.874	ng/ul	98
77) Carbazole	17.875	167	468345	4.625	ng/ul	99
78) Di-n-butylphthalate	18.386	149	531286	4.205	ng/ul	99
80) Fluoranthene	19.463	202	618739	4.388	ng/ul	97
82) Pyrene	19.816	202	664162	4.560	ng/ul	99
83) Butylbenzylphthalate	20.675	149	230962	3.677	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.475	252	196572	4.303	ng/ul	98
85) Benzo(a)anthracene	21.539	228	680654	4.623	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.416	149	342984	3.739	ng/ul	98
87) Chrysene	21.592	228	627836	4.695	ng/ul	98
89) Di-n-octyl phthalate	22.398	149	568799	3.293	ng/ul	100
90) Benzo(b)fluoranthene	23.321	252	637161	4.350	ng/ul	98
91) Benzo(k)fluoranthene	23.374	252	640881	4.425	ng/ul	99
93) Benzo(a)pyrene	24.010	252	581486	4.389	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.862	276	610860	4.513	ng/ul	95
95) Dibenzo(a,h)anthracene	26.892	278	504858	4.490	ng/ul	97
96) Benzo(g,h,i)perylene	27.721	276	473565	4.614	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 25 00:53:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 25 00:53:07 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.947 ng/uL
RT: 3.423 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion: 88 Resp: 17280
Ion Ratio Lower Upper
88 100
43 29.0 24.0 36.0
58 71.2 59.8 89.8



#5
Pyridine
Concen: 4.392 ng/uL
RT: 3.846 min Scan# 95
Delta R.T. 0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion: 79 Resp: 112684
Ion Ratio Lower Upper
79 100
52 72.3 59.6 89.4
51 38.3 28.2 42.2









#6
Benzaldehyde
Concen: 5.098 ng/ul
RT: 7.193 min Scan# 664
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion: 77 Resp: 83515
Ion Ratio Lower Upper
77 100
105 91.9 75.4 113.2
106 88.1 73.1 109.7



#8
Phenol
Concen: 4.067 ng/ul
RT: 7.205 min Scan# 666
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion: 94 Resp: 127491
Ion Ratio Lower Upper
94 100
65 29.8 22.4 33.6
66 37.7 28.6 42.8









#10
Bis(2-Chloroethyl)ether
Concen: 4.378 ng/ul
RT: 7.469 min Scan# 711
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion: 93 Resp: 110779
Ion Ratio Lower Upper
93 100
63 80.2 62.9 94.3
95 32.3 26.3 39.5



#12
2-Chlorophenol
Concen: 4.402 ng/ul
RT: 7.587 min Scan# 731
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion: 128 Resp: 101346
Ion Ratio Lower Upper
128 100
64 49.6 40.0 60.0
130 31.8 26.4 39.6





#13

2-Methylphenol

Concen: 3.882 ng/ul

RT: 8.475 min Scan# 882

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :





SSTD005631

Tgt Ion:108 Resp: 91300

Ion Ratio Lower Upper

108 100

107 88.0 71.4 107.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 2.937 ng/ul

RT: 8.552 min Scan# 895

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion: 45 Resp: 108415


Ion Ratio Lower Upper

45 100

77 14.6 11.1 16.7

79 11.3 8.2 12.4









#16
Acetophenone
Concen: 4.333 ng/ul
RT: 8.887 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:105 Resp: 164314
Ion Ratio Lower Upper
105 100
77 77.3 64.2 96.4
51 26.2 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 4.093 ng/ul
RT: 8.852 min Scan# 946
Delta R.T. -0.012 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion: 70 Resp: 81787
Ion Ratio Lower Upper
70 100
42 60.9 46.0 69.0
101 8.7 7.6 11.4
130 21.2 16.1 24.1









#18
4-Methylphenol
Concen: 3.901 ng/ul
RT: 8.805 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:108 Resp: 99902
Ion Ratio Lower Upper
108 100
107 113.1 89.8 134.6



#19
Hexachloroethane
Concen: 4.574 ng/ul
RT: 9.110 min Scan# 990
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:117 Resp: 43809
Ion Ratio Lower Upper
117 100
201 99.1 75.4 113.2
199 65.6 47.8 71.6









#22
Nitrobenzene
Concen: 4.354 ng/ul
RT: 9.275 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion: 77 Resp: 120161
Ion Ratio Lower Upper
77 100
123 41.9 34.9 52.3
65 12.9 10.6 15.8



#23
Isophorone
Concen: 4.201 ng/ul
RT: 9.793 min Scan# 1106
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion: 82 Resp: 225159
Ion Ratio Lower Upper
82 100
95 7.9 5.4 8.0
138 16.8 13.5 20.3









#25
2-Nitrophenol
Concen: 3.932 ng/ul
RT: 9.981 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:139 Resp: 49292
Ion Ratio Lower Upper
139 100
65 53.8 42.0 63.0
109 30.7 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 4.349 ng/ul
RT: 10.016 min Scan# 1144
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:107 Resp: 109594
Ion Ratio Lower Upper
107 100
121 49.7 40.7 61.1
122 89.6 69.7 104.5





#27

Bis(2-Chloroethoxy)methane

Concen: 4.537 ng/ul

RT: 10.275 min Scan# 1189

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion: 93 Resp: 150031

Ion Ratio Lower Upper

93 100

95 31.3 26.3 39.5

123 11.7 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 4.239 ng/ul

RT: 10.505 min Scan# 1227

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:162 Resp: 90000


Ion Ratio Lower Upper

162 100

164 64.4 52.2 78.2

98 35.1 29.3 43.9









#30
Naphthalene
Concen: 4.791 ng/ul
RT: 10.922 min Scan# 1298
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:128 Resp: 353110
Ion Ratio Lower Upper
128 100
129 10.4 8.8 13.2
127 12.6 10.3 15.5



#32
4-Chloroaniline
Concen: 4.300 ng/ul
RT: 11.052 min Scan# 1320
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:127 Resp: 136989
Ion Ratio Lower Upper
127 100
129 31.8 25.9 38.9









#33
Hexachlorobutadiene
Concen: 5.057 ng/ul
RT: 11.152 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:225 Resp: 66142
Ion Ratio Lower Upper
225 100
223 63.4 51.0 76.4
227 66.8 52.5 78.7



#34
Caprolactam
Concen: 2.946 ng/ul
RT: 11.852 min Scan# 1456
Delta R.T. -0.024 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:113 Resp: 23266
Ion Ratio Lower Upper
113 100
55 178.9 140.6 210.8
56 134.6 104.5 156.7









#35
4-Chloro-3-methylphenol
Concen: 3.925 ng/ul
RT: 12.140 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:107 Resp: 93461
Ion Ratio Lower Upper
107 100
144 25.0 20.6 31.0
142 84.4 63.7 95.5



#36
2-Methylnaphthalene
Concen: 4.631 ng/ul
RT: 12.522 min Scan# 1570
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:142 Resp: 235584
Ion Ratio Lower Upper
142 100
141 91.8 71.4 107.0









#37
1-Methylnaphthalene
Concen: 4.650 ng/ul
RT: 12.740 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:142 Resp: 238547
Ion Ratio Lower Upper
142 100
141 90.4 73.0 109.4
116 3.7 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.766 ng/ul
RT: 12.869 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:216 Resp: 125315
Ion Ratio Lower Upper
216 100
214 77.1 63.4 95.2
179 21.3 17.4 26.2
108 17.6 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 3.371 ng/ul
RT: 12.816 min Scan# 1670
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:237 Resp: 65181
Ion Ratio Lower Upper
237 100
235 68.2 51.3 76.9
272 14.4 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 3.994 ng/ul
RT: 13.116 min Scan# 1671
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:196 Resp: 70122
Ion Ratio Lower Upper
196 100
198 96.1 76.6 115.0
200 29.3 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 4.077 ng/ul
RT: 13.187 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:196 Resp: 77451
Ion Ratio Lower Upper
196 100
198 94.4 77.8 116.8
200 28.6 24.8 37.2



#43
1,1'-Biphenyl
Concen: 4.713 ng/ul
RT: 13.522 min Scan# 1740
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:154 Resp: 323232
Ion Ratio Lower Upper
154 100
153 40.8 32.7 49.1
76 13.6 11.3 16.9









#44
2-Chloronaphthalene
Concen: 4.657 ng/ul
RT: 13.575 min Scan# 1749
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:162 Resp: 252132
Ion Ratio Lower Upper
162 100
164 32.5 26.1 39.1
127 37.0 30.8 46.2



#45
2-Nitroaniline
Concen: 3.570 ng/ul
RT: 13.798 min Scan# 1787
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion: 65 Resp: 58623
Ion Ratio Lower Upper
65 100
92 69.5 53.8 80.8
138 96.5 80.7 121.1





#47

Dimethylphthalate

Concen: 4.689 ng/ul

RT: 14.146 min Scan# 1846

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:163 Resp: 330366

Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8

164 10.0 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 0.005 ng/ul

RT: 14.181 min Scan# 1852

Delta R.T. -0.112 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:165 Resp: 75


Ion Ratio Lower Upper

165 100

89 55.5 48.0 72.0

121 21.3 16.9 25.3





#50

Acenaphthylene

Concen: 4.591 ng/ul

RT: 14.416 min Scan# 1892

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:152 Resp: 413283

Ion Ratio Lower Upper

152 100

151 19.3 16.1 24.1

153 13.1 10.6 15.8



#51

3-Nitroaniline

Concen: 3.679 ng/ul

RT: 14.628 min Scan# 1928

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:138 Resp: 52888


Ion Ratio Lower Upper

138 100

108 14.1 8.2 12.4#

92 117.7 95.4 143.0





#52

Acenaphthene

Concen: 4.771 ng/ul

RT: 14.751 min Scan# 1950

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:153 Resp: 284118

Ion Ratio Lower Upper

153 100

152 48.9 38.5 57.7

154 86.2 70.6 105.8



#53

2,4-Dinitrophenol

Concen: 1.981 ng/ul

RT: 14.828 min Scan# 1962

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:184 Resp: 19875


Ion Ratio Lower Upper

184 100

63 64.8 55.3 82.9

154 63.3 50.3 75.5









#55
4-Nitrophenol
Concen: 3.298 ng/ul
RT: 14.910 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:109 Resp: 35819
Ion Ratio Lower Upper
109 100
139 130.6 108.6 163.0
65 134.3 107.8 161.6



#56
Dibenzofuran
Concen: 4.856 ng/ul
RT: 15.087 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:168 Resp: 396765
Ion Ratio Lower Upper
168 100
139 37.1 29.2 43.8





#57

2,4-Dinitrotoluene

Concen: 3.659 ng/ul

RT: 15.069 min Scan# 2003

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:165 Resp: 74358

Ion Ratio Lower Upper

165 100

63 50.1 39.0 58.4

182 3.2 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.291 ng/ul

RT: 15.304 min Scan# 2043

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Tgt Ion:232 Resp: 69401




Ion Ratio Lower Upper


232 100

131 46.2 38.1 57.1

130 2.2 1.9 2.9

166 32.0 26.5 39.7









#59
Diethylphthalate
Concen: 4.555 ng/ul
RT: 15.498 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:149 Resp: 326037
Ion Ratio Lower Upper
149 100
177 22.1 17.3 25.9
150 12.1 10.2 15.2



#61
Fluorene
Concen: 4.785 ng/ul
RT: 15.740 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:166 Resp: 327052
Ion Ratio Lower Upper
166 100
165 98.6 78.8 118.2
167 13.9 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 4.775 ng/ul
RT: 15.734 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:204 Resp: 160230
Ion Ratio Lower Upper
204 100
206 33.5 26.7 40.1
141 57.0 46.8 70.2



#63
4-Nitroaniline
Concen: 3.590 ng/ul
RT: 15.792 min Scan# 2126
Delta R.T. -0.012 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:138 Resp: 49111
Ion Ratio Lower Upper
138 100
92 54.0 41.6 62.4
108 87.4 74.9 112.3









#66
4,6-Dinitro-2-methylphenol
Concen: 2.988 ng/ul
RT: 15.816 min Scan# 2130
Delta R.T. -0.012 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:198 Resp: 40518
Ion Ratio Lower Upper
198 100
182 8.1 3.2 4.8#
77 33.0 20.7 31.1#



#67
N-Nitrosodiphenylamine
Concen: 4.579 ng/ul
RT: 15.957 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:169 Resp: 267512
Ion Ratio Lower Upper
169 100
168 67.9 54.2 81.2
167 37.4 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 4.806 ng/ul
RT: 16.634 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SSTD005631


Tgt Ion:248 Resp: 97341
Ion Ratio Lower Upper
248 100
250 101.7 75.4 113.2
141 75.0 58.7 88.1



#69
Hexachlorobenzene
Concen: 5.141 ng/ul
RT: 16.716 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BP016727.D
Acq: 24 Aug 2023 13:23

Tgt Ion:284 Resp: 116567
Ion Ratio Lower Upper
284 100
142 35.6 29.8 44.8
249 28.9 25.9 38.9





#70

Atrazine

Concen: 4.214 ng/ul

RT: 16.904 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:200 Resp: 92825

Ion Ratio Lower Upper

200 100

173 23.6 21.4 32.2

215 46.5 37.3 55.9



#71

Pentachlorophenol

Concen: 3.588 ng/ul

RT: 17.081 min Scan# 2345

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:266 Resp: 60388


Ion Ratio Lower Upper

266 100

264 64.8 50.7 76.1

268 63.0 50.8 76.2





#72

Phenanthrene

Concen: 4.822 ng/ul

RT: 17.498 min Scan# 2416

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:178 Resp: 536246

Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 19.1 15.6 23.4



#74

Anthracene

Concen: 4.718 ng/ul

RT: 17.592 min Scan# 2432

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:178 Resp: 530555


Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 18.4 15.3 22.9





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.565 ng/uL

RT: 13.481 min Scan# 1733

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:216 Resp: 129101

Ion Ratio Lower Upper

216 100

214 76.8 64.0 96.0

218 49.4 38.2 57.2



#76

Pentachlorobenzene

Concen: 4.874 ng/uL

RT: 14.981 min Scan# 1988

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:250 Resp: 123204


Ion Ratio Lower Upper

250 100

248 61.6 50.9 76.3

252 63.7 52.2 78.4





#77

Carbazole

Concen: 4.625 ng/ul

RT: 17.875 min Scan# 2480

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:167 Resp: 468345

Ion Ratio Lower Upper

167 100

166 22.2 17.4 26.2

139 12.4 10.4 15.6



#78

Di-n-butylphthalate

Concen: 4.205 ng/ul

RT: 18.386 min Scan# 2567

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:149 Resp: 531286


Ion Ratio Lower Upper

149 100

150 8.7 7.5 11.3

104 5.2 4.1 6.1





#80

Fluoranthene

Concen: 4.388 ng/ul

RT: 19.463 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:202 Resp: 618739

Ion Ratio Lower Upper

202 100

101 11.0 9.7 14.5

100 8.0 7.4 11.2



#82

Pyrene

Concen: 4.560 ng/ul

RT: 19.816 min Scan# 2810

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:202 Resp: 664162


Ion Ratio Lower Upper

202 100

101 13.7 11.4 17.2

100 11.8 9.4 14.2





#83

Butylbenzylphthalate

Concen: 3.677 ng/ul

RT: 20.675 min Scan# 2956

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:149 Resp: 230962

Ion Ratio Lower Upper

149 100

91 72.6 58.5 87.7

206 20.0 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 4.303 ng/ul

RT: 21.475 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:252 Resp: 196572


Ion Ratio Lower Upper

252 100

254 66.8 52.1 78.1

126 12.3 9.6 14.4





#85

Benzo(a)anthracene

Concen: 4.623 ng/ul

RT: 21.539 min Scan# 3103

Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:228 Resp: 680654

Ion Ratio Lower Upper

228 100

229 19.6 16.2 24.4

226 27.4 22.9 34.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.739 ng/ul

RT: 21.416 min Scan# 3082

Delta R.T. 0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:149 Resp: 342984


Ion Ratio Lower Upper

149 100

167 26.3 22.1 33.1

279 4.0 3.0 4.4





#87

Chrysene

Concen: 4.695 ng/ul

RT: 21.592 min Scan# 3113

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:228 Resp: 627836

Ion Ratio Lower Upper

228 100

226 30.0 25.0 37.4

229 19.4 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 3.293 ng/ul




RT: 22.398 min Scan# 3249


Delta R.T. -0.000 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Tgt Ion:149 Resp: 568799





#90

Benzo(b)fluoranthene

Concen: 4.350 ng/ul

RT: 23.321 min Scan# 3406

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:252 Resp: 637161

Ion Ratio Lower Upper

252 100

253 22.8 17.7 26.5

125 10.3 9.0 13.4



#91

Benzo(k)fluoranthene

Concen: 4.425 ng/ul

RT: 23.374 min Scan# 3415

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:252 Resp: 640881


Ion Ratio Lower Upper

252 100

253 22.2 18.1 27.1

125 10.2 8.9 13.3





#93

Benzo(a)pyrene

Concen: 4.389 ng/ul

RT: 24.010 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:252 Resp: 581486

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8

125 11.3 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.513 ng/ul

RT: 26.862 min Scan# 4008

Delta R.T. -0.012 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23




Tgt Ion:276 Resp: 610860


Ion Ratio Lower Upper

276 100

138 21.3 20.9 31.3

277 24.4 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 4.490 ng/ul

RT: 26.892 min Scan# 4014

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SSTD005631





Tgt Ion:278 Resp: 504858

Ion Ratio Lower Upper

278 100

139 16.5 15.0 22.6

279 24.0 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 4.614 ng/ul

RT: 27.721 min Scan# 4154

Delta R.T. -0.006 min

Lab File: BP016727.D

Acq: 24 Aug 2023 13:23

Tgt Ion:276 Resp: 473565

Ion Ratio Lower Upper

276 100

138 21.1 20.2 30.4

277 23.8 19.3 28.9

