

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082423\
 Data File : BP016728.D
 Acq On : 24 Aug 2023 14:31
 Operator : MA/JU
 Sample : SSTD01032
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Quant Time: Aug 25 00:54:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 25 00:53:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	562069	20.000	ng/u1	0.00
20) Naphthalene-d8	10.869	136	2524926	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.692	164	1596906	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.457	188	3501870	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	3250994	20.000	ng/u1	0.00
88) Perylene-d12	24.127	264	3140457	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	50158	3.688	ng/uL	0.00
4) Pyridine-d5	3.822	84	375932	9.046	ng/u1	0.00
7) Phenol-d5	7.175	99	449866	9.001	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.375	67	307948	9.667	ng/u1	0.00
11) 2-Chlorophenol-d4	7.557	132	350849	9.337	ng/u1	0.00
15) 4-Methylphenol-d8	8.740	113	372325	9.126	ng/u1	0.00
21) Nitrobenzene-d5	9.234	128	176746	9.258	ng/u1	0.00
24) 2-Nitrophenol-d4	9.951	143	179820	8.936	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	342867	9.139	ng/u1	0.00
31) 4-Chloroaniline-d4	11.022	131	529914	9.211	ng/u1	0.00
46) Dimethylphthalate-d6	14.098	166	1135264	9.427	ng/u1	0.00
49) Acenaphthylene-d8	14.392	160	1295281	9.514	ng/u1	0.00
54) 4-Nitrophenol-d4	14.892	143	210470	8.455	ng/u1	0.00
60) Fluorene-d10	15.686	176	950315	9.371	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	163035	7.818	ng/u1	0.00
73) Anthracene-d10	17.557	188	1468167	9.458	ng/u1	0.00
81) Pyrene-d10	19.786	212	1693436	9.484	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	1464085	9.307	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	56755	3.832	ng/uL	95
5) Pyridine	3.846	79	388870	9.084	ng/u1	98
6) Benzaldehyde	7.193	77	312525	11.434	ng/u1	99
8) Phenol	7.205	94	478246	9.143	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.469	93	405627	9.607	ng/u1	99
12) 2-Chlorophenol	7.587	128	363397	9.461	ng/u1	99
13) 2-Methylphenol	8.475	108	355247	9.054	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.551	45	391615	6.358	ng/u1	100
16) Acetophenone	8.887	105	613192	9.693	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.857	70	329626	9.887	ng/u1	100
18) 4-Methylphenol	8.804	108	391816	9.171	ng/u1	100
19) Hexachloroethane	9.110	117	152735	9.557	ng/u1	97
22) Nitrobenzene	9.281	77	468729	9.570	ng/u1	99
23) Isophorone	9.793	82	903281	9.496	ng/u1	99
25) 2-Nitrophenol	9.981	139	206076	9.262	ng/u1	97
26) 2,4-Dimethylphenol	10.022	107	425554	9.516	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.281	93	569915	9.711	ng/u1	99
29) 2,4-Dichlorophenol	10.504	162	350831	9.310	ng/u1	99
30) Naphthalene	10.922	128	1247103	9.535	ng/u1	99
32) 4-Chloroaniline	11.045	127	528272	9.344	ng/u1	98
33) Hexachlorobutadiene	11.151	225	213495	9.197	ng/u1	98
34) Caprolactam	11.863	113	119499	8.526	ng/u1	97
35) 4-Chloro-3-methylphenol	12.140	107	396254	9.377	ng/u1	99
36) 2-Methylnaphthalene	12.522	142	865198	9.583	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082423\
 Data File : BP016728.D
 Acq On : 24 Aug 2023 14:31
 Operator : MA/JU
 Sample : SSTD01032
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Quant Time: Aug 25 00:54:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 25 00:53:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.740	142	869012	9.545	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.869	216	423948	9.315	ng/ul	99
40) Hexachlorocyclopentadiene	12.822	237	429167	12.822	ng/ul	98
41) 2,4,6-Trichlorophenol	13.116	196	276842	9.109	ng/ul	97
42) 2,4,5-Trichlorophenol	13.187	196	297796	9.055	ng/ul	99
43) 1,1'-Biphenyl	13.528	154	1132606	9.539	ng/ul	100
44) 2-Chloronaphthalene	13.575	162	898349	9.585	ng/ul	99
45) 2-Nitroaniline	13.804	65	262218	9.225	ng/ul	97
47) Dimethylphthalate	14.151	163	1146602	9.401	ng/ul	99
48) 2,6-Dinitrotoluene	14.292	165	214365	8.971	ng/ul	98
50) Acenaphthylene	14.416	152	1497518	9.610	ng/ul	100
51) 3-Nitroaniline	14.628	138	226468	9.100	ng/ul	99
52) Acenaphthene	14.757	153	980413	9.511	ng/ul	99
53) 2,4-Dinitrophenol	14.828	184	234737	13.517	ng/ul	98
55) 4-Nitrophenol	14.910	109	163975	8.723	ng/ul	99
56) Dibenzofuran	15.092	168	1343952	9.502	ng/ul	99
57) 2,4-Dinitrotoluene	15.069	165	324274	9.218	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.304	232	249456	8.909	ng/ul	99
59) Diethylphthalate	15.504	149	1167116	9.420	ng/ul	99
61) Fluorene	15.739	166	1121104	9.476	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.733	204	542550	9.340	ng/ul	99
63) 4-Nitroaniline	15.792	138	221999	9.375	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.822	198	186970	8.316	ng/ul	96
67) N-Nitrosodiphenylamine	15.957	169	935939	9.663	ng/ul	99
68) 4-Bromophenyl-phenylether	16.633	248	307443	9.157	ng/ul	95
69) Hexachlorobenzene	16.716	284	340216	9.050	ng/ul	95
70) Atrazine	16.910	200	339971	9.310	ng/ul	99
71) Pentachlorophenol	17.080	266	212666	7.621	ng/ul	98
72) Phenanthrene	17.498	178	1734659	9.409	ng/ul	100
74) Anthracene	17.592	178	1790835	9.607	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.481	216	447445	9.543	ng/uL	99
76) Pentachlorobenzene	14.981	250	391953	9.354	ng/uL	99
77) Carbazole	17.875	167	1634419	9.735	ng/ul	99
78) Di-n-butylphthalate	18.386	149	2011583	9.604	ng/ul	99
80) Fluoranthene	19.463	202	2048418	9.550	ng/ul	99
82) Pyrene	19.816	202	2153963	9.722	ng/ul	99
83) Butylbenzylphthalate	20.674	149	905144	9.473	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.474	252	600550	8.643	ng/ul	99
85) Benzo(a)anthracene	21.539	228	2133360	9.527	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.410	149	1324334	9.492	ng/ul	98
87) Chrysene	21.592	228	1927080	9.474	ng/ul	100
89) Di-n-octyl phthalate	22.392	149	2149580	9.088	ng/ul	100
90) Benzo(b)fluoranthene	23.321	252	1888471	9.417	ng/ul	99
91) Benzo(k)fluoranthene	23.374	252	1880590	9.482	ng/ul	99
93) Benzo(a)pyrene	24.009	252	1697606	9.359	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.856	276	1658824	8.949	ng/ul	99
95) Dibenzo(a,h)anthracene	26.886	278	1378180	8.952	ng/ul	100
96) Benzo(g,h,i)perylene	27.715	276	1259651	8.964	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 25 00:54:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 25 00:53:07 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 3.832 ng/uL
 RT: 3.422 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BP016728.D
 Acq: 24 Aug 2023 14:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Tgt Ion: 88 Resp: 56755
 Ion Ratio Lower Upper
 88 100
 43 30.3 24.0 36.0
 58 68.5 59.8 89.8



#5
 Pyridine
 Concen: 9.084 ng/uL
 RT: 3.846 min Scan# 95
 Delta R.T. 0.006 min
 Lab File: BP016728.D
 Acq: 24 Aug 2023 14:31

Tgt Ion: 79 Resp: 388870
 Ion Ratio Lower Upper
 79 100
 52 71.9 59.6 89.4
 51 34.8 28.2 42.2









#6
Benzaldehyde
Concen: 11.434 ng/ul
RT: 7.193 min Scan# 664
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 77 Resp: 312525
Ion Ratio Lower Upper
77 100
105 95.8 75.4 113.2
106 90.7 73.1 109.7



#8
Phenol
Concen: 9.143 ng/ul
RT: 7.205 min Scan# 666
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion: 94 Resp: 478246
Ion Ratio Lower Upper
94 100
65 28.0 22.4 33.6
66 37.1 28.6 42.8









#10
Bis(2-Chloroethyl)ether
Concen: 9.607 ng/ul
RT: 7.469 min Scan# 711
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 93 Resp: 405627
Ion Ratio Lower Upper
93 100
63 78.2 62.9 94.3
95 32.3 26.3 39.5



#12
2-Chlorophenol
Concen: 9.461 ng/ul
RT: 7.587 min Scan# 731
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion: 128 Resp: 363397
Ion Ratio Lower Upper
128 100
64 50.9 40.0 60.0
130 32.6 26.4 39.6





#13

2-Methylphenol

Concen: 9.054 ng/ul

RT: 8.475 min Scan# 882

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :





SSTD010632

Tgt Ion:108 Resp: 355247

Ion Ratio Lower Upper

108 100

107 90.2 71.4 107.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 6.358 ng/ul

RT: 8.551 min Scan# 895

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion: 45 Resp: 391615


Ion Ratio Lower Upper

45 100

77 13.8 11.1 16.7

79 10.0 8.2 12.4









#16
Acetophenone
Concen: 9.693 ng/ul
RT: 8.887 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:105 Resp: 613192
Ion Ratio Lower Upper
105 100
77 79.5 64.2 96.4
51 28.1 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 9.887 ng/ul
RT: 8.857 min Scan# 947
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion: 70 Resp: 329626
Ion Ratio Lower Upper
70 100
42 57.8 46.0 69.0
101 9.2 7.6 11.4
130 20.2 16.1 24.1









#18
4-Methylphenol
Concen: 9.171 ng/ul
RT: 8.804 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:108 Resp: 391816
Ion Ratio Lower Upper
108 100
107 112.5 89.8 134.6



#19
Hexachloroethane
Concen: 9.557 ng/ul
RT: 9.110 min Scan# 990
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:117 Resp: 152735
Ion Ratio Lower Upper
117 100
201 91.0 75.4 113.2
199 57.1 47.8 71.6









#22
Nitrobenzene
Concen: 9.570 ng/ul
RT: 9.281 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 77 Resp: 468729
Ion Ratio Lower Upper
77 100
123 42.9 34.9 52.3
65 13.1 10.6 15.8



#23
Isophorone
Concen: 9.496 ng/ul
RT: 9.793 min Scan# 1106
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion: 82 Resp: 903281
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.0
138 16.6 13.5 20.3









#25
2-Nitrophenol
Concen: 9.262 ng/ul
RT: 9.981 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:139 Resp: 206076
Ion Ratio Lower Upper
139 100
65 55.8 42.0 63.0
109 29.8 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 9.516 ng/ul
RT: 10.022 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:107 Resp: 425554
Ion Ratio Lower Upper
107 100
121 49.2 40.7 61.1
122 84.4 69.7 104.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.711 ng/ul
 RT: 10.281 min Scan# 1189
 Delta R.T. -0.000 min
 Lab File: BP016728.D
 Acq: 24 Aug 2023 14:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Tgt Ion: 93 Resp: 569915
 Ion Ratio Lower Upper
 93 100
 95 32.4 26.3 39.5
 123 10.3 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 9.310 ng/ul
 RT: 10.504 min Scan# 1227
 Delta R.T. -0.000 min
 Lab File: BP016728.D
 Acq: 24 Aug 2023 14:31

Tgt Ion:162 Resp: 350831
 Ion Ratio Lower Upper
 162 100
 164 64.8 52.2 78.2
 98 37.6 29.3 43.9









#30
Naphthalene
Concen: 9.535 ng/ul
RT: 10.922 min Scan# 1298
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:128 Resp: 1247103
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 9.344 ng/ul
RT: 11.045 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:127 Resp: 528272
Ion Ratio Lower Upper
127 100
129 31.5 25.9 38.9









#33
Hexachlorobutadiene
Concen: 9.197 ng/ul
RT: 11.151 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:225 Resp: 213495
Ion Ratio Lower Upper
225 100
223 64.2 51.0 76.4
227 63.5 52.5 78.7



#34
Caprolactam
Concen: 8.526 ng/ul
RT: 11.863 min Scan# 1458
Delta R.T. -0.012 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:113 Resp: 119499
Ion Ratio Lower Upper
113 100
55 180.8 140.6 210.8
56 133.0 104.5 156.7









#35
4-Chloro-3-methylphenol
Concen: 9.377 ng/ul
RT: 12.140 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:107 Resp: 396254
Ion Ratio Lower Upper
107 100
144 24.6 20.6 31.0
142 79.4 63.7 95.5



#36
2-Methylnaphthalene
Concen: 9.583 ng/ul
RT: 12.522 min Scan# 1570
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:142 Resp: 865198
Ion Ratio Lower Upper
142 100
141 88.4 71.4 107.0









#37
1-Methylnaphthalene
Concen: 9.545 ng/ul
RT: 12.740 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:142 Resp: 869012
Ion Ratio Lower Upper
142 100
141 92.1 73.0 109.4
116 3.7 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.315 ng/ul
RT: 12.869 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:216 Resp: 423948
Ion Ratio Lower Upper
216 100
214 78.0 63.4 95.2
179 21.4 17.4 26.2
108 18.1 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 12.822 ng/ul
RT: 12.822 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:237 Resp: 429167
Ion Ratio Lower Upper
237 100
235 62.3 51.3 76.9
272 12.3 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 9.109 ng/ul
RT: 13.116 min Scan# 1671
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:196 Resp: 276842
Ion Ratio Lower Upper
196 100
198 93.2 76.6 115.0
200 30.3 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 9.055 ng/ul

RT: 13.187 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:196 Resp: 297796

Ion Ratio Lower Upper

196 100

198 96.5 77.8 116.8

200 31.3 24.8 37.2



#43

1,1'-Biphenyl

Concen: 9.539 ng/ul

RT: 13.528 min Scan# 1741

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:154 Resp: 1132606


Ion Ratio Lower Upper

154 100

153 40.8 32.7 49.1

76 14.3 11.3 16.9









#44
2-Chloronaphthalene
Concen: 9.585 ng/ul
RT: 13.575 min Scan# 1749
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:162 Resp: 898349
Ion Ratio Lower Upper
162 100
164 33.6 26.1 39.1
127 38.8 30.8 46.2



#45
2-Nitroaniline
Concen: 9.225 ng/ul
RT: 13.804 min Scan# 1788
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion: 65 Resp: 262218
Ion Ratio Lower Upper
65 100
92 70.6 53.8 80.8
138 103.1 80.7 121.1









#47
Dimethylphthalate
Concen: 9.401 ng/ul
RT: 14.151 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:163 Resp: 1146602
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 9.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 8.971 ng/ul
RT: 14.292 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:165 Resp: 214365
Ion Ratio Lower Upper
165 100
89 61.4 48.0 72.0
121 22.4 16.9 25.3









#50
Acenaphthylene
Concen: 9.610 ng/ul
RT: 14.416 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:152 Resp: 1497518
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1
153 13.4 10.6 15.8



#51
3-Nitroaniline
Concen: 9.100 ng/ul
RT: 14.628 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:138 Resp: 226468
Ion Ratio Lower Upper
138 100
108 11.2 8.2 12.4
92 120.2 95.4 143.0





#52

Acenaphthene

Concen: 9.511 ng/ul

RT: 14.757 min Scan# 1950

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:153 Resp: 980413

Ion Ratio Lower Upper

153 100

152 48.7 38.5 57.7

154 88.8 70.6 105.8



#53

2,4-Dinitrophenol

Concen: 13.517 ng/ul

RT: 14.828 min Scan# 1962

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:184 Resp: 234737


Ion Ratio Lower Upper

184 100

63 68.5 55.3 82.9

154 60.6 50.3 75.5









#55
4-Nitrophenol
Concen: 8.723 ng/ul
RT: 14.910 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:109 Resp: 163975
Ion Ratio Lower Upper
109 100
139 133.0 108.6 163.0
65 135.3 107.8 161.6



#56
Dibenzofuran
Concen: 9.502 ng/ul
RT: 15.092 min Scan# 2007
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:168 Resp: 1343952
Ion Ratio Lower Upper
168 100
139 37.0 29.2 43.8









#57
2,4-Dinitrotoluene
Concen: 9.218 ng/ul
RT: 15.069 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:165 Resp: 324274
Ion Ratio Lower Upper
165 100
63 48.4 39.0 58.4
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.909 ng/ul
RT: 15.304 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:232 Resp: 249456
Ion Ratio Lower Upper
232 100
131 48.5 38.1 57.1
130 2.6 1.9 2.9
166 32.8 26.5 39.7









#59
Diethylphthalate
Concen: 9.420 ng/ul
RT: 15.504 min Scan# 2077
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:149 Resp: 1167116
Ion Ratio Lower Upper
149 100
177 22.1 17.3 25.9
150 12.2 10.2 15.2



#61
Fluorene
Concen: 9.476 ng/ul
RT: 15.739 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:166 Resp: 1121104
Ion Ratio Lower Upper
166 100
165 98.8 78.8 118.2
167 13.3 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 9.340 ng/ul
RT: 15.733 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:204 Resp: 542550
Ion Ratio Lower Upper
204 100
206 32.1 26.7 40.1
141 58.8 46.8 70.2



#63
4-Nitroaniline
Concen: 9.375 ng/ul
RT: 15.792 min Scan# 2126
Delta R.T. -0.012 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:138 Resp: 221999
Ion Ratio Lower Upper
138 100
92 54.7 41.6 62.4
108 86.0 74.9 112.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 8.316 ng/ul
 RT: 15.822 min Scan# 2131
 Delta R.T. -0.006 min
 Lab File: BP016728.D
 Acq: 24 Aug 2023 14:31

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Tgt Ion:198 Resp: 186970




Ion	Ratio	Lower	Upper
198	100		
182	4.3	3.2	4.8
77	28.1	20.7	31.1




#67
 N-Nitrosodiphenylamine
 Concen: 9.663 ng/ul
 RT: 15.957 min Scan# 2154
 Delta R.T. -0.000 min
 Lab File: BP016728.D
 Acq: 24 Aug 2023 14:31

Tgt Ion:169 Resp: 935939

Ion	Ratio	Lower	Upper
169	100		
168	67.2	54.2	81.2
167	36.9	29.4	44.2









#68
4-Bromophenyl-phenylether
Concen: 9.157 ng/ul
RT: 16.633 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:248 Resp: 307443
Ion Ratio Lower Upper
248 100
250 97.4 75.4 113.2
141 79.5 58.7 88.1



#69
Hexachlorobenzene
Concen: 9.050 ng/ul
RT: 16.716 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:284 Resp: 340216
Ion Ratio Lower Upper
284 100
142 42.5 29.8 44.8
249 33.2 25.9 38.9





#70

Atrazine

Concen: 9.310 ng/ul

RT: 16.910 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:200 Resp: 339971

Ion Ratio Lower Upper

200 100

173 26.5 21.4 32.2

215 47.2 37.3 55.9



#71

Pentachlorophenol

Concen: 7.621 ng/ul

RT: 17.080 min Scan# 2345

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:266 Resp: 212666


Ion Ratio Lower Upper

266 100

264 65.0 50.7 76.1

268 64.6 50.8 76.2





#72

Phenanthrene

Concen: 9.409 ng/ul

RT: 17.498 min Scan# 2416

Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:178 Resp: 1734659

Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6

176 19.5 15.6 23.4



#74

Anthracene

Concen: 9.607 ng/ul

RT: 17.592 min Scan# 2432

Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:178 Resp: 1790835


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 18.4 15.3 22.9





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.543 ng/uL

RT: 13.481 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:216 Resp: 447445

Ion Ratio Lower Upper

216 100

214 78.4 64.0 96.0

218 48.2 38.2 57.2



#76

Pentachlorobenzene

Concen: 9.354 ng/uL

RT: 14.981 min Scan# 1988

Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:250 Resp: 391953


Ion Ratio Lower Upper

250 100

248 63.9 50.9 76.3

252 63.8 52.2 78.4





#77

Carbazole

Concen: 9.735 ng/ul

RT: 17.875 min Scan# 2480

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:167 Resp: 1634419

Ion Ratio Lower Upper

167 100

166 21.6 17.4 26.2

139 12.7 10.4 15.6



#78

Di-n-butylphthalate

Concen: 9.604 ng/ul

RT: 18.386 min Scan# 2567

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:149 Resp: 2011583


Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3

104 5.1 4.1 6.1





#80

Fluoranthene

Concen: 9.550 ng/ul

RT: 19.463 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:202 Resp: 2048418

Ion Ratio Lower Upper

202 100

101 11.7 9.7 14.5

100 9.0 7.4 11.2



#82

Pyrene

Concen: 9.722 ng/ul

RT: 19.816 min Scan# 2810

Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:202 Resp: 2153963


Ion Ratio Lower Upper

202 100

101 14.8 11.4 17.2

100 12.1 9.4 14.2





#83

Butylbenzylphthalate

Concen: 9.473 ng/ul

RT: 20.674 min Scan# 2956

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:149 Resp: 905144

Ion Ratio Lower Upper

149 100

91 74.3 58.5 87.7

206 19.7 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 8.643 ng/ul

RT: 21.474 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:252 Resp: 600550


Ion Ratio Lower Upper

252 100

254 64.2 52.1 78.1

126 12.2 9.6 14.4





#85

Benzo(a)anthracene

Concen: 9.527 ng/ul

RT: 21.539 min Scan# 3103

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:228 Resp: 2133360

Ion Ratio Lower Upper

228 100

229 19.6 16.2 24.4

226 28.0 22.9 34.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.492 ng/ul

RT: 21.410 min Scan# 3081

Delta R.T. -0.000 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:149 Resp: 1324334


Ion Ratio Lower Upper

149 100

167 26.4 22.1 33.1

279 3.2 3.0 4.4





#87

Chrysene

Concen: 9.474 ng/ul

RT: 21.592 min Scan# 3113

Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:228 Resp: 1927080

Ion Ratio Lower Upper

228 100

226 31.0 25.0 37.4

229 19.8 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 9.088 ng/ul




RT: 22.392 min Scan# 3248


Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Tgt Ion:149 Resp: 2149580









#90
Benzo(b)fluoranthene
Concen: 9.417 ng/ul
RT: 23.321 min Scan# 3406
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:252 Resp: 1888471
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 11.9 9.0 13.4



#91
Benzo(k)fluoranthene
Concen: 9.482 ng/ul
RT: 23.374 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BP016728.D
Acq: 24 Aug 2023 14:31

Tgt Ion:252 Resp: 1880590
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 11.4 8.9 13.3





#93

Benzo(a)pyrene

Concen: 9.359 ng/ul

RT: 24.009 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:252 Resp: 1697606

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 12.8 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.949 ng/ul

RT: 26.856 min Scan# 4007

Delta R.T. -0.018 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31




Tgt Ion:276 Resp: 1658824


Ion Ratio Lower Upper

276 100

138 25.2 20.9 31.3

277 24.1 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 8.952 ng/ul

RT: 26.886 min Scan# 4014

Delta R.T. -0.012 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:278 Resp: 1378180

Ion Ratio Lower Upper

278 100

139 18.5 15.0 22.6

279 24.2 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 8.964 ng/ul

RT: 27.715 min Scan# 4153

Delta R.T. -0.012 min

Lab File: BP016728.D

Acq: 24 Aug 2023 14:31

Tgt Ion:276 Resp: 1259651

Ion Ratio Lower Upper

276 100

138 25.7 20.2 30.4

277 23.4 19.3 28.9

