

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082523\
 Data File : BP016764.D
 Acq On : 26 Aug 2023 01:44
 Operator : MA/JU
 Sample : 04105-09
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BHAB1

Quant Time: Aug 26 02:21:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 25 01:47:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	351639	20.000	ng/ul	0.00
20) Naphthalene-d8	10.863	136	1539147	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.687	164	967874	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.457	188	2110583	20.000	ng/ul	0.00
79) Chrysene-d12	21.551	240	1692158	20.000	ng/ul	0.00
88) Perylene-d12	24.121	264	1464989	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	46748	5.494	ng/uL	0.00
4) Pyridine-d5	3.823	84	255381	9.822	ng/ul	0.00
7) Phenol-d5	7.175	99	248043	7.933	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	636759	31.952	ng/ul	0.00
11) 2-Chlorophenol-d4	7.558	132	626262	26.639	ng/ul	0.00
15) 4-Methylphenol-d8	8.740	113	428244	16.777	ng/ul	0.00
21) Nitrobenzene-d5	9.234	128	399835	34.358	ng/ul	0.00
24) 2-Nitrophenol-d4	9.952	143	416991	33.994	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	704906	30.824	ng/ul	0.00
31) 4-Chloroaniline-d4	11.022	131	1074934	30.651	ng/ul	0.00
46) Dimethylphthalate-d6	14.098	166	2442135	33.458	ng/ul	0.00
49) Acenaphthylene-d8	14.387	160	2874604	34.838	ng/ul	0.00
54) 4-Nitrophenol-d4	14.898	143	77241	5.120	ng/ul	0.00
60) Fluorene-d10	15.687	176	2165694	35.234	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	301031	23.950	ng/ul	0.00
73) Anthracene-d10	17.557	188	3548679	37.932	ng/ul	0.00
81) Pyrene-d10	19.792	212	3914377	42.119	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.957	264	2707027	36.889	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082523\
Data File : BP016764.D
Acq On : 26 Aug 2023 01:44
Operator : MA/JU
Sample : 04105-09
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BHAB1

Quant Time: Aug 26 02:21:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 25 01:47:30 2023
Response via : Initial Calibration

