

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082523\
 Data File : BP016772.D
 Acq On : 26 Aug 2023 06:50
 Operator : MA/JU
 Sample : 04105-12
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH927

Quant Time: Aug 26 13:23:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 25 01:47:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	322276	20.000	ng/ul	0.00
20) Naphthalene-d8	10.863	136	1438401	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.687	164	903510	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.457	188	1987728	20.000	ng/ul	0.00
79) Chrysene-d12	21.551	240	1696731	20.000	ng/ul	0.00
88) Perylene-d12	24.121	264	1525521	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	51731	6.634	ng/uL	0.00
4) Pyridine-d5	3.822	84	353906	14.852	ng/ul	0.00
7) Phenol-d5	7.175	99	295167	10.300	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	776182	42.497	ng/ul	0.00
11) 2-Chlorophenol-d4	7.552	132	776107	36.020	ng/ul	0.00
15) 4-Methylphenol-d8	8.734	113	533668	22.812	ng/ul	-0.01
21) Nitrobenzene-d5	9.234	128	491720	45.213	ng/ul	0.00
24) 2-Nitrophenol-d4	9.946	143	514474	44.878	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.469	165	873279	40.861	ng/ul	0.00
31) 4-Chloroaniline-d4	11.022	131	775405	23.659	ng/ul	0.00
46) Dimethylphthalate-d6	14.098	166	2969318	43.579	ng/ul	0.00
49) Acenaphthylene-d8	14.387	160	3526444	45.782	ng/ul	0.00
54) 4-Nitrophenol-d4	14.892	143	96424	6.847	ng/ul	0.00
60) Fluorene-d10	15.681	176	2638949	45.992	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	412735	34.867	ng/ul	0.00
73) Anthracene-d10	17.557	188	4162542	47.244	ng/ul	0.00
81) Pyrene-d10	19.786	212	4723608	50.689	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.957	264	3636279	47.586	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082523\
Data File : BP016772.D
Acq On : 26 Aug 2023 06:50
Operator : MA/JU
Sample : 04105-12
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH927

Quant Time: Aug 26 13:23:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 25 01:47:30 2023
Response via : Initial Calibration

