

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082525\
 Data File : BP025567.D
 Acq On : 25 Aug 2025 16:25
 Operator : CG/JU
 Sample : Q2932-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ARS20-0002-CONCRETE

Quant Time: Aug 25 17:00:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.778	152	201999	20.000	ng	-0.02
21) Naphthalene-d8	10.543	136	798125	20.000	ng	-0.02
39) Acenaphthene-d10	14.395	164	501586	20.000	ng	-0.01
64) Phenanthrene-d10	17.195	188	872813	20.000	ng	# 0.00
76) Chrysene-d12	21.630	240	1022336	20.000	ng	-0.02
86) Perylene-d12	24.995	264	1200704	20.000	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.402	112	1372530	112.840	ng	0.00
7) Phenol-d6	6.960	99	1829638	117.325	ng	0.00
23) Nitrobenzene-d5	8.919	82	1228254	77.847	ng	-0.02
42) 2,4,6-Tribromophenol	15.907	330	729729	108.086	ng	0.00
45) 2-Fluorobiphenyl	13.001	172	2940629	80.124	ng	-0.02
79) Terphenyl-d14	19.919	244	4220139	75.524	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.548	149	105057	2.370	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082525\
Data File : BP025567.D
Acq On : 25 Aug 2025 16:25
Operator : CG/JU
Sample : Q2932-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ARS20-0002-CONCRETE

Quant Time: Aug 25 17:00:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration

