




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082525\
Data File : BP025560.D
Acq On : 25 Aug 2025 10:24
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 25 12:41:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration



TIC: BP025560.D\data.ms

(70) Pentachlorophenol (C)




16.831min (-0.012) 30402.24 ng

response 788586

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.80	63.60
264.00	65.00	62.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082525\
Data File : BP025560.D
Acq On : 25 Aug 2025 10:24
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 25 12:41:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration



TIC: BP025560.D\data.ms

(77) Benzidine

19.513min (+ 0.000) 18802.33 ng

response 4079827

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.93
183.00	11.80	12.00
0.00	0.00	0.00